

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 22:00 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|-----------|------|------------|
|                      |     |            |           | FROM      | TO   |            |
| Aroclor-1016-1       | (1) | 4.78       | 4.78      | 4.68      | 4.88 | 0.00       |
| Aroclor-1016-2       | (2) | 4.80       | 4.80      | 4.70      | 4.90 | 0.00       |
| Aroclor-1016-3       | (3) | 4.85       | 4.86      | 4.76      | 4.96 | 0.01       |
| Aroclor-1016-4       | (4) | 4.97       | 4.98      | 4.88      | 5.08 | 0.01       |
| Aroclor-1016-5       | (5) | 5.23       | 5.23      | 5.13      | 5.33 | 0.00       |
| Aroclor-1260-1       | (1) | 6.27       | 6.27      | 6.17      | 6.37 | 0.00       |
| Aroclor-1260-2       | (2) | 6.46       | 6.46      | 6.36      | 6.56 | 0.00       |
| Aroclor-1260-3       | (3) | 6.83       | 6.83      | 6.73      | 6.93 | 0.01       |
| Aroclor-1260-4       | (4) | 7.09       | 7.09      | 6.99      | 7.19 | 0.00       |
| Aroclor-1260-5       | (5) | 7.33       | 7.33      | 7.23      | 7.43 | 0.00       |
| Tetrachloro-m-xylene |     | 3.69       | 3.69      | 3.59      | 3.79 | 0.00       |
| Decachlorobiphenyl   |     | 8.73       | 8.73      | 8.63      | 8.83 | 0.00       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 22:00 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1       | (1) | 4.76       | 4.77      | 4.67                   | 4.87 | 0.01       |
| Aroclor-1016-2       | (2) | 4.78       | 4.78      | 4.68                   | 4.88 | 0.00       |
| Aroclor-1016-3       | (3) | 4.96       | 4.96      | 4.86                   | 5.06 | 0.00       |
| Aroclor-1016-4       | (4) | 5.00       | 5.00      | 4.90                   | 5.10 | 0.00       |
| Aroclor-1016-5       | (5) | 5.21       | 5.22      | 5.12                   | 5.32 | 0.01       |
| Aroclor-1260-1       | (1) | 6.24       | 6.25      | 6.15                   | 6.35 | 0.01       |
| Aroclor-1260-2       | (2) | 6.43       | 6.43      | 6.33                   | 6.53 | 0.00       |
| Aroclor-1260-3       | (3) | 6.58       | 6.59      | 6.49                   | 6.69 | 0.01       |
| Aroclor-1260-4       | (4) | 7.05       | 7.06      | 6.96                   | 7.16 | 0.01       |
| Aroclor-1260-5       | (5) | 7.30       | 7.30      | 7.20                   | 7.40 | 0.00       |
| Tetrachloro-m-xylene |     | 3.68       | 3.69      | 3.59                   | 3.79 | 0.01       |
| Decachlorobiphenyl   |     | 8.68       | 8.68      | 8.58                   | 8.78 | 0.00       |