

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 22:55 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1248-2	(2)	5.02	5.02	4.92	5.12	0.01
Aroclor-1248-3	(3)	5.23	5.23	5.13	5.33	0.00
Aroclor-1248-4	(4)	5.58	5.59	5.49	5.69	0.01
Aroclor-1248-5	(5)	5.63	5.63	5.53	5.73	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.73	8.73	8.63	8.83	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 22:55 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1248-2	(2)	5.00	5.00	4.90	5.10	0.00
Aroclor-1248-3	(3)	5.04	5.04	4.94	5.14	0.00
Aroclor-1248-4	(4)	5.21	5.22	5.12	5.32	0.01
Aroclor-1248-5	(5)	5.60	5.61	5.51	5.71	0.01
Tetrachloro-m-xylene		3.68	3.69	3.59	3.79	0.01
Decachlorobiphenyl		8.68	8.69	8.59	8.79	0.01