

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 18:03 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |       | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|-------|------------|
| Aroclor-1016-1       | (1) | 5.67       | 5.67      | 5.57                   | 5.77  | 0.01       |
| Aroclor-1016-2       | (2) | 5.69       | 5.69      | 5.59                   | 5.79  | 0.00       |
| Aroclor-1016-3       | (3) | 5.75       | 5.75      | 5.65                   | 5.85  | 0.00       |
| Aroclor-1016-4       | (4) | 5.85       | 5.85      | 5.75                   | 5.95  | 0.00       |
| Aroclor-1016-5       | (5) | 6.14       | 6.15      | 6.05                   | 6.25  | 0.01       |
| Aroclor-1260-1       | (1) | 7.26       | 7.27      | 7.17                   | 7.37  | 0.01       |
| Aroclor-1260-2       | (2) | 7.51       | 7.52      | 7.42                   | 7.62  | 0.01       |
| Aroclor-1260-3       | (3) | 7.87       | 7.88      | 7.78                   | 7.98  | 0.01       |
| Aroclor-1260-4       | (4) | 8.09       | 8.10      | 8.00                   | 8.20  | 0.01       |
| Aroclor-1260-5       | (5) | 8.41       | 8.42      | 8.32                   | 8.52  | 0.01       |
| Tetrachloro-m-xylene |     | 4.51       | 4.52      | 4.42                   | 4.62  | 0.01       |
| Decachlorobiphenyl   |     | 10.23      | 10.24     | 10.14                  | 10.34 | 0.01       |

### CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 18:03 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             |     | CCAL<br>RT | AVG<br>RT | RT WINDOW<br>FROM   TO |      | DIFF<br>RT |
|----------------------|-----|------------|-----------|------------------------|------|------------|
| Aroclor-1016-1       | (1) | 4.90       | 4.90      | 4.80                   | 5.00 | 0.01       |
| Aroclor-1016-2       | (2) | 4.91       | 4.92      | 4.82                   | 5.02 | 0.01       |
| Aroclor-1016-3       | (3) | 5.09       | 5.09      | 4.99                   | 5.19 | 0.00       |
| Aroclor-1016-4       | (4) | 5.13       | 5.14      | 5.04                   | 5.24 | 0.01       |
| Aroclor-1016-5       | (5) | 5.35       | 5.35      | 5.25                   | 5.45 | 0.00       |
| Aroclor-1260-1       | (1) | 6.38       | 6.39      | 6.29                   | 6.49 | 0.01       |
| Aroclor-1260-2       | (2) | 6.57       | 6.57      | 6.47                   | 6.67 | 0.00       |
| Aroclor-1260-3       | (3) | 6.72       | 6.73      | 6.63                   | 6.83 | 0.01       |
| Aroclor-1260-4       | (4) | 7.19       | 7.20      | 7.10                   | 7.30 | 0.01       |
| Aroclor-1260-5       | (5) | 7.43       | 7.44      | 7.34                   | 7.54 | 0.01       |
| Tetrachloro-m-xylene |     | 3.81       | 3.81      | 3.71                   | 3.91 | 0.00       |
| Decachlorobiphenyl   |     | 8.84       | 8.85      | 8.75                   | 8.95 | 0.01       |