

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 19:09 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	6.52	6.52	6.42	6.62	0.01
Aroclor-1254-2	(2)	6.73	6.74	6.64	6.84	0.01
Aroclor-1254-3	(3)	7.09	7.10	7.00	7.20	0.01
Aroclor-1254-4	(4)	7.38	7.38	7.28	7.48	0.00
Aroclor-1254-5	(5)	7.79	7.80	7.70	7.90	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.23	10.24	10.14	10.34	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 19:09 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1254-1	(1)	5.70	5.70	5.60	5.80	0.00
Aroclor-1254-2	(2)	5.85	5.85	5.75	5.95	0.00
Aroclor-1254-3	(3)	6.25	6.25	6.15	6.35	0.00
Aroclor-1254-4	(4)	6.48	6.48	6.38	6.58	0.00
Aroclor-1254-5	(5)	6.90	6.90	6.80	7.00	0.00
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01