

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 21:55 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.67	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.00
Aroclor-1016-4	(4)	5.85	5.85	5.75	5.95	0.00
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.41	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.23	10.24	10.14	10.34	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 21:55 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.35	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-2	(2)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01