

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110846.D Time Analyzed: 09:24

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.775	4.679	4.879	554.440	500.000	10.9
Aroclor-1016-2	4.795	4.698	4.898	558.040	500.000	11.6
Aroclor-1016-3	4.851	4.755	4.955	542.460	500.000	8.5
Aroclor-1016-4	4.971	4.875	5.075	553.630	500.000	10.7
Aroclor-1016-5	5.228	5.132	5.332	543.180	500.000	8.6
Aroclor-1260-1	6.268	6.172	6.372	535.720	500.000	7.1
Aroclor-1260-2	6.457	6.360	6.560	519.830	500.000	4.0
Aroclor-1260-3	6.824	6.729	6.929	522.530	500.000	4.5
Aroclor-1260-4	7.085	6.988	7.188	513.390	500.000	2.7
Aroclor-1260-5	7.326	7.230	7.430	515.480	500.000	3.1
Decachlorobiphenyl	8.727	8.632	8.832	47.400	50.000	-5.2
Tetrachloro-m-xylene	3.686	3.588	3.788	56.210	50.000	12.4

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110846.D Time Analyzed: 09:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	520.260	500.000	4.1
Aroclor-1016-2	4.781	4.684	4.884	519.530	500.000	3.9
Aroclor-1016-3	4.956	4.860	5.060	502.490	500.000	0.5
Aroclor-1016-4	4.998	4.902	5.102	493.190	500.000	-1.4
Aroclor-1016-5	5.211	5.115	5.315	490.700	500.000	-1.9
Aroclor-1260-1	6.242	6.146	6.346	492.670	500.000	-1.5
Aroclor-1260-2	6.430	6.334	6.534	484.910	500.000	-3.0
Aroclor-1260-3	6.582	6.487	6.687	480.540	500.000	-3.9
Aroclor-1260-4	7.053	6.957	7.157	480.890	500.000	-3.8
Aroclor-1260-5	7.294	7.197	7.397	482.240	500.000	-3.6
Decachlorobiphenyl	8.679	8.584	8.784	46.500	50.000	-7.0
Tetrachloro-m-xylene	3.683	3.586	3.786	53.130	50.000	6.3