

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP071572.D Time Analyzed: 23:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.661	5.571	5.771	502.500	500.000	0.5
Aroclor-1016-2	5.683	5.593	5.793	488.520	500.000	-2.3
Aroclor-1016-3	5.745	5.654	5.854	492.490	500.000	-1.5
Aroclor-1016-4	5.843	5.752	5.952	494.900	500.000	-1.0
Aroclor-1016-5	6.135	6.045	6.245	486.570	500.000	-2.7
Aroclor-1260-1	7.254	7.165	7.365	498.210	500.000	-0.4
Aroclor-1260-2	7.508	7.418	7.618	488.990	500.000	-2.2
Aroclor-1260-3	7.866	7.777	7.977	505.180	500.000	1.0
Aroclor-1260-4	8.090	8.002	8.202	488.020	500.000	-2.4
Aroclor-1260-5	8.410	8.322	8.522	497.350	500.000	-0.5
Decachlorobiphenyl	10.223	10.137	10.337	50.350	50.000	0.7
Tetrachloro-m-xylene	4.508	4.417	4.617	49.040	50.000	-1.9

CALIBRATION VERIFICATION SUMMARY

Contract: ROYF02

Lab Code: CHEM Case No.: Q1884 SAS No.: Q1884 SDG NO.: Q1884

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP071572.D Time Analyzed: 23:31

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.893	4.798	4.998	521.320	500.000	4.3
Aroclor-1016-2	4.912	4.816	5.016	520.920	500.000	4.2
Aroclor-1016-3	5.088	4.993	5.193	527.030	500.000	5.4
Aroclor-1016-4	5.130	5.036	5.236	522.470	500.000	4.5
Aroclor-1016-5	5.344	5.250	5.450	552.040	500.000	10.4
Aroclor-1260-1	6.378	6.285	6.485	531.810	500.000	6.4
Aroclor-1260-2	6.567	6.473	6.673	528.120	500.000	5.6
Aroclor-1260-3	6.720	6.627	6.827	531.940	500.000	6.4
Aroclor-1260-4	7.191	7.098	7.298	532.870	500.000	6.6
Aroclor-1260-5	7.432	7.339	7.539	540.330	500.000	8.1
Decachlorobiphenyl	8.840	8.749	8.949	51.350	50.000	2.7
Tetrachloro-m-xylene	3.809	3.712	3.912	51.300	50.000	2.6