

Report of Analysis

| | | | | | |
|--------------------|------------------------|----------|--------------------|----------|-----------|
| Client: | Weston Solutions, Inc. | | Date Collected: | 04/24/25 | |
| Project: | RFP 911 | | Date Received: | 04/25/25 | |
| Client Sample ID: | P001-SS037-01DL | | SDG No.: | Q1884 | |
| Lab Sample ID: | Q1884-01DL | | Matrix: | SOIL | |
| Analytical Method: | SW8082A | | % Solid: | 94.2 | Decanted: |
| Sample Wt/Vol: | 30.07 | Units: g | Final Vol: | 10000 | uL |
| Soil Aliquot Vol: | | uL | Test: | PCB | |
| Extraction Type: | | | Injection Volume : | | |
| GPC Factor : | 1.0 | PH : | | | |
| Prep Method : | SW3541B | | | | |

| | | | | |
|-------------------|-----------|----------------|----------------|---------------|
| File ID/Qc Batch: | Dilution: | Prep Date | Date Analyzed | Prep Batch ID |
| PP071605.D | 20 | 04/28/25 09:05 | 04/29/25 12:12 | PB167765 |

| CAS Number | Parameter | Conc. | Qualifier | MDL | LOQ / CRQL | Units(Dry Weight) |
|-------------------|----------------------|-------|-----------|----------|------------|-------------------|
| TARGETS | | | | | | |
| 12674-11-2 | Aroclor-1016 | 83.7 | UD | 83.7 | 360 | ug/kg |
| 11104-28-2 | Aroclor-1221 | 85.4 | UD | 85.4 | 360 | ug/kg |
| 11141-16-5 | Aroclor-1232 | 78.8 | UD | 78.8 | 360 | ug/kg |
| 53469-21-9 | Aroclor-1242 | 84.9 | UD | 84.9 | 360 | ug/kg |
| 12672-29-6 | Aroclor-1248 | 3000 | D | 125 | 360 | ug/kg |
| 11097-69-1 | Aroclor-1254 | 3400 | D | 68.0 | 360 | ug/kg |
| 37324-23-5 | Aroclor-1262 | 106 | UD | 106 | 360 | ug/kg |
| 11100-14-4 | Aroclor-1268 | 76.3 | UD | 76.3 | 360 | ug/kg |
| 11096-82-5 | Aroclor-1260 | 2900 | D | 68.4 | 360 | ug/kg |
| SURROGATES | | | | | | |
| 877-09-8 | Tetrachloro-m-xylene | 33.4 | * | 32 - 144 | 167% | SPK: 20 |
| 2051-24-3 | Decachlorobiphenyl | 93.4 | * | 32 - 175 | 467% | SPK: 20 |

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit