

ANALYTICAL RESULTS SUMMARY

SEMI-VOLATILE ORGANICS

PROJECT NAME : NWIRP BETHPAGE - RW5B CTO WE13 112G08005

TETRA TECH NUS, INC.

661 Andersen Drive

Suite 200

Pittsburgh, PA - 15220-2745

Phone No: 412-921-7090

ORDER ID : Q1886

ATTENTION : Ernie Wu



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) SVOC-SIMGroup1- Case Narrative	4
3) Qualifier Page	6
4) QA Checklist	7
5) SVOC-SIMGroup1 Data	8
6) Shipping Document	33
6.1) CHAIN OF CUSTODY	34
6.2) Lab Certificate	35

Cover Page

Order ID : Q1886

Project ID : NWIRP Bethpage - RW5B CTO WE13 112G08005

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q1886-01
Q1886-02
Q1886-03

Client Sample Number

RW5-SP100-20250423
RW5-SP201-20250423
RW5-SP303-20250423

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 5/3/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage - RW5B CTO WE13 112G08005

Project Manager: Ernie Wu

Chemtech Project # Q1886

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

3 Water samples were received on 04/25/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank Spike Duplicate for {PB167760BSD} with File ID: BN036953.D met requirements for all samples except for 1,4-Dioxane[68%], recovery marginally failed therefore no corrective action was taken.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

The Tuning criteria met requirements.

Sample RW5-SP100-20250423 was diluted due to high concentration.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is)."

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1886

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 05/03/2025

LAB CHRONICLE

OrderID:	Q1886	OrderDate:	4/25/2025 10:16:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage - RW5B CTO WE13 112G08005
Contact:	Ernie Wu	Location:	L51

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1886-01	RW5-SP100-2025042 3	Water			04/23/25			04/25/25
			SVOC-SIMGroup1	8270-Modified		04/28/25	04/30/25	
Q1886-01DL	RW5-SP100-2025042 3DL	Water			04/23/25			04/25/25
			SVOC-SIMGroup1	8270-Modified		04/28/25	04/30/25	
Q1886-02	RW5-SP201-2025042 3	Water			04/23/25			04/25/25
			SVOC-SIMGroup1	8270-Modified		04/28/25	04/30/25	
Q1886-03	RW5-SP303-2025042 3	Water			04/23/25			04/25/25
			SVOC-SIMGroup1	8270-Modified		04/28/25	04/30/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: Q1886

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : RW5-SP100-20250423								
Q1886-01	RW5-SP100-20250423	WATER	1,4-Dioxane	6.200	EQ	0.07	0.2	0.2 ug/L
			Total Svoc :	6.20				
			Total Concentration:	6.20				
Client ID : RW5-SP100-20250423DL								
Q1886-01DL	RW5-SP100-20250423DI	WATER	1,4-Dioxane	6.400	DQ	0.13	0.4	0.4 ug/L
			Total Svoc :	6.40				
			Total Concentration:	6.40				



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	04/23/25
Project:	NWIRP Bethpage - RW5B CTO WE13 112G08005	Date Received:	04/25/25
Client Sample ID:	RW5-SP100-20250423	SDG No.:	Q1886
Lab Sample ID:	Q1886-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036947.D	1	04/28/25 08:50	04/30/25 15:16	PB167760

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	6.20	EQ	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		76%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.40		30 - 150		99%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		79%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.31		53 - 106		78%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.36		58 - 132		89%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1740	7.633				
1146-65-2	Naphthalene-d8	4360	10.415				
15067-26-2	Acenaphthene-d10	2480	14.277				
1517-22-2	Phenanthrene-d10	5000	17.021				
1719-03-5	Chrysene-d12	4500	21.215				
1520-96-3	Perylene-d12	4170	23.424				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	04/23/25
Project:	NWIRP Bethpage - RW5B CTO WE13 112G08005	Date Received:	04/25/25
Client Sample ID:	RW5-SP100-20250423DL	SDG No.:	Q1886
Lab Sample ID:	Q1886-01DL	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036951.D	2	04/28/25 08:50	04/30/25 17:41	PB167760

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	6.40	DQ	0.13	0.40	0.40	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		76%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.38		30 - 150		94%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		79%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.28		53 - 106		69%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.38		58 - 132		95%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1640	7.633				
1146-65-2	Naphthalene-d8	4150	10.415				
15067-26-2	Acenaphthene-d10	2380	14.277				
1517-22-2	Phenanthrene-d10	4970	17.021				
1719-03-5	Chrysene-d12	3860	21.215				
1520-96-3	Perylene-d12	3390	23.424				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	04/23/25
Project:	NWIRP Bethpage - RW5B CTO WE13 112G08005	Date Received:	04/25/25
Client Sample ID:	RW5-SP201-20250423	SDG No.:	Q1886
Lab Sample ID:	Q1886-02	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036948.D	1	04/28/25 08:50	04/30/25 15:52	PB167760

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	UQ	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		77%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.35		30 - 150		86%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		80%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.33		53 - 106		83%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.40		58 - 132		100%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1780	7.633				
1146-65-2	Naphthalene-d8	4540	10.415				
15067-26-2	Acenaphthene-d10	2510	14.277				
1517-22-2	Phenanthrene-d10	5100	17.021				
1719-03-5	Chrysene-d12	4280	21.215				
1520-96-3	Perylene-d12	3780	23.424				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	04/23/25
Project:	NWIRP Bethpage - RW5B CTO WE13 112G08005	Date Received:	04/25/25
Client Sample ID:	RW5-SP303-20250423	SDG No.:	Q1886
Lab Sample ID:	Q1886-03	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036949.D	1	04/28/25 08:50	04/30/25 16:28	PB167760

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	UQ	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.29		30 - 150		72%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.37		30 - 150		91%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		76%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.26		53 - 106		66%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.33		58 - 132		83%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1940	7.633				
1146-65-2	Naphthalene-d8	4850	10.415				
15067-26-2	Acenaphthene-d10	2710	14.277				
1517-22-2	Phenanthrene-d10	5600	17.021				
1719-03-5	Chrysene-d12	4530	21.216				
1520-96-3	Perylene-d12	3930	23.424				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q1886

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167760BL	PB167760BL	2-Methylnaphthalene-d10	0.4	0.27	68		30	150
		Fluoranthene-d10	0.4	0.28	71		30	150
		Nitrobenzene-d5	0.4	0.28	70		55	111
		2-Fluorobiphenyl	0.4	0.29	73		53	106
		Terphenyl-d14	0.4	0.31	77		58	132
PB167760BS	PB167760BS	2-Methylnaphthalene-d10	0.4	0.36	91		30	150
		Fluoranthene-d10	0.4	0.35	88		30	150
		Nitrobenzene-d5	0.4	0.35	86		55	111
		2-Fluorobiphenyl	0.4	0.31	79		53	106
		Terphenyl-d14	0.4	0.36	90		58	132
PB167760BSD	PB167760BSD	2-Methylnaphthalene-d10	0.4	0.36	89		30	150
		Fluoranthene-d10	0.4	0.34	84		30	150
		Nitrobenzene-d5	0.4	0.34	84		55	111
		2-Fluorobiphenyl	0.4	0.31	77		53	106
		Terphenyl-d14	0.4	0.36	90		58	132
Q1886-01	RW5-SP100-20250423	2-Methylnaphthalene-d10	0.4	0.31	76		30	150
		Fluoranthene-d10	0.4	0.40	99		30	150
		Nitrobenzene-d5	0.4	0.32	79		55	111
		2-Fluorobiphenyl	0.4	0.31	78		53	106
		Terphenyl-d14	0.4	0.36	89		58	132
Q1886-01DL	RW5-SP100-20250423DL	2-Methylnaphthalene-d10	0.4	0.31	76		30	150
		Fluoranthene-d10	0.4	0.38	94		30	150
		Nitrobenzene-d5	0.4	0.32	79		55	111
		2-Fluorobiphenyl	0.4	0.28	69		53	106
		Terphenyl-d14	0.4	0.38	95		58	132
Q1886-02	RW5-SP201-20250423	2-Methylnaphthalene-d10	0.4	0.31	77		30	150
		Fluoranthene-d10	0.4	0.35	86		30	150
		Nitrobenzene-d5	0.4	0.32	80		55	111
		2-Fluorobiphenyl	0.4	0.33	83		53	106
		Terphenyl-d14	0.4	0.40	100		58	132
Q1886-03	RW5-SP303-20250423	2-Methylnaphthalene-d10	0.4	0.29	72		30	150
		Fluoranthene-d10	0.4	0.37	91		30	150
		Nitrobenzene-d5	0.4	0.31	76		55	111
		2-Fluorobiphenyl	0.4	0.26	66		53	106
		Terphenyl-d14	0.4	0.33	83		58	132

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1886

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN036952.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB167760BS	1,4-Dioxane	0.4	0.29	ug/L	73				70	130	

A
B
C
D
E
F
G

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1886

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN036953.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB167760BSD	1,4-Dioxane	0.4	0.27	ug/L	68	7	*		70	130	20

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167760BL

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q1886

SAS No.: Q1886 SDG NO.: Q1886

Lab File ID: BN036942.D

Lab Sample ID: PB167760BL

Instrument ID: BNA_N

Date Extracted: 04/28/2025

Matrix: (soil/water) Water

Date Analyzed: 04/30/2025

Level: (low/med) LOW

Time Analyzed: 11:38

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB167760BS	PB167760BS	BN036952.D	04/30/2025
RW5-SP100-20250423	Q1886-01	BN036947.D	04/30/2025
RW5-SP201-20250423	Q1886-02	BN036948.D	04/30/2025
RW5-SP303-20250423	Q1886-03	BN036949.D	04/30/2025
PB167760BSD	PB167760BSD	BN036953.D	04/30/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1886 SDG NO.: Q1886

Lab File ID: BN036922.D

DFTPP Injection Date: 04/28/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:56

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	67.4
68	Less than 2.0% of mass 69	0.8 (1.4) 1
69	Mass 69 relative abundance	58.8
70	Less than 2.0% of mass 69	0.2 (0.4) 1
127	10.0 - 80.0% of mass 198	54.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
275	10.0 - 60.0% of mass 198	23.7
365	Greater than 1% of mass 198	3.8
441	Present, but less than mass 443	8.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	9.3 (19.4) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036923.D	04/28/2025	11:35
SSTDICC0.2	SSTDICC0.2	BN036924.D	04/28/2025	12:11
SSTDICCC0.4	SSTDICCC0.4	BN036925.D	04/28/2025	12:47
SSTDICC0.8	SSTDICC0.8	BN036926.D	04/28/2025	13:24
SSTDICC1.6	SSTDICC1.6	BN036927.D	04/28/2025	14:00
SSTDICC3.2	SSTDICC3.2	BN036928.D	04/28/2025	14:36
SSTDICC5.0	SSTDICC5.0	BN036929.D	04/28/2025	15:12

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1886 SDG NO.: Q1886

Lab File ID: BN036940.D

DFTPP Injection Date: 04/30/2025

Instrument ID: BNA_N

DFTPP Injection Time: 10:20

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	72
68	Less than 2.0% of mass 69	0.6 (0.9) 1
69	Mass 69 relative abundance	60.6
70	Less than 2.0% of mass 69	0.3 (0.5) 1
127	10.0 - 80.0% of mass 198	54.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	24.7
365	Greater than 1% of mass 198	4
441	Present, but less than mass 443	8.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	9.9 (19.5) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036941.D	04/30/2025	10:59
PB167760BL	PB167760BL	BN036942.D	04/30/2025	11:38
RW5-SP100-20250423	Q1886-01	BN036947.D	04/30/2025	15:16
RW5-SP201-20250423	Q1886-02	BN036948.D	04/30/2025	15:52
RW5-SP303-20250423	Q1886-03	BN036949.D	04/30/2025	16:28
RW5-SP100-20250423DL	Q1886-01DL	BN036951.D	04/30/2025	17:41
PB167760BS	PB167760BS	BN036952.D	04/30/2025	18:17
PB167760BSD	PB167760BSD	BN036953.D	04/30/2025	18:53
SSTDCCC0.4EC	SSTDCCC0.4	BN036954.D	04/30/2025	19:29

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1886 SAS No.: Q1886 SDG NO.: Q1886

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 04/30/2025

Lab File ID: BN036941.D Time Analyzed: 10:59

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2089	7.625	5327	10.40	3019	14.28
UPPER LIMIT	4178	8.125	10654	10.904	6038	14.777
LOWER LIMIT	1044.5	7.125	2663.5	9.904	1509.5	13.777
EPA SAMPLE NO.						
01 PB167760BL	2641	7.63	6297	10.42	3237	14.28
02 RW5-SP100-20250423	1739	7.63	4356	10.42	2481	14.28
03 PB167760BS	2072	7.63	5162	10.40	2727	14.28
04 PB167760BSD	2106	7.63	5174	10.40	2716	14.28
05 RW5-SP100-20250423DL	1638	7.63	4145	10.42	2384	14.28
06 RW5-SP201-20250423	1784	7.63	4537	10.42	2513	14.28
07 RW5-SP303-20250423	1937	7.63	4854	10.42	2711	14.28

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1886 SAS No.: Q1886 SDG NO.: Q1886

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 04/30/2025

Lab File ID: BN036941.D Time Analyzed: 10:59

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	6059	17.021	4788	21.215	4103	23.427
UPPER LIMIT	12118	17.521	9576	21.715	8206	23.927
LOWER LIMIT	3029.5	16.521	2394	20.715	2051.5	22.927
EPA SAMPLE NO.						
01 PB167760BL	6245	17.02	4196	21.22	3575	23.43
02 RW5-SP100-20250423	5000	17.02	4495	21.22	4168	23.42
03 PB167760BS	5423	17.02	4126	21.22	3488	23.42
04 PB167760BSD	5366	17.02	3877	21.22	3193	23.42
05 RW5-SP100-20250423DL	4971	17.02	3859	21.22	3394	23.42
06 RW5-SP201-20250423	5104	17.02	4282	21.22	3782	23.42
07 RW5-SP303-20250423	5602	17.02	4527	21.22	3928	23.42

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:		
Project:	NWIRP Bethpage - RW5B CTO WE13 112G08005		Date Received:		
Client Sample ID:	PB167760BL		SDG No.:	Q1886	
Lab Sample ID:	PB167760BL		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036942.D	1	04/28/25 08:50	04/30/25 11:38	PB167760

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.27		30 - 150		68%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.28		30 - 150		71%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.28		55 - 111		70%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.29		53 - 106		73%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.31		58 - 132		77%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2640	7.633				
1146-65-2	Naphthalene-d8	6300	10.415				
15067-26-2	Acenaphthene-d10	3240	14.277				
1517-22-2	Phenanthrene-d10	6250	17.021				
1719-03-5	Chrysene-d12	4200	21.224				
1520-96-3	Perylene-d12	3580	23.43				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage - RW5B CTO WE13 112G08005		Date Received:	
Client Sample ID:	PB167760BS		SDG No.:	Q1886
Lab Sample ID:	PB167760BS		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036952.D	1	04/28/25 08:50	04/30/25 18:17	PB167760

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.29		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.36		30 - 150		91%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.35		30 - 150		88%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.35		55 - 111		86%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.31		53 - 106		79%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.36		58 - 132		90%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2070	7.626				
1146-65-2	Naphthalene-d8	5160	10.404				
15067-26-2	Acenaphthene-d10	2730	14.277				
1517-22-2	Phenanthrene-d10	5420	17.021				
1719-03-5	Chrysene-d12	4130	21.216				
1520-96-3	Perylene-d12	3490	23.424				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:		
Project:	NWIRP Bethpage - RW5B CTO WE13 112G08005		Date Received:		
Client Sample ID:	PB167760BSD		SDG No.:	Q1886	
Lab Sample ID:	PB167760BSD		Matrix:	Water	
Analytical Method:	SW8270ESIM		% Solid:	0	
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000	uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1	
Extraction Type :		Decanted : N	Level :	LOW	
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N	PH :
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036953.D	1	04/28/25 08:50	04/30/25 18:53	PB167760

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.27		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.36		30 - 150		89%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.34		30 - 150		84%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.34		55 - 111		84%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.31		53 - 106		77%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.36		58 - 132		90%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	2110	7.625				
1146-65-2	Naphthalene-d8	5170	10.404				
15067-26-2	Acenaphthene-d10	2720	14.277				
1517-22-2	Phenanthrene-d10	5370	17.021				
1719-03-5	Chrysene-d12	3880	21.215				
1520-96-3	Perylene-d12	3190	23.424				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN042825.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Apr 28 15:35:03 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN036923.D 0.2 =BN036924.D 0.4 =BN036925.D 0.8 =BN036926.D 1.6 =BN036927.D 3.2 =BN036928.D 5.0 =BN036929.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.452	0.489	0.551	0.506	0.537	0.489	0.465	0.498	7.23
3)	n-Nitrosodimet...	0.903	0.998	1.010	0.957	1.034	0.952	0.918	0.967	5.01
4) S	2-Fluorophenol	1.050	1.056	1.118	0.946	1.040	0.982	0.970	1.023	5.86
5) S	Phenol-d6	1.270	1.237	1.337	1.151	1.294	1.255	1.272	1.259	4.57
6)	bis(2-Chloroet...	1.174	1.123	1.170	1.139	1.240	1.162	1.162	1.167	3.17
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.400	0.401	0.411	0.404	0.446	0.432	0.436	0.418	4.52
9)	Naphthalene	1.155	1.147	1.155	1.132	1.225	1.170	1.165	1.164	2.56
10)	Hexachlorobuta...	0.260	0.250	0.253	0.249	0.262	0.248	0.240	0.252	2.99
11) SURR2	Methylnaphth...	0.540	0.532	0.541	0.543	0.596	0.575	0.589	0.559	4.75
12)	2-Methylnaphth...	0.716	0.713	0.719	0.735	0.804	0.782	0.798	0.753	5.41
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.156	0.173	0.177	0.175	0.187	0.184	0.196	0.178	7.18
15) S	2-Fluorobiphenyl	1.877	1.975	2.055	1.690	2.023	1.986	1.928	1.933	6.32
16)	Acenaphthylene	1.876	1.850	1.907	1.884	2.067	2.035	2.066	1.955	4.93
17)	Acenaphthene	1.264	1.270	1.275	1.248	1.333	1.295	1.305	1.284	2.22
18)	Fluorene	1.604	1.612	1.624	1.658	1.788	1.720	1.752	1.680	4.39
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.083	0.090	0.096	0.113	0.120	0.134	0.106		18.55
21)	4-Bromophenyl-...	0.260	0.263	0.262	0.260	0.282	0.272	0.270	0.267	3.11
22)	Hexachlorobenzene	0.301	0.289	0.300	0.280	0.303	0.293	0.282	0.293	3.18
23)	Atrazine	0.193	0.198	0.199	0.217	0.227	0.226	0.248	0.215	9.20
24)	Pentachlorophenol	0.160	0.136	0.144	0.145	0.163	0.168	0.181	0.157	10.06
25)	Phenanthrene	1.309	1.274	1.299	1.280	1.387	1.346	1.347	1.320	3.13
26)	Anthracene	1.131	1.108	1.147	1.138	1.275	1.261	1.299	1.194	6.74
27) SURR	Fluoranthene-d10	0.993	1.004	0.991	1.016	1.087	1.053	1.115	1.037	4.74
28)	Fluoranthene	1.387	1.380	1.399	1.471	1.578	1.530	1.613	1.480	6.46
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.919	1.942	1.958	1.802	2.073	1.969	1.823	1.927	4.77
31) S	Terphenyl-d14	0.974	0.942	0.946	0.893	1.005	0.952	0.897	0.944	4.22
32)	Benzo(a)anthra...	1.402	1.407	1.429	1.422	1.583	1.509	1.561	1.473	5.19
33)	Chrysene	1.517	1.576	1.637	1.582	1.700	1.572	1.536	1.589	3.91
34)	Bis(2-ethylhex...	0.949	0.847	0.834	0.784	0.804	0.782	0.866	0.838	6.96
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN042825.M

36)	Indeno(1,2,3-c...	1.595	1.571	1.712	1.503	1.720	1.724	1.609	1.634	5.29
37)	Benzo(b)fluora...	1.580	1.552	1.634	1.628	1.796	1.758	1.825	1.682	6.50
38)	Benzo(k)fluora...	1.601	1.569	1.648	1.641	1.812	1.784	1.785	1.691	5.89
39) C	Benzo(a)pyrene	1.315	1.301	1.361	1.315	1.463	1.447	1.477	1.383	5.57
40)	Dibenzo(a,h)an...	1.229	1.241	1.349	1.176	1.357	1.379	1.268	1.286	5.96
41)	Benzo(g,h,i)pe...	1.459	1.405	1.515	1.305	1.495	1.470	1.339	1.427	5.61

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1886 SAS No.: Q1886 SDG No.: Q1886

Instrument ID: BNA_N Calibration Date/Time: 04/30/2025 10:59

Lab File ID: BN036941.D Init. Calib. Date(s): 04/28/2025 04/28/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 11:35 15:12

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.559	0.553		-1.1	20.0
Fluoranthene-d10	1.037	1.024		-1.3	20.0
2-Fluorophenol	1.023	1.039		1.6	20.0
Phenol-d6	1.259	1.276		1.4	20.0
Nitrobenzene-d5	0.418	0.412		-1.4	20.0
2-Fluorobiphenyl	1.933	2.002		3.6	20.0
2,4,6-Tribromophenol	0.178	0.171		-3.9	20.0
Terphenyl-d14	0.944	0.910		-3.6	20.0
1,4-Dioxane	0.498	0.530		6.4	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1886 SAS No.: Q1886 SDG No.: Q1886

Instrument ID: BNA_N Calibration Date/Time: 04/30/2025 19:29

Lab File ID: BN036954.D Init. Calib. Date(s): 04/28/2025 04/28/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 11:35 15:12

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.559	0.551		-1.4	50.0
Fluoranthene-d10	1.037	1.031		-0.6	50.0
2-Fluorophenol	1.023	1.071		4.7	50.0
Phenol-d6	1.259	1.249		-0.8	50.0
Nitrobenzene-d5	0.418	0.418		0.0	50.0
2-Fluorobiphenyl	1.933	1.735		-10.2	50.0
2,4,6-Tribromophenol	0.178	0.181		1.7	50.0
Terphenyl-d14	0.944	0.934		-1.1	50.0
1,4-Dioxane	0.498	0.513		3.0	50.0

All other compounds must meet a minimum RRF of 0.010.



SHIPPING DOCUMENTS

CHEMTECH CHAIN OF CUSTODY RECORD		284 Sheffield Street, Mountainside, NJ 07092 (908) 789-8900 Fax: (908) 78-8922 www.chemtech.net		Chemtech Project Number: Q1886 COC Number:										
CLIENT INFORMATION		PROJECT INFORMATION		BILLING INFORMATION										
COMPANY: Tetra Tech		PROJECT NAME: NWIRP Bethpage		BILL TO: PO#										
ADDRESS: 4433 Corporation Ln, Suite 300		PROJECT #: 112G08005-WE13 LOCATION: RW5B		ADDRESS:										
CITY: Virginia Beach STATE: VA ZIP: 23462		PROJECT MANAGER: Ernie Wu		CITY: STATE: ZIP:										
ATTENTION: Ernie Wu		E-MAIL: ernie.wu@tetratech.com		ATTENTION: PHONE:										
PHONE: 757-466-4901 FAX: 757-461-4148		PHONE: 757-466-4901 FAX: 757-461-4148												
DATA TURNAROUND INFORMATION		DATA DELIVERABLE INFORMATION		ANALYSIS										
FAX: _____ 10 _____ DAYS* HARD COPY: _____ 10 _____ DAYS* EDD _____ 10 _____ DAYS* * TO BE APPROVED BY CHEMTECH STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS		☐ RESEULTS ONLY ☐ USEPA CLP ☐ RESULTS + QC ☐ New York State ASP "B" ☐ New Jersey REDUCED ☐ New York State ASP "A" ☐ New Jersey CLP ☐ Other _____ ☐ EDD Format _____		1,4-Dioxane SW846 9270 SIM										
				<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td>1</td><td>2</td><td>3</td><td>4</td><td>5</td><td>6</td><td>7</td><td>8</td><td>9</td> </tr> </table>		1	2	3	4	5	6	7	8	9
1	2	3	4	5	6	7	8	9						
				PRESERVATIVES										
				<table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td>1</td><td>2</td><td>3</td><td>4</td><td>5</td><td>6</td><td>7</td><td>8</td><td>9</td> </tr> </table>		1	2	3	4	5	6	7	8	9
1	2	3	4	5	6	7	8	9						
				COMMENTS										
				<-- Specify Preservatives A-HCl B-HNO3 C-H2SO4 D-NaOH E-ICE F-Other										
CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP GRAB	SAMPLE COLLECTION DATE TIME	# of Bottles									
1.	RW5-SP100-20250423	GW		4/23/25 10:45	1									
2.	RW5-SP201-20250423	GW		4/23/25 10:47	1									
3.	RW5-SP303-20250423	GW		4/23/25 10:53	1									
4.														
5.														
6.														
7.														
8.														
9.														
10.														

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER	DATE/TIME	RECEIVED BY	Conditions of bottles or coolers at receipt: ☐ Compliant ☐ Non Compliant ☐ Cooler Temp <u>26°C</u> MeOH extraction requires an additional 4oz. Jar for percent solid Comments:
1.	4/24/25/1600	1.	☐ Ice in Cooler?: <u>Y</u> Adjust Factor +1 IR (N/A #)
RELINQUISHED BY	DATE/TIME	RECEIVED BY	
2.	4-25-25 0930	2.	
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY	SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight CHEMTECH: ☐ Picked Up ☐ Overnight
3.		3.	Shipment Complete ☐ YES ☐ NO

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT
YELLOW - CHEMTECH COPY
PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488