

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN042825

Review By	Rahul	Review On	4/29/2025 8:54:47 AM
Supervise By	Jagrut	Supervise On	4/29/2025 11:57:54 AM
SubDirectory	BN042825	HP Acquire Method	BNA_N, 8270_HP Processing Method BN042825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731
CCC	SP6735
Internal Standard/PEM	SP6740,1ul/100ul sample
ICV/I.BLK	SP6768
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN036922.D	28 Apr 2025 10:56		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN036923.D	28 Apr 2025 11:35	Compound#20 removed from 0.1 ppm	RC/JU	Ok,M
3	SSTDICC0.2	SSTDICC0.2	BN036924.D	28 Apr 2025 12:11		RC/JU	Ok,M
4	SSTDICCC0.4	SSTDICCC0.4	BN036925.D	28 Apr 2025 12:47		RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN036926.D	28 Apr 2025 13:24		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN036927.D	28 Apr 2025 14:00		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN036928.D	28 Apr 2025 14:36		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN036929.D	28 Apr 2025 15:12		RC/JU	Ok
9	SSTDICV0.4	ICVBN042825	BN036930.D	28 Apr 2025 15:51		RC/JU	Ok
10	PB167430BL	PB167430BL	BN036931.D	28 Apr 2025 17:39	Internal Standard Fail, Analyzed for contamination check	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN043025

Review By	Rahul	Review On	5/1/2025 11:31:01 AM
Supervise By	Jagrut	Supervise On	5/1/2025 1:21:29 PM
SubDirectory	BN043025	HP Acquire Method	BNA_N, 8270_HP Processing Method BN042825

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731
CCC	SP6735
Internal Standard/PEM	SP6740,1ul/100ul sample
ICV/I.BLK	SP6768
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN036940.D	30 Apr 2025 10:20		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN036941.D	30 Apr 2025 10:59		RC/JU	Ok
3	PB167760BL	PB167760BL	BN036942.D	30 Apr 2025 11:38		RC/JU	Ok
4	Q1885-01	RW7-SP100-20250423	BN036943.D	30 Apr 2025 12:16		RC/JU	Ok
5	Q1885-02	RW7-SP201-20250423	BN036944.D	30 Apr 2025 12:52		RC/JU	Ok,M
6	Q1885-03	RW7-SP302-20250423	BN036945.D	30 Apr 2025 13:28		RC/JU	Ok
7	Q1870-01DL3	38072-010925DL3	BN036946.D	30 Apr 2025 14:40		RC/JU	Ok
8	Q1886-01	RW5-SP100-20250423	BN036947.D	30 Apr 2025 15:16	Need 2X Dilution	RC/JU	Dilution
9	Q1886-02	RW5-SP201-20250423	BN036948.D	30 Apr 2025 15:52		RC/JU	Ok
10	Q1886-03	RW5-SP303-20250423	BN036949.D	30 Apr 2025 16:28		RC/JU	Ok
11	Q1885-04	RW7-SP303-20250423	BN036950.D	30 Apr 2025 17:05		RC/JU	Ok
12	Q1886-01DL	RW5-SP100-20250423	BN036951.D	30 Apr 2025 17:41		RC/JU	Ok
13	PB167760BS	PB167760BS	BN036952.D	30 Apr 2025 18:17		RC/JU	Ok,M
14	PB167760BSD	PB167760BSD	BN036953.D	30 Apr 2025 18:53		RC/JU	Ok,M
15	SSTDCCC0.4	SSTDCCC0.4EC	BN036954.D	30 Apr 2025 19:29		RC/JU	Ok

M : Manual Integration