

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1888 SAS No.: Q1888 SDG NO.: Q1888

EPA Sample No.: SSTD020367 Date Analyzed: 05/02/2025

Lab File ID: BN036956.D Time Analyzed: 13:16

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	304492	17.022	335761	21.257	363826	24.145
UPPER LIMIT	608984	17.522	671522	21.757	727652	24.645
LOWER LIMIT	152246	16.522	167881	20.757	181913	23.645
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	263314	17.02	254237	21.25	240799	24.15

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.