

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: Q1888 SAS No.: Q1888 SDG No.: Q1888
Instrument ID: MSVOA_X Calibration Date/Time: 04/25/2025 09:42
Lab File ID: VX045932.D Init. Calib. Date(s): 04/01/2025 04/01/2025
Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02
GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
t-1,3-Dichloropropene	50.000	53.16	-6.316002	20