

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1888 SAS No.: Q1888 SDG NO.: Q1888
 Lab File ID: VX045932.D Date Analyzed: 04/25/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:42
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	61250	12.018				
UPPER LIMIT	122500	12.518				
LOWER LIMIT	30625	11.518				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	52236	12.02				
STORAGE-BLANK-WATER-REF-5	48670	12.02				
STORAGE-BLANK-WATER-REF-4	49096	12.02				
STORAGE-BLANK-SAMPLE-MANAGEMENT	49502	12.02				
VX0425WBL01	56095	12.02				
VX0425WBS01	61645	12.02				
VX0425WBSD01	60088	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.