

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q1889

Project: Mitchell School

RunNo.: LB135573

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1						
Bromide	mg/L	10.2	10	102	90-110	04/22/2025
Chloride	mg/L	3.1	3	103	90-110	04/22/2025
Fluoride	mg/L	2	2	100	90-110	04/22/2025
Nitrite	mg/L	3.1	3	103	90-110	04/22/2025
Nitrate	mg/L	2.6	2.5	104	90-110	04/22/2025
Sulfate	mg/L	15.1	15	101	90-110	04/22/2025
Orthophosphate as P	mg/L	5.2	5	104	90-110	04/22/2025
Sample ID: CCV1						
Bromide	mg/L	10.5	10	105	90-110	04/28/2025
Chloride	mg/L	3.2	3	107	90-110	04/28/2025
Fluoride	mg/L	2.1	2	105	90-110	04/28/2025
Nitrite	mg/L	3.1	3	103	90-110	04/28/2025
Nitrate	mg/L	2.7	2.5	108	90-110	04/28/2025
Sulfate	mg/L	15.4	15	103	90-110	04/28/2025
Orthophosphate as P	mg/L	5.1	5	102	90-110	04/28/2025
Sample ID: CCV2						
Bromide	mg/L	10.6	10	106	90-110	04/28/2025
Chloride	mg/L	3.2	3	107	90-110	04/28/2025
Fluoride	mg/L	2.1	2	105	90-110	04/28/2025
Nitrite	mg/L	3.2	3	107	90-110	04/28/2025
Nitrate	mg/L	2.7	2.5	108	90-110	04/28/2025
Sulfate	mg/L	15.4	15	103	90-110	04/28/2025
Orthophosphate as P	mg/L	5.3	5	106	90-110	04/28/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q1889

Project: Mitchell School

RunNo.: LB135578

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.498	0.5	100	90-110	04/28/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.490	0.5	98	90-110	04/28/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.490	0.5	98	90-110	04/28/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.502	0.5	100	90-110	04/28/2025
Sample ID: CCV4 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	04/28/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q1889

Project: Mitchell School

RunNo.: LB135612

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	0.92	1	92	90-110	04/30/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.94	1	94	90-110	04/30/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	04/30/2025
Sample ID: CCV3 Ammonia as N	mg/L	0.97	1	97	90-110	04/30/2025

Initial and Continuing Calibration Verification

Client: Kleinfelder

SDG No.: Q1889

Project: Mitchell School

RunNo.: LB135612

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------