

Duplicate Sample Summary

Client:	Kleinfelder	SDG No.:	Q1889
Project:	Mitchell School	Sample ID:	Q1878-01
Client ID:	TR-4-042425DUP	Percent Solids for Spike Sample:	98.9

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Hexavalent Chromium	mg/Kg	+/-20	0.070	U	0.070	U	1	0		04/28/2025

Duplicate Sample Summary

Client: Kleinfelder	SDG No.: Q1889
Project: Mitchell School	Sample ID: Q1889-01
Client ID: COMP-1DUP	Percent Solids for Spike Sample: 82.5

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/Kg	+/-20	5.50	J	5.60	J	1	2		04/30/2025

Duplicate Sample Summary

Client: Kleinfelder Project: Mitchell School Client ID: COMP-1MSD	SDG No.: Q1889 Sample ID: Q1889-01 Percent Solids for Spike Sample: 82.5
--	---

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
Ammonia as N	mg/Kg	+/-20	63.1		65.5		1	4		04/30/2025

Duplicate Sample Summary

Client: Kleinfelder	SDG No.: Q1889
Project: Mitchell School	Sample ID: Q1889-03
Client ID: COMP-3MSD	Percent Solids for Spike Sample: 80.7

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/AD	Qual	Analysis Date
Bromide	mg/Kg	+/-15	252		252		1	0		04/28/2025
Chloride	mg/Kg	+/-15	79.3		79.3		1	0		04/28/2025
Nitrate	mg/Kg	+/-15	63.3		63.3		1	0		04/28/2025
Nitrite	mg/Kg	+/-15	75.1		75.2		1	0		04/28/2025
Orthophosphate as P	mg/Kg	+/-15	98.8		99.1		1	0		04/28/2025
Sulfate	mg/Kg	+/-15	393		394		1	0		04/28/2025
Fluoride	mg/Kg	+/-15	53.2		53.7		1	1		04/28/2025