

Laboratory Control Sample Summary

Client: Kleinfelder
Project: Mitchell School

SDG No.: Q1889
Run No.: LB135573

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB135573BSS							
Bromide	mg/Kg	200	211		106	1	90-110	04/28/2025
Chloride	mg/Kg	60	63.6		106	1	90-110	04/28/2025
Fluoride	mg/Kg	40	41.5		104	1	90-110	04/28/2025
Nitrite	mg/Kg	60	63.2		105	1	90-110	04/28/2025
Nitrate	mg/Kg	50	53.6		107	1	90-110	04/28/2025
Sulfate	mg/Kg	300	308		103	1	90-110	04/28/2025
Orthophosphate as P	mg/Kg	100	105		105	1	90-110	04/28/2025

Laboratory Control Sample Summary

Client: Kleinfelder
Project: Mitchell School

SDG No.: Q1889
Run No.: LB135578

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167730BS							
Hexavalent Chromium	mg/Kg	20	20.0		100	1	84-110	04/28/2025

Laboratory Control Sample Summary

Client: Kleinfelder
Project: Mitchell School

SDG No.: Q1889
Run No.: LB135612

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB167793BS							
Ammonia as N	mg/Kg	50	47.5		95	1	90-110	04/30/2025