

RETENTION TIMES OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Instrument ID: ECD_P Calibration Date(s): 04/22/2025 04/22/2025

Calibration Times: 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID: RT 1000 = PP071389.D RT 750 = PP071390.D

RT 500 = PP071391.D RT 250 = PP071392.D RT 050 = PP071393.D

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	5.66	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1016-2	(2)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Aroclor-1016-3	(3)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Aroclor-1016-4	(4)	5.85	5.85	5.85	5.85	5.85	5.85	5.75	5.95
Aroclor-1016-5	(5)	6.14	6.14	6.15	6.14	6.14	6.14	6.04	6.24
Aroclor-1260-1	(1)	7.26	7.26	7.27	7.26	7.26	7.26	7.16	7.36
Aroclor-1260-2	(2)	7.51	7.51	7.52	7.52	7.52	7.52	7.42	7.62
Aroclor-1260-3	(3)	7.87	7.87	7.88	7.87	7.88	7.87	7.77	7.97
Aroclor-1260-4	(4)	8.10	8.10	8.10	8.10	8.10	8.10	8.00	8.20
Aroclor-1260-5	(5)	8.42	8.42	8.42	8.42	8.42	8.42	8.32	8.52
Decachlorobiphenyl		10.23	10.23	10.24	10.23	10.24	10.23	10.13	10.33
Tetrachloro-m-xylene		4.51	4.51	4.52	4.51	4.52	4.51	4.41	4.61
Aroclor-1254-1	(1)	6.52	6.52	6.52	6.52	6.52	6.52	6.42	6.62
Aroclor-1254-2	(2)	6.74	6.74	6.74	6.73	6.74	6.74	6.64	6.84
Aroclor-1254-3	(3)	7.10	7.10	7.10	7.10	7.10	7.10	7.00	7.20
Aroclor-1254-4	(4)	7.38	7.38	7.38	7.38	7.38	7.38	7.28	7.48
Aroclor-1254-5	(5)	7.80	7.80	7.80	7.80	7.80	7.80	7.70	7.90
Decachlorobiphenyl		10.24	10.24	10.24	10.23	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.52	4.52	4.52	4.51	4.51	4.51	4.41	4.61



Contract:	<u>POWE02</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1889</u>	SAS No.:	<u>Q1889</u>	SDG NO.:	<u>Q1889</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):		<u>04/22/2025</u>	<u>04/22/2025</u>		
		Calibration Times:		<u>10:29</u>	<u>17:49</u>		

LAB FILE ID:	RT 1000 =	<u>PP071389.D</u>	RT 750 =	<u>PP071390.D</u>
RT 500 =	PP071391.D	RT 250 =	PP071392.D	RT 050 =
				PP071393.D

[illegible]