

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q1889 **SAS No.:** Q1889 **SDG NO.:** Q1889

Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP071389.D</u>		CF 750 = <u>PP071390.D</u>			
CF 500 = <u>PP071391.D</u>		CF 250 = <u>PP071392.D</u>		CF 050 = <u>PP071393.D</u>			
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	% RSD
Aroclor-1016-1	(1)	64709894	67652695	70083532	71128656	61674340	67049823
Aroclor-1016-2	(2)	98977911	102364897	107369074	108813892	95615840	102628323
Aroclor-1016-3	(3)	59905708	62127143	64296308	65849928	56056440	61647105
Aroclor-1016-4	(4)	50563747	52419789	53619076	54941428	44946700	51298148
Aroclor-1016-5	(5)	47772963	49344036	50670668	50888564	42515680	48238382
Aroclor-1260-1	(1)	90879847	94857525	98814986	99659744	95910940	96024608
Aroclor-1260-2	(2)	137163459	142386411	147498372	149887700	138450980	143077384
Aroclor-1260-3	(3)	110350751	114445481	117702804	126515524	91505860	112104084
Aroclor-1260-4	(4)	105570732	110780085	111211068	131260612	105501660	112864831
Aroclor-1260-5	(5)	220536830	229070229	228919678	255108180	199231240	226573231
Decachlorobiphenyl		1409816860	1467002307	1522583160	1539948480	1300700600	1448010281
Tetrachloro-m-xylene		1947183160	2005072840	2066603680	2100352680	1793966200	1982635712
Aroclor-1254-1	(1)	78574372	81246615	85577334	88598188	75759400	81951182
Aroclor-1254-2	(2)	121624435	125816099	132146328	137115708	118678260	127076166
Aroclor-1254-3	(3)	123432630	127799839	132981654	137623252	117036460	127774767
Aroclor-1254-4	(4)	112512019	116259037	121737436	124392856	112665620	117513394
Aroclor-1254-5	(5)	106399089	105922557	113785120	116153044	96391680	107730298
Decachlorobiphenyl		1410101810	1469202520	1526389740	1549706840	1328725400	1456825262
Tetrachloro-m-xylene		1888241120	1936990333	2012368600	2071565920	1722515200	1926336235

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM **Case No.:** Q1889 **SAS No.:** Q1889 **SDG NO.:** Q1889

Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PP071389.D	CF 750 =	PP071390.D			
CF 500 =		PP071391.D	CF 250 =	PP071392.D	CF 050 =	PP071393.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	48121828	49414308	53806128	56180596	60145020	53533576	9
Aroclor-1016-2	(2)	69736709	72096436	77213744	78520924	84250740	76363711	7
Aroclor-1016-3	(3)	38517048	40455577	43831736	45082616	43165080	42210411	6
Aroclor-1016-4	(4)	31206636	32814844	36123020	37095092	35190100	34485938	7
Aroclor-1016-5	(5)	39458687	41985771	45850608	46345748	48595640	44447291	8
Aroclor-1260-1	(1)	63219917	66123699	73727666	76350368	82074520	72299234	11
Aroclor-1260-2	(2)	77338708	77548395	89921718	91492956	101005300	87461415	12
Aroclor-1260-3	(3)	72062678	70944365	81698234	79344004	87888600	78387576	9
Aroclor-1260-4	(4)	57333188	56779915	65331750	68083052	68025140	63110609	9
Aroclor-1260-5	(5)	143809415	139161576	155842678	157512280	157560640	150777318	6
Decachlorobiphenyl		831147600	875337973	902319620	934998560	841460400	877052831	5
Tetrachloro-m-xylene		1380433570	1378340613	1415112020	1448341640	1420974200	1408640409	2
Aroclor-1254-1	(1)	81690839	87361264	89755404	97285536	100908960	91400401	8
Aroclor-1254-2	(2)	70045476	75469849	77317914	83154376	87573280	78712179	9
Aroclor-1254-3	(3)	109725744	114103376	120096818	124380164	124235440	118508308	5
Aroclor-1254-4	(4)	70169900	74068720	77810162	81877428	81901880	77165618	7
Aroclor-1254-5	(5)	92229111	96109599	104690794	107384400	106924600	101467701	7
Decachlorobiphenyl		846307500	875335867	916104900	955261880	868223800	892246789	5
Tetrachloro-m-xylene		1390553300	1501013893	1450901500	1518608160	1468503000	1465915971	3