

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** POWE02

**Lab Code:** CHEM      **Case No.:** Q1889      **SAS No.:** Q1889      **SDG NO.:** Q1889

**Continuing Calib Date:** 04/29/2025      **Initial Calibration Date(s):** 04/22/2025      04/22/2025

**Continuing Calib Time:** 05:31      **Initial Calibration Time(s):** 10:29      17:49

**GC Column:** ZB-MR1      **ID:** 0.32      (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Tetrachloro-m-xylene | 4.51       | 4.51      | 4.41      | 4.61  | 0.00       |
| Decachlorobiphenyl   | 10.23      | 10.24     | 10.14     | 10.34 | 0.01       |

**CALIBRATION VERIFICATION SUMMARY**

**Contract:** POWE02

**Lab Code:** CHEM      **Case No.:** Q1889      **SAS No.:** Q1889      **SDG NO.:** Q1889

**Continuing Calib Date:** 04/29/2025      **Initial Calibration Date(s):** 04/22/2025      04/22/2025

**Continuing Calib Time:** 05:31      **Initial Calibration Time(s):** 10:29      17:49

**GC Column:** ZB-MR2      **ID:** 0.32      (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Tetrachloro-m-xylene | 3.81       | 3.81      | 3.71      | 3.91 | 0.00       |
| Decachlorobiphenyl   | 8.84       | 8.85      | 8.75      | 8.95 | 0.01       |