

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP071557.D Time Analyzed: 18:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.665	5.571	5.771	479.350	500.000	-4.1
Aroclor-1016-2	5.686	5.593	5.793	482.270	500.000	-3.5
Aroclor-1016-3	5.748	5.654	5.854	479.730	500.000	-4.1
Aroclor-1016-4	5.846	5.752	5.952	487.050	500.000	-2.6
Aroclor-1016-5	6.139	6.045	6.245	473.110	500.000	-5.4
Aroclor-1260-1	7.257	7.165	7.365	468.870	500.000	-6.2
Aroclor-1260-2	7.511	7.418	7.618	472.380	500.000	-5.5
Aroclor-1260-3	7.869	7.777	7.977	486.210	500.000	-2.8
Aroclor-1260-4	8.094	8.002	8.202	464.380	500.000	-7.1
Aroclor-1260-5	8.414	8.322	8.522	483.210	500.000	-3.4
Decachlorobiphenyl	10.226	10.137	10.337	49.680	50.000	-0.6
Tetrachloro-m-xylene	4.512	4.417	4.617	48.490	50.000	-3.0

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP071557.D Time Analyzed: 18:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.895	4.798	4.998	508.180	500.000	1.6
Aroclor-1016-2	4.913	4.816	5.016	513.170	500.000	2.6
Aroclor-1016-3	5.090	4.993	5.193	522.490	500.000	4.5
Aroclor-1016-4	5.131	5.036	5.236	523.600	500.000	4.7
Aroclor-1016-5	5.346	5.250	5.450	524.280	500.000	4.9
Aroclor-1260-1	6.380	6.285	6.485	500.820	500.000	0.2
Aroclor-1260-2	6.569	6.473	6.673	499.190	500.000	-0.2
Aroclor-1260-3	6.721	6.627	6.827	511.360	500.000	2.3
Aroclor-1260-4	7.193	7.098	7.298	524.290	500.000	4.9
Aroclor-1260-5	7.434	7.339	7.539	517.900	500.000	3.6
Decachlorobiphenyl	8.842	8.749	8.949	50.210	50.000	0.4
Tetrachloro-m-xylene	3.810	3.712	3.912	47.590	50.000	-4.8