

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071587.D Time Analyzed: 04:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.661	5.571	5.771	499.460	500.000	-0.1
Aroclor-1016-2	5.683	5.593	5.793	497.330	500.000	-0.5
Aroclor-1016-3	5.745	5.654	5.854	491.000	500.000	-1.8
Aroclor-1016-4	5.842	5.752	5.952	494.630	500.000	-1.1
Aroclor-1016-5	6.135	6.045	6.245	504.640	500.000	0.9
Aroclor-1260-1	7.254	7.165	7.365	524.920	500.000	5.0
Aroclor-1260-2	7.507	7.418	7.618	503.170	500.000	0.6
Aroclor-1260-3	7.866	7.777	7.977	514.880	500.000	3.0
Aroclor-1260-4	8.090	8.002	8.202	502.310	500.000	0.5
Aroclor-1260-5	8.409	8.322	8.522	515.560	500.000	3.1
Decachlorobiphenyl	10.222	10.137	10.337	52.270	50.000	4.5
Tetrachloro-m-xylene	4.508	4.417	4.617	49.750	50.000	-0.5

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071587.D Time Analyzed: 04:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.892	4.798	4.998	520.170	500.000	4.0
Aroclor-1016-2	4.911	4.816	5.016	523.850	500.000	4.8
Aroclor-1016-3	5.088	4.993	5.193	534.650	500.000	6.9
Aroclor-1016-4	5.130	5.036	5.236	526.290	500.000	5.3
Aroclor-1016-5	5.344	5.250	5.450	567.310	500.000	13.5
Aroclor-1260-1	6.378	6.285	6.485	527.630	500.000	5.5
Aroclor-1260-2	6.567	6.473	6.673	513.280	500.000	2.7
Aroclor-1260-3	6.719	6.627	6.827	518.650	500.000	3.7
Aroclor-1260-4	7.190	7.098	7.298	511.600	500.000	2.3
Aroclor-1260-5	7.432	7.339	7.539	517.450	500.000	3.5
Decachlorobiphenyl	8.840	8.749	8.949	51.260	50.000	2.5
Tetrachloro-m-xylene	3.808	3.712	3.912	50.380	50.000	0.8