

Cover Page

Order ID : Q1889

Project ID : Mitchell School

Client : Kleinfelder

Lab Sample Number

Q1889-01
Q1889-02
Q1889-03

Client Sample Number

COMP-1
COMP-2
COMP-3

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 5/2/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Kleinfelder

Project Name: Mitchell School

Project # N/A

Chemtech Project # Q1889

Test Name: PCB Group1

A. Number of Samples and Date of Receipt:

3 Solid samples were received on 04/25/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: Ammonia, Anions Group1, Hexavalent Chromium, Mercury, Metals Group1, Metals ICP-Group1, PCB Group1, PESTICIDE Group1, SVOCMS Group1, Trivalent Chromium and VOCMS Group1. This data package contains results for PCB Group1.

C. Analytical Techniques:

The analyses were performed on instrument GCECD_P. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11. The analyses were performed on instrument GCECD_O. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df, Catalogue # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 µm; Catalogue # 7HM-G017-11. The analysis of PCB Group1s was based on method 8082A and extraction was done based on method 3541.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria.

The Retention Times were acceptable for all samples.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the acceptable requirements.

The RPD met criteria.

The Blank Spike met requirements for all samples.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

E. Additional Comments:

The soil samples results are based on a dry weight basis.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1889

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: PRATIK PATEL

Date: 05/02/2025

LAB CHRONICLE

OrderID: Q1889	OrderDate: 4/25/2025 11:06:00 AM
Client: Kleinfelder	Project: Mitchell School
Contact: Mark Warchol	Location: L51,VOA Ref. #2 Soil

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1889-01	COMP-1	SOIL			04/24/25			04/25/25
			PCB Group1	8082A		04/28/25	04/29/25	
			PESTICIDE Group1	8081B		04/28/25	04/28/25	
Q1889-02	COMP-2	SOIL			04/24/25			04/25/25
			PCB Group1	8082A		04/28/25	04/28/25	
			PESTICIDE Group1	8081B		04/28/25	04/28/25	
Q1889-03	COMP-3	SOIL			04/24/25			04/25/25
			PCB Group1	8082A		04/28/25	04/29/25	
			PESTICIDE Group1	8081B		04/28/25	04/28/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1889

Order ID: Q1889

Client: Kleinfelder

Project ID: Mitchell School

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
-----------	-----------	--------	-----------	---------------	---	-----	-----	-------

Client ID :

Total Concentration: 0.000



QC SUMMARY

Surrogate Summary

SDG No.: Q1889

Client: Kleinfelder

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO110348.D	PIBLK-PO110348.D	Tetrachloro-m-xylene	1	20	18.1	91		60	140
		Decachlorobiphenyl	1	20	19.2	96		60	140
		Tetrachloro-m-xylene	2	20	18.2	91		60	140
		Decachlorobiphenyl	2	20	20.2	101		60	140
I.BLK-PO110850.D	PIBLK-PO110850.D	Tetrachloro-m-xylene	1	20	18.0	90		60	140
		Decachlorobiphenyl	1	20	16.9	85		60	140
		Tetrachloro-m-xylene	2	20	16.6	83		60	140
		Decachlorobiphenyl	2	20	16.7	84		60	140
Q1889-01	COMP-1	Tetrachloro-m-xylene	1	20	21.5	108		32	144
		Decachlorobiphenyl	1	20	16.8	84		32	175
		Tetrachloro-m-xylene	2	20	19.7	99		32	144
		Decachlorobiphenyl	2	20	16.5	83		32	175
Q1889-01MS	COMP-1MS	Tetrachloro-m-xylene	1	20	19.4	97		32	144
		Decachlorobiphenyl	1	20	16.4	82		32	175
		Tetrachloro-m-xylene	2	20	17.8	89		32	144
		Decachlorobiphenyl	2	20	16.1	81		32	175
Q1889-01MSD	COMP-1MSD	Tetrachloro-m-xylene	1	20	18.9	95		32	144
		Decachlorobiphenyl	1	20	16.6	83		32	175
		Tetrachloro-m-xylene	2	20	17.2	86		32	144
		Decachlorobiphenyl	2	20	16.3	82		32	175
I.BLK-PO110865.D	PIBLK-PO110865.D	Tetrachloro-m-xylene	1	20	18.4	92		60	140
		Decachlorobiphenyl	1	20	17.5	88		60	140
		Tetrachloro-m-xylene	2	20	16.9	84		60	140
		Decachlorobiphenyl	2	20	16.9	85		60	140
I.BLK-PP071388.D	PIBLK-PP071388.D	Tetrachloro-m-xylene	1	20	17.1	85		60	140
		Decachlorobiphenyl	1	20	17.6	88		60	140
		Tetrachloro-m-xylene	2	20	17.2	86		60	140
		Decachlorobiphenyl	2	20	17.6	88		60	140
I.BLK-PP071561.D	PIBLK-PP071561.D	Tetrachloro-m-xylene	1	20	15.5	78		60	140
		Decachlorobiphenyl	1	20	16.3	81		60	140
		Tetrachloro-m-xylene	2	20	16.4	82		60	140
		Decachlorobiphenyl	2	20	17.4	87		60	140
PB167765BL	PB167765BL	Tetrachloro-m-xylene	1	20	18.4	92		32	144
		Decachlorobiphenyl	1	20	18.9	95		32	175
		Tetrachloro-m-xylene	2	20	20.5	103		32	144
		Decachlorobiphenyl	2	20	19.9	100		32	175
PB167765BS	PB167765BS	Tetrachloro-m-xylene	1	20	18.1	91		32	144
		Decachlorobiphenyl	1	20	18.7	94		32	175
		Tetrachloro-m-xylene	2	20	17.7	89		32	144
		Decachlorobiphenyl	2	20	19.8	99		32	175
Q1889-02	COMP-2	Tetrachloro-m-xylene	1	20	19.8	99		32	144

Surrogate Summary

SDG No.: Q1889

Client: Kleinfelder

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1889-02	COMP-2	Decachlorobiphenyl	1	20	17.9	90		32	175
		Tetrachloro-m-xylene	2	20	22.5	113		32	144
I.BLK-PP071576.D	PIBLK-PP071576.D	Decachlorobiphenyl	2	20	18.8	94		32	175
		Tetrachloro-m-xylene	1	20	15.8	79		60	140
		Decachlorobiphenyl	1	20	16.4	82		60	140
		Tetrachloro-m-xylene	2	20	16.9	84		60	140
		Decachlorobiphenyl	2	20	17.5	87		60	140
Q1889-03	COMP-3	Tetrachloro-m-xylene	1	20	19.6	98		32	144
		Decachlorobiphenyl	1	20	19.0	95		32	175
		Tetrachloro-m-xylene	2	20	21.8	109		32	144
		Decachlorobiphenyl	2	20	19.8	99		32	175
I.BLK-PP071591.D	PIBLK-PP071591.D	Tetrachloro-m-xylene	1	20	16.1	80		60	140
		Decachlorobiphenyl	1	20	16.7	83		60	140
		Tetrachloro-m-xylene	2	20	16.8	84		60	140
		Decachlorobiphenyl	2	20	17.9	89		60	140

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1889

Client: Kleinfelder

Analytical Method: 8082A

DataFile : PO110853.D

Lab Sample ID:	Parameter	Spike	Sample		Units	Rec	Rec		RPD		Limits	RPD
			Result	Result			Qual	RPD	Qual	Low	High	
Client Sample ID: Q1889-01MS	COMP-IMS											
	AR1016	201.6	0	189	ug/kg	94				55	146	
	AR1260	201.6	0	169	ug/kg	84				31	146	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1889

Client: Kleinfelder

Analytical Method: 8082A

DataFile : PO110854.D

Lab Sample ID:	Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Client Sample ID: Q1889-01MSD	COMP-1MSD											
	AR1016	201.8	0	180	ug/kg	89		5		55	146	20
	AR1260	201.8	0	158	ug/kg	78		7		31	146	20



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1889

Client: Kleinfelder

Analytical Method: **8082A** **Datafile :** PP071565.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB167765BS	AR1016	166.5	143	ug/kg	86				71	120	
	AR1260	166.5	147	ug/kg	88				65	130	

4C

PESTICIDE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB167765BL

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM Case No.: Q1889

SAS No.: Q1889 SDG NO.: Q1889

Lab Sample ID: PB167765BL

Lab File ID: PP071564.D

Matrix: (soil/water) Solid

Extraction: (Type) SOXH

Sulfur Cleanup: (Y/N) N

Date Extracted: 04/28/2025

Date Analyzed (1): 04/28/2025

Date Analyzed (2): 04/28/2025

Time Analyzed (1): 20:15

Time Analyzed (2): 20:15

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column (1): ZB-MR1 ID: 0.32 (mm)

GC Column (2): ZB-MR2 ID: 0.32 (mm)

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED 1	DATE ANALYZED 2
COMP-1	Q1889-01	PO110852.D	04/29/2025	04/29/2025
COMP-1MS	Q1889-01MS	PO110853.D	04/29/2025	04/29/2025
COMP-1MSD	Q1889-01MSD	PO110854.D	04/29/2025	04/29/2025
PB167765BS	PB167765BS	PP071565.D	04/28/2025	04/28/2025
COMP-2	Q1889-02	PP071571.D	04/28/2025	04/28/2025
COMP-3	Q1889-03	PP071577.D	04/29/2025	04/29/2025

COMMENTS: _____



SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/24/25
Project:	Mitchell School	Date Received:	04/25/25
Client Sample ID:	COMP-1	SDG No.:	Q1889
Lab Sample ID:	Q1889-01	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	82.5 Decanted:
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB Group1
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110852.D	1	04/28/25 09:05	04/29/25 11:16	PB167765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.80	U	4.80	20.6	ug/kg
11097-69-1	Aroclor-1254	3.90	U	3.90	20.6	ug/kg
11096-82-5	Aroclor-1260	3.90	U	3.90	20.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.5		32 - 144	108%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.8		32 - 175	84%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042925\
 Data File : PO110852.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2025 11:16
 Operator : YP/AJ
 Sample : Q1889-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 COMP-1

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 02:23:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 11 02:12:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.685	3.682	188.5E6	98486312	21.543	19.738
2) SA Decachlor...	8.725	8.676	132.3E6	31825906	16.753	16.533

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042925\
Data File : PO110852.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2025 11:16
Operator : YP/AJ
Sample : Q1889-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
COMP-1

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 02:23:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 02:12:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.685 min
Delta R.T.: -0.003 min
Response: 188484157
Conc: 21.54 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1




#1 Tetrachloro-m-xylene

R.T.: 3.682 min
Delta R.T.: -0.004 min
Response: 98486312
Conc: 19.74 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.725 min
Delta R.T.: -0.007 min
Response: 132263417
Conc: 16.75 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.676 min
Delta R.T.: -0.008 min
Response: 31825906
Conc: 16.53 ng/ml

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/24/25
Project:	Mitchell School	Date Received:	04/25/25
Client Sample ID:	COMP-2	SDG No.:	Q1889
Lab Sample ID:	Q1889-02	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	80.5 Decanted:
Sample Wt/Vol:	30.05 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB Group1
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071571.D	1	04/28/25 09:05	04/28/25 22:09	PB167765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	21.1	ug/kg
11097-69-1	Aroclor-1254	4.00	U	4.00	21.1	ug/kg
11096-82-5	Aroclor-1260	4.00	U	4.00	21.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	22.5		32 - 144	113%	SPK: 20
2051-24-3	Decachlorobiphenyl	18.8		32 - 175	94%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
 Data File : PP071571.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Apr 2025 22:09
 Operator : YP\AJ
 Sample : Q1889-02
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 COMP-2

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 01:14:02 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.809	39178643	31692469	19.761	22.499
2) SA Decachlor...	10.228	8.840	25940976	16521456	17.915	18.837

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
Data File : PP071571.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2025 22:09
Operator : YP\AJ
Sample : Q1889-02
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
COMP-2

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 29 01:14:02 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: -0.004 min
Response: 39178643
Conc: 19.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
COMP-2




#1 Tetrachloro-m-xylene

R.T.: 3.809 min
Delta R.T.: -0.003 min
Response: 31692469
Conc: 22.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.228 min
Delta R.T.: -0.010 min
Response: 25940976
Conc: 17.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.840 min
Delta R.T.: -0.009 min
Response: 16521456
Conc: 18.84 ng/ml

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/24/25
Project:	Mitchell School	Date Received:	04/25/25
Client Sample ID:	COMP-3	SDG No.:	Q1889
Lab Sample ID:	Q1889-03	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	80.7
Sample Wt/Vol:	30.02	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071577.D	1	04/28/25 09:05	04/29/25 01:09	PB167765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	4.90	U	4.90	21.1	ug/kg
11097-69-1	Aroclor-1254	4.00	U	4.00	21.1	ug/kg
11096-82-5	Aroclor-1260	4.00	U	4.00	21.1	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	21.8		32 - 144	109%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.8		32 - 175	99%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
 Data File : PP071577.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2025 01:09
 Operator : YP\AJ
 Sample : Q1889-03
 Misc :
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 COMP-3

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 06:28:56 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.510	3.808	38952618	30730616	19.647	21.816
2) SA Decachlor...	10.225	8.839	27532884	17382152	19.014	19.819

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
Data File : PP071577.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2025 01:09
Operator : YP\AJ
Sample : Q1889-03
Misc :
ALS Vial : 36 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
COMP-3

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 29 06:28:56 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.510 min
Delta R.T.: -0.007 min
Response: 38952618
Conc: 19.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
COMP-3




#1 Tetrachloro-m-xylene

R.T.: 3.808 min
Delta R.T.: -0.004 min
Response: 30730616
Conc: 21.82 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.225 min
Delta R.T.: -0.013 min
Response: 27532884
Conc: 19.01 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.839 min
Delta R.T.: -0.010 min
Response: 17382152
Conc: 19.82 ng/ml



CALIBRATION SUMMARY

GC Column: ZB-MR1 ID: 0.32 (mm)

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: POWE02
Lab Code: CHEM **Case No.:** Q1889 **SAS No.:** Q1889 **SDG NO.:** Q1889
Instrument ID: ECD_O **Calibration Date(s):** 04/10/2025 04/10/2025
Calibration Times: 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PO110349.D</u>	RT 750 = <u>PO110350.D</u>
RT 500 = <u>PO110351.D</u>	RT 250 = <u>PO110352.D</u>	RT 050 = <u>PO110353.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	4.77	4.77	4.77	4.77	4.77	4.77	4.67	4.87
Aroclor-1016-2	(2)	4.78	4.79	4.78	4.78	4.79	4.78	4.68	4.88
Aroclor-1016-3	(3)	4.96	4.96	4.96	4.96	4.96	4.96	4.86	5.06
Aroclor-1016-4	(4)	5.00	5.00	5.00	5.00	5.00	5.00	4.90	5.10
Aroclor-1016-5	(5)	5.22	5.22	5.22	5.21	5.22	5.22	5.12	5.32
Aroclor-1260-1	(1)	6.25	6.25	6.25	6.25	6.25	6.25	6.15	6.35
Aroclor-1260-2	(2)	6.43	6.43	6.43	6.43	6.43	6.43	6.33	6.53
Aroclor-1260-3	(3)	6.59	6.59	6.59	6.59	6.59	6.59	6.49	6.69
Aroclor-1260-4	(4)	7.06	7.06	7.06	7.06	7.06	7.06	6.96	7.16
Aroclor-1260-5	(5)	7.30	7.30	7.30	7.30	7.30	7.30	7.20	7.40
Decachlorobiphenyl		8.69	8.69	8.68	8.68	8.68	8.68	8.58	8.78
Tetrachloro-m-xylene		3.69	3.69	3.69	3.69	3.69	3.69	3.59	3.79
Aroclor-1254-1	(1)	5.57	5.57	5.57	5.57	5.57	5.57	5.47	5.67
Aroclor-1254-2	(2)	5.71	5.71	5.71	5.71	5.71	5.71	5.61	5.81
Aroclor-1254-3	(3)	6.12	6.12	6.12	6.12	6.12	6.12	6.02	6.22
Aroclor-1254-4	(4)	6.34	6.34	6.34	6.34	6.34	6.34	6.24	6.44
Aroclor-1254-5	(5)	6.76	6.76	6.76	6.76	6.76	6.76	6.66	6.86
Decachlorobiphenyl		8.68	8.68	8.68	8.68	8.68	8.68	8.58	8.78
Tetrachloro-m-xylene		3.69	3.69	3.69	3.68	3.69	3.69	3.59	3.79



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Instrument ID: ECD_O

Calibration Date(s): 04/10/2025 04/10/2025

Calibration Times: 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:		CF 1000 =	PO110349.D	CF 750 =	PO110350.D			
CF 500 =		PO110351.D	CF 250 =	PO110352.D	CF 050 =	PO110353.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	301539199	312147691	325366094	341425388	360239300	328143534	7
Aroclor-1016-2	(2)	424202495	441346851	452587054	472466840	484138100	454948268	5
Aroclor-1016-3	(3)	286857785	300177247	315792572	336640016	373170260	322527576	10
Aroclor-1016-4	(4)	227624644	238137328	247842320	261218376	270999820	249164498	7
Aroclor-1016-5	(5)	238190199	250655776	262830422	281583724	311798080	269011640	11
Aroclor-1260-1	(1)	429982139	447081743	462005040	490865312	532292280	472445303	9
Aroclor-1260-2	(2)	524274151	548276752	560001580	596374412	703195060	586424391	12
Aroclor-1260-3	(3)	446012749	466468717	482076654	511871688	557932460	492872454	9
Aroclor-1260-4	(4)	382143422	400404183	417623168	445374396	476569960	424423026	9
Aroclor-1260-5	(5)	1002895230	1031868237	1046125762	1065122080	1071346720	1043471606	3
Decachlorobiphenyl		7251274760	7526590267	7768647820	8211713920	8716546400	7894954633	7
Tetrachloro-m-xylene		8603197440	8796106920	8951138380	8737007280	8657843200	8749058644	2
Aroclor-1254-1	(1)	517202507	536393799	555537214	581626948	599185020	557989098	6
Aroclor-1254-2	(2)	450798901	468206104	477861830	506770124	524875880	485702568	6
Aroclor-1254-3	(3)	746169545	769125461	779743060	811121676	809948840	783221716	4
Aroclor-1254-4	(4)	460149954	473941677	490156778	505394424	508383480	487605263	4
Aroclor-1254-5	(5)	657635451	678198725	698380664	724566284	745718260	700899877	5
Decachlorobiphenyl		7612990370	7886723680	8142983840	8479059160	8719027400	8168156890	5
Tetrachloro-m-xylene		8665092200	8825966600	8709112240	8974669960	8357494600	8706467120	3

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02
Lab Code: CHEM **Case No.:** Q1889 **SAS No.:** Q1889 **SDG NO.:** Q1889
Instrument ID: ECD_O **Calibration Date(s):** 04/10/2025 04/10/2025
Calibration Times: 09:36 17:52

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PO110349.D	CF 750 =	PO110350.D			
CF 500 =		PO110351.D	CF 250 =	PO110352.D	CF 050 =	PO110353.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	159213491	166153001	173161426	184455932	194563640	175509498	8
Aroclor-1016-2	(2)	235577328	244073360	249915638	259769828	267988640	251464959	5
Aroclor-1016-3	(3)	123754519	129314112	134837594	142342816	148305560	135710920	7
Aroclor-1016-4	(4)	100154968	106815488	112688448	121621132	130559280	114367863	11
Aroclor-1016-5	(5)	131736277	138471012	145266388	155239544	174656140	149073872	11
Aroclor-1260-1	(1)	221282381	231306280	240804372	256804492	279620260	245963557	9
Aroclor-1260-2	(2)	260045501	271771868	281572488	298549944	346814040	291750768	12
Aroclor-1260-3	(3)	243412958	253822376	261595986	277903536	317685740	270884119	11
Aroclor-1260-4	(4)	177247956	186943988	195746612	210913364	232348220	200640028	11
Aroclor-1260-5	(5)	430279301	445092429	452357666	469475868	497083160	458857685	6
Decachlorobiphenyl		1704536950	1791384040	1872045360	2025627840	2231357400	1924990318	11
Tetrachloro-m-xylene		4868987290	4975470093	5049711440	5093085920	4961663200	4989783589	2
Aroclor-1254-1	(1)	275528987	285381728	296386294	310788612	339409960	301499116	8
Aroclor-1254-2	(2)	239912515	249193828	258146844	274680708	301492780	264685335	9
Aroclor-1254-3	(3)	379979735	392440955	403528434	417422392	423708940	403416091	4
Aroclor-1254-4	(4)	215870098	222432417	229245508	238627400	242389020	229712889	5
Aroclor-1254-5	(5)	307722172	319481015	330728230	346867608	356659680	332291741	6
Decachlorobiphenyl		1672742580	1794748347	1901326780	2085791440	2132484800	1917418789	10
Tetrachloro-m-xylene		4865625930	4959437227	4872791400	5058078960	4761547200	4903496143	2

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110349.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 09:36
 Operator : YP/AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 10:56:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 10:56:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	860.3E6	486.9E6	98.018	98.178
2) SA Decachlor...	8.732	8.685	725.1E6	170.5E6	96.555	95.317
Target Compounds						
3) L1 AR-1016-1	4.779	4.765	301.5E6	159.2E6	961.993	958.036
4) L1 AR-1016-2	4.798	4.784	424.2E6	235.6E6	967.627	970.466
5) L1 AR-1016-3	4.855	4.960	286.9E6	123.8E6	951.987	957.141
6) L1 AR-1016-4	4.975	5.002	227.6E6	100.2E6	957.478	941.114
7) L1 AR-1016-5	5.232	5.215	238.2E6	131.7E6	950.820	951.155
31) L7 AR-1260-1	6.272	6.246	430.0E6	221.3E6	964.099	957.753
32) L7 AR-1260-2	6.461	6.434	524.3E6	260.0E6	967.049	960.254
33) L7 AR-1260-3	6.828	6.586	446.0E6	243.4E6	961.142	963.995
34) L7 AR-1260-4	7.089	7.058	382.1E6	177.2E6	955.637	950.405
35) L7 AR-1260-5	7.330	7.299	1002.9E6	430.3E6	978.902	974.986



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110349.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 09:36
Operator : YP/AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 10:56:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 10:56:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 860319744
Conc: 98.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 486898729
Conc: 98.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 725127476
Conc: 96.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.685 min
Delta R.T.: 0.000 min
Response: 170453695
Conc: 95.32 ng/ml



#3 AR-1016-1


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 301539199
Conc: 961.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 159213491
Conc: 958.04 ng/ml




#4 AR-1016-2

R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 424202495
Conc: 967.63 ng/ml



#4 AR-1016-2


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 235577328
Conc: 970.47 ng/ml



#5 AR-1016-3


R.T.: 4.855 min
Delta R.T.: 0.000 min
Response: 286857785
Conc: 951.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 4.960 min
Delta R.T.: 0.000 min
Response: 123754519
Conc: 957.14 ng/ml




#6 AR-1016-4

R.T.: 4.975 min
Delta R.T.: 0.000 min
Response: 227624644
Conc: 957.48 ng/ml



#6 AR-1016-4


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 100154968
Conc: 941.11 ng/ml



#7 AR-1016-5


R.T.: 5.232 min
Delta R.T.: 0.000 min
Response: 238190199
Conc: 950.82 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 5.215 min
Delta R.T.: 0.000 min
Response: 131736277
Conc: 951.16 ng/ml




#31 AR-1260-1

R.T.: 6.272 min
Delta R.T.: 0.000 min
Response: 429982139
Conc: 964.10 ng/ml



#31 AR-1260-1


R.T.: 6.246 min
Delta R.T.: 0.000 min
Response: 221282381
Conc: 957.75 ng/ml



#32 AR-1260-2


R.T.: 6.461 min
Delta R.T.: 0.000 min
Response: 524274151
Conc: 967.05 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 6.434 min
Delta R.T.: 0.000 min
Response: 260045501
Conc: 960.25 ng/ml




#33 AR-1260-3

R.T.: 6.828 min
Delta R.T.: 0.000 min
Response: 446012749
Conc: 961.14 ng/ml



#33 AR-1260-3


R.T.: 6.586 min
Delta R.T.: 0.000 min
Response: 243412958
Conc: 963.99 ng/ml



#34 AR-1260-4


R.T.: 7.089 min
Delta R.T.: 0.000 min
Response: 382143422
Conc: 955.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 7.058 min
Delta R.T.: 0.000 min
Response: 177247956
Conc: 950.41 ng/ml



#35 AR-1260-5

R.T.: 7.330 min
Delta R.T.: 0.000 min
Response: 1002895230
Conc: 978.90 ng/ml



#35 AR-1260-5

R.T.: 7.299 min
Delta R.T.: 0.000 min
Response: 430279301
Conc: 974.99 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110350.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 09:54
 Operator : YP/AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 10:59:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 10:56:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	659.7E6	373.2E6	75.108	75.162
2) SA Decachlor...	8.733	8.685	564.5E6	134.4E6	75.111	75.086
Target Compounds						
3) L1 AR-1016-1	4.779	4.765	234.1E6	124.6E6	747.916	749.896
4) L1 AR-1016-2	4.799	4.785	331.0E6	183.1E6	753.359	752.728
5) L1 AR-1016-3	4.855	4.960	225.1E6	96985584	748.093	750.070
6) L1 AR-1016-4	4.975	5.002	178.6E6	80111616	750.849	751.848
7) L1 AR-1016-5	5.233	5.215	188.0E6	103.9E6	750.290	749.891
31) L7 AR-1260-1	6.273	6.246	335.3E6	173.5E6	751.219	750.569
32) L7 AR-1260-2	6.462	6.434	411.2E6	203.8E6	755.640	751.776
33) L7 AR-1260-3	6.829	6.586	349.9E6	190.4E6	752.607	752.605
34) L7 AR-1260-4	7.090	7.058	300.3E6	140.2E6	750.651	751.197
35) L7 AR-1260-5	7.331	7.298	773.9E6	333.8E6	753.582	754.264



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110350.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 09:54
Operator : YP/AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 10:59:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 10:56:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 659708019
Conc: 75.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 373160257
Conc: 75.16 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 564494270
Conc: 75.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.685 min
Delta R.T.: 0.000 min
Response: 134353803
Conc: 75.09 ng/ml



#3 AR-1016-1


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 234110768
Conc: 747.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#3 AR-1016-1

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 124614751
Conc: 749.90 ng/ml




#4 AR-1016-2

R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 331010138
Conc: 753.36 ng/ml



#4 AR-1016-2


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 183055020
Conc: 752.73 ng/ml



#5 AR-1016-3


R.T.: 4.855 min
Delta R.T.: 0.000 min
Response: 225132935
Conc: 748.09 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#5 AR-1016-3

R.T.: 4.960 min
Delta R.T.: 0.000 min
Response: 96985584
Conc: 750.07 ng/ml




#6 AR-1016-4

R.T.: 4.975 min
Delta R.T.: 0.000 min
Response: 178602996
Conc: 750.85 ng/ml



#6 AR-1016-4


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 80111616
Conc: 751.85 ng/ml



#7 AR-1016-5


R.T.: 5.233 min
Delta R.T.: 0.000 min
Response: 187991832
Conc: 750.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#7 AR-1016-5

R.T.: 5.215 min
Delta R.T.: 0.000 min
Response: 103853259
Conc: 749.89 ng/ml




#31 AR-1260-1

R.T.: 6.273 min
Delta R.T.: 0.000 min
Response: 335311307
Conc: 751.22 ng/ml



#31 AR-1260-1


R.T.: 6.246 min
Delta R.T.: 0.000 min
Response: 173479710
Conc: 750.57 ng/ml



#32 AR-1260-2


R.T.: 6.462 min
Delta R.T.: 0.000 min
Response: 411207564
Conc: 755.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#32 AR-1260-2

R.T.: 6.434 min
Delta R.T.: 0.000 min
Response: 203828901
Conc: 751.78 ng/ml




#33 AR-1260-3

R.T.: 6.829 min
Delta R.T.: 0.000 min
Response: 349851538
Conc: 752.61 ng/ml



#33 AR-1260-3


R.T.: 6.586 min
Delta R.T.: 0.000 min
Response: 190366782
Conc: 752.61 ng/ml



#34 AR-1260-4


R.T.: 7.090 min
Delta R.T.: 0.000 min
Response: 300303137
Conc: 750.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC750




#34 AR-1260-4

R.T.: 7.058 min
Delta R.T.: 0.000 min
Response: 140207991
Conc: 751.20 ng/ml



#35 AR-1260-5

R.T.: 7.331 min
Delta R.T.: 0.000 min
Response: 773901178
Conc: 753.58 ng/ml



#35 AR-1260-5

R.T.: 7.298 min
Delta R.T.: 0.000 min
Response: 333819322
Conc: 754.26 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110351.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 10:13
 Operator : YP/AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 10:53:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 10:52:45 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.686	447.6E6	252.5E6	50.000	50.000
2) SA Decachlor...	8.732	8.684	388.4E6	93602268	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	4.779	4.766	162.7E6	86580713	500.000	500.000
4) L1 AR-1016-2	4.798	4.784	226.3E6	125.0E6	500.000	500.000
5) L1 AR-1016-3	4.855	4.960	157.9E6	67418797	500.000	500.000
6) L1 AR-1016-4	4.975	5.002	123.9E6	56344224	500.000	500.000
7) L1 AR-1016-5	5.232	5.215	131.4E6	72633194	500.000	500.000
31) L7 AR-1260-1	6.272	6.246	231.0E6	120.4E6	500.000	500.000
32) L7 AR-1260-2	6.460	6.434	280.0E6	140.8E6	500.000	500.000
33) L7 AR-1260-3	6.829	6.587	241.0E6	130.8E6	500.000	500.000
34) L7 AR-1260-4	7.088	7.057	208.8E6	97873306	500.000	500.000
35) L7 AR-1260-5	7.330	7.297	523.1E6	226.2E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110351.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 10:13
Operator : YP/AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 10:53:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 10:52:45 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 447556919
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.686 min
Delta R.T.: 0.000 min
Response: 252485572
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 388432391
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 93602268
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 162683047
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.766 min
Delta R.T.: 0.000 min
Response: 86580713
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 226293527
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 124957819
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 4.855 min
Delta R.T.: 0.000 min
Response: 157896286
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 4.960 min
Delta R.T.: 0.000 min
Response: 67418797
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 4.975 min
Delta R.T.: 0.000 min
Response: 123921160
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 56344224
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 5.232 min
Delta R.T.: 0.000 min
Response: 131415211
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.215 min
Delta R.T.: 0.000 min
Response: 72633194
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 6.272 min
Delta R.T.: 0.000 min
Response: 231002520
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.246 min
Delta R.T.: 0.000 min
Response: 120402186
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 6.460 min
Delta R.T.: 0.000 min
Response: 280000790
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.434 min
Delta R.T.: 0.000 min
Response: 140786244
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 6.829 min
Delta R.T.: 0.000 min
Response: 241038327
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.587 min
Delta R.T.: 0.000 min
Response: 130797993
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 7.088 min
Delta R.T.: 0.000 min
Response: 208811584
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 7.057 min
Delta R.T.: 0.000 min
Response: 97873306
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.330 min
Delta R.T.: 0.000 min
Response: 523062881
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.297 min
Delta R.T.: 0.000 min
Response: 226178833
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110352.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 10:31
 Operator : YP/AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 11:01:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 10:56:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	218.4E6	127.3E6	24.901	25.482
2) SA Decachlor...	8.734	8.683	205.3E6	50640696	26.698	27.397
Target Compounds						
3) L1 AR-1016-1	4.780	4.765	85356347	46113983	266.639	270.074
4) L1 AR-1016-2	4.798	4.784	118.1E6	64942457	263.859	262.570
5) L1 AR-1016-3	4.855	4.959	84160004	35585704	271.600	268.445
6) L1 AR-1016-4	4.975	5.002	65304594	30405283	267.965	275.610
7) L1 AR-1016-5	5.232	5.214	70395931	38809886	272.520	272.010
31) L7 AR-1260-1	6.272	6.245	122.7E6	64201123	268.242	270.264
32) L7 AR-1260-2	6.461	6.433	149.1E6	74637486	267.561	268.495
33) L7 AR-1260-3	6.828	6.586	128.0E6	69475884	268.498	268.057
34) L7 AR-1260-4	7.089	7.057	111.3E6	52728341	270.655	273.611
35) L7 AR-1260-5	7.331	7.297	266.3E6	117.4E6	256.903	261.226



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110352.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 10:31
Operator : YP/AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 11:01:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 10:56:01 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 218425182
Conc: 24.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 127327148
Conc: 25.48 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.734 min
Delta R.T.: 0.000 min
Response: 205292848
Conc: 26.70 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.683 min
Delta R.T.: 0.000 min
Response: 50640696
Conc: 27.40 ng/ml



#3 AR-1016-1


R.T.: 4.780 min
Delta R.T.: 0.000 min
Response: 85356347
Conc: 266.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250




#3 AR-1016-1

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 46113983
Conc: 270.07 ng/ml




#4 AR-1016-2

R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 118116710
Conc: 263.86 ng/ml



#4 AR-1016-2


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 64942457
Conc: 262.57 ng/ml



#5 AR-1016-3


R.T.: 4.855 min
Delta R.T.: 0.000 min
Response: 84160004
Conc: 271.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250




#5 AR-1016-3

R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 35585704
Conc: 268.45 ng/ml




#6 AR-1016-4

R.T.: 4.975 min
Delta R.T.: 0.000 min
Response: 65304594
Conc: 267.97 ng/ml



#6 AR-1016-4


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 30405283
Conc: 275.61 ng/ml



#7 AR-1016-5


R.T.: 5.232 min
Delta R.T.: 0.000 min
Response: 70395931
Conc: 272.52 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250




#7 AR-1016-5

R.T.: 5.214 min
Delta R.T.: 0.000 min
Response: 38809886
Conc: 272.01 ng/ml




#31 AR-1260-1

R.T.: 6.272 min
Delta R.T.: 0.000 min
Response: 122716328
Conc: 268.24 ng/ml



#31 AR-1260-1


R.T.: 6.245 min
Delta R.T.: 0.000 min
Response: 64201123
Conc: 270.26 ng/ml



#32 AR-1260-2


R.T.: 6.461 min
Delta R.T.: 0.000 min
Response: 149093603
Conc: 267.56 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250




#32 AR-1260-2

R.T.: 6.433 min
Delta R.T.: 0.000 min
Response: 74637486
Conc: 268.49 ng/ml




#33 AR-1260-3

R.T.: 6.828 min
Delta R.T.: 0.000 min
Response: 127967922
Conc: 268.50 ng/ml



#33 AR-1260-3


R.T.: 6.586 min
Delta R.T.: 0.000 min
Response: 69475884
Conc: 268.06 ng/ml



#34 AR-1260-4


R.T.: 7.089 min
Delta R.T.: 0.000 min
Response: 111343599
Conc: 270.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC250




#34 AR-1260-4

R.T.: 7.057 min
Delta R.T.: 0.000 min
Response: 52728341
Conc: 273.61 ng/ml



#35 AR-1260-5

R.T.: 7.331 min
Delta R.T.: 0.000 min
Response: 266280520
Conc: 256.90 ng/ml



#35 AR-1260-5

R.T.: 7.297 min
Delta R.T.: 0.000 min
Response: 117368967
Conc: 261.23 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110353.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 10:49
 Operator : YP/AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 11:04:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 10:56:01 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm


Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml


System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	43289216	24808316	4.952	4.972
2) SA Decachlor...	8.733	8.683	43582732	11156787	5.520	5.796
Target Compounds						
3) L1 AR-1016-1	4.779	4.765	18011965	9728182	54.891	55.428
4) L1 AR-1016-2	4.798	4.785	24206905	13399432	53.208	53.285
5) L1 AR-1016-3	4.855	4.960	18658513	7415278	57.851	54.640
6) L1 AR-1016-4	4.975	5.002	13549991	6527964	54.382	57.079
7) L1 AR-1016-5	5.233	5.215	15589904	8732807	57.953	58.580
31) L7 AR-1260-1	6.272	6.246	26614614	13981013	56.334	56.842
32) L7 AR-1260-2	6.461	6.433	35159753	17340702	59.956	59.437
33) L7 AR-1260-3	6.829	6.586	27896623	15884287	56.600	58.639
34) L7 AR-1260-4	7.088	7.057	23828498	11617411	56.143	57.902
35) L7 AR-1260-5	7.331	7.297	53567336	24854158	51.336	54.165

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 43289216
Conc: 4.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 24808316
Conc: 4.97 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 43582732
Conc: 5.52 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.683 min
Delta R.T.: 0.000 min
Response: 11156787
Conc: 5.80 ng/ml



#3 AR-1016-1


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 18011965
Conc: 54.89 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#3 AR-1016-1

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 9728182
Conc: 55.43 ng/ml




#4 AR-1016-2

R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 24206905
Conc: 53.21 ng/ml



#4 AR-1016-2


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 13399432
Conc: 53.29 ng/ml



#5 AR-1016-3


R.T.: 4.855 min
Delta R.T.: 0.000 min
Response: 18658513
Conc: 57.85 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#5 AR-1016-3

R.T.: 4.960 min
Delta R.T.: 0.000 min
Response: 7415278
Conc: 54.64 ng/ml




#6 AR-1016-4

R.T.: 4.975 min
Delta R.T.: 0.000 min
Response: 13549991
Conc: 54.38 ng/ml



#6 AR-1016-4


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 6527964
Conc: 57.08 ng/ml



#7 AR-1016-5


R.T.: 5.233 min
Delta R.T.: 0.000 min
Response: 15589904
Conc: 57.95 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#7 AR-1016-5

R.T.: 5.215 min
Delta R.T.: 0.000 min
Response: 8732807
Conc: 58.58 ng/ml




#31 AR-1260-1

R.T.: 6.272 min
Delta R.T.: 0.000 min
Response: 26614614
Conc: 56.33 ng/ml



#31 AR-1260-1


R.T.: 6.246 min
Delta R.T.: 0.000 min
Response: 13981013
Conc: 56.84 ng/ml



#32 AR-1260-2


R.T.: 6.461 min
Delta R.T.: 0.000 min
Response: 35159753
Conc: 59.96 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#32 AR-1260-2

R.T.: 6.433 min
Delta R.T.: 0.000 min
Response: 17340702
Conc: 59.44 ng/ml




#33 AR-1260-3

R.T.: 6.829 min
Delta R.T.: 0.000 min
Response: 27896623
Conc: 56.60 ng/ml



#33 AR-1260-3


R.T.: 6.586 min
Delta R.T.: 0.000 min
Response: 15884287
Conc: 58.64 ng/ml



#34 AR-1260-4


R.T.: 7.088 min
Delta R.T.: 0.000 min
Response: 23828498
Conc: 56.14 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660ICC050




#34 AR-1260-4

R.T.: 7.057 min
Delta R.T.: 0.000 min
Response: 11617411
Conc: 57.90 ng/ml



#35 AR-1260-5

R.T.: 7.331 min
Delta R.T.: 0.000 min
Response: 53567336
Conc: 51.34 ng/ml



#35 AR-1260-5

R.T.: 7.297 min
Delta R.T.: 0.000 min
Response: 24854158
Conc: 54.17 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110354.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 11:08
 Operator : YP/AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 11:55:47 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 11:55:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	417.4E6	233.4E6	50.000	50.000
2) SA Decachlor...	8.732	8.684	379.4E6	91675770	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	3.901	3.895	59017332	33386157	500.000	500.000
9) L2 AR-1221-2	3.987	3.981	44171792	25048140	500.000	500.000
10) L2 AR-1221-3	4.064	4.057	131.6E6	74585161	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110354.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 11:08
Operator : YP/AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 11:55:47 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 11:55:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 417415920
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 233406278
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 379375296
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 91675770
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 3.901 min
Delta R.T.: 0.000 min
Response: 59017332
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 3.895 min
Delta R.T.: 0.000 min
Response: 33386157
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 3.987 min
Delta R.T.: 0.000 min
Response: 44171792
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 3.981 min
Delta R.T.: 0.000 min
Response: 25048140
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 4.064 min
Delta R.T.: 0.000 min
Response: 131562252
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 4.057 min
Delta R.T.: 0.000 min
Response: 74585161
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110355.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 11:26
 Operator : YP/AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 11:58:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 11:55:23 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	398.8E6	234.5E6	50.000	50.000
2) SA Decachlor...	8.732	8.684	375.1E6	90279792	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	4.063	4.056	102.4E6	57511970	500.000	500.000
12) L3 AR-1232-2	4.556	4.784	53354286	55434068	500.000	500.000
13) L3 AR-1232-3	4.798	4.959	99890710	28601554	500.000	500.000
14) L3 AR-1232-4	4.975	5.044	51149202	26976681	500.000	500.000
15) L3 AR-1232-5	5.016	5.214	34451183	28567949	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110355.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 11:26
Operator : YP/AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 11:58:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 11:55:23 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 398843258
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 234466810
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 375127785
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 90279792
Conc: 50.00 ng/ml



#11 AR-1232-1


R.T.: 4.063 min
Delta R.T.: 0.000 min
Response: 102411775
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500




#11 AR-1232-1

R.T.: 4.056 min
Delta R.T.: 0.000 min
Response: 57511970
Conc: 500.00 ng/ml




#12 AR-1232-2

R.T.: 4.556 min
Delta R.T.: 0.000 min
Response: 53354286
Conc: 500.00 ng/ml



#12 AR-1232-2


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 55434068
Conc: 500.00 ng/ml



#13 AR-1232-3


R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 99890710
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500




#13 AR-1232-3

R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 28601554
Conc: 500.00 ng/ml




#14 AR-1232-4

R.T.: 4.975 min
Delta R.T.: 0.000 min
Response: 51149202
Conc: 500.00 ng/ml



#14 AR-1232-4


R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 26976681
Conc: 500.00 ng/ml



#15 AR-1232-5

R.T.: 5.016 min
Delta R.T.: 0.000 min
Response: 34451183
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1232ICC500



#15 AR-1232-5

R.T.: 5.214 min
Delta R.T.: 0.000 min
Response: 28567949
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110356.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 11:44
 Operator : YP/AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 12:34:26 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 12:31:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	827.2E6	466.8E6	96.341	96.381
2) SA Decachlor...	8.733	8.684	723.7E6	161.9E6	95.968	92.585
Target Compounds						
16) L4 AR-1242-1	4.780	4.766	251.4E6	135.4E6	944.813	952.306
17) L4 AR-1242-2	4.799	4.785	357.5E6	196.0E6	959.280	955.179
18) L4 AR-1242-3	4.855	4.960	243.8E6	104.3E6	938.223	947.136
19) L4 AR-1242-4	4.976	5.044	192.5E6	104.2E6	946.796	934.477
20) L4 AR-1242-5	5.628	5.565	201.5E6	127.6E6	943.266	944.974



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110356.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 11:44
Operator : YP/AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 12:34:26 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 12:31:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 827152171
Conc: 96.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 466802602
Conc: 96.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 723668193
Conc: 95.97 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 161857670
Conc: 92.58 ng/ml



#16 AR-1242-1


R.T.: 4.780 min
Delta R.T.: 0.000 min
Response: 251425090
Conc: 944.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#16 AR-1242-1

R.T.: 4.766 min
Delta R.T.: 0.000 min
Response: 135406625
Conc: 952.31 ng/ml




#17 AR-1242-2

R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 357466275
Conc: 959.28 ng/ml



#17 AR-1242-2


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 196039746
Conc: 955.18 ng/ml



#18 AR-1242-3


R.T.: 4.855 min
Delta R.T.: 0.000 min
Response: 243771373
Conc: 938.22 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000




#18 AR-1242-3

R.T.: 4.960 min
Delta R.T.: 0.000 min
Response: 104255216
Conc: 947.14 ng/ml




#19 AR-1242-4

R.T.: 4.976 min
Delta R.T.: 0.000 min
Response: 192501706
Conc: 946.80 ng/ml



#19 AR-1242-4


R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 104208953
Conc: 934.48 ng/ml



#20 AR-1242-5

R.T.: 5.628 min
Delta R.T.: 0.000 min
Response: 201519138
Conc: 943.27 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC1000



#20 AR-1242-5

R.T.: 5.565 min
Delta R.T.: 0.000 min
Response: 127551049
Conc: 944.97 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110357.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 12:03
 Operator : YP/AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 12:36:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 12:31:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	658.2E6	370.5E6	76.100	75.993
2) SA Decachlor...	8.733	8.684	567.8E6	132.4E6	75.200	75.507
Target Compounds						
16) L4 AR-1242-1	4.780	4.766	198.9E6	107.1E6	748.410	752.372
17) L4 AR-1242-2	4.798	4.785	286.0E6	156.3E6	761.574	757.700
18) L4 AR-1242-3	4.855	4.960	217.2E6	83520623	805.293	755.822
19) L4 AR-1242-4	4.975	5.045	154.0E6	84537972	754.983	755.368
20) L4 AR-1242-5	5.627	5.565	161.1E6	102.1E6	752.686	754.354



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110357.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 12:03
Operator : YP/AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 12:36:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 12:31:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 658203968
Conc: 76.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 370506533
Conc: 75.99 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 567817183
Conc: 75.20 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 132448865
Conc: 75.51 ng/ml



#16 AR-1242-1


R.T.: 4.780 min
Delta R.T.: 0.000 min
Response: 198949228
Conc: 748.41 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#16 AR-1242-1

R.T.: 4.766 min
Delta R.T.: 0.000 min
Response: 107147844
Conc: 752.37 ng/ml




#17 AR-1242-2

R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 285999640
Conc: 761.57 ng/ml



#17 AR-1242-2


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 156311902
Conc: 757.70 ng/ml



#18 AR-1242-3


R.T.: 4.855 min
Delta R.T.: 0.000 min
Response: 217241363
Conc: 805.29 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750




#18 AR-1242-3

R.T.: 4.960 min
Delta R.T.: 0.000 min
Response: 83520623
Conc: 755.82 ng/ml




#19 AR-1242-4

R.T.: 4.975 min
Delta R.T.: 0.000 min
Response: 154014183
Conc: 754.98 ng/ml



#19 AR-1242-4


R.T.: 5.045 min
Delta R.T.: 0.000 min
Response: 84537972
Conc: 755.37 ng/ml



#20 AR-1242-5

R.T.: 5.627 min
Delta R.T.: 0.000 min
Response: 161092202
Conc: 752.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 5.565 min
Delta R.T.: 0.000 min
Response: 102118034
Conc: 754.35 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110358.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 12:21
 Operator : YP/AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 12:31:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 12:31:05 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.684	445.0E6	250.9E6	50.000	50.000
2) SA Decachlor...	8.732	8.683	392.2E6	93891957	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	4.779	4.765	140.4E6	74484875	500.000	500.000
17) L4 AR-1242-2	4.798	4.783	193.9E6	107.2E6	500.000	500.000
18) L4 AR-1242-3	4.855	4.959	137.9E6	57946521	500.000	500.000
19) L4 AR-1242-4	4.975	5.044	107.1E6	59411329	500.000	500.000
20) L4 AR-1242-5	5.628	5.564	112.9E6	71202904	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110358.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 12:21
Operator : YP/AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 12:31:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 12:31:05 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.687 min
Delta R.T.: 0.000 min
Response: 444994526
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 250927877
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 392237800
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.683 min
Delta R.T.: 0.000 min
Response: 93891957
Conc: 50.00 ng/ml



#16 AR-1242-1


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 140398326
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500




#16 AR-1242-1

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 74484875
Conc: 500.00 ng/ml




#17 AR-1242-2

R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 193906918
Conc: 500.00 ng/ml



#17 AR-1242-2


R.T.: 4.783 min
Delta R.T.: 0.000 min
Response: 107218849
Conc: 500.00 ng/ml



#18 AR-1242-3


R.T.: 4.855 min
Delta R.T.: 0.000 min
Response: 137936830
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500




#18 AR-1242-3

R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 57946521
Conc: 500.00 ng/ml




#19 AR-1242-4

R.T.: 4.975 min
Delta R.T.: 0.000 min
Response: 107068171
Conc: 500.00 ng/ml



#19 AR-1242-4


R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 59411329
Conc: 500.00 ng/ml



#20 AR-1242-5

R.T.: 5.628 min
Delta R.T.: 0.000 min
Response: 112880165
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC500



#20 AR-1242-5

R.T.: 5.564 min
Delta R.T.: 0.000 min
Response: 71202904
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110359.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 12:39
 Operator : YP/AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 12:51:54 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 12:51:44 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	224.9E6	127.8E6	25.741	25.901
2) SA Decachlor...	8.732	8.684	204.7E6	50443721	26.551	27.716
Target Compounds						
16) L4 AR-1242-1	4.779	4.766	74380905	39467591	271.708	269.813
17) L4 AR-1242-2	4.799	4.785	100.8E6	56229093	263.544	266.549
18) L4 AR-1242-3	4.855	4.960	73089348	30705869	265.380	270.338
19) L4 AR-1242-4	4.975	5.045	56349611	32190259	269.168	277.198
20) L4 AR-1242-5	5.627	5.565	59802630	38560049	271.435	275.255



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110359.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 12:39
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 12:51:54 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 12:51:44 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 224861129
Conc: 25.74 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 127817855
Conc: 25.90 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 204715334
Conc: 26.55 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 50443721
Conc: 27.72 ng/ml



#16 AR-1242-1


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 74380905
Conc: 271.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.766 min
Delta R.T.: 0.000 min
Response: 39467591
Conc: 269.81 ng/ml




#17 AR-1242-2

R.T.: 4.799 min
Delta R.T.: 0.000 min
Response: 100791009
Conc: 263.54 ng/ml



#17 AR-1242-2


R.T.: 4.785 min
Delta R.T.: 0.000 min
Response: 56229093
Conc: 266.55 ng/ml



#18 AR-1242-3


R.T.: 4.855 min
Delta R.T.: 0.000 min
Response: 73089348
Conc: 265.38 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 4.960 min
Delta R.T.: 0.000 min
Response: 30705869
Conc: 270.34 ng/ml




#19 AR-1242-4

R.T.: 4.975 min
Delta R.T.: 0.000 min
Response: 56349611
Conc: 269.17 ng/ml



#19 AR-1242-4


R.T.: 5.045 min
Delta R.T.: 0.000 min
Response: 32190259
Conc: 277.20 ng/ml



#20 AR-1242-5

R.T.: 5.627 min
Delta R.T.: 0.000 min
Response: 59802630
Conc: 271.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.565 min
Delta R.T.: 0.000 min
Response: 38560049
Conc: 275.25 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110360.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 12:58
 Operator : YP/AJ
 Sample : AR1242ICCO50
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICCO50

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 13:08:06 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 13:07:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.684	40144674	23927704	4.671	4.878
2) SA Decachlor...	8.732	8.683	41703697	10219031	5.322	5.480
Target Compounds						
16) L4 AR-1242-1	4.779	4.765	14632790	8120107	52.724	54.314
17) L4 AR-1242-2	4.798	4.784	19954532	11150308	51.726	52.260
18) L4 AR-1242-3	4.854	4.959	14467594	6178953	52.004	53.459
19) L4 AR-1242-4	4.975	5.044	10972399	6799320	51.912	56.614
20) L4 AR-1242-5	5.628	5.564	12534403	8256494	55.366	56.903



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
Data File : P0110360.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 12:58
Operator : YP/AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 13:08:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 13:07:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.687 min
Delta R.T.: 0.000 min
Response: 40144674
Conc: 4.67 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 23927704
Conc: 4.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 41703697
Conc: 5.32 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.683 min
Delta R.T.: 0.000 min
Response: 10219031
Conc: 5.48 ng/ml



#16 AR-1242-1


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 14632790
Conc: 52.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050




#16 AR-1242-1

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 8120107
Conc: 54.31 ng/ml




#17 AR-1242-2

R.T.: 4.798 min
Delta R.T.: 0.000 min
Response: 19954532
Conc: 51.73 ng/ml



#17 AR-1242-2


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 11150308
Conc: 52.26 ng/ml



#18 AR-1242-3


R.T.: 4.854 min
Delta R.T.: 0.000 min
Response: 14467594
Conc: 52.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050




#18 AR-1242-3

R.T.: 4.959 min
Delta R.T.: 0.000 min
Response: 6178953
Conc: 53.46 ng/ml




#19 AR-1242-4

R.T.: 4.975 min
Delta R.T.: 0.000 min
Response: 10972399
Conc: 51.91 ng/ml



#19 AR-1242-4


R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 6799320
Conc: 56.61 ng/ml



#20 AR-1242-5

R.T.: 5.628 min
Delta R.T.: 0.000 min
Response: 12534403
Conc: 55.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC050



#20 AR-1242-5

R.T.: 5.564 min
Delta R.T.: 0.000 min
Response: 8256494
Conc: 56.90 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110361.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 13:16
 Operator : YP/AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 14:06:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 14:03:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.685	857.2E6	482.8E6	97.792	97.864
2) SA Decachlor...	8.731	8.684	742.3E6	170.8E6	96.063	94.678
Target Compounds						
21) L5 AR-1248-1	4.779	4.765	197.4E6	105.8E6	950.625	949.399
22) L5 AR-1248-2	5.017	5.001	269.7E6	147.0E6	946.388	943.970
23) L5 AR-1248-3	5.233	5.043	336.7E6	157.5E6	949.716	945.079
24) L5 AR-1248-4	5.586	5.214	480.1E6	185.6E6	956.437	952.264
25) L5 AR-1248-5	5.627	5.606	340.6E6	181.9E6	955.211	957.967



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110361.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 13:16
Operator : YP/AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 14:06:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 14:03:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.687 min
Delta R.T.: 0.000 min
Response: 857217740
Conc: 97.79 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 482790827
Conc: 97.86 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.731 min
Delta R.T.: 0.000 min
Response: 742336789
Conc: 96.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 170828178
Conc: 94.68 ng/ml



#21 AR-1248-1


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 197377890
Conc: 950.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#21 AR-1248-1

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 105821161
Conc: 949.40 ng/ml




#22 AR-1248-2

R.T.: 5.017 min
Delta R.T.: 0.000 min
Response: 269669918
Conc: 946.39 ng/ml



#22 AR-1248-2


R.T.: 5.001 min
Delta R.T.: 0.000 min
Response: 147033899
Conc: 943.97 ng/ml



#23 AR-1248-3


R.T.: 5.233 min
Delta R.T.: 0.000 min
Response: 336662132
Conc: 949.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC1000




#23 AR-1248-3

R.T.: 5.043 min
Delta R.T.: 0.000 min
Response: 157542784
Conc: 945.08 ng/ml



#24 AR-1248-4

R.T.: 5.586 min
Delta R.T.: 0.000 min
Response: 480139637
Conc: 956.44 ng/ml



#24 AR-1248-4

R.T.: 5.214 min
Delta R.T.: 0.000 min
Response: 185573916
Conc: 952.26 ng/ml

Response_

Signal: PO110361.D\ECD1A.ch

#25 AR-1248-5

R.T.: 5.627 min

Delta R.T.: 0.000 min

Response: 340550967


Conc: 955.21 ng/ml

Instrument :

ECD_O

ClientSampleId :

AR1248ICC1000



Time

Response_

Signal: PO110361.D\ECD2B.ch


#25 AR-1248-5

R.T.: 5.606 min

Delta R.T.: 0.000 min

Response: 181941821

Conc: 957.97 ng/ml



Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110362.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 13:35
 Operator : YP/AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 14:08:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 14:03:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	650.6E6	365.5E6	74.481	74.396
2) SA Decachlor...	8.733	8.685	575.0E6	131.4E6	74.605	73.547
Target Compounds						
21) L5 AR-1248-1	4.779	4.765	154.3E6	83273178	745.481	748.067
22) L5 AR-1248-2	5.017	5.002	210.8E6	115.1E6	743.133	742.489
23) L5 AR-1248-3	5.232	5.044	264.1E6	123.2E6	746.588	742.651
24) L5 AR-1248-4	5.586	5.215	372.9E6	144.5E6	745.252	744.242
25) L5 AR-1248-5	5.627	5.606	265.1E6	140.5E6	745.654	743.111



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110362.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 13:35
Operator : YP/AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 14:08:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 14:03:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 650631933
Conc: 74.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 365548178
Conc: 74.40 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 574998503
Conc: 74.60 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.685 min
Delta R.T.: 0.000 min
Response: 131428252
Conc: 73.55 ng/ml



#21 AR-1248-1


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 154318833
Conc: 745.48 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 83273178
Conc: 748.07 ng/ml




#22 AR-1248-2

R.T.: 5.017 min
Delta R.T.: 0.000 min
Response: 210788058
Conc: 743.13 ng/ml



#22 AR-1248-2


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 115074831
Conc: 742.49 ng/ml



#23 AR-1248-3


R.T.: 5.232 min
Delta R.T.: 0.000 min
Response: 264055417
Conc: 746.59 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 123194788
Conc: 742.65 ng/ml



#24 AR-1248-4

R.T.: 5.586 min
Delta R.T.: 0.000 min
Response: 372942670
Conc: 745.25 ng/ml



#24 AR-1248-4

R.T.: 5.215 min
Delta R.T.: 0.000 min
Response: 144480618
Conc: 744.24 ng/ml

Response_

Signal: PO110362.D\ECD1A.ch

#25 AR-1248-5

R.T.: 5.627 min

Delta R.T.: 0.000 min

Response: 265071795


Conc: 745.65 ng/ml

Instrument :

ECD_O

ClientSampleId :

AR1248ICC750



Time

Response_

Signal: PO110362.D\ECD2B.ch


#25 AR-1248-5

R.T.: 5.606 min

Delta R.T.: 0.000 min

Response: 140490252

Conc: 743.11 ng/ml



Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110363.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 13:53
 Operator : YP/AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 14:03:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 14:03:21 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.685	448.0E6	251.9E6	50.000	50.000
2) SA Decachlor...	8.732	8.685	401.6E6	95016625	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	4.779	4.765	108.9E6	58550593	500.000	500.000
22) L5 AR-1248-2	5.017	5.002	150.1E6	82244273	500.000	500.000
23) L5 AR-1248-3	5.232	5.044	186.2E6	87926534	500.000	500.000
24) L5 AR-1248-4	5.586	5.215	261.9E6	102.1E6	500.000	500.000
25) L5 AR-1248-5	5.628	5.606	186.2E6	98954091	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110363.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 13:53
Operator : YP/AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 14:03:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 14:03:21 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.687 min
Delta R.T.: 0.000 min
Response: 447959162
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 251933187
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 401588257
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.685 min
Delta R.T.: 0.000 min
Response: 95016625
Conc: 50.00 ng/ml



#21 AR-1248-1


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 108940585
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#21 AR-1248-1

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 58550593
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 5.017 min
Delta R.T.: 0.000 min
Response: 150111647
Conc: 500.00 ng/ml



#22 AR-1248-2


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 82244273
Conc: 500.00 ng/ml



#23 AR-1248-3


R.T.: 5.232 min
Delta R.T.: 0.000 min
Response: 186156063
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500




#23 AR-1248-3

R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 87926534
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 5.586 min
Delta R.T.: 0.000 min
Response: 261938735
Conc: 500.00 ng/ml



#24 AR-1248-4


R.T.: 5.215 min
Delta R.T.: 0.000 min
Response: 102089672
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 5.628 min
Delta R.T.: 0.000 min
Response: 186243693
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 5.606 min
Delta R.T.: 0.000 min
Response: 98954091
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110364.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 14:11
 Operator : YP/AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 14:22:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 14:22:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	221.9E6	124.9E6	25.300	25.319
2) SA Decachlor...	8.732	8.684	209.6E6	50390082	26.607	27.324
Target Compounds						
21) L5 AR-1248-1	4.779	4.766	56940091	30514286	268.339	267.663
22) L5 AR-1248-2	5.018	5.002	78187177	43024876	268.755	270.149
23) L5 AR-1248-3	5.232	5.044	100.5E6	46712689	274.798	272.972
24) L5 AR-1248-4	5.586	5.214	138.1E6	54859460	269.040	273.671
25) L5 AR-1248-5	5.628	5.607	98878039	52576565	270.532	270.499



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110364.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 14:11
Operator : YP/AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 14:22:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 14:22:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 221895724
Conc: 25.30 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 124938500
Conc: 25.32 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 209558937
Conc: 26.61 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 50390082
Conc: 27.32 ng/ml



#21 AR-1248-1


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 56940091
Conc: 268.34 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#21 AR-1248-1

R.T.: 4.766 min
Delta R.T.: 0.000 min
Response: 30514286
Conc: 267.66 ng/ml




#22 AR-1248-2

R.T.: 5.018 min
Delta R.T.: 0.000 min
Response: 78187177
Conc: 268.76 ng/ml



#22 AR-1248-2


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 43024876
Conc: 270.15 ng/ml



#23 AR-1248-3


R.T.: 5.232 min
Delta R.T.: 0.000 min
Response: 100514618
Conc: 274.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250




#23 AR-1248-3

R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 46712689
Conc: 272.97 ng/ml




#24 AR-1248-4

R.T.: 5.586 min
Delta R.T.: 0.000 min
Response: 138141162
Conc: 269.04 ng/ml



#24 AR-1248-4


R.T.: 5.214 min
Delta R.T.: 0.000 min
Response: 54859460
Conc: 273.67 ng/ml



#25 AR-1248-5

R.T.: 5.628 min
Delta R.T.: 0.000 min
Response: 98878039
Conc: 270.53 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC250



#25 AR-1248-5

R.T.: 5.607 min
Delta R.T.: 0.000 min
Response: 52576565
Conc: 270.50 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110365.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 14:30
 Operator : YP/AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1248ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 14:40:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 14:40:24 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.684	39393478	22189412	4.585	4.589
2) SA Decachlor...	8.731	8.683	41107420	10123504	5.174	5.384
Target Compounds						
21) L5 AR-1248-1	4.778	4.764	10770182	5982472	50.603	51.962
22) L5 AR-1248-2	5.017	5.001	15796625	8867896	53.380	54.443
23) L5 AR-1248-3	5.231	5.043	19335025	9738708	52.262	55.379
24) L5 AR-1248-4	5.585	5.214	27315770	10828083	52.527	53.163
25) L5 AR-1248-5	5.627	5.605	19257467	10475122	52.128	53.067



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
Data File : P0110365.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 14:30
Operator : YP/AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 14:40:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 14:40:24 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.687 min
Delta R.T.: 0.000 min
Response: 39393478
Conc: 4.58 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 22189412
Conc: 4.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.731 min
Delta R.T.: 0.000 min
Response: 41107420
Conc: 5.17 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.683 min
Delta R.T.: 0.000 min
Response: 10123504
Conc: 5.38 ng/ml



#21 AR-1248-1


R.T.: 4.778 min
Delta R.T.: 0.000 min
Response: 10770182
Conc: 50.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050




#21 AR-1248-1

R.T.: 4.764 min
Delta R.T.: 0.000 min
Response: 5982472
Conc: 51.96 ng/ml




#22 AR-1248-2

R.T.: 5.017 min
Delta R.T.: 0.000 min
Response: 15796625
Conc: 53.38 ng/ml



#22 AR-1248-2


R.T.: 5.001 min
Delta R.T.: 0.000 min
Response: 8867896
Conc: 54.44 ng/ml



#23 AR-1248-3


R.T.: 5.231 min
Delta R.T.: 0.000 min
Response: 19335025
Conc: 52.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050




#23 AR-1248-3

R.T.: 5.043 min
Delta R.T.: 0.000 min
Response: 9738708
Conc: 55.38 ng/ml




#24 AR-1248-4

R.T.: 5.585 min
Delta R.T.: 0.000 min
Response: 27315770
Conc: 52.53 ng/ml



#24 AR-1248-4


R.T.: 5.214 min
Delta R.T.: 0.000 min
Response: 10828083
Conc: 53.16 ng/ml



#25 AR-1248-5

R.T.: 5.627 min
Delta R.T.: 0.000 min
Response: 19257467
Conc: 52.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1248ICC050



#25 AR-1248-5

R.T.: 5.605 min
Delta R.T.: 0.000 min
Response: 10475122
Conc: 53.07 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110366.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 14:48
 Operator : YP/AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 15:41:12 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 15:38:18 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	866.5E6	486.6E6	99.747	99.926
2) SA Decachlor...	8.733	8.684	761.3E6	167.3E6	96.636	93.604
Target Compounds						
26) L6 AR-1254-1	5.587	5.566	517.2E6	275.5E6	964.265	963.531
27) L6 AR-1254-2	5.736	5.713	450.8E6	239.9E6	970.858	963.389
28) L6 AR-1254-3	6.142	6.117	746.2E6	380.0E6	977.998	969.945
29) L6 AR-1254-4	6.371	6.344	460.1E6	215.9E6	968.424	969.951
30) L6 AR-1254-5	6.790	6.761	657.6E6	307.7E6	969.952	963.966



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110366.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 14:48
Operator : YP/AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 15:41:12 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 15:38:18 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 866509220
Conc: 99.75 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 486562593
Conc: 99.93 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 761299037
Conc: 96.64 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 167274258
Conc: 93.60 ng/ml



#26 AR-1254-1


R.T.: 5.587 min
Delta R.T.: 0.000 min
Response: 517202507
Conc: 964.26 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#26 AR-1254-1

R.T.: 5.566 min
Delta R.T.: 0.000 min
Response: 275528987
Conc: 963.53 ng/ml




#27 AR-1254-2

R.T.: 5.736 min
Delta R.T.: 0.000 min
Response: 450798901
Conc: 970.86 ng/ml



#27 AR-1254-2


R.T.: 5.713 min
Delta R.T.: 0.000 min
Response: 239912515
Conc: 963.39 ng/ml



#28 AR-1254-3


R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 746169545
Conc: 978.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC1000




#28 AR-1254-3

R.T.: 6.117 min
Delta R.T.: 0.000 min
Response: 379979735
Conc: 969.94 ng/ml




#29 AR-1254-4

R.T.: 6.371 min
Delta R.T.: 0.000 min
Response: 460149954
Conc: 968.42 ng/ml



#29 AR-1254-4

R.T.: 6.344 min
Delta R.T.: 0.000 min
Response: 215870098
Conc: 969.95 ng/ml



#30 AR-1254-5

R.T.: 6.790 min

Delta R.T.: 0.000 min

Response: 657635451


Conc: 969.95 ng/ml

Instrument :

ECD_O

ClientSampleId :

AR1254ICC1000



#30 AR-1254-5

R.T.: 6.761 min

Delta R.T.: 0.000 min

Response: 307722172

Conc: 963.97 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110367.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 15:06
 Operator : YP/AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 15:44:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 15:38:18 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	661.9E6	372.0E6	75.795	75.921
2) SA Decachlor...	8.732	8.684	591.5E6	134.6E6	75.055	75.216
Target Compounds						
26) L6 AR-1254-1	5.587	5.566	402.3E6	214.0E6	750.022	748.992
27) L6 AR-1254-2	5.736	5.714	351.2E6	186.9E6	754.162	750.330
28) L6 AR-1254-3	6.141	6.116	576.8E6	294.3E6	754.032	750.876
29) L6 AR-1254-4	6.370	6.344	355.5E6	166.8E6	748.724	749.718
30) L6 AR-1254-5	6.790	6.761	508.6E6	239.6E6	750.141	750.401



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110367.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 15:06
Operator : YP/AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 15:44:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 15:38:18 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 661947495
Conc: 75.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 371957792
Conc: 75.92 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 591504276
Conc: 75.06 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 134606126
Conc: 75.22 ng/ml



#26 AR-1254-1


R.T.: 5.587 min
Delta R.T.: 0.000 min
Response: 402295349
Conc: 750.02 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750




#26 AR-1254-1

R.T.: 5.566 min
Delta R.T.: 0.000 min
Response: 214036296
Conc: 748.99 ng/ml




#27 AR-1254-2

R.T.: 5.736 min
Delta R.T.: 0.000 min
Response: 351154578
Conc: 754.16 ng/ml



#27 AR-1254-2


R.T.: 5.714 min
Delta R.T.: 0.000 min
Response: 186895371
Conc: 750.33 ng/ml



#28 AR-1254-3


R.T.: 6.141 min
Delta R.T.: 0.000 min
Response: 576844096
Conc: 754.03 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC750




#28 AR-1254-3

R.T.: 6.116 min
Delta R.T.: 0.000 min
Response: 294330716
Conc: 750.88 ng/ml




#29 AR-1254-4

R.T.: 6.370 min
Delta R.T.: 0.000 min
Response: 355456258
Conc: 748.72 ng/ml



#29 AR-1254-4

R.T.: 6.344 min
Delta R.T.: 0.000 min
Response: 166824313
Conc: 749.72 ng/ml



#30 AR-1254-5

R.T.: 6.790 min

Delta R.T.: 0.000 min

Response: 508649044


Conc: 750.14 ng/ml

Instrument :

ECD_O

ClientSampleId :

AR1254ICC750



#30 AR-1254-5

R.T.: 6.761 min

Delta R.T.: 0.000 min

Response: 239610761

Conc: 750.40 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110368.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 15:25
 Operator : YP/AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 15:38:32 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 15:38:18 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	435.5E6	243.6E6	50.000	50.000
2) SA Decachlor...	8.733	8.684	407.1E6	95066339	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	5.587	5.566	277.8E6	148.2E6	500.000	500.000
27) L6 AR-1254-2	5.735	5.714	238.9E6	129.1E6	500.000	500.000
28) L6 AR-1254-3	6.141	6.115	389.9E6	201.8E6	500.000	500.000
29) L6 AR-1254-4	6.370	6.343	245.1E6	114.6E6	500.000	500.000
30) L6 AR-1254-5	6.790	6.761	349.2E6	165.4E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110368.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 15:25
Operator : YP/AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 15:38:32 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 15:38:18 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 435455612
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 243639570
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 407149192
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 95066339
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 5.587 min
Delta R.T.: 0.000 min
Response: 277768607
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 5.566 min
Delta R.T.: 0.000 min
Response: 148193147
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 5.735 min
Delta R.T.: 0.000 min
Response: 238930915
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.714 min
Delta R.T.: 0.000 min
Response: 129073422
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 6.141 min
Delta R.T.: 0.000 min
Response: 389871530
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 6.115 min
Delta R.T.: 0.000 min
Response: 201764217
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 6.370 min
Delta R.T.: 0.000 min
Response: 245078389
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.343 min
Delta R.T.: 0.000 min
Response: 114622754
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 6.790 min
Delta R.T.: 0.000 min
Response: 349190332
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 6.761 min
Delta R.T.: 0.000 min
Response: 165364115
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110369.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 15:43
 Operator : YP/AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 16:25:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 16:25:25 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.684	224.4E6	126.5E6	25.514	25.603
2) SA Decachlor...	8.732	8.683	212.0E6	52144786	26.397	27.980
Target Compounds						
26) L6 AR-1254-1	5.586	5.565	145.4E6	77697153	265.491	266.067
27) L6 AR-1254-2	5.735	5.713	126.7E6	68670177	266.212	268.785
28) L6 AR-1254-3	6.140	6.115	202.8E6	104.4E6	261.133	261.974
29) L6 AR-1254-4	6.370	6.343	126.3E6	59656850	261.911	263.335
30) L6 AR-1254-5	6.789	6.760	181.1E6	86716902	262.640	265.840



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110369.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 15:43
Operator : YP/AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 16:25:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 16:25:25 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.687 min
Delta R.T.: 0.000 min
Response: 224366749
Conc: 25.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 126451974
Conc: 25.60 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 211976479
Conc: 26.40 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.683 min
Delta R.T.: 0.000 min
Response: 52144786
Conc: 27.98 ng/ml



#26 AR-1254-1


R.T.: 5.586 min
Delta R.T.: 0.000 min
Response: 145406737
Conc: 265.49 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254\ICC250




#26 AR-1254-1

R.T.: 5.565 min
Delta R.T.: 0.000 min
Response: 77697153
Conc: 266.07 ng/ml




#27 AR-1254-2

R.T.: 5.735 min
Delta R.T.: 0.000 min
Response: 126692531
Conc: 266.21 ng/ml



#27 AR-1254-2


R.T.: 5.713 min
Delta R.T.: 0.000 min
Response: 68670177
Conc: 268.79 ng/ml



#28 AR-1254-3


R.T.: 6.140 min
Delta R.T.: 0.000 min
Response: 202780419
Conc: 261.13 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 6.115 min
Delta R.T.: 0.000 min
Response: 104355598
Conc: 261.97 ng/ml




#29 AR-1254-4

R.T.: 6.370 min
Delta R.T.: 0.000 min
Response: 126348606
Conc: 261.91 ng/ml



#29 AR-1254-4


R.T.: 6.343 min
Delta R.T.: 0.000 min
Response: 59656850
Conc: 263.33 ng/ml



#30 AR-1254-5

R.T.: 6.789 min
Delta R.T.: 0.000 min
Response: 181141571
Conc: 262.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 86716902
Conc: 265.84 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110370.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 16:02
 Operator : YP/AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1254ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 16:28:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 16:28:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	41787473	23807736	4.800	4.855
2) SA Decachlor...	8.732	8.682	43595137	10662424	5.337	5.561
Target Compounds						
26) L6 AR-1254-1	5.587	5.566	29959251	16970498	53.691	56.287
27) L6 AR-1254-2	5.736	5.713	26243794	15074639	54.033	56.953
28) L6 AR-1254-3	6.141	6.115	40497442	21185447	51.706	52.515
29) L6 AR-1254-4	6.370	6.343	25419174	12119451	52.131	52.759
30) L6 AR-1254-5	6.790	6.760	37285913	17832984	53.197	53.667



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110370.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 16:02
Operator : YP/AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 16:28:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 16:28:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 41787473
Conc: 4.80 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 23807736
Conc: 4.86 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 43595137
Conc: 5.34 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.682 min
Delta R.T.: 0.000 min
Response: 10662424
Conc: 5.56 ng/ml



#26 AR-1254-1


R.T.: 5.587 min
Delta R.T.: 0.000 min
Response: 29959251
Conc: 53.69 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#26 AR-1254-1

R.T.: 5.566 min
Delta R.T.: 0.000 min
Response: 16970498
Conc: 56.29 ng/ml




#27 AR-1254-2

R.T.: 5.736 min
Delta R.T.: 0.000 min
Response: 26243794
Conc: 54.03 ng/ml



#27 AR-1254-2


R.T.: 5.713 min
Delta R.T.: 0.000 min
Response: 15074639
Conc: 56.95 ng/ml



#28 AR-1254-3


R.T.: 6.141 min
Delta R.T.: 0.000 min
Response: 40497442
Conc: 51.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050




#28 AR-1254-3

R.T.: 6.115 min
Delta R.T.: 0.000 min
Response: 21185447
Conc: 52.52 ng/ml




#29 AR-1254-4

R.T.: 6.370 min
Delta R.T.: 0.000 min
Response: 25419174
Conc: 52.13 ng/ml



#29 AR-1254-4


R.T.: 6.343 min
Delta R.T.: 0.000 min
Response: 12119451
Conc: 52.76 ng/ml



#30 AR-1254-5

R.T.: 6.790 min
Delta R.T.: 0.000 min
Response: 37285913
Conc: 53.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1254ICC050



#30 AR-1254-5

R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 17832984
Conc: 53.67 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110371.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 16:20
 Operator : YP/AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 16:33:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 16:33:00 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.685	436.1E6	244.3E6	50.000	50.000
2) SA Decachlor...	8.732	8.684	395.3E6	92706545	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	6.829	6.799	350.7E6	164.7E6	500.000	500.000
37) L8 AR-1262-2	7.331	7.298	603.9E6	254.2E6	500.000	500.000
38) L8 AR-1262-3	7.615	7.581	254.6E6	96815138	500.000	500.000
39) L8 AR-1262-4	7.679	7.644	448.1E6	173.5E6	500.000	500.000
40) L8 AR-1262-5	8.175	8.137	201.5E6	62297516	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110371.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 16:20
Operator : YP/AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 16:33:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 16:33:00 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.687 min
Delta R.T.: 0.000 min
Response: 436096183
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 244331831
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 395295598
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 92706545
Conc: 50.00 ng/ml



#36 AR-1262-1


R.T.: 6.829 min
Delta R.T.: 0.000 min
Response: 350729274
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500




#36 AR-1262-1

R.T.: 6.799 min
Delta R.T.: 0.000 min
Response: 164669723
Conc: 500.00 ng/ml




#37 AR-1262-2

R.T.: 7.331 min
Delta R.T.: 0.000 min
Response: 603868004
Conc: 500.00 ng/ml



#37 AR-1262-2


R.T.: 7.298 min
Delta R.T.: 0.000 min
Response: 254228551
Conc: 500.00 ng/ml



#38 AR-1262-3


R.T.: 7.615 min
Delta R.T.: 0.000 min
Response: 254551820
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500




#38 AR-1262-3

R.T.: 7.581 min
Delta R.T.: 0.000 min
Response: 96815138
Conc: 500.00 ng/ml




#39 AR-1262-4

R.T.: 7.679 min
Delta R.T.: 0.000 min
Response: 448136894
Conc: 500.00 ng/ml



#39 AR-1262-4


R.T.: 7.644 min
Delta R.T.: 0.000 min
Response: 173526656
Conc: 500.00 ng/ml



#40 AR-1262-5

R.T.: 8.175 min
Delta R.T.: 0.000 min
Response: 201497456
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1262ICC500



#40 AR-1262-5

R.T.: 8.137 min
Delta R.T.: 0.000 min
Response: 62297516
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110372.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 16:38
 Operator : YP/AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 18:27:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 18:23:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.684	884.3E6	497.4E6	102.204	102.064
2) SA Decachlor...	8.732	8.684	1400.9E6	301.8E6	98.671	95.869
Target Compounds						
41) L9 AR-1268-1	7.615	7.581	1383.6E6	525.2E6	990.863	980.668
42) L9 AR-1268-2	7.680	7.646	1284.7E6	480.6E6	996.216	982.242
43) L9 AR-1268-3	7.887	7.852	1040.5E6	351.9E6	994.813	974.012
44) L9 AR-1268-4	8.175	8.137	443.2E6	132.4E6	1001.524	967.863
45) L9 AR-1268-5	8.471	8.429	3289.1E6	751.7E6	1000.454	973.760



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110372.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 16:38
Operator : YP/AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 18:27:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 18:23:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 884275602
Conc: 102.20 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 497380509
Conc: 102.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 1400878204
Conc: 98.67 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 301839224
Conc: 95.87 ng/ml



#41 AR-1268-1


R.T.: 7.615 min
Delta R.T.: 0.000 min
Response: 1383597022
Conc: 990.86 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#41 AR-1268-1

R.T.: 7.581 min
Delta R.T.: 0.000 min
Response: 525182929
Conc: 980.67 ng/ml




#42 AR-1268-2

R.T.: 7.680 min
Delta R.T.: 0.000 min
Response: 1284666762
Conc: 996.22 ng/ml



#42 AR-1268-2


R.T.: 7.646 min
Delta R.T.: 0.000 min
Response: 480641047
Conc: 982.24 ng/ml



#43 AR-1268-3


R.T.: 7.887 min
Delta R.T.: 0.000 min
Response: 1040477449
Conc: 994.81 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000




#43 AR-1268-3

R.T.: 7.852 min
Delta R.T.: 0.000 min
Response: 351934435
Conc: 974.01 ng/ml




#44 AR-1268-4

R.T.: 8.175 min
Delta R.T.: 0.000 min
Response: 443191011
Conc: 1001.52 ng/ml



#44 AR-1268-4


R.T.: 8.137 min
Delta R.T.: 0.000 min
Response: 132381746
Conc: 967.86 ng/ml



#45 AR-1268-5

R.T.: 8.471 min
Delta R.T.: 0.000 min
Response: 3289098269
Conc: 1000.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC1000



#45 AR-1268-5

R.T.: 8.429 min
Delta R.T.: 0.000 min
Response: 751733306
Conc: 973.76 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110373.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 16:57
 Operator : YP/AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 18:31:42 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 18:23:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.684	651.4E6	366.7E6	75.195	75.170
2) SA Decachlor...	8.730	8.684	1067.4E6	237.5E6	75.123	75.295
Target Compounds						
41) L9 AR-1268-1	7.614	7.580	1047.9E6	404.2E6	750.320	753.161
42) L9 AR-1268-2	7.679	7.645	973.0E6	367.9E6	752.993	751.271
43) L9 AR-1268-3	7.886	7.851	784.9E6	272.0E6	750.332	751.923
44) L9 AR-1268-4	8.174	8.137	317.2E6	102.1E6	727.594	747.479
45) L9 AR-1268-5	8.472	8.428	2465.3E6	581.7E6	749.920	752.311



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110373.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 16:57
Operator : YP/AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 18:31:42 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 18:23:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.687 min
Delta R.T.: 0.000 min
Response: 651437343
Conc: 75.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 366737670
Conc: 75.17 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.730 min
Delta R.T.: 0.000 min
Response: 1067430568
Conc: 75.12 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 237530036
Conc: 75.30 ng/ml



#41 AR-1268-1


R.T.: 7.614 min
Delta R.T.: 0.000 min
Response: 1047935900
Conc: 750.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750




#41 AR-1268-1

R.T.: 7.580 min
Delta R.T.: 0.000 min
Response: 404196449
Conc: 753.16 ng/ml




#42 AR-1268-2

R.T.: 7.679 min
Delta R.T.: 0.000 min
Response: 972960332
Conc: 752.99 ng/ml



#42 AR-1268-2


R.T.: 7.645 min
Delta R.T.: 0.000 min
Response: 367932036
Conc: 751.27 ng/ml



#43 AR-1268-3


R.T.: 7.886 min
Delta R.T.: 0.000 min
Response: 784947577
Conc: 750.33 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC750




#43 AR-1268-3

R.T.: 7.851 min
Delta R.T.: 0.000 min
Response: 272037162
Conc: 751.92 ng/ml



#44 AR-1268-4

R.T.: 8.174 min
Delta R.T.: 0.000 min
Response: 317233528
Conc: 727.59 ng/ml




#44 AR-1268-4

R.T.: 8.137 min
Delta R.T.: 0.000 min
Response: 102066560
Conc: 747.48 ng/ml

Response_

Signal: PO110373.D\ECD1A.ch

#45 AR-1268-5



R.T.: 8.472 min

Delta R.T.: 0.000 min

Response: 2465307486

Conc: 749.92 ng/ml

Instrument :

ECD_O

ClientSampleId :


AR1268ICC750

Time

Response_

Signal: PO110373.D\ECD2B.ch

#45 AR-1268-5



R.T.: 8.428 min

Delta R.T.: 0.000 min

Response: 581672709

Conc: 752.31 ng/ml

Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110374.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 17:15
 Operator : YP/AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 18:24:15 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 18:23:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	423.1E6	238.6E6	50.000	50.000
2) SA Decachlor...	8.732	8.684	719.3E6	163.9E6	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	7.615	7.580	704.6E6	272.9E6	500.000	500.000
42) L9 AR-1268-2	7.680	7.645	647.2E6	249.0E6	500.000	500.000
43) L9 AR-1268-3	7.887	7.851	525.7E6	185.4E6	500.000	500.000
44) L9 AR-1268-4	8.175	8.137	220.9E6	70586461	500.000	500.000
45) L9 AR-1268-5	8.472	8.429	1643.1E6	396.1E6	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110374.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 17:15
Operator : YP/AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 18:24:15 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 18:23:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 423067129
Conc: 50.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 238629925
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 719312182
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 163925502
Conc: 50.00 ng/ml



#41 AR-1268-1


R.T.: 7.615 min
Delta R.T.: 0.000 min
Response: 704556317
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500




#41 AR-1268-1

R.T.: 7.580 min
Delta R.T.: 0.000 min
Response: 272944304
Conc: 500.00 ng/ml




#42 AR-1268-2

R.T.: 7.680 min
Delta R.T.: 0.000 min
Response: 647213588
Conc: 500.00 ng/ml



#42 AR-1268-2


R.T.: 7.645 min
Delta R.T.: 0.000 min
Response: 249010122
Conc: 500.00 ng/ml



#43 AR-1268-3


R.T.: 7.887 min
Delta R.T.: 0.000 min
Response: 525663675
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500




#43 AR-1268-3

R.T.: 7.851 min
Delta R.T.: 0.000 min
Response: 185357416
Conc: 500.00 ng/ml




#44 AR-1268-4

R.T.: 8.175 min
Delta R.T.: 0.000 min
Response: 220920910
Conc: 500.00 ng/ml



#44 AR-1268-4


R.T.: 8.137 min
Delta R.T.: 0.000 min
Response: 70586461
Conc: 500.00 ng/ml



#45 AR-1268-5

R.T.: 8.472 min
Delta R.T.: 0.000 min
Response: 1643055400
Conc: 500.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC500



#45 AR-1268-5

R.T.: 8.429 min
Delta R.T.: 0.000 min
Response: 396123334
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110375.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 17:33
 Operator : YP/AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 18:34:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 18:23:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.685	223.2E6	125.8E6	25.567	25.592
2) SA Decachlor...	8.733	8.684	374.5E6	87516933	26.002	27.002
Target Compounds						
41) L9 AR-1268-1	7.615	7.581	363.7E6	142.0E6	257.715	260.832
42) L9 AR-1268-2	7.681	7.646	329.0E6	129.2E6	253.440	260.179
43) L9 AR-1268-3	7.888	7.852	270.5E6	97320040	256.351	263.982
44) L9 AR-1268-4	8.176	8.137	118.1E6	38142084	265.411	271.372
45) L9 AR-1268-5	8.472	8.429	825.0E6	203.5E6	250.707	259.740



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110375.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 17:33
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 18:34:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 18:23:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.687 min
Delta R.T.: 0.000 min
Response: 223177656
Conc: 25.57 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 125848954
Conc: 25.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 374472641
Conc: 26.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 87516933
Conc: 27.00 ng/ml



#41 AR-1268-1


R.T.: 7.615 min
Delta R.T.: 0.000 min
Response: 363679668
Conc: 257.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#41 AR-1268-1

R.T.: 7.581 min
Delta R.T.: 0.000 min
Response: 142031158
Conc: 260.83 ng/ml




#42 AR-1268-2

R.T.: 7.681 min
Delta R.T.: 0.000 min
Response: 328985593
Conc: 253.44 ng/ml



#42 AR-1268-2


R.T.: 7.646 min
Delta R.T.: 0.000 min
Response: 129174850
Conc: 260.18 ng/ml



#43 AR-1268-3


R.T.: 7.888 min
Delta R.T.: 0.000 min
Response: 270468128
Conc: 256.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#43 AR-1268-3

R.T.: 7.852 min
Delta R.T.: 0.000 min
Response: 97320040
Conc: 263.98 ng/ml




#44 AR-1268-4

R.T.: 8.176 min
Delta R.T.: 0.000 min
Response: 118147663
Conc: 265.41 ng/ml



#44 AR-1268-4

R.T.: 8.137 min
Delta R.T.: 0.000 min
Response: 38142084
Conc: 271.37 ng/ml



#45 AR-1268-5

R.T.: 8.472 min

Delta R.T.: 0.000 min

Response: 824957315


Conc: 250.71 ng/ml

Instrument :

ECD_O

ClientSampleId :

AR1268ICC250



#45 AR-1268-5

R.T.: 8.429 min

Delta R.T.: 0.000 min

Response: 203468755

Conc: 259.74 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110376.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 17:52
 Operator : YP/AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 18:37:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 18:23:57 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.684	39467832	22436384	4.610	4.644
2) SA Decachlor...	8.732	8.684	71324202	17516545	4.962	5.318
Target Compounds						
41) L9 AR-1268-1	7.616	7.580	68808575	28029289	49.003	51.172
42) L9 AR-1268-2	7.680	7.645	60320169	25075562	47.135	50.404
43) L9 AR-1268-3	7.887	7.852	49726337	19160308	47.678	51.566
44) L9 AR-1268-4	8.175	8.136	22009230	7174009	49.553	50.830
45) L9 AR-1268-5	8.473	8.428	146.9E6	40130945	45.608	50.979



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110376.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 17:52
Operator : YP/AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 18:37:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 18:23:57 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.687 min
Delta R.T.: 0.000 min
Response: 39467832
Conc: 4.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 22436384
Conc: 4.64 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 71324202
Conc: 4.96 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 17516545
Conc: 5.32 ng/ml



#41 AR-1268-1


R.T.: 7.616 min
Delta R.T.: 0.000 min
Response: 68808575
Conc: 49.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#41 AR-1268-1

R.T.: 7.580 min
Delta R.T.: 0.000 min
Response: 28029289
Conc: 51.17 ng/ml




#42 AR-1268-2

R.T.: 7.680 min
Delta R.T.: 0.000 min
Response: 60320169
Conc: 47.13 ng/ml



#42 AR-1268-2


R.T.: 7.645 min
Delta R.T.: 0.000 min
Response: 25075562
Conc: 50.40 ng/ml



#43 AR-1268-3


R.T.: 7.887 min
Delta R.T.: 0.000 min
Response: 49726337
Conc: 47.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050




#43 AR-1268-3

R.T.: 7.852 min
Delta R.T.: 0.000 min
Response: 19160308
Conc: 51.57 ng/ml




#44 AR-1268-4

R.T.: 8.175 min
Delta R.T.: 0.000 min
Response: 22009230
Conc: 49.55 ng/ml



#44 AR-1268-4


R.T.: 8.136 min
Delta R.T.: 0.000 min
Response: 7174009
Conc: 50.83 ng/ml



#45 AR-1268-5

R.T.: 8.473 min
Delta R.T.: 0.000 min
Response: 146850912
Conc: 45.61 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC050



#45 AR-1268-5

R.T.: 8.428 min
Delta R.T.: 0.000 min
Response: 40130945
Conc: 50.98 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110377.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 18:09
 Operator : YP/AJ
 Sample : P0041025ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO041025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 18:45:43 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 18:44:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.686	3.685	421.6E6	244.8E6	48.193	49.064
2) SA Decachlor...	8.731	8.685	398.8E6	93206294	50.509	48.419
Target Compounds						
3) L1 AR-1016-1	4.778	4.766	160.3E6	85578058	488.595	487.598
4) L1 AR-1016-2	4.797	4.784	224.6E6	123.9E6	493.781	492.855
5) L1 AR-1016-3	4.853	4.960	157.2E6	66551398	487.462	490.391
6) L1 AR-1016-4	4.973	5.002	122.3E6	55570470	491.017	485.892
7) L1 AR-1016-5	5.231	5.215	133.8E6	73527628	497.315	493.229
31) L7 AR-1260-1	6.270	6.246	237.6E6	121.3E6	502.988	493.308
32) L7 AR-1260-2	6.460	6.434	287.8E6	140.8E6	490.780	482.474
33) L7 AR-1260-3	6.827	6.586	245.6E6	132.0E6	498.218	487.409
34) L7 AR-1260-4	7.087	7.058	213.8E6	98678673	503.635	491.819
35) L7 AR-1260-5	7.329	7.298	535.6E6	226.2E6	513.241	492.883



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110377.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 18:09
Operator : YP/AJ
Sample : PO041025ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO041025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 18:45:43 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 18:44:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.686 min
Delta R.T.: -0.002 min
Response: 421644591
Conc: 48.19 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: -0.001 min
Response: 244816340
Conc: 49.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.731 min
Delta R.T.: 0.000 min
Response: 398769118
Conc: 50.51 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.685 min
Delta R.T.: 0.000 min
Response: 93206294
Conc: 48.42 ng/ml



#3 AR-1016-1


R.T.: 4.778 min
Delta R.T.: 0.000 min
Response: 160329385
Conc: 488.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025




#3 AR-1016-1

R.T.: 4.766 min
Delta R.T.: 0.000 min
Response: 85578058
Conc: 487.60 ng/ml




#4 AR-1016-2

R.T.: 4.797 min
Delta R.T.: -0.002 min
Response: 224644909
Conc: 493.78 ng/ml



#4 AR-1016-2


R.T.: 4.784 min
Delta R.T.: 0.000 min
Response: 123935759
Conc: 492.85 ng/ml



#5 AR-1016-3


R.T.: 4.853 min
Delta R.T.: -0.002 min
Response: 157219813
Conc: 487.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025




#5 AR-1016-3

R.T.: 4.960 min
Delta R.T.: 0.000 min
Response: 66551398
Conc: 490.39 ng/ml




#6 AR-1016-4

R.T.: 4.973 min
Delta R.T.: -0.002 min
Response: 122344041
Conc: 491.02 ng/ml



#6 AR-1016-4


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 55570470
Conc: 485.89 ng/ml



#7 AR-1016-5


R.T.: 5.231 min
Delta R.T.: 0.000 min
Response: 133783641
Conc: 497.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025




#7 AR-1016-5

R.T.: 5.215 min
Delta R.T.: 0.000 min
Response: 73527628
Conc: 493.23 ng/ml




#31 AR-1260-1

R.T.: 6.270 min
Delta R.T.: -0.002 min
Response: 237634519
Conc: 502.99 ng/ml



#31 AR-1260-1


R.T.: 6.246 min
Delta R.T.: 0.000 min
Response: 121335820
Conc: 493.31 ng/ml



#32 AR-1260-2


R.T.: 6.460 min
Delta R.T.: 0.000 min
Response: 287805163
Conc: 490.78 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025




#32 AR-1260-2

R.T.: 6.434 min
Delta R.T.: 0.000 min
Response: 140762216
Conc: 482.47 ng/ml




#33 AR-1260-3

R.T.: 6.827 min
Delta R.T.: -0.001 min
Response: 245558046
Conc: 498.22 ng/ml



#33 AR-1260-3


R.T.: 6.586 min
Delta R.T.: 0.000 min
Response: 132031272
Conc: 487.41 ng/ml



#34 AR-1260-4


R.T.: 7.087 min
Delta R.T.: -0.001 min
Response: 213754206
Conc: 503.63 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025




#34 AR-1260-4

R.T.: 7.058 min
Delta R.T.: 0.000 min
Response: 98678673
Conc: 491.82 ng/ml



#35 AR-1260-5

R.T.: 7.329 min
Delta R.T.: 0.000 min
Response: 535552109
Conc: 513.24 ng/ml



#35 AR-1260-5

R.T.: 7.298 min
Delta R.T.: 0.000 min
Response: 226163331
Conc: 492.88 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110378.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 18:46
 Operator : YP/AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO041025AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 19:22:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 18:44:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.684	443.8E6	250.6E6	50.720	50.218
2) SA Decachlor...	8.732	8.684	397.2E6	92891086	50.313	48.255
Target Compounds						
16) L4 AR-1242-1	4.778	4.765	140.4E6	74475295	505.918	498.154
17) L4 AR-1242-2	4.797	4.784	194.4E6	106.8E6	503.798	500.644
18) L4 AR-1242-3	4.854	4.959	138.3E6	57809929	496.987	500.162
19) L4 AR-1242-4	4.975	5.043	107.3E6	59712665	507.656	497.194
20) L4 AR-1242-5	5.627	5.564	113.7E6	71571546	502.212	493.267



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110378.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 18:46
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 19:22:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 18:44:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.687 min
Delta R.T.: -0.001 min
Response: 443754118
Conc: 50.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: -0.002 min
Response: 250576991
Conc: 50.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 397220716
Conc: 50.31 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.684 min
Delta R.T.: 0.000 min
Response: 92891086
Conc: 48.26 ng/ml



#16 AR-1242-1


R.T.: 4.778 min
Delta R.T.: -0.002 min
Response: 140409130
Conc: 505.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1242




#16 AR-1242-1

R.T.: 4.765 min
Delta R.T.: -0.001 min
Response: 74475295
Conc: 498.15 ng/ml




#17 AR-1242-2

R.T.: 4.797 min
Delta R.T.: -0.001 min
Response: 194352025
Conc: 503.80 ng/ml



#17 AR-1242-2


R.T.: 4.784 min
Delta R.T.: -0.001 min
Response: 106818938
Conc: 500.64 ng/ml



#18 AR-1242-3


R.T.: 4.854 min
Delta R.T.: -0.001 min
Response: 138262793
Conc: 496.99 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1242




#18 AR-1242-3

R.T.: 4.959 min
Delta R.T.: -0.001 min
Response: 57809929
Conc: 500.16 ng/ml




#19 AR-1242-4

R.T.: 4.975 min
Delta R.T.: 0.000 min
Response: 107301967
Conc: 507.66 ng/ml



#19 AR-1242-4


R.T.: 5.043 min
Delta R.T.: -0.001 min
Response: 59712665
Conc: 497.19 ng/ml



#20 AR-1242-5

R.T.: 5.627 min
Delta R.T.: -0.001 min
Response: 113697523
Conc: 502.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1242



#20 AR-1242-5

R.T.: 5.564 min
Delta R.T.: -0.001 min
Response: 71571546
Conc: 493.27 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110379.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 19:22
 Operator : YP/AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO041025AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 01:39:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 11 01:37:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	439.1E6	246.8E6	51.110	51.053
2) SA Decachlor...	8.732	8.685	400.1E6	92657947	50.360	49.279
Target Compounds						
21) L5 AR-1248-1	4.779	4.765	106.0E6	57397514	498.124	498.537
22) L5 AR-1248-2	5.017	5.002	147.4E6	80796045	498.235	496.039
23) L5 AR-1248-3	5.232	5.044	183.5E6	86306627	495.938	490.780
24) L5 AR-1248-4	5.587	5.214	257.3E6	100.2E6	494.824	491.802
25) L5 AR-1248-5	5.627	5.606	183.1E6	97158423	495.645	492.201



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110379.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 19:22
Operator : YP/AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 33 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 01:39:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 01:37:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 439149263
Conc: 51.11 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 246849981
Conc: 51.05 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 400112620
Conc: 50.36 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.685 min
Delta R.T.: 0.000 min
Response: 92657947
Conc: 49.28 ng/ml



#21 AR-1248-1


R.T.: 4.779 min
Delta R.T.: 0.000 min
Response: 106018952
Conc: 498.12 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1248




#21 AR-1248-1

R.T.: 4.765 min
Delta R.T.: 0.000 min
Response: 57397514
Conc: 498.54 ng/ml




#22 AR-1248-2

R.T.: 5.017 min
Delta R.T.: 0.000 min
Response: 147440286
Conc: 498.24 ng/ml



#22 AR-1248-2


R.T.: 5.002 min
Delta R.T.: 0.000 min
Response: 80796045
Conc: 496.04 ng/ml



#23 AR-1248-3


R.T.: 5.232 min
Delta R.T.: 0.000 min
Response: 183478000
Conc: 495.94 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1248




#23 AR-1248-3

R.T.: 5.044 min
Delta R.T.: 0.000 min
Response: 86306627
Conc: 490.78 ng/ml




#24 AR-1248-4

R.T.: 5.587 min
Delta R.T.: 0.000 min
Response: 257323549
Conc: 494.82 ng/ml



#24 AR-1248-4


R.T.: 5.214 min
Delta R.T.: 0.000 min
Response: 100169669
Conc: 491.80 ng/ml



#25 AR-1248-5

R.T.: 5.627 min
Delta R.T.: 0.000 min
Response: 183103878
Conc: 495.64 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1248



#25 AR-1248-5

R.T.: 5.606 min
Delta R.T.: 0.000 min
Response: 97158423
Conc: 492.20 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110380.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 19:58
 Operator : YP/AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO041025AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 02:09:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 11 02:07:00 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.686	3.684	444.9E6	251.2E6	51.100	51.235
2) SA Decachlor...	8.730	8.683	404.8E6	94520608	49.556	49.296
Target Compounds						
26) L6 AR-1254-1	5.585	5.565	274.4E6	147.3E6	491.771	488.501
27) L6 AR-1254-2	5.734	5.713	239.5E6	129.3E6	493.094	488.639
28) L6 AR-1254-3	6.139	6.115	390.3E6	201.0E6	498.314	498.355
29) L6 AR-1254-4	6.369	6.343	241.8E6	114.4E6	495.833	498.111
30) L6 AR-1254-5	6.788	6.760	346.0E6	164.5E6	493.657	494.897



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110380.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 19:58
Operator : YP/AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 02:09:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 02:07:00 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.686 min
Delta R.T.: -0.002 min
Response: 444900239
Conc: 51.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: 0.000 min
Response: 251232379
Conc: 51.24 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.730 min
Delta R.T.: -0.003 min
Response: 404777959
Conc: 49.56 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.683 min
Delta R.T.: 0.000 min
Response: 94520608
Conc: 49.30 ng/ml



#26 AR-1254-1


R.T.: 5.585 min
Delta R.T.: -0.002 min
Response: 274402754
Conc: 491.77 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1254




#26 AR-1254-1

R.T.: 5.565 min
Delta R.T.: 0.000 min
Response: 147282616
Conc: 488.50 ng/ml




#27 AR-1254-2

R.T.: 5.734 min
Delta R.T.: -0.002 min
Response: 239497012
Conc: 493.09 ng/ml



#27 AR-1254-2


R.T.: 5.713 min
Delta R.T.: 0.000 min
Response: 129335526
Conc: 488.64 ng/ml



#28 AR-1254-3


R.T.: 6.139 min
Delta R.T.: -0.002 min
Response: 390290001
Conc: 498.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1254




#28 AR-1254-3

R.T.: 6.115 min
Delta R.T.: 0.000 min
Response: 201044539
Conc: 498.36 ng/ml




#29 AR-1254-4

R.T.: 6.369 min
Delta R.T.: -0.001 min
Response: 241770989
Conc: 495.83 ng/ml



#29 AR-1254-4


R.T.: 6.343 min
Delta R.T.: 0.000 min
Response: 114422519
Conc: 498.11 ng/ml



#30 AR-1254-5

R.T.: 6.788 min
Delta R.T.: -0.002 min
Response: 346004179
Conc: 493.66 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1254



#30 AR-1254-5

R.T.: 6.760 min
Delta R.T.: 0.000 min
Response: 164450224
Conc: 494.90 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : P0110381.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 20:35
 Operator : YP/AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO041025AR1268

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 11 02:13:00 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 11 02:12:26 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.688	3.685	447.9E6	251.5E6	52.934	52.706
2) SA Decachlor...	8.733	8.685	720.6E6	163.2E6	50.087	49.769
Target Compounds						
41) L9 AR-1268-1	7.615	7.581	703.2E6	273.9E6	499.070	501.677
42) L9 AR-1268-2	7.680	7.646	652.7E6	249.2E6	504.213	500.302
43) L9 AR-1268-3	7.887	7.852	529.5E6	185.9E6	503.681	501.520
44) L9 AR-1268-4	8.176	8.137	229.7E6	72282546	519.956	512.014
45) L9 AR-1268-5	8.472	8.430	1642.7E6	393.7E6	499.897	496.919



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110381.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 20:35
Operator : YP/AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1268

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 11 02:13:00 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 02:12:26 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.688 min
Delta R.T.: 0.000 min
Response: 447894257
Conc: 52.93 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1268




#1 Tetrachloro-m-xylene

R.T.: 3.685 min
Delta R.T.: 0.000 min
Response: 251543531
Conc: 52.71 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.733 min
Delta R.T.: 0.001 min
Response: 720557931
Conc: 50.09 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.685 min
Delta R.T.: 0.001 min
Response: 163169409
Conc: 49.77 ng/ml



#41 AR-1268-1


R.T.: 7.615 min
Delta R.T.: 0.000 min
Response: 703246486
Conc: 499.07 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1268




#41 AR-1268-1

R.T.: 7.581 min
Delta R.T.: 0.001 min
Response: 273859995
Conc: 501.68 ng/ml




#42 AR-1268-2

R.T.: 7.680 min
Delta R.T.: 0.000 min
Response: 652667543
Conc: 504.21 ng/ml



#42 AR-1268-2


R.T.: 7.646 min
Delta R.T.: 0.000 min
Response: 249160771
Conc: 500.30 ng/ml



#43 AR-1268-3


R.T.: 7.887 min
Delta R.T.: 0.000 min
Response: 529533142
Conc: 503.68 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1268




#43 AR-1268-3

R.T.: 7.852 min
Delta R.T.: 0.000 min
Response: 185920985
Conc: 501.52 ng/ml




#44 AR-1268-4

R.T.: 8.176 min
Delta R.T.: 0.000 min
Response: 229738508
Conc: 519.96 ng/ml



#44 AR-1268-4


R.T.: 8.137 min
Delta R.T.: 0.000 min
Response: 72282546
Conc: 512.01 ng/ml



#45 AR-1268-5

R.T.: 8.472 min
Delta R.T.: 0.000 min
Response: 1642717623
Conc: 499.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO041025AR1268



#45 AR-1268-5

R.T.: 8.430 min
Delta R.T.: 0.000 min
Response: 393682568
Conc: 496.92 ng/ml

RETENTION TIMES OF INITIAL CALIBRATION

Contract: POWE02
Lab Code: CHEM **Case No.:** Q1889 **SAS No.:** Q1889 **SDG NO.:** Q1889
Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

LAB FILE ID:	RT 1000 = <u>PP071389.D</u>	RT 750 = <u>PP071390.D</u>
RT 500 = <u>PP071391.D</u>	RT 250 = <u>PP071392.D</u>	RT 050 = <u>PP071393.D</u>

COMPOUND		RT 1000	RT 750	RT 500	RT 250	RT 050	MEAN RT	RT WINDOW	
								FROM	TO
Aroclor-1016-1	(1)	5.66	5.67	5.67	5.67	5.67	5.67	5.57	5.77
Aroclor-1016-2	(2)	5.69	5.69	5.69	5.69	5.69	5.69	5.59	5.79
Aroclor-1016-3	(3)	5.75	5.75	5.75	5.75	5.75	5.75	5.65	5.85
Aroclor-1016-4	(4)	5.85	5.85	5.85	5.85	5.85	5.85	5.75	5.95
Aroclor-1016-5	(5)	6.14	6.14	6.15	6.14	6.14	6.14	6.04	6.24
Aroclor-1260-1	(1)	7.26	7.26	7.27	7.26	7.26	7.26	7.16	7.36
Aroclor-1260-2	(2)	7.51	7.51	7.52	7.52	7.52	7.52	7.42	7.62
Aroclor-1260-3	(3)	7.87	7.87	7.88	7.87	7.88	7.87	7.77	7.97
Aroclor-1260-4	(4)	8.10	8.10	8.10	8.10	8.10	8.10	8.00	8.20
Aroclor-1260-5	(5)	8.42	8.42	8.42	8.42	8.42	8.42	8.32	8.52
Decachlorobiphenyl		10.23	10.23	10.24	10.23	10.24	10.23	10.13	10.33
Tetrachloro-m-xylene		4.51	4.51	4.52	4.51	4.52	4.51	4.41	4.61
Aroclor-1254-1	(1)	6.52	6.52	6.52	6.52	6.52	6.52	6.42	6.62
Aroclor-1254-2	(2)	6.74	6.74	6.74	6.73	6.74	6.74	6.64	6.84
Aroclor-1254-3	(3)	7.10	7.10	7.10	7.10	7.10	7.10	7.00	7.20
Aroclor-1254-4	(4)	7.38	7.38	7.38	7.38	7.38	7.38	7.28	7.48
Aroclor-1254-5	(5)	7.80	7.80	7.80	7.80	7.80	7.80	7.70	7.90
Decachlorobiphenyl		10.24	10.24	10.24	10.23	10.24	10.24	10.14	10.34
Tetrachloro-m-xylene		4.52	4.52	4.52	4.51	4.51	4.51	4.41	4.61

Contract:	<u>POWE02</u>						
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1889</u>	SAS No.:	<u>Q1889</u>	SDG NO.:	<u>Q1889</u>
Instrument ID:	<u>ECD_P</u>	Calibration Date(s):		<u>04/22/2025</u>	<u>04/22/2025</u>		
		Calibration Times:		<u>10:29</u>	<u>17:49</u>		

LAB FILE ID:		RT 1000 =	<u>PP071389.D</u>	RT 750 =	<u>PP071390.D</u>
RT 500 =	PP071391.D	RT 250 =	PP071392.D	RT 050 =	PP071393.D

[illegible]

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02
Lab Code: CHEM **Case No.:** Q1889 **SAS No.:** Q1889 **SDG NO.:** Q1889
Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49

GC Column: ZB-MR1 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 =	PP071389.D	CF 750 =	PP071390.D			
CF 500 =		PP071391.D	CF 250 =	PP071392.D	CF 050 =	PP071393.D		
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	64709894	67652695	70083532	71128656	61674340	67049823	6
Aroclor-1016-2	(2)	98977911	102364897	107369074	108813892	95615840	102628323	5
Aroclor-1016-3	(3)	59905708	62127143	64296308	65849928	56056440	61647105	6
Aroclor-1016-4	(4)	50563747	52419789	53619076	54941428	44946700	51298148	8
Aroclor-1016-5	(5)	47772963	49344036	50670668	50888564	42515680	48238382	7
Aroclor-1260-1	(1)	90879847	94857525	98814986	99659744	95910940	96024608	4
Aroclor-1260-2	(2)	137163459	142386411	147498372	149887700	138450980	143077384	4
Aroclor-1260-3	(3)	110350751	114445481	117702804	126515524	91505860	112104084	12
Aroclor-1260-4	(4)	105570732	110780085	111211068	131260612	105501660	112864831	9
Aroclor-1260-5	(5)	220536830	229070229	228919678	255108180	199231240	226573231	9
Decachlorobiphenyl		1409816860	1467002307	1522583160	1539948480	1300700600	1448010281	7
Tetrachloro-m-xylene		1947183160	2005072840	2066603680	2100352680	1793966200	1982635712	6
Aroclor-1254-1	(1)	78574372	81246615	85577334	88598188	75759400	81951182	6
Aroclor-1254-2	(2)	121624435	125816099	132146328	137115708	118678260	127076166	6
Aroclor-1254-3	(3)	123432630	127799839	132981654	137623252	117036460	127774767	6
Aroclor-1254-4	(4)	112512019	116259037	121737436	124392856	112665620	117513394	5
Aroclor-1254-5	(5)	106399089	105922557	113785120	116153044	96391680	107730298	7
Decachlorobiphenyl		1410101810	1469202520	1526389740	1549706840	1328725400	1456825262	6
Tetrachloro-m-xylene		1888241120	1936990333	2012368600	2071565920	1722515200	1926336235	7

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: POWE02
Lab Code: CHEM **Case No.:** Q1889 **SAS No.:** Q1889 **SDG NO.:** Q1889
Instrument ID: ECD_P **Calibration Date(s):** 04/22/2025 04/22/2025
Calibration Times: 10:29 17:49

GC Column: ZB-MR2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 1000 = <u>PP071389.D</u>		CF 750 = <u>PP071390.D</u>				
CF 500 = <u>PP071391.D</u>		CF 250 = <u>PP071392.D</u>		CF 050 = <u>PP071393.D</u>				
COMPOUND		CF 1000	CF 750	CF 500	CF 250	CF 050	CF	% RSD
Aroclor-1016-1	(1)	48121828	49414308	53806128	56180596	60145020	53533576	9
Aroclor-1016-2	(2)	69736709	72096436	77213744	78520924	84250740	76363711	7
Aroclor-1016-3	(3)	38517048	40455577	43831736	45082616	43165080	42210411	6
Aroclor-1016-4	(4)	31206636	32814844	36123020	37095092	35190100	34485938	7
Aroclor-1016-5	(5)	39458687	41985771	45850608	46345748	48595640	44447291	8
Aroclor-1260-1	(1)	63219917	66123699	73727666	76350368	82074520	72299234	11
Aroclor-1260-2	(2)	77338708	77548395	89921718	91492956	101005300	87461415	12
Aroclor-1260-3	(3)	72062678	70944365	81698234	79344004	87888600	78387576	9
Aroclor-1260-4	(4)	57333188	56779915	65331750	68083052	68025140	63110609	9
Aroclor-1260-5	(5)	143809415	139161576	155842678	157512280	157560640	150777318	6
Decachlorobiphenyl		831147600	875337973	902319620	934998560	841460400	877052831	5
Tetrachloro-m-xylene		1380433570	1378340613	1415112020	1448341640	1420974200	1408640409	2
Aroclor-1254-1	(1)	81690839	87361264	89755404	97285536	100908960	91400401	8
Aroclor-1254-2	(2)	70045476	75469849	77317914	83154376	87573280	78712179	9
Aroclor-1254-3	(3)	109725744	114103376	120096818	124380164	124235440	118508308	5
Aroclor-1254-4	(4)	70169900	74068720	77810162	81877428	81901880	77165618	7
Aroclor-1254-5	(5)	92229111	96109599	104690794	107384400	106924600	101467701	7
Decachlorobiphenyl		846307500	875335867	916104900	955261880	868223800	892246789	5
Tetrachloro-m-xylene		1390553300	1501013893	1450901500	1518608160	1468503000	1465915971	3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071389.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 10:29
 Operator : YP\AJ
 Sample : AR1660ICC1000
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:18:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.511	3.811	194.7E6	138.0E6	97.025	98.760
2) SA Decachlor...	10.230	8.849	141.0E6	83114760	96.154	95.894
Target Compounds						
3) L1 AR-1016-1	5.664	4.897	64709894	48121828	960.134	944.232
4) L1 AR-1016-2	5.686	4.915	98977911	69736709	959.335	949.119
5) L1 AR-1016-3	5.749	5.092	59905708	38517048	964.650	935.461
6) L1 AR-1016-4	5.846	5.134	50563747	31206636	970.673	926.980
7) L1 AR-1016-5	6.139	5.349	47772963	39458687	970.565	925.074
31) L7 AR-1260-1	7.259	6.384	90879847	63219917	958.169	923.272
32) L7 AR-1260-2	7.512	6.572	137.2E6	77338708	963.694	924.770
33) L7 AR-1260-3	7.872	6.726	110.4E6	72062678	967.762	937.334
34) L7 AR-1260-4	8.095	7.197	105.6E6	57333188	973.982	934.793
35) L7 AR-1260-5	8.416	7.439	220.5E6	143.8E6	981.349	959.843



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071389.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:29
Operator : YP\AJ
Sample : AR1660ICC1000
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:18:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.511 min
Delta R.T.: 0.000 min
Response: 194718316
Conc: 97.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.811 min
Delta R.T.: 0.000 min
Response: 138043357
Conc: 98.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.230 min
Delta R.T.: 0.000 min
Response: 140981686
Conc: 96.15 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 83114760
Conc: 95.89 ng/ml



#3 AR-1016-1


R.T.: 5.664 min
Delta R.T.: 0.000 min
Response: 64709894
Conc: 960.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#3 AR-1016-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 48121828
Conc: 944.23 ng/ml




#4 AR-1016-2

R.T.: 5.686 min
Delta R.T.: 0.000 min
Response: 98977911
Conc: 959.33 ng/ml



#4 AR-1016-2


R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 69736709
Conc: 949.12 ng/ml



#5 AR-1016-3


R.T.: 5.749 min
Delta R.T.: 0.000 min
Response: 59905708
Conc: 964.65 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#5 AR-1016-3

R.T.: 5.092 min
Delta R.T.: 0.000 min
Response: 38517048
Conc: 935.46 ng/ml




#6 AR-1016-4

R.T.: 5.846 min
Delta R.T.: 0.000 min
Response: 50563747
Conc: 970.67 ng/ml



#6 AR-1016-4


R.T.: 5.134 min
Delta R.T.: 0.000 min
Response: 31206636
Conc: 926.98 ng/ml



#7 AR-1016-5


R.T.: 6.139 min
Delta R.T.: 0.000 min
Response: 47772963
Conc: 970.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#7 AR-1016-5

R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 39458687
Conc: 925.07 ng/ml




#31 AR-1260-1

R.T.: 7.259 min
Delta R.T.: 0.000 min
Response: 90879847
Conc: 958.17 ng/ml



#31 AR-1260-1


R.T.: 6.384 min
Delta R.T.: 0.000 min
Response: 63219917
Conc: 923.27 ng/ml



#32 AR-1260-2


R.T.: 7.512 min
Delta R.T.: 0.000 min
Response: 137163459
Conc: 963.69 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#32 AR-1260-2

R.T.: 6.572 min
Delta R.T.: 0.000 min
Response: 77338708
Conc: 924.77 ng/ml




#33 AR-1260-3

R.T.: 7.872 min
Delta R.T.: 0.000 min
Response: 110350751
Conc: 967.76 ng/ml



#33 AR-1260-3


R.T.: 6.726 min
Delta R.T.: 0.000 min
Response: 72062678
Conc: 937.33 ng/ml



#34 AR-1260-4


R.T.: 8.095 min
Delta R.T.: 0.000 min
Response: 105570732
Conc: 973.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC1000




#34 AR-1260-4

R.T.: 7.197 min
Delta R.T.: 0.000 min
Response: 57333188
Conc: 934.79 ng/ml



#35 AR-1260-5

R.T.: 8.416 min
Delta R.T.: 0.000 min
Response: 220536830
Conc: 981.35 ng/ml



#35 AR-1260-5

R.T.: 7.439 min
Delta R.T.: 0.000 min
Response: 143809415
Conc: 959.84 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071390.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 10:45
 Operator : YP\AJ
 Sample : AR1660ICC750
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:20:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.811	150.4E6	103.4E6	74.955	74.302
2) SA Decachlor...	10.233	8.848	110.0E6	65650348	75.027	75.495
Target Compounds						
3) L1 AR-1016-1	5.668	4.897	50739521	37060731	751.897	734.641
4) L1 AR-1016-2	5.690	4.915	76773673	54072327	746.071	740.558
5) L1 AR-1016-3	5.752	5.092	46595357	30341683	750.210	741.220
6) L1 AR-1016-4	5.849	5.134	39314842	24611133	753.145	737.269
7) L1 AR-1016-5	6.142	5.349	37008027	31489328	751.240	742.118
31) L7 AR-1260-1	7.261	6.383	71143144	49592774	750.053	732.641
32) L7 AR-1260-2	7.514	6.572	106.8E6	58161296	750.195	712.735
33) L7 AR-1260-3	7.873	6.725	85834111	53208274	751.834	710.374
34) L7 AR-1260-4	8.097	7.197	83085064	42584936	760.941	711.945
35) L7 AR-1260-5	8.418	7.438	171.8E6	104.4E6	759.599	713.546



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071390.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:45
Operator : YP\AJ
Sample : AR1660ICC750
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:20:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 150380463
Conc: 74.95 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.811 min
Delta R.T.: 0.000 min
Response: 103375546
Conc: 74.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: 0.000 min
Response: 110025173
Conc: 75.03 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.848 min
Delta R.T.: 0.000 min
Response: 65650348
Conc: 75.49 ng/ml



#3 AR-1016-1


R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 50739521
Conc: 751.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#3 AR-1016-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 37060731
Conc: 734.64 ng/ml




#4 AR-1016-2

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 76773673
Conc: 746.07 ng/ml



#4 AR-1016-2


R.T.: 4.915 min
Delta R.T.: 0.000 min
Response: 54072327
Conc: 740.56 ng/ml



#5 AR-1016-3


R.T.: 5.752 min
Delta R.T.: 0.000 min
Response: 46595357
Conc: 750.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#5 AR-1016-3

R.T.: 5.092 min
Delta R.T.: 0.000 min
Response: 30341683
Conc: 741.22 ng/ml




#6 AR-1016-4

R.T.: 5.849 min
Delta R.T.: 0.000 min
Response: 39314842
Conc: 753.15 ng/ml



#6 AR-1016-4


R.T.: 5.134 min
Delta R.T.: 0.000 min
Response: 24611133
Conc: 737.27 ng/ml



#7 AR-1016-5


R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 37008027
Conc: 751.24 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#7 AR-1016-5

R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 31489328
Conc: 742.12 ng/ml




#31 AR-1260-1

R.T.: 7.261 min
Delta R.T.: 0.000 min
Response: 71143144
Conc: 750.05 ng/ml



#31 AR-1260-1


R.T.: 6.383 min
Delta R.T.: 0.000 min
Response: 49592774
Conc: 732.64 ng/ml



#32 AR-1260-2


R.T.: 7.514 min
Delta R.T.: 0.000 min
Response: 106789808
Conc: 750.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#32 AR-1260-2

R.T.: 6.572 min
Delta R.T.: 0.000 min
Response: 58161296
Conc: 712.74 ng/ml




#33 AR-1260-3

R.T.: 7.873 min
Delta R.T.: 0.000 min
Response: 85834111
Conc: 751.83 ng/ml



#33 AR-1260-3


R.T.: 6.725 min
Delta R.T.: 0.000 min
Response: 53208274
Conc: 710.37 ng/ml



#34 AR-1260-4


R.T.: 8.097 min
Delta R.T.: 0.000 min
Response: 83085064
Conc: 760.94 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC750




#34 AR-1260-4

R.T.: 7.197 min
Delta R.T.: 0.000 min
Response: 42584936
Conc: 711.94 ng/ml



#35 AR-1260-5

R.T.: 8.418 min
Delta R.T.: 0.000 min
Response: 171802672
Conc: 759.60 ng/ml



#35 AR-1260-5

R.T.: 7.438 min
Delta R.T.: 0.000 min
Response: 104371182
Conc: 713.55 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071391.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:02
 Operator : YP\AJ
 Sample : AR1660ICC500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:15:13 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:14:53 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.517	3.812	103.3E6	70755601	50.000	50.000
2) SA Decachlor...	10.237	8.849	76129158	45115981	50.000	50.000
Target Compounds						
3) L1 AR-1016-1	5.671	4.898	35041766	26903064	500.000	500.000
4) L1 AR-1016-2	5.693	4.916	53684537	38606872	500.000	500.000
5) L1 AR-1016-3	5.754	5.093	32148154	21915868	500.000	500.000
6) L1 AR-1016-4	5.852	5.136	26809538	18061510	500.000	500.000
7) L1 AR-1016-5	6.145	5.350	25335334	22925304	500.000	500.000
31) L7 AR-1260-1	7.265	6.385	49407493	36863833	500.000	500.000
32) L7 AR-1260-2	7.518	6.573	73749186	44960859	500.000	500.000
33) L7 AR-1260-3	7.877	6.727	58851402	40849117	500.000	500.000
34) L7 AR-1260-4	8.102	7.198	55605534	32665875	500.000	500.000
35) L7 AR-1260-5	8.422	7.439	114.5E6	77921339	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071391.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:02
Operator : YP\AJ
Sample : AR1660ICC500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:15:13 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:14:53 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.517 min
Delta R.T.: 0.000 min
Response: 103330184
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 70755601
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.237 min
Delta R.T.: 0.000 min
Response: 76129158
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 45115981
Conc: 50.00 ng/ml



#3 AR-1016-1


R.T.: 5.671 min
Delta R.T.: 0.000 min
Response: 35041766
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#3 AR-1016-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 26903064
Conc: 500.00 ng/ml




#4 AR-1016-2

R.T.: 5.693 min
Delta R.T.: 0.000 min
Response: 53684537
Conc: 500.00 ng/ml



#4 AR-1016-2


R.T.: 4.916 min
Delta R.T.: 0.000 min
Response: 38606872
Conc: 500.00 ng/ml



#5 AR-1016-3


R.T.: 5.754 min
Delta R.T.: 0.000 min
Response: 32148154
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#5 AR-1016-3

R.T.: 5.093 min
Delta R.T.: 0.000 min
Response: 21915868
Conc: 500.00 ng/ml




#6 AR-1016-4

R.T.: 5.852 min
Delta R.T.: 0.000 min
Response: 26809538
Conc: 500.00 ng/ml



#6 AR-1016-4


R.T.: 5.136 min
Delta R.T.: 0.000 min
Response: 18061510
Conc: 500.00 ng/ml



#7 AR-1016-5


R.T.: 6.145 min
Delta R.T.: 0.000 min
Response: 25335334
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#7 AR-1016-5

R.T.: 5.350 min
Delta R.T.: 0.000 min
Response: 22925304
Conc: 500.00 ng/ml




#31 AR-1260-1

R.T.: 7.265 min
Delta R.T.: 0.000 min
Response: 49407493
Conc: 500.00 ng/ml



#31 AR-1260-1


R.T.: 6.385 min
Delta R.T.: 0.000 min
Response: 36863833
Conc: 500.00 ng/ml



#32 AR-1260-2


R.T.: 7.518 min
Delta R.T.: 0.000 min
Response: 73749186
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#32 AR-1260-2

R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 44960859
Conc: 500.00 ng/ml




#33 AR-1260-3

R.T.: 7.877 min
Delta R.T.: 0.000 min
Response: 58851402
Conc: 500.00 ng/ml



#33 AR-1260-3


R.T.: 6.727 min
Delta R.T.: 0.000 min
Response: 40849117
Conc: 500.00 ng/ml



#34 AR-1260-4


R.T.: 8.102 min
Delta R.T.: 0.000 min
Response: 55605534
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC500




#34 AR-1260-4

R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 32665875
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 8.422 min
Delta R.T.: 0.000 min
Response: 114459839
Conc: 500.00 ng/ml



#35 AR-1260-5

R.T.: 7.439 min
Delta R.T.: 0.000 min
Response: 77921339
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071392.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:18
 Operator : YP\AJ
 Sample : AR1660ICC250
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:30:03 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:29:52 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	52508817	36208541	25.869	25.761
2) SA Decachlor...	10.233	8.850	38498712	23374964	25.928	26.384
Target Compounds						
3) L1 AR-1016-1	5.668	4.899	17782164	14045149	259.997	270.720
4) L1 AR-1016-2	5.689	4.917	27203473	19630231	260.616	263.876
5) L1 AR-1016-3	5.752	5.094	16462482	11270654	261.124	268.530
6) L1 AR-1016-4	5.849	5.136	13735357	9273773	259.716	270.294
7) L1 AR-1016-5	6.142	5.351	12722141	11586437	256.138	266.906
31) L7 AR-1260-1	7.261	6.386	24914936	19087592	259.387	273.244
32) L7 AR-1260-2	7.515	6.574	37471925	22873239	259.800	272.056
33) L7 AR-1260-3	7.873	6.727	31628881	19836001	269.748	260.958
34) L7 AR-1260-4	8.098	7.199	32815153	17020763	286.081	275.052
35) L7 AR-1260-5	8.417	7.441	63777045	39378070	273.242	264.138



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071392.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:18
Operator : YP\AJ
Sample : AR1660ICC250
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:30:03 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:29:52 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 52508817
Conc: 25.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 36208541
Conc: 25.76 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: 0.000 min
Response: 38498712
Conc: 25.93 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 23374964
Conc: 26.38 ng/ml



#3 AR-1016-1


R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 17782164
Conc: 260.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#3 AR-1016-1

R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 14045149
Conc: 270.72 ng/ml




#4 AR-1016-2

R.T.: 5.689 min
Delta R.T.: 0.000 min
Response: 27203473
Conc: 260.62 ng/ml



#4 AR-1016-2


R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 19630231
Conc: 263.88 ng/ml



#5 AR-1016-3


R.T.: 5.752 min
Delta R.T.: 0.000 min
Response: 16462482
Conc: 261.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#5 AR-1016-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 11270654
Conc: 268.53 ng/ml




#6 AR-1016-4

R.T.: 5.849 min
Delta R.T.: 0.000 min
Response: 13735357
Conc: 259.72 ng/ml



#6 AR-1016-4


R.T.: 5.136 min
Delta R.T.: 0.000 min
Response: 9273773
Conc: 270.29 ng/ml



#7 AR-1016-5


R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 12722141
Conc: 256.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#7 AR-1016-5

R.T.: 5.351 min
Delta R.T.: 0.000 min
Response: 11586437
Conc: 266.91 ng/ml




#31 AR-1260-1

R.T.: 7.261 min
Delta R.T.: 0.000 min
Response: 24914936
Conc: 259.39 ng/ml



#31 AR-1260-1


R.T.: 6.386 min
Delta R.T.: 0.000 min
Response: 19087592
Conc: 273.24 ng/ml



#32 AR-1260-2


R.T.: 7.515 min
Delta R.T.: 0.000 min
Response: 37471925
Conc: 259.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#32 AR-1260-2

R.T.: 6.574 min
Delta R.T.: 0.000 min
Response: 22873239
Conc: 272.06 ng/ml




#33 AR-1260-3

R.T.: 7.873 min
Delta R.T.: 0.000 min
Response: 31628881
Conc: 269.75 ng/ml



#33 AR-1260-3


R.T.: 6.727 min
Delta R.T.: 0.000 min
Response: 19836001
Conc: 260.96 ng/ml



#34 AR-1260-4


R.T.: 8.098 min
Delta R.T.: 0.000 min
Response: 32815153
Conc: 286.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC250




#34 AR-1260-4

R.T.: 7.199 min
Delta R.T.: 0.000 min
Response: 17020763
Conc: 275.05 ng/ml



#35 AR-1260-5

R.T.: 8.417 min
Delta R.T.: 0.000 min
Response: 63777045
Conc: 273.24 ng/ml



#35 AR-1260-5

R.T.: 7.441 min
Delta R.T.: 0.000 min
Response: 39378070
Conc: 264.14 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071393.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:34
 Operator : YP\AJ
 Sample : AR1660ICC050
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 11:47:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 11:46:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	8969831	7104871	4.524	5.044
2) SA Decachlor...	10.235	8.850	6503503	4207302	4.491	4.797
Target Compounds						
3) L1 AR-1016-1	5.669	4.898	3083717	3007251	45.991	56.175
4) L1 AR-1016-2	5.691	4.916	4780792	4212537	46.584	55.164
5) L1 AR-1016-3	5.753	5.094	2802822	2158254	45.466	51.131
6) L1 AR-1016-4	5.851	5.135	2247335	1759505	43.809	51.021
7) L1 AR-1016-5	6.144	5.349	2125784	2429782	44.068	54.667
31) L7 AR-1260-1	7.262	6.384	4795547	4103726	49.941	56.760
32) L7 AR-1260-2	7.516	6.573	6922549	5050265	48.383	57.743
33) L7 AR-1260-3	7.875	6.726	4575293	4394430	40.810m	56.060 #
34) L7 AR-1260-4	8.098	7.197	5275083	3401257	46.807m	53.894
35) L7 AR-1260-5	8.417	7.439	9961562	7878032	43.897m	52.249

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071393.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:34
Operator : YP\AJ
Sample : AR1660ICC050
Misc :
ALS Vial : 7 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 11:47:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 11:46:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 8969831
Conc: 4.52 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 7104871
Conc: 5.04 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: 0.000 min
Response: 6503503
Conc: 4.49 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 4207302
Conc: 4.80 ng/ml




#3 AR-1016-1

R.T.: 5.669 min
Delta R.T.: 0.000 min
Response: 3083717
Conc: 45.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#3 AR-1016-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 3007251
Conc: 56.18 ng/ml




#4 AR-1016-2

R.T.: 5.691 min
Delta R.T.: 0.000 min
Response: 4780792
Conc: 46.58 ng/ml



#4 AR-1016-2

R.T.: 4.916 min
Delta R.T.: 0.000 min
Response: 4212537
Conc: 55.16 ng/ml




#5 AR-1016-3

R.T.: 5.753 min
Delta R.T.: 0.000 min
Response: 2802822
Conc: 45.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#5 AR-1016-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 2158254
Conc: 51.13 ng/ml




#6 AR-1016-4

R.T.: 5.851 min
Delta R.T.: 0.000 min
Response: 2247335
Conc: 43.81 ng/ml



#6 AR-1016-4

R.T.: 5.135 min
Delta R.T.: 0.000 min
Response: 1759505
Conc: 51.02 ng/ml




#7 AR-1016-5

R.T.: 6.144 min
Delta R.T.: 0.000 min
Response: 2125784
Conc: 44.07 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#7 AR-1016-5

R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 2429782
Conc: 54.67 ng/ml




#31 AR-1260-1

R.T.: 7.262 min
Delta R.T.: 0.000 min
Response: 4795547
Conc: 49.94 ng/ml



#31 AR-1260-1

R.T.: 6.384 min
Delta R.T.: 0.000 min
Response: 4103726
Conc: 56.76 ng/ml




#32 AR-1260-2

R.T.: 7.516 min
Delta R.T.: 0.000 min
Response: 6922549
Conc: 48.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#32 AR-1260-2

R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 5050265
Conc: 57.74 ng/ml




#33 AR-1260-3

R.T.: 7.875 min
Delta R.T.: 0.000 min
Response: 4575293
Conc: 40.81 ng/ml m



#33 AR-1260-3

R.T.: 6.726 min
Delta R.T.: 0.000 min
Response: 4394430
Conc: 56.06 ng/ml




#34 AR-1260-4

R.T.: 8.098 min
Delta R.T.: 0.000 min
Response: 5275083
Conc: 46.81 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#34 AR-1260-4

R.T.: 7.197 min
Delta R.T.: 0.000 min
Response: 3401257
Conc: 53.89 ng/ml



#35 AR-1260-5

R.T.: 8.417 min
Delta R.T.: 0.000 min
Response: 9961562
Conc: 43.90 ng/ml m



#35 AR-1260-5

R.T.: 7.439 min
Delta R.T.: 0.000 min
Response: 7878032
Conc: 52.25 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071394.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 11:51
 Operator : YP\AJ
 Sample : AR1221ICC500
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1221ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 12:26:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 12:26:37 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.812	102.0E6	70886063	50.000	50.000
2) SA Decachlor...	10.233	8.849	73129811	43684216	50.000	50.000
Target Compounds						
8) L2 AR-1221-1	4.715	4.022	12058427	10997465	500.000	500.000
9) L2 AR-1221-2	4.800	4.109	9008471	8280856	500.000	500.000
10) L2 AR-1221-3	4.876	4.185	28305615	24238496	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071394.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 11:51
Operator : YP\AJ
Sample : AR1221ICC500
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 12:26:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 12:26:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 101953072
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 70886063
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: 0.000 min
Response: 73129811
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 43684216
Conc: 50.00 ng/ml



#8 AR-1221-1


R.T.: 4.715 min
Delta R.T.: 0.000 min
Response: 12058427
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500




#8 AR-1221-1

R.T.: 4.022 min
Delta R.T.: 0.000 min
Response: 10997465
Conc: 500.00 ng/ml




#9 AR-1221-2

R.T.: 4.800 min
Delta R.T.: 0.000 min
Response: 9008471
Conc: 500.00 ng/ml



#9 AR-1221-2


R.T.: 4.109 min
Delta R.T.: 0.000 min
Response: 8280856
Conc: 500.00 ng/ml



#10 AR-1221-3

R.T.: 4.876 min
Delta R.T.: 0.000 min
Response: 28305615
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1221ICC500



#10 AR-1221-3

R.T.: 4.185 min
Delta R.T.: 0.000 min
Response: 24238496
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071395.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:07
 Operator : YP\AJ
 Sample : AR1232ICC500
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1232ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 12:29:45 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 12:26:37 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	97587735	67524433	50.000	50.000
2) SA Decachlor...	10.233	8.851	71248479	42941843	50.000	50.000
Target Compounds						
11) L3 AR-1232-1	4.878	4.186	22456038	19158895	500.000	500.000
12) L3 AR-1232-2	5.403	4.917	11506547	19093695	500.000	500.000
13) L3 AR-1232-3	5.690	5.094	23608778	10587643	500.000	500.000
14) L3 AR-1232-4	5.850	5.178	11589674	9270206	500.000	500.000
15) L3 AR-1232-5	5.940	5.351	8072861	9950926	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071395.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:07
Operator : YP\AJ
Sample : AR1232ICC500
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 12:29:45 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 12:26:37 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 97587735
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 67524433
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: 0.000 min
Response: 71248479
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.851 min
Delta R.T.: 0.000 min
Response: 42941843
Conc: 50.00 ng/ml



#11 AR-1232-1


R.T.: 4.878 min
Delta R.T.: 0.000 min
Response: 22456038
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#11 AR-1232-1

R.T.: 4.186 min
Delta R.T.: 0.000 min
Response: 19158895
Conc: 500.00 ng/ml




#12 AR-1232-2

R.T.: 5.403 min
Delta R.T.: 0.000 min
Response: 11506547
Conc: 500.00 ng/ml



#12 AR-1232-2


R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 19093695
Conc: 500.00 ng/ml



#13 AR-1232-3


R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 23608778
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500




#13 AR-1232-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 10587643
Conc: 500.00 ng/ml




#14 AR-1232-4

R.T.: 5.850 min
Delta R.T.: 0.000 min
Response: 11589674
Conc: 500.00 ng/ml



#14 AR-1232-4


R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 9270206
Conc: 500.00 ng/ml



#15 AR-1232-5

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 8072861
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1232ICC500



#15 AR-1232-5

R.T.: 5.351 min
Delta R.T.: 0.000 min
Response: 9950926
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071396.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:23
 Operator : YP\AJ
 Sample : AR1242ICC1000
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:20:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:13:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.812	186.9E6	129.5E6	96.183	92.543
2) SA Decachlor...	10.233	8.849	136.9E6	81561508	95.790	95.248
Target Compounds						
16) L4 AR-1242-1	5.668	4.897	53870259	42717527	958.842	942.603
17) L4 AR-1242-2	5.689	4.916	81881116	62346958	951.456	954.464
18) L4 AR-1242-3	5.751	5.093	49423643	34614031	962.179	947.245
19) L4 AR-1242-4	5.848	5.177	41638031	33518120	960.813	943.289
20) L4 AR-1242-5	6.580	5.701	44955701	41214500	963.821	956.409



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071396.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:23
Operator : YP\AJ
Sample : AR1242ICC1000
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:20:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:13:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 186894535
Conc: 96.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 129515962
Conc: 92.54 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: 0.000 min
Response: 136854767
Conc: 95.79 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 81561508
Conc: 95.25 ng/ml



#16 AR-1242-1


R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 53870259
Conc: 958.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#16 AR-1242-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 42717527
Conc: 942.60 ng/ml




#17 AR-1242-2

R.T.: 5.689 min
Delta R.T.: 0.000 min
Response: 81881116
Conc: 951.46 ng/ml



#17 AR-1242-2


R.T.: 4.916 min
Delta R.T.: 0.000 min
Response: 62346958
Conc: 954.46 ng/ml



#18 AR-1242-3


R.T.: 5.751 min
Delta R.T.: 0.000 min
Response: 49423643
Conc: 962.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000




#18 AR-1242-3

R.T.: 5.093 min
Delta R.T.: 0.000 min
Response: 34614031
Conc: 947.24 ng/ml




#19 AR-1242-4

R.T.: 5.848 min
Delta R.T.: 0.000 min
Response: 41638031
Conc: 960.81 ng/ml



#19 AR-1242-4


R.T.: 5.177 min
Delta R.T.: 0.000 min
Response: 33518120
Conc: 943.29 ng/ml



#20 AR-1242-5

R.T.: 6.580 min
Delta R.T.: 0.000 min
Response: 44955701
Conc: 963.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC1000



#20 AR-1242-5

R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 41214500
Conc: 956.41 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071397.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:39
 Operator : YP\AJ
 Sample : AR1242ICC750
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:23:22 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:13:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	144.5E6	99405951	74.571	72.305
2) SA Decachlor...	10.236	8.850	106.6E6	63879480	74.739	74.732
Target Compounds						
16) L4 AR-1242-1	5.669	4.899	41763307	33495158	745.553	742.700
17) L4 AR-1242-2	5.691	4.917	64231120	48675819	747.572	746.776
18) L4 AR-1242-3	5.753	5.094	38470772	26777904	749.299	738.446
19) L4 AR-1242-4	5.850	5.179	32141516	25994749	744.431	737.606
20) L4 AR-1242-5	6.581	5.703	34794164	32016406	747.305	745.293



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071397.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:39
Operator : YP\AJ
Sample : AR1242ICC750
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:23:22 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:13:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 144487060
Conc: 74.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 99405951
Conc: 72.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 106593076
Conc: 74.74 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 63879480
Conc: 74.73 ng/ml



#16 AR-1242-1


R.T.: 5.669 min
Delta R.T.: 0.000 min
Response: 41763307
Conc: 745.55 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#16 AR-1242-1

R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 33495158
Conc: 742.70 ng/ml




#17 AR-1242-2

R.T.: 5.691 min
Delta R.T.: 0.000 min
Response: 64231120
Conc: 747.57 ng/ml



#17 AR-1242-2


R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 48675819
Conc: 746.78 ng/ml



#18 AR-1242-3


R.T.: 5.753 min
Delta R.T.: 0.000 min
Response: 38470772
Conc: 749.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750




#18 AR-1242-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 26777904
Conc: 738.45 ng/ml




#19 AR-1242-4

R.T.: 5.850 min
Delta R.T.: 0.000 min
Response: 32141516
Conc: 744.43 ng/ml



#19 AR-1242-4


R.T.: 5.179 min
Delta R.T.: 0.000 min
Response: 25994749
Conc: 737.61 ng/ml



#20 AR-1242-5

R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 34794164
Conc: 747.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC750



#20 AR-1242-5

R.T.: 5.703 min
Delta R.T.: 0.000 min
Response: 32016406
Conc: 745.29 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071398.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 12:56
 Operator : YP\AJ
 Sample : AR1242ICC500
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:13:34 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:13:15 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	100.9E6	75193817	50.000	50.000
2) SA Decachlor...	10.235	8.850	74441902	44850133	50.000	50.000
Target Compounds						
16) L4 AR-1242-1	5.669	4.898	29247490	23959903	500.000	500.000
17) L4 AR-1242-2	5.690	4.916	45118235	34147953	500.000	500.000
18) L4 AR-1242-3	5.752	5.093	26654530	19234787	500.000	500.000
19) L4 AR-1242-4	5.849	5.178	22517234	18774167	500.000	500.000
20) L4 AR-1242-5	6.581	5.701	24165350	22485728	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071398.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 12:56
Operator : YP\AJ
Sample : AR1242ICC500
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:13:34 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:13:15 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 100863617
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 75193817
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: 0.000 min
Response: 74441902
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 44850133
Conc: 50.00 ng/ml



#16 AR-1242-1


R.T.: 5.669 min
Delta R.T.: 0.000 min
Response: 29247490
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#16 AR-1242-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 23959903
Conc: 500.00 ng/ml




#17 AR-1242-2

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 45118235
Conc: 500.00 ng/ml



#17 AR-1242-2


R.T.: 4.916 min
Delta R.T.: 0.000 min
Response: 34147953
Conc: 500.00 ng/ml



#18 AR-1242-3


R.T.: 5.752 min
Delta R.T.: 0.000 min
Response: 26654530
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500




#18 AR-1242-3

R.T.: 5.093 min
Delta R.T.: 0.000 min
Response: 19234787
Conc: 500.00 ng/ml




#19 AR-1242-4

R.T.: 5.849 min
Delta R.T.: 0.000 min
Response: 22517234
Conc: 500.00 ng/ml



#19 AR-1242-4


R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 18774167
Conc: 500.00 ng/ml



#20 AR-1242-5

R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 24165350
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC500



#20 AR-1242-5

R.T.: 5.701 min
Delta R.T.: 0.000 min
Response: 22485728
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071399.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 13:12
 Operator : YP\AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:34:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:34:27 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	50759754	35812201	25.888	25.778
2) SA Decachlor...	10.233	8.850	37582587	22937032	26.000	26.351
Target Compounds						
16) L4 AR-1242-1	5.668	4.899	14669196	12394168	258.800	268.164
17) L4 AR-1242-2	5.689	4.917	22436050	17569793	258.254	264.383
18) L4 AR-1242-3	5.751	5.094	13459217	10184944	259.001	272.457
19) L4 AR-1242-4	5.849	5.178	11164111	9920754	256.375	272.906
20) L4 AR-1242-5	6.580	5.702	11989115	11814664	255.584	268.312



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071399.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 13:12
Operator : YP\AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:34:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:34:27 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 50759754
Conc: 25.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 35812201
Conc: 25.78 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: 0.000 min
Response: 37582587
Conc: 26.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 22937032
Conc: 26.35 ng/ml



#16 AR-1242-1


R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 14669196
Conc: 258.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 12394168
Conc: 268.16 ng/ml




#17 AR-1242-2

R.T.: 5.689 min
Delta R.T.: 0.000 min
Response: 22436050
Conc: 258.25 ng/ml



#17 AR-1242-2


R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 17569793
Conc: 264.38 ng/ml



#18 AR-1242-3


R.T.: 5.751 min
Delta R.T.: 0.000 min
Response: 13459217
Conc: 259.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 10184944
Conc: 272.46 ng/ml




#19 AR-1242-4

R.T.: 5.849 min
Delta R.T.: 0.000 min
Response: 11164111
Conc: 256.37 ng/ml



#19 AR-1242-4


R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 9920754
Conc: 272.91 ng/ml



#20 AR-1242-5

R.T.: 6.580 min
Delta R.T.: 0.000 min
Response: 11989115
Conc: 255.58 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 11814664
Conc: 268.31 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071400.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 13:28
 Operator : YP\AJ
 Sample : AR1242ICC050
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1242ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 13:44:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 13:44:19 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	8065732	6945226	4.265	4.999
2) SA Decachlor...	10.237	8.850	5895463	3833854	4.235	4.512
Target Compounds						
16) L4 AR-1242-1	5.670	4.899	2356519	2317190	43.025	50.108
17) L4 AR-1242-2	5.691	4.917	3869967	3214527	45.539	48.688
18) L4 AR-1242-3	5.754	5.094	2387637	1635706	46.703	44.877
19) L4 AR-1242-4	5.851	5.179	1842925	1657188	43.662	46.406
20) L4 AR-1242-5	6.582	5.702	2190873	2070356	47.329	47.586



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071400.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 13:28
Operator : YP\AJ
Sample : AR1242ICC050
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 13:44:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 13:44:19 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 8065732
Conc: 4.26 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 6945226
Conc: 5.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.237 min
Delta R.T.: 0.000 min
Response: 5895463
Conc: 4.23 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 3833854
Conc: 4.51 ng/ml



#16 AR-1242-1


R.T.: 5.670 min
Delta R.T.: 0.000 min
Response: 2356519
Conc: 43.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#16 AR-1242-1

R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 2317190
Conc: 50.11 ng/ml




#17 AR-1242-2

R.T.: 5.691 min
Delta R.T.: 0.000 min
Response: 3869967
Conc: 45.54 ng/ml



#17 AR-1242-2


R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 3214527
Conc: 48.69 ng/ml



#18 AR-1242-3


R.T.: 5.754 min
Delta R.T.: 0.000 min
Response: 2387637
Conc: 46.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050




#18 AR-1242-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 1635706
Conc: 44.88 ng/ml




#19 AR-1242-4

R.T.: 5.851 min
Delta R.T.: 0.000 min
Response: 1842925
Conc: 43.66 ng/ml



#19 AR-1242-4


R.T.: 5.179 min
Delta R.T.: 0.000 min
Response: 1657188
Conc: 46.41 ng/ml



#20 AR-1242-5

R.T.: 6.582 min
Delta R.T.: 0.000 min
Response: 2190873
Conc: 47.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1242ICC050



#20 AR-1242-5

R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 2070356
Conc: 47.59 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071401.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 13:45
 Operator : YP\AJ
 Sample : AR1248ICC1000
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:30:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:27:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	187.9E6	132.5E6	97.145	92.221
2) SA Decachlor...	10.237	8.850	137.9E6	82637510	96.283	95.713
Target Compounds						
21) L5 AR-1248-1	5.669	4.898	42447397	30996932	976.497	917.952
22) L5 AR-1248-2	5.941	5.136	57932834	43624561	965.902	945.881
23) L5 AR-1248-3	6.144	5.178	65030901	45791659	967.700	948.403
24) L5 AR-1248-4	6.543	5.350	80777360	53871444	963.427	957.270
25) L5 AR-1248-5	6.582	5.743	76029464	52150607	969.727	950.272



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071401.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 13:45
Operator : YP\AJ
Sample : AR1248ICC1000
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:30:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:27:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 187919144
Conc: 97.15 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 132471223
Conc: 92.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.237 min
Delta R.T.: 0.000 min
Response: 137913077
Conc: 96.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 82637510
Conc: 95.71 ng/ml



#21 AR-1248-1


R.T.: 5.669 min
Delta R.T.: 0.000 min
Response: 42447397
Conc: 976.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#21 AR-1248-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 30996932
Conc: 917.95 ng/ml




#22 AR-1248-2

R.T.: 5.941 min
Delta R.T.: 0.000 min
Response: 57932834
Conc: 965.90 ng/ml



#22 AR-1248-2


R.T.: 5.136 min
Delta R.T.: 0.000 min
Response: 43624561
Conc: 945.88 ng/ml



#23 AR-1248-3


R.T.: 6.144 min
Delta R.T.: 0.000 min
Response: 65030901
Conc: 967.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000




#23 AR-1248-3

R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 45791659
Conc: 948.40 ng/ml




#24 AR-1248-4

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 80777360
Conc: 963.43 ng/ml



#24 AR-1248-4


R.T.: 5.350 min
Delta R.T.: 0.000 min
Response: 53871444
Conc: 957.27 ng/ml



#25 AR-1248-5

R.T.: 6.582 min
Delta R.T.: 0.000 min
Response: 76029464
Conc: 969.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC1000



#25 AR-1248-5

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 52150607
Conc: 950.27 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071402.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:01
 Operator : YP\AJ
 Sample : AR1248ICC750
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC750

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:33:31 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:27:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.516	3.813	144.4E6	113.6E6	74.770	77.657
2) SA Decachlor...	10.238	8.850	108.0E6	64556701	75.275	74.847
Target Compounds						
21) L5 AR-1248-1	5.670	4.898	32527847	26959361	748.865	781.576
22) L5 AR-1248-2	5.941	5.135	45161422	35482042	751.976	762.779
23) L5 AR-1248-3	6.144	5.177	51040114	37424945	756.312	766.560
24) L5 AR-1248-4	6.543	5.350	63330814	43921936	753.553	770.043
25) L5 AR-1248-5	6.583	5.743	58877047	42217581	750.636	762.741



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071402.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:01
Operator : YP\AJ
Sample : AR1248ICC750
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:33:31 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:27:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.516 min
Delta R.T.: 0.000 min
Response: 144413867
Conc: 74.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 113561766
Conc: 77.66 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.238 min
Delta R.T.: 0.000 min
Response: 108020118
Conc: 75.28 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 64556701
Conc: 74.85 ng/ml



#21 AR-1248-1


R.T.: 5.670 min
Delta R.T.: 0.000 min
Response: 32527847
Conc: 748.87 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#21 AR-1248-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 26959361
Conc: 781.58 ng/ml




#22 AR-1248-2

R.T.: 5.941 min
Delta R.T.: 0.000 min
Response: 45161422
Conc: 751.98 ng/ml



#22 AR-1248-2


R.T.: 5.135 min
Delta R.T.: 0.000 min
Response: 35482042
Conc: 762.78 ng/ml



#23 AR-1248-3


R.T.: 6.144 min
Delta R.T.: 0.000 min
Response: 51040114
Conc: 756.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750




#23 AR-1248-3

R.T.: 5.177 min
Delta R.T.: 0.000 min
Response: 37424945
Conc: 766.56 ng/ml




#24 AR-1248-4

R.T.: 6.543 min
Delta R.T.: 0.000 min
Response: 63330814
Conc: 753.55 ng/ml



#24 AR-1248-4


R.T.: 5.350 min
Delta R.T.: 0.000 min
Response: 43921936
Conc: 770.04 ng/ml



#25 AR-1248-5

R.T.: 6.583 min
Delta R.T.: 0.000 min
Response: 58877047
Conc: 750.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC750



#25 AR-1248-5

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 42217581
Conc: 762.74 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071403.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:17
 Operator : YP\AJ
 Sample : AR1248ICC500
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:28:25 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:27:59 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	99482021	77410182	50.000	50.000
2) SA Decachlor...	10.237	8.850	74280984	45020002	50.000	50.000
Target Compounds						
21) L5 AR-1248-1	5.668	4.898	22245336	18269007	500.000	500.000
22) L5 AR-1248-2	5.940	5.136	31011521	24308259	500.000	500.000
23) L5 AR-1248-3	6.142	5.178	34686045	25387087	500.000	500.000
24) L5 AR-1248-4	6.542	5.350	43455122	29340392	500.000	500.000
25) L5 AR-1248-5	6.581	5.743	40388237	28804357	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071403.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:17
Operator : YP\AJ
Sample : AR1248ICC500
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:28:25 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:27:59 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 99482021
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 77410182
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.237 min
Delta R.T.: 0.000 min
Response: 74280984
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 45020002
Conc: 50.00 ng/ml



#21 AR-1248-1


R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 22245336
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#21 AR-1248-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 18269007
Conc: 500.00 ng/ml




#22 AR-1248-2

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 31011521
Conc: 500.00 ng/ml



#22 AR-1248-2


R.T.: 5.136 min
Delta R.T.: 0.000 min
Response: 24308259
Conc: 500.00 ng/ml



#23 AR-1248-3


R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 34686045
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500




#23 AR-1248-3

R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 25387087
Conc: 500.00 ng/ml




#24 AR-1248-4

R.T.: 6.542 min
Delta R.T.: 0.000 min
Response: 43455122
Conc: 500.00 ng/ml



#24 AR-1248-4


R.T.: 5.350 min
Delta R.T.: 0.000 min
Response: 29340392
Conc: 500.00 ng/ml



#25 AR-1248-5

R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 40388237
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC500



#25 AR-1248-5

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 28804357
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071404.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:33
 Operator : YP\AJ
 Sample : AR1248ICC250
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 14:46:35 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 14:46:22 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	50940076	35871474	26.017	24.646
2) SA Decachlor...	10.236	8.850	38029945	23509396	26.110	26.655
Target Compounds						
21) L5 AR-1248-1	5.668	4.898	11434087	9814021	259.799	275.024
22) L5 AR-1248-2	5.940	5.135	15821478	13376378	259.947	277.150
23) L5 AR-1248-3	6.143	5.178	17579514	14000544	257.788	276.598
24) L5 AR-1248-4	6.542	5.350	22285729	16222320	261.208	274.950
25) L5 AR-1248-5	6.581	5.743	20888056	16317809	262.033	282.168



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071404.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:33
Operator : YP\AJ
Sample : AR1248ICC250
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 14:46:35 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 14:46:22 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 50940076
Conc: 26.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 35871474
Conc: 24.65 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 38029945
Conc: 26.11 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 23509396
Conc: 26.66 ng/ml



#21 AR-1248-1


R.T.: 5.668 min
Delta R.T.: 0.000 min
Response: 11434087
Conc: 259.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#21 AR-1248-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 9814021
Conc: 275.02 ng/ml




#22 AR-1248-2

R.T.: 5.940 min
Delta R.T.: 0.000 min
Response: 15821478
Conc: 259.95 ng/ml



#22 AR-1248-2


R.T.: 5.135 min
Delta R.T.: 0.000 min
Response: 13376378
Conc: 277.15 ng/ml



#23 AR-1248-3


R.T.: 6.143 min
Delta R.T.: 0.000 min
Response: 17579514
Conc: 257.79 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250




#23 AR-1248-3

R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 14000544
Conc: 276.60 ng/ml




#24 AR-1248-4

R.T.: 6.542 min
Delta R.T.: 0.000 min
Response: 22285729
Conc: 261.21 ng/ml



#24 AR-1248-4


R.T.: 5.350 min
Delta R.T.: 0.000 min
Response: 16222320
Conc: 274.95 ng/ml



#25 AR-1248-5

R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 20888056
Conc: 262.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC250



#25 AR-1248-5

R.T.: 5.743 min
Delta R.T.: 0.000 min
Response: 16317809
Conc: 282.17 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071405.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 14:50
 Operator : YP\AJ
 Sample : AR1248ICC050
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1248ICC050

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:04:57 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:04:44 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.813	9059747	7084653	4.697	4.893
2) SA Decachlor...	10.236	8.850	6791716	4608858	4.727	5.179
Target Compounds						
21) L5 AR-1248-1	5.667	4.899	1932478	1889804	45.005	52.340
22) L5 AR-1248-2	5.939	5.136	2935274	2587801	48.571	52.853
23) L5 AR-1248-3	6.142	5.178	3057380	2647979	45.780	51.834
24) L5 AR-1248-4	6.541	5.350	4089143	3364119	48.329	55.461
25) L5 AR-1248-5	6.581	5.744	3919031	3006686	49.328	51.581



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071405.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 14:50
Operator : YP\AJ
Sample : AR1248ICC050
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:04:57 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:04:44 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 9059747
Conc: 4.70 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 7084653
Conc: 4.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 6791716
Conc: 4.73 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 4608858
Conc: 5.18 ng/ml



#21 AR-1248-1


R.T.: 5.667 min
Delta R.T.: 0.000 min
Response: 1932478
Conc: 45.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050




#21 AR-1248-1

R.T.: 4.899 min
Delta R.T.: 0.000 min
Response: 1889804
Conc: 52.34 ng/ml




#22 AR-1248-2

R.T.: 5.939 min
Delta R.T.: 0.000 min
Response: 2935274
Conc: 48.57 ng/ml



#22 AR-1248-2


R.T.: 5.136 min
Delta R.T.: 0.000 min
Response: 2587801
Conc: 52.85 ng/ml



#23 AR-1248-3


R.T.: 6.142 min
Delta R.T.: 0.000 min
Response: 3057380
Conc: 45.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050




#23 AR-1248-3

R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 2647979
Conc: 51.83 ng/ml




#24 AR-1248-4

R.T.: 6.541 min
Delta R.T.: 0.000 min
Response: 4089143
Conc: 48.33 ng/ml



#24 AR-1248-4


R.T.: 5.350 min
Delta R.T.: 0.000 min
Response: 3364119
Conc: 55.46 ng/ml



#25 AR-1248-5

R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 3919031
Conc: 49.33 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1248ICC050



#25 AR-1248-5

R.T.: 5.744 min
Delta R.T.: 0.000 min
Response: 3006686
Conc: 51.58 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071406.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:06
 Operator : YP\AJ
 Sample : AR1254ICC1000
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC1000

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:51:39 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	188.8E6	139.1E6	96.818	97.876
2) SA Decachlor...	10.236	8.850	141.0E6	84630750	96.040	96.040
Target Compounds						
26) L6 AR-1254-1	6.519	5.702	78574372	81690839	957.338	952.962
27) L6 AR-1254-2	6.735	5.851	121.6E6	70045476	958.538	950.650
28) L6 AR-1254-3	7.099	6.254	123.4E6	109.7E6	962.759	954.874
29) L6 AR-1254-4	7.381	6.483	112.5E6	70169900	960.617	948.370
30) L6 AR-1254-5	7.798	6.901	106.4E6	92229111	966.455	936.717



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071406.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:06
Operator : YP\AJ
Sample : AR1254ICC1000
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:51:39 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 188824112
Conc: 96.82 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 139055330
Conc: 97.88 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 141010181
Conc: 96.04 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 84630750
Conc: 96.04 ng/ml



#26 AR-1254-1


R.T.: 6.519 min
Delta R.T.: 0.000 min
Response: 78574372
Conc: 957.34 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#26 AR-1254-1

R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 81690839
Conc: 952.96 ng/ml




#27 AR-1254-2

R.T.: 6.735 min
Delta R.T.: 0.000 min
Response: 121624435
Conc: 958.54 ng/ml



#27 AR-1254-2


R.T.: 5.851 min
Delta R.T.: 0.000 min
Response: 70045476
Conc: 950.65 ng/ml



#28 AR-1254-3


R.T.: 7.099 min
Delta R.T.: 0.000 min
Response: 123432630
Conc: 962.76 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000




#28 AR-1254-3

R.T.: 6.254 min
Delta R.T.: 0.000 min
Response: 109725744
Conc: 954.87 ng/ml




#29 AR-1254-4

R.T.: 7.381 min
Delta R.T.: 0.000 min
Response: 112512019
Conc: 960.62 ng/ml



#29 AR-1254-4


R.T.: 6.483 min
Delta R.T.: 0.000 min
Response: 70169900
Conc: 948.37 ng/ml



#30 AR-1254-5

R.T.: 7.798 min
Delta R.T.: 0.000 min
Response: 106399089
Conc: 966.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC1000



#30 AR-1254-5

R.T.: 6.901 min
Delta R.T.: 0.000 min
Response: 92229111
Conc: 936.72 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071407.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:22
 Operator : YP\AJ
 Sample : AR1254ICC750
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:54:33 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	145.3E6	112.6E6	74.658	77.773
2) SA Decachlor...	10.236	8.850	110.2E6	65650190	75.033	74.666
Target Compounds						
26) L6 AR-1254-1	6.519	5.703	60934961	65520948	744.931	759.494
27) L6 AR-1254-2	6.735	5.851	94362074	56602387	745.775	762.037
28) L6 AR-1254-3	7.098	6.255	95849879	85577532	748.410	746.476
29) L6 AR-1254-4	7.381	6.483	87194278	55551540	746.295	750.532
30) L6 AR-1254-5	7.797	6.901	79441918	72082199	721.000m	737.969

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071407.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:22
Operator : YP\AJ
Sample : AR1254ICC750
Misc :
ALS Vial : 21 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:54:33 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 145274275
Conc: 74.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 112576042
Conc: 77.77 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 110190189
Conc: 75.03 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 65650190
Conc: 74.67 ng/ml




#26 AR-1254-1

R.T.: 6.519 min
Delta R.T.: 0.000 min
Response: 60934961
Conc: 744.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#26 AR-1254-1

R.T.: 5.703 min
Delta R.T.: 0.000 min
Response: 65520948
Conc: 759.49 ng/ml




#27 AR-1254-2

R.T.: 6.735 min
Delta R.T.: 0.000 min
Response: 94362074
Conc: 745.77 ng/ml



#27 AR-1254-2

R.T.: 5.851 min
Delta R.T.: 0.000 min
Response: 56602387
Conc: 762.04 ng/ml




#28 AR-1254-3

R.T.: 7.098 min
Delta R.T.: 0.000 min
Response: 95849879
Conc: 748.41 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#28 AR-1254-3

R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 85577532
Conc: 746.48 ng/ml




#29 AR-1254-4

R.T.: 7.381 min
Delta R.T.: 0.000 min
Response: 87194278
Conc: 746.30 ng/ml



#29 AR-1254-4

R.T.: 6.483 min
Delta R.T.: 0.000 min
Response: 55551540
Conc: 750.53 ng/ml




#30 AR-1254-5

R.T.: 7.797 min
Delta R.T.: 0.000 min
Response: 79441918
Conc: 721.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#30 AR-1254-5

R.T.: 6.901 min
Delta R.T.: 0.000 min
Response: 72082199
Conc: 737.97 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071408.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:38
 Operator : YP\AJ
 Sample : AR1254ICC500
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 15:48:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 15:48:40 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.812	100.6E6	72545075	50.000	50.000
2) SA Decachlor...	10.237	8.849	76319487	45805245	50.000	50.000
Target Compounds						
26) L6 AR-1254-1	6.519	5.702	42788667	44877702	500.000	500.000
27) L6 AR-1254-2	6.735	5.851	66073164	38658957	500.000	500.000
28) L6 AR-1254-3	7.098	6.254	66490827	60048409	500.000	500.000
29) L6 AR-1254-4	7.381	6.483	60868718	38905081	500.000	500.000
30) L6 AR-1254-5	7.797	6.901	56892560	52345397	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071408.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:38
Operator : YP\AJ
Sample : AR1254ICC500
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 15:48:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 15:48:40 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 100618430
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 72545075
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.237 min
Delta R.T.: 0.000 min
Response: 76319487
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 45805245
Conc: 50.00 ng/ml



#26 AR-1254-1


R.T.: 6.519 min
Delta R.T.: 0.000 min
Response: 42788667
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#26 AR-1254-1

R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 44877702
Conc: 500.00 ng/ml




#27 AR-1254-2

R.T.: 6.735 min
Delta R.T.: 0.000 min
Response: 66073164
Conc: 500.00 ng/ml



#27 AR-1254-2


R.T.: 5.851 min
Delta R.T.: 0.000 min
Response: 38658957
Conc: 500.00 ng/ml



#28 AR-1254-3


R.T.: 7.098 min
Delta R.T.: 0.000 min
Response: 66490827
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500




#28 AR-1254-3

R.T.: 6.254 min
Delta R.T.: 0.000 min
Response: 60048409
Conc: 500.00 ng/ml




#29 AR-1254-4

R.T.: 7.381 min
Delta R.T.: 0.000 min
Response: 60868718
Conc: 500.00 ng/ml



#29 AR-1254-4


R.T.: 6.483 min
Delta R.T.: 0.000 min
Response: 38905081
Conc: 500.00 ng/ml



#30 AR-1254-5

R.T.: 7.797 min
Delta R.T.: 0.000 min
Response: 56892560
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC500



#30 AR-1254-5

R.T.: 6.901 min
Delta R.T.: 0.000 min
Response: 52345397
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071409.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 15:55
 Operator : YP\AJ
 Sample : AR1254ICC250
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 16:07:19 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 16:07:08 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.813	51789148	37965204	26.192	25.910
2) SA Decachlor...	10.234	8.850	38742671	23881547	26.022	26.587
Target Compounds						
26) L6 AR-1254-1	6.517	5.703	22149547	24321384	265.267	273.203
27) L6 AR-1254-2	6.733	5.851	34278927	20788594	265.367	271.757
28) L6 AR-1254-3	7.096	6.255	34405813	31095041	263.728	265.596
29) L6 AR-1254-4	7.379	6.483	31098214	20469357	261.934	269.399
30) L6 AR-1254-5	7.796	6.901	29038261	26846100	262.635	268.183



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071409.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 15:55
Operator : YP\AJ
Sample : AR1254ICC250
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 16:07:19 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 16:07:08 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 51789148
Conc: 26.19 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 37965204
Conc: 25.91 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: 0.000 min
Response: 38742671
Conc: 26.02 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 23881547
Conc: 26.59 ng/ml



#26 AR-1254-1


R.T.: 6.517 min
Delta R.T.: 0.000 min
Response: 22149547
Conc: 265.27 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#26 AR-1254-1

R.T.: 5.703 min
Delta R.T.: 0.000 min
Response: 24321384
Conc: 273.20 ng/ml




#27 AR-1254-2

R.T.: 6.733 min
Delta R.T.: 0.000 min
Response: 34278927
Conc: 265.37 ng/ml



#27 AR-1254-2


R.T.: 5.851 min
Delta R.T.: 0.000 min
Response: 20788594
Conc: 271.76 ng/ml



#28 AR-1254-3


R.T.: 7.096 min
Delta R.T.: 0.000 min
Response: 34405813
Conc: 263.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250




#28 AR-1254-3

R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 31095041
Conc: 265.60 ng/ml




#29 AR-1254-4

R.T.: 7.379 min
Delta R.T.: 0.000 min
Response: 31098214
Conc: 261.93 ng/ml



#29 AR-1254-4


R.T.: 6.483 min
Delta R.T.: 0.000 min
Response: 20469357
Conc: 269.40 ng/ml



#30 AR-1254-5

R.T.: 7.796 min
Delta R.T.: 0.000 min
Response: 29038261
Conc: 262.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC250



#30 AR-1254-5

R.T.: 6.901 min
Delta R.T.: 0.000 min
Response: 26846100
Conc: 268.18 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071410.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 16:11
 Operator : YP\AJ
 Sample : AR1254ICC050
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1254ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 16:54:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 16:54:36 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	8612576	7342515	4.471m	5.009
2) SA Decachlor...	10.236	8.850	6643627	4341119	4.560	4.865
Target Compounds						
26) L6 AR-1254-1	6.520	5.703	3787970	5045448	46.222	55.202
27) L6 AR-1254-2	6.735	5.852	5933913	4378664	46.696	55.629
28) L6 AR-1254-3	7.099	6.255	5851823	6211772	45.798	52.416
29) L6 AR-1254-4	7.382	6.484	5633281	4095094	47.937	53.069
30) L6 AR-1254-5	7.798	6.901	4819584	5346230	44.737	52.689

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071410.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 16:11
Operator : YP\AJ
Sample : AR1254ICC050
Misc :
ALS Vial : 24 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 16:54:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 16:54:36 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: 0.000 min
Response: 8612576
Conc: 4.47 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 7342515
Conc: 5.01 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 6643627
Conc: 4.56 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 4341119
Conc: 4.87 ng/ml




#26 AR-1254-1

R.T.: 6.520 min
Delta R.T.: 0.000 min
Response: 3787970
Conc: 46.22 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#26 AR-1254-1

R.T.: 5.703 min
Delta R.T.: 0.000 min
Response: 5045448
Conc: 55.20 ng/ml




#27 AR-1254-2

R.T.: 6.735 min
Delta R.T.: 0.000 min
Response: 5933913
Conc: 46.70 ng/ml



#27 AR-1254-2

R.T.: 5.852 min
Delta R.T.: 0.000 min
Response: 4378664
Conc: 55.63 ng/ml




#28 AR-1254-3

R.T.: 7.099 min
Delta R.T.: 0.000 min
Response: 5851823
Conc: 45.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#28 AR-1254-3

R.T.: 6.255 min
Delta R.T.: 0.000 min
Response: 6211772
Conc: 52.42 ng/ml




#29 AR-1254-4

R.T.: 7.382 min
Delta R.T.: 0.000 min
Response: 5633281
Conc: 47.94 ng/ml



#29 AR-1254-4

R.T.: 6.484 min
Delta R.T.: 0.000 min
Response: 4095094
Conc: 53.07 ng/ml




#30 AR-1254-5

R.T.: 7.798 min
Delta R.T.: 0.000 min
Response: 4819584
Conc: 44.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1254ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#30 AR-1254-5

R.T.: 6.901 min
Delta R.T.: 0.000 min
Response: 5346230
Conc: 52.69 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071411.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 16:27
 Operator : YP\AJ
 Sample : AR1262ICC500
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1262ICC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 17:08:30 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 17:08:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	98220466	71170259	50.000	50.000
2) SA Decachlor...	10.235	8.850	74643123	44870081	50.000	50.000
Target Compounds						
36) L8 AR-1262-1	8.100	6.939	72245642	56806132	500.000	500.000
37) L8 AR-1262-2	8.420	7.198	137.1E6	46169678	500.000	500.000
38) L8 AR-1262-3	8.737	7.721	91438860	37847231	500.000	500.000
39) L8 AR-1262-4	8.822	7.786	67105487	62412938	500.000	500.000
40) L8 AR-1262-5	9.478	8.286	44080983	27746588	500.000	500.000



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071411.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 16:27
Operator : YP\AJ
Sample : AR1262ICC500
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 17:08:30 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 98220466
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 71170259
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: 0.000 min
Response: 74643123
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 44870081
Conc: 50.00 ng/ml



#36 AR-1262-1


R.T.: 8.100 min
Delta R.T.: 0.000 min
Response: 72245642
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500




#36 AR-1262-1

R.T.: 6.939 min
Delta R.T.: 0.000 min
Response: 56806132
Conc: 500.00 ng/ml




#37 AR-1262-2

R.T.: 8.420 min
Delta R.T.: 0.000 min
Response: 137093050
Conc: 500.00 ng/ml



#37 AR-1262-2


R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 46169678
Conc: 500.00 ng/ml



#38 AR-1262-3


R.T.: 8.737 min
Delta R.T.: 0.000 min
Response: 91438860
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1262ICC500




#38 AR-1262-3

R.T.: 7.721 min
Delta R.T.: 0.000 min
Response: 37847231
Conc: 500.00 ng/ml



#39 AR-1262-4

R.T.: 8.822 min
Delta R.T.: 0.000 min
Response: 67105487
Conc: 500.00 ng/ml




#39 AR-1262-4

R.T.: 7.786 min
Delta R.T.: 0.000 min
Response: 62412938
Conc: 500.00 ng/ml

Response_

Signal: PP071411.D\ECD1A.ch

#40 AR-1262-5



R.T.: 9.478 min
Delta R.T.: 0.000 min
Response: 44080983
Conc: 500.00 ng/ml


Instrument :
ECD_P
ClientSampleId :
AR1262ICC500

Time

Response_

Signal: PP071411.D\ECD2B.ch

#40 AR-1262-5



R.T.: 8.286 min
Delta R.T.: 0.000 min
Response: 27746588
Conc: 500.00 ng/ml

Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071412.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 16:44
 Operator : YP\AJ
 Sample : AR1268ICC1000
 Misc :
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC1000

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:44:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	188.4E6	147.6E6	95.054	100.837
2) SA Decachlor...	10.235	8.850	248.7E6	149.9E6	93.152	93.528
Target Compounds						
41) L9 AR-1268-1	8.733	7.722	299.6E6	181.9E6	936.979	895.922m
42) L9 AR-1268-2	8.826	7.788	251.1E6	154.8E6	937.116	896.075
43) L9 AR-1268-3	9.058	7.993	211.9E6	128.5E6	940.656	898.014
44) L9 AR-1268-4	9.477	8.287	93231632	56964280	941.667	896.773
45) L9 AR-1268-5	9.894	8.588	590.1E6	367.0E6	950.117	952.869

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071412.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 16:44
Operator : YP\AJ
Sample : AR1268ICC1000
Misc :
ALS Vial : 26 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

AR1268ICC1000

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 04/23/2025

Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:44:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: -0.001 min
Response: 188397795
Conc: 95.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: -0.002 min
Response: 147647099
Conc: 100.84 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: 0.000 min
Response: 248662366
Conc: 93.15 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: -0.001 min
Response: 149905120
Conc: 93.53 ng/ml




#41 AR-1268-1

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 299594266
Conc: 936.98 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#41 AR-1268-1

R.T.: 7.722 min
Delta R.T.: 0.000 min
Response: 181860271
Conc: 895.92 ng/ml




#42 AR-1268-2

R.T.: 8.826 min
Delta R.T.: 0.000 min
Response: 251109850
Conc: 937.12 ng/ml



#42 AR-1268-2

R.T.: 7.788 min
Delta R.T.: 0.000 min
Response: 154801411
Conc: 896.07 ng/ml




#43 AR-1268-3

R.T.: 9.058 min
Delta R.T.: 0.000 min
Response: 211915080
Conc: 940.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#43 AR-1268-3

R.T.: 7.993 min
Delta R.T.: 0.000 min
Response: 128491160
Conc: 898.01 ng/ml




#44 AR-1268-4

R.T.: 9.477 min
Delta R.T.: 0.000 min
Response: 93231632
Conc: 941.67 ng/ml



#44 AR-1268-4

R.T.: 8.287 min
Delta R.T.: 0.000 min
Response: 56964280
Conc: 896.77 ng/ml




#45 AR-1268-5

R.T.: 9.894 min
Delta R.T.: 0.000 min
Response: 590082071
Conc: 950.12 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC1000

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#45 AR-1268-5

R.T.: 8.588 min
Delta R.T.: 0.000 min
Response: 367021219
Conc: 952.87 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071413.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:00
 Operator : YP\AJ
 Sample : AR1268ICC750
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC750

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:45:40 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	144.9E6	108.4E6	73.091	74.065
2) SA Decachlor...	10.234	8.849	191.8E6	116.0E6	71.833	72.344
Target Compounds						
41) L9 AR-1268-1	8.733	7.721	231.8E6	141.3E6	725.026	696.327m
42) L9 AR-1268-2	8.826	7.787	194.3E6	119.6E6	725.052	692.534
43) L9 AR-1268-3	9.058	7.991	163.8E6	99816904	726.929	697.612
44) L9 AR-1268-4	9.476	8.285	72323269	44771879	730.486	704.831
45) L9 AR-1268-5	9.894	8.587	452.8E6	283.3E6	729.082	735.437

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071413.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:00
Operator : YP\AJ
Sample : AR1268ICC750
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :

ECD_P

ClientSampleId :

AR1268ICC750

Manual Integrations



APPROVED


Reviewed By :Yogesh Patel 04/23/2025

Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:45:40 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 144867571
Conc: 73.09 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: -0.001 min
Response: 108446461
Conc: 74.06 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: -0.001 min
Response: 191753222
Conc: 71.83 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.849 min
Delta R.T.: -0.002 min
Response: 115953007
Conc: 72.34 ng/ml




#41 AR-1268-1

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 231823386
Conc: 725.03 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#41 AR-1268-1

R.T.: 7.721 min
Delta R.T.: -0.002 min
Response: 141345145
Conc: 696.33 ng/ml m




#42 AR-1268-2

R.T.: 8.826 min
Delta R.T.: 0.000 min
Response: 194285253
Conc: 725.05 ng/ml



#42 AR-1268-2

R.T.: 7.787 min
Delta R.T.: -0.001 min
Response: 119638783
Conc: 692.53 ng/ml




#43 AR-1268-3

R.T.: 9.058 min
Delta R.T.: 0.000 min
Response: 163765585
Conc: 726.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#43 AR-1268-3

R.T.: 7.991 min
Delta R.T.: -0.002 min
Response: 99816904
Conc: 697.61 ng/ml




#44 AR-1268-4

R.T.: 9.476 min
Delta R.T.: -0.002 min
Response: 72323269
Conc: 730.49 ng/ml



#44 AR-1268-4

R.T.: 8.285 min
Delta R.T.: -0.002 min
Response: 44771879
Conc: 704.83 ng/ml




#45 AR-1268-5

R.T.: 9.894 min
Delta R.T.: 0.000 min
Response: 452805185
Conc: 729.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC750

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#45 AR-1268-5

R.T.: 8.587 min
Delta R.T.: -0.002 min
Response: 283271897
Conc: 735.44 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071414.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:16
 Operator : YP\AJ
 Sample : AR1268ICC500
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC500

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:46:05 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.814	99100413	73210719	50.000	50.000
2) SA Decachlor...	10.236	8.851	133.5E6	80139523	50.000	50.000
Target Compounds						
41) L9 AR-1268-1	8.733	7.722	159.9E6	101.7E6	500.000	501.224m
42) L9 AR-1268-2	8.826	7.788	134.0E6	86377504	500.000	500.000
43) L9 AR-1268-3	9.059	7.993	112.6E6	71541846	500.000	500.000
44) L9 AR-1268-4	9.478	8.287	49503520	31760699	500.000	500.000
45) L9 AR-1268-5	9.894	8.589	310.5E6	192.6E6	500.000	500.000

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071414.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:16
Operator : YP\AJ
Sample : AR1268ICC500
Misc :
ALS Vial : 28 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:46:05 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 99100413
Conc: 50.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.814 min
Delta R.T.: 0.000 min
Response: 73210719
Conc: 50.00 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 133470748
Conc: 50.00 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.851 min
Delta R.T.: 0.000 min
Response: 80139523
Conc: 50.00 ng/ml




#41 AR-1268-1

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 159872402
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#41 AR-1268-1

R.T.: 7.722 min
Delta R.T.: 0.000 min
Response: 101741890
Conc: 501.22 ng/ml




#42 AR-1268-2

R.T.: 8.826 min
Delta R.T.: 0.000 min
Response: 133980177
Conc: 500.00 ng/ml



#42 AR-1268-2

R.T.: 7.788 min
Delta R.T.: 0.000 min
Response: 86377504
Conc: 500.00 ng/ml




#43 AR-1268-3

R.T.: 9.059 min
Delta R.T.: 0.000 min
Response: 112642140
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#43 AR-1268-3

R.T.: 7.993 min
Delta R.T.: 0.000 min
Response: 71541846
Conc: 500.00 ng/ml




#44 AR-1268-4

R.T.: 9.478 min
Delta R.T.: 0.000 min
Response: 49503520
Conc: 500.00 ng/ml



#44 AR-1268-4

R.T.: 8.287 min
Delta R.T.: 0.000 min
Response: 31760699
Conc: 500.00 ng/ml




#45 AR-1268-5

R.T.: 9.894 min
Delta R.T.: 0.000 min
Response: 310531224
Conc: 500.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC500

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#45 AR-1268-5

R.T.: 8.589 min
Delta R.T.: 0.000 min
Response: 192587534
Conc: 500.00 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071415.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:33
 Operator : YP\AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC250

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:47:10 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.517	3.813	50174786	36443440	25.315	24.889
2) SA Decachlor...	10.236	8.850	67366681	41318494	25.236	25.779
Target Compounds						
41) L9 AR-1268-1	8.735	7.721	82956522	51353583	259.446	252.990m
42) L9 AR-1268-2	8.828	7.788	68964227	43318102	257.367	250.749
43) L9 AR-1268-3	9.061	7.991	57876424	36193296	256.904	252.952
44) L9 AR-1268-4	9.480	8.286	24514996	16044448	247.609	252.583
45) L9 AR-1268-5	9.897	8.587	157.8E6	96869753	254.135	251.495

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071415.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:33
Operator : YP\AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:47:10 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.517 min
Delta R.T.: 0.002 min
Response: 50174786
Conc: 25.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: -0.001 min
Response: 36443440
Conc: 24.89 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.236 min
Delta R.T.: 0.000 min
Response: 67366681
Conc: 25.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: -0.002 min
Response: 41318494
Conc: 25.78 ng/ml




#41 AR-1268-1

R.T.: 8.735 min
Delta R.T.: 0.002 min
Response: 82956522
Conc: 259.45 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#41 AR-1268-1

R.T.: 7.721 min
Delta R.T.: -0.001 min
Response: 51353583
Conc: 252.99 ng/ml m




#42 AR-1268-2

R.T.: 8.828 min
Delta R.T.: 0.002 min
Response: 68964227
Conc: 257.37 ng/ml



#42 AR-1268-2

R.T.: 7.788 min
Delta R.T.: 0.000 min
Response: 43318102
Conc: 250.75 ng/ml




#43 AR-1268-3

R.T.: 9.061 min
Delta R.T.: 0.003 min
Response: 57876424
Conc: 256.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#43 AR-1268-3

R.T.: 7.991 min
Delta R.T.: -0.002 min
Response: 36193296
Conc: 252.95 ng/ml




#44 AR-1268-4

R.T.: 9.480 min
Delta R.T.: 0.002 min
Response: 24514996
Conc: 247.61 ng/ml



#44 AR-1268-4

R.T.: 8.286 min
Delta R.T.: -0.001 min
Response: 16044448
Conc: 252.58 ng/ml




#45 AR-1268-5

R.T.: 9.897 min
Delta R.T.: 0.003 min
Response: 157833758
Conc: 254.14 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC250

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#45 AR-1268-5

R.T.: 8.587 min
Delta R.T.: -0.002 min
Response: 96869753
Conc: 251.50 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071416.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 17:49
 Operator : YP\AJ
 Sample : AR1268ICC050
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1268ICC050

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:47:36 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.813	7954830	6504410	4.014	4.442
2) SA Decachlor...	10.235	8.850	11316360	7324768	4.239	4.570
Target Compounds						
41) L9 AR-1268-1	8.731	7.721	14103190	10095232	44.108m	49.733m
42) L9 AR-1268-2	8.825	7.787	11406198	8528068	42.567m	49.365
43) L9 AR-1268-3	9.058	7.992	10823544	6773456	48.044	47.339
44) L9 AR-1268-4	9.476	8.287	4136815	2789666	41.783	43.917
45) L9 AR-1268-5	9.894	8.588	26835395	17143092	43.209	44.507

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071416.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 17:49
Operator : YP\AJ
Sample : AR1268ICC050
Misc :
ALS Vial : 30 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:47:36 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.514 min
Delta R.T.: -0.001 min
Response: 7954830
Conc: 4.01 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: -0.001 min
Response: 6504410
Conc: 4.44 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.235 min
Delta R.T.: 0.000 min
Response: 11316360
Conc: 4.24 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: 0.000 min
Response: 7324768
Conc: 4.57 ng/ml




#41 AR-1268-1

R.T.: 8.731 min
Delta R.T.: -0.001 min
Response: 14103190
Conc: 44.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#41 AR-1268-1

R.T.: 7.721 min
Delta R.T.: -0.001 min
Response: 10095232
Conc: 49.73 ng/ml




#42 AR-1268-2

R.T.: 8.825 min
Delta R.T.: -0.001 min
Response: 11406198
Conc: 42.57 ng/ml



#42 AR-1268-2

R.T.: 7.787 min
Delta R.T.: -0.001 min
Response: 8528068
Conc: 49.37 ng/ml




#43 AR-1268-3

R.T.: 9.058 min
Delta R.T.: 0.000 min
Response: 10823544
Conc: 48.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#43 AR-1268-3

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 6773456
Conc: 47.34 ng/ml




#44 AR-1268-4

R.T.: 9.476 min
Delta R.T.: -0.002 min
Response: 4136815
Conc: 41.78 ng/ml



#44 AR-1268-4

R.T.: 8.287 min
Delta R.T.: 0.000 min
Response: 2789666
Conc: 43.92 ng/ml




#45 AR-1268-5

R.T.: 9.894 min
Delta R.T.: 0.000 min
Response: 26835395
Conc: 43.21 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1268ICC050

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#45 AR-1268-5

R.T.: 8.588 min
Delta R.T.: 0.000 min
Response: 17143092
Conc: 44.51 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071417.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 18:05
 Operator : YP\AJ
 Sample : PP042225ICV500
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 19:27:41 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 17:08:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	98003244	70310326	49.431	49.914
2) SA Decachlor...	10.233	8.849	73424396	44681766	50.707	50.945
Target Compounds						
3) L1 AR-1016-1	5.668	4.898	33546404	27255669	500.321	509.132
4) L1 AR-1016-2	5.690	4.916	50377176	38783246	490.870	507.875
5) L1 AR-1016-3	5.752	5.093	30221252	21921068	490.230	519.328
6) L1 AR-1016-4	5.850	5.135	25444869	18087794	496.019	524.498
7) L1 AR-1016-5	6.142	5.349	24049539	22819712	498.556	513.411
31) L7 AR-1260-1	7.262	6.384	46974832	35744650	489.196	494.399
32) L7 AR-1260-2	7.515	6.573	70731391	43416292	494.358	496.405
33) L7 AR-1260-3	7.875	6.726	58007061	39241127	517.439	500.604
34) L7 AR-1260-4	8.098	7.198	56235072	31343464	498.252	496.643
35) L7 AR-1260-5	8.418	7.439	114.7E6	75416996	506.086	500.188



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071417.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 18:05
Operator : YP\AJ
Sample : PP042225ICV500
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 19:27:41 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: -0.003 min
Response: 98003244
Conc: 49.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 70310326
Conc: 49.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: -0.004 min
Response: 73424396
Conc: 50.71 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 44681766
Conc: 50.95 ng/ml

Response_ Signal: PP071417.D\ECD1A.ch

#3 AR-1016-1

R.T.: 5.668 min
Delta R.T.: -0.003 min
Response: 33546404
Conc: 500.32 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225

Time

Response_ Signal: PP071417.D\ECD2B.ch

#3 AR-1016-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 27255669
Conc: 509.13 ng/ml

Time

Response_ Signal: PP071417.D\ECD1A.ch

#4 AR-1016-2

R.T.: 5.690 min
Delta R.T.: -0.003 min
Response: 50377176
Conc: 490.87 ng/ml


Time

Response_ Signal: PP071417.D\ECD2B.ch

#4 AR-1016-2

R.T.: 4.916 min
Delta R.T.: 0.000 min
Response: 38783246
Conc: 507.88 ng/ml


Time



#5 AR-1016-3


R.T.: 5.752 min
Delta R.T.: -0.002 min
Response: 30221252
Conc: 490.23 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225




#5 AR-1016-3

R.T.: 5.093 min
Delta R.T.: 0.000 min
Response: 21921068
Conc: 519.33 ng/ml




#6 AR-1016-4

R.T.: 5.850 min
Delta R.T.: -0.002 min
Response: 25444869
Conc: 496.02 ng/ml



#6 AR-1016-4


R.T.: 5.135 min
Delta R.T.: -0.001 min
Response: 18087794
Conc: 524.50 ng/ml



#7 AR-1016-5


R.T.: 6.142 min
Delta R.T.: -0.003 min
Response: 24049539
Conc: 498.56 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225




#7 AR-1016-5

R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 22819712
Conc: 513.41 ng/ml




#31 AR-1260-1

R.T.: 7.262 min
Delta R.T.: -0.003 min
Response: 46974832
Conc: 489.20 ng/ml



#31 AR-1260-1

R.T.: 6.384 min
Delta R.T.: 0.000 min
Response: 35744650
Conc: 494.40 ng/ml



R.T.: 7.515 min
Delta R.T.: -0.003 min
Response: 70731391
Conc: 494.36 ng/ml


Instrument :
ECD_P
ClientSampleId :
ICVPP042225

Time

Response_

Signal: PP071417.D\ECD2B.ch

#32 AR-1260-2




R.T.: 6.573 min
Delta R.T.: 0.000 min
Response: 43416292
Conc: 496.41 ng/ml

Time

Response_

Signal: PP071417.D\ECD1A.ch

#33 AR-1260-3




R.T.: 7.875 min
Delta R.T.: -0.003 min
Response: 58007061
Conc: 517.44 ng/ml

Time


Response_

Signal: PP071417.D\ECD2B.ch

#33 AR-1260-3




R.T.: 6.726 min
Delta R.T.: -0.001 min
Response: 39241127
Conc: 500.60 ng/ml



#34 AR-1260-4


R.T.: 8.098 min
Delta R.T.: -0.003 min
Response: 56235072
Conc: 498.25 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225




#34 AR-1260-4

R.T.: 7.198 min
Delta R.T.: 0.000 min
Response: 31343464
Conc: 496.64 ng/ml



#35 AR-1260-5

R.T.: 8.418 min
Delta R.T.: -0.003 min
Response: 114665502
Conc: 506.09 ng/ml



#35 AR-1260-5

R.T.: 7.439 min
Delta R.T.: 0.000 min
Response: 75416996
Conc: 500.19 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071418.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 18:38
 Operator : YP\AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 02:23:37 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 02:23:12 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.515	3.813	98322267	73653264	48.740	48.976
2) SA Decachlor...	10.234	8.849	73973652	44775920	49.685	49.917
Target Compounds						
16) L4 AR-1242-1	5.669	4.898	28608484	23663233	489.076	493.809
17) L4 AR-1242-2	5.690	4.917	43367430	33449196	480.598	489.769
18) L4 AR-1242-3	5.753	5.094	25764272	18726327	483.300	486.783
19) L4 AR-1242-4	5.850	5.178	21798483	18331184	484.040	488.202
20) L4 AR-1242-5	6.581	5.702	23472650	21942094	485.668	487.912



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071418.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 18:38
Operator : YP\AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 02:23:37 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 02:23:12 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.515 min
Delta R.T.: 0.000 min
Response: 98322267
Conc: 48.74 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 73653264
Conc: 48.98 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: -0.001 min
Response: 73973652
Conc: 49.69 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: 0.000 min
Response: 44775920
Conc: 49.92 ng/ml



#16 AR-1242-1


R.T.: 5.669 min
Delta R.T.: 0.000 min
Response: 28608484
Conc: 489.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1242




#16 AR-1242-1

R.T.: 4.898 min
Delta R.T.: 0.000 min
Response: 23663233
Conc: 493.81 ng/ml




#17 AR-1242-2

R.T.: 5.690 min
Delta R.T.: 0.000 min
Response: 43367430
Conc: 480.60 ng/ml



#17 AR-1242-2


R.T.: 4.917 min
Delta R.T.: 0.000 min
Response: 33449196
Conc: 489.77 ng/ml



#18 AR-1242-3


R.T.: 5.753 min
Delta R.T.: 0.000 min
Response: 25764272
Conc: 483.30 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1242




#18 AR-1242-3

R.T.: 5.094 min
Delta R.T.: 0.000 min
Response: 18726327
Conc: 486.78 ng/ml




#19 AR-1242-4

R.T.: 5.850 min
Delta R.T.: 0.000 min
Response: 21798483
Conc: 484.04 ng/ml



#19 AR-1242-4


R.T.: 5.178 min
Delta R.T.: 0.000 min
Response: 18331184
Conc: 488.20 ng/ml



#20 AR-1242-5

R.T.: 6.581 min
Delta R.T.: 0.000 min
Response: 23472650
Conc: 485.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1242



#20 AR-1242-5

R.T.: 5.702 min
Delta R.T.: 0.000 min
Response: 21942094
Conc: 487.91 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071419.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 19:27
 Operator : YP\AJ
 Sample : AR1248ICV500
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1248

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: May 01 07:42:53 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.514	3.812	98804591	74071948	49.835	52.584
2) SA Decachlor...	10.234	8.848	74699675	45534064	51.588	51.917
Target Compounds						
21) L5 AR-1248-1	5.667	4.897	22002964	19168563	512.425	530.888
22) L5 AR-1248-2	5.938	5.135	30310191	26184570	501.555	534.789
23) L5 AR-1248-3	6.141	5.177	34304891	27346403	513.666	535.307
24) L5 AR-1248-4	6.541	5.349	43159659	31642146	510.096	521.654
25) L5 AR-1248-5	6.580	5.742	40310718	30029077	507.383	515.159



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071419.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 19:27
Operator : YP\AJ
Sample : AR1248ICV500
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1248

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: May 01 07:42:53 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.514 min
Delta R.T.: -0.003 min
Response: 98804591
Conc: 49.83 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1248




#1 Tetrachloro-m-xylene

R.T.: 3.812 min
Delta R.T.: 0.000 min
Response: 74071948
Conc: 52.58 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: -0.004 min
Response: 74699675
Conc: 51.59 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.848 min
Delta R.T.: -0.001 min
Response: 45534064
Conc: 51.92 ng/ml



#21 AR-1248-1


R.T.: 5.667 min
Delta R.T.: 0.000 min
Response: 22002964
Conc: 512.43 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1248




#21 AR-1248-1

R.T.: 4.897 min
Delta R.T.: 0.000 min
Response: 19168563
Conc: 530.89 ng/ml




#22 AR-1248-2

R.T.: 5.938 min
Delta R.T.: -0.002 min
Response: 30310191
Conc: 501.55 ng/ml



#22 AR-1248-2


R.T.: 5.135 min
Delta R.T.: -0.001 min
Response: 26184570
Conc: 534.79 ng/ml



#23 AR-1248-3


R.T.: 6.141 min
Delta R.T.: 0.000 min
Response: 34304891
Conc: 513.67 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1248




#23 AR-1248-3

R.T.: 5.177 min
Delta R.T.: 0.000 min
Response: 27346403
Conc: 535.31 ng/ml



#24 AR-1248-4

R.T.: 6.541 min
Delta R.T.: 0.000 min
Response: 43159659
Conc: 510.10 ng/ml



#24 AR-1248-4


R.T.: 5.349 min
Delta R.T.: 0.000 min
Response: 31642146
Conc: 521.65 ng/ml

Response_

Signal: PP071419.D\ECD1A.ch

#25 AR-1248-5

ICAL Form



R.T.: 6.580 min
Delta R.T.: 0.000 min
Response: 40310718
Conc: 507.38 ng/ml


Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1248

Time

Response_

Signal: PP071419.D\ECD2B.ch

#25 AR-1248-5



R.T.: 5.742 min
Delta R.T.: -0.001 min
Response: 30029077
Conc: 515.16 ng/ml

Time

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071420.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 19:43
 Operator : YP\AJ
 Sample : AR1254ICV500
 Misc :
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1254

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 01:28:28 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 19:27:22 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.813	98111190	74345146	50.931	50.716
2) SA Decachlor...	10.232	8.849	73988546	45066273	50.788	50.509
Target Compounds						
26) L6 AR-1254-1	6.517	5.702	41728109	42206013	509.183	461.771
27) L6 AR-1254-2	6.733	5.850	64122746	35938664	504.601	456.583
28) L6 AR-1254-3	7.097	6.254	64574948	56055056	505.381	473.005
29) L6 AR-1254-4	7.379	6.482	58986526	36616929	501.956	474.524
30) L6 AR-1254-5	7.796	6.900	55244441	47442452	512.803	467.562



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071420.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 19:43
Operator : YP\AJ
Sample : AR1254ICV500
Misc :
ALS Vial : 34 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1254

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 01:28:28 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 19:27:22 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: 0.000 min
Response: 98111190
Conc: 50.93 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1254




#1 Tetrachloro-m-xylene

R.T.: 3.813 min
Delta R.T.: 0.000 min
Response: 74345146
Conc: 50.72 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.232 min
Delta R.T.: -0.004 min
Response: 73988546
Conc: 50.79 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.849 min
Delta R.T.: -0.001 min
Response: 45066273
Conc: 50.51 ng/ml



#26 AR-1254-1


R.T.: 6.517 min
Delta R.T.: -0.003 min
Response: 41728109
Conc: 509.18 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1254




#26 AR-1254-1

R.T.: 5.702 min
Delta R.T.: -0.001 min
Response: 42206013
Conc: 461.77 ng/ml




#27 AR-1254-2

R.T.: 6.733 min
Delta R.T.: -0.002 min
Response: 64122746
Conc: 504.60 ng/ml



#27 AR-1254-2


R.T.: 5.850 min
Delta R.T.: -0.001 min
Response: 35938664
Conc: 456.58 ng/ml



#28 AR-1254-3


R.T.: 7.097 min
Delta R.T.: -0.002 min
Response: 64574948
Conc: 505.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1254




#28 AR-1254-3

R.T.: 6.254 min
Delta R.T.: -0.001 min
Response: 56055056
Conc: 473.01 ng/ml




#29 AR-1254-4

R.T.: 7.379 min
Delta R.T.: -0.003 min
Response: 58986526
Conc: 501.96 ng/ml



#29 AR-1254-4


R.T.: 6.482 min
Delta R.T.: -0.002 min
Response: 36616929
Conc: 474.52 ng/ml



#30 AR-1254-5

R.T.: 7.796 min
Delta R.T.: -0.002 min
Response: 55244441
Conc: 512.80 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1254



#30 AR-1254-5

R.T.: 6.900 min
Delta R.T.: -0.002 min
Response: 47442452
Conc: 467.56 ng/ml

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071421.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 20:16
 Operator : YP\AJ
 Sample : AR1268ICV500
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 ICVPP042225AR1268

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 04/23/2025
 Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 23 04:47:52 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 04:41:50 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.516	3.814	98428428	75098106	49.661	51.289
2) SA Decachlor...	10.234	8.850	131.8E6	79464849	49.365	49.579
Target Compounds						
41) L9 AR-1268-1	8.733	7.722	157.9E6	98709054	493.783	486.283m
42) L9 AR-1268-2	8.827	7.788	132.3E6	84218105	493.704	487.500
43) L9 AR-1268-3	9.058	7.992	112.3E6	69254815	498.379	484.016
44) L9 AR-1268-4	9.477	8.286	48904008	30724632	493.945	483.689
45) L9 AR-1268-5	9.894	8.588	309.0E6	189.8E6	497.455	492.818

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071421.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 20:16
Operator : YP\AJ
Sample : AR1268ICV500
Misc :
ALS Vial : 35 Sample Multiplier: 1



Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1268


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 23 04:47:52 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 04:41:50 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene

R.T.: 4.516 min
Delta R.T.: 0.001 min
Response: 98428428
Conc: 49.66 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1268


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#1 Tetrachloro-m-xylene

R.T.: 3.814 min
Delta R.T.: 0.000 min
Response: 75098106
Conc: 51.29 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.234 min
Delta R.T.: -0.002 min
Response: 131775212
Conc: 49.36 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.850 min
Delta R.T.: -0.001 min
Response: 79464849
Conc: 49.58 ng/ml




#41 AR-1268-1

R.T.: 8.733 min
Delta R.T.: 0.000 min
Response: 157884565
Conc: 493.78 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1268


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#41 AR-1268-1

R.T.: 7.722 min
Delta R.T.: 0.000 min
Response: 98709054
Conc: 486.28 ng/ml




#42 AR-1268-2

R.T.: 8.827 min
Delta R.T.: 0.000 min
Response: 132293136
Conc: 493.70 ng/ml



#42 AR-1268-2

R.T.: 7.788 min
Delta R.T.: 0.000 min
Response: 84218105
Conc: 487.50 ng/ml




#43 AR-1268-3

R.T.: 9.058 min
Delta R.T.: 0.000 min
Response: 112276971
Conc: 498.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1268


Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025




#43 AR-1268-3

R.T.: 7.992 min
Delta R.T.: 0.000 min
Response: 69254815
Conc: 484.02 ng/ml




#44 AR-1268-4

R.T.: 9.477 min
Delta R.T.: 0.000 min
Response: 48904008
Conc: 493.94 ng/ml



#44 AR-1268-4

R.T.: 8.286 min
Delta R.T.: 0.000 min
Response: 30724632
Conc: 483.69 ng/ml




#45 AR-1268-5

R.T.: 9.894 min
Delta R.T.: 0.000 min
Response: 308950692
Conc: 497.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
ICVPP042225AR1268

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 04/23/2025
Supervised By :mohammad ahmed 04/24/2025



#45 AR-1268-5

R.T.: 8.588 min
Delta R.T.: 0.000 min
Response: 189821225
Conc: 492.82 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 09:24 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.85	4.86	4.76	4.96	0.01
Aroclor-1016-4	(4)	4.97	4.98	4.88	5.08	0.01
Aroclor-1016-5	(5)	5.23	5.23	5.13	5.33	0.00
Aroclor-1260-1	(1)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-2	(2)	6.46	6.46	6.36	6.56	0.00
Aroclor-1260-3	(3)	6.82	6.83	6.73	6.93	0.01
Aroclor-1260-4	(4)	7.09	7.09	6.99	7.19	0.00
Aroclor-1260-5	(5)	7.33	7.33	7.23	7.43	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.73	8.73	8.63	8.83	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 09:24 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-4	(4)	5.00	5.00	4.90	5.10	0.00
Aroclor-1016-5	(5)	5.21	5.22	5.12	5.32	0.01
Aroclor-1260-1	(1)	6.24	6.25	6.15	6.35	0.01
Aroclor-1260-2	(2)	6.43	6.43	6.33	6.53	0.00
Aroclor-1260-3	(3)	6.58	6.59	6.49	6.69	0.01
Aroclor-1260-4	(4)	7.05	7.06	6.96	7.16	0.01
Aroclor-1260-5	(5)	7.29	7.30	7.20	7.40	0.01
Tetrachloro-m-xylene		3.68	3.69	3.59	3.79	0.01
Decachlorobiphenyl		8.68	8.68	8.58	8.78	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110846.D Time Analyzed: 09:24

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.775	4.679	4.879	554.440	500.000	10.9
Aroclor-1016-2	4.795	4.698	4.898	558.040	500.000	11.6
Aroclor-1016-3	4.851	4.755	4.955	542.460	500.000	8.5
Aroclor-1016-4	4.971	4.875	5.075	553.630	500.000	10.7
Aroclor-1016-5	5.228	5.132	5.332	543.180	500.000	8.6
Aroclor-1260-1	6.268	6.172	6.372	535.720	500.000	7.1
Aroclor-1260-2	6.457	6.360	6.560	519.830	500.000	4.0
Aroclor-1260-3	6.824	6.729	6.929	522.530	500.000	4.5
Aroclor-1260-4	7.085	6.988	7.188	513.390	500.000	2.7
Aroclor-1260-5	7.326	7.230	7.430	515.480	500.000	3.1
Decachlorobiphenyl	8.727	8.632	8.832	47.400	50.000	-5.2
Tetrachloro-m-xylene	3.686	3.588	3.788	56.210	50.000	12.4



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL01 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110846.D Time Analyzed: 09:24

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.762	4.666	4.866	520.260	500.000	4.1
Aroclor-1016-2	4.781	4.684	4.884	519.530	500.000	3.9
Aroclor-1016-3	4.956	4.860	5.060	502.490	500.000	0.5
Aroclor-1016-4	4.998	4.902	5.102	493.190	500.000	-1.4
Aroclor-1016-5	5.211	5.115	5.315	490.700	500.000	-1.9
Aroclor-1260-1	6.242	6.146	6.346	492.670	500.000	-1.5
Aroclor-1260-2	6.430	6.334	6.534	484.910	500.000	-3.0
Aroclor-1260-3	6.582	6.487	6.687	480.540	500.000	-3.9
Aroclor-1260-4	7.053	6.957	7.157	480.890	500.000	-3.8
Aroclor-1260-5	7.294	7.197	7.397	482.240	500.000	-3.6
Decachlorobiphenyl	8.679	8.584	8.784	46.500	50.000	-7.0
Tetrachloro-m-xylene	3.683	3.586	3.786	53.130	50.000	6.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042925\
 Data File : PO110846.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2025 09:24
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 02:21:06 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 11 02:12:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.686	3.683	491.8E6	265.1E6	56.209	53.129
2) SA Decachlor...	8.727	8.679	374.2E6	89516389	47.402	46.502
Target Compounds						
3) L1 AR-1016-1	4.775	4.762	181.9E6	91310307	554.436	520.258
4) L1 AR-1016-2	4.795	4.781	253.9E6	130.6E6	558.043	519.530
5) L1 AR-1016-3	4.851	4.956	175.0E6	68193575	542.456	502.491
6) L1 AR-1016-4	4.971	4.998	137.9E6	56404722	553.628	493.187
7) L1 AR-1016-5	5.228	5.211	146.1E6	73150003	543.178	490.696
31) L7 AR-1260-1	6.268	6.242	253.1E6	121.2E6	535.716	492.669
32) L7 AR-1260-2	6.457	6.430	304.8E6	141.5E6	519.834	484.907
33) L7 AR-1260-3	6.824	6.582	257.5E6	130.2E6	522.531	480.536
34) L7 AR-1260-4	7.085	7.053	217.9E6	96484957	513.388	480.886
35) L7 AR-1260-5	7.326	7.294	537.9E6	221.3E6	515.484	482.241



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042925\
Data File : PO110846.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2025 09:24
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 02:21:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 02:12:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.686 min
Delta R.T.: -0.002 min
Response: 491775700
Conc: 56.21 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.683 min
Delta R.T.: -0.003 min
Response: 265100883
Conc: 53.13 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.727 min
Delta R.T.: -0.005 min
Response: 374237629
Conc: 47.40 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.679 min
Delta R.T.: -0.006 min
Response: 89516389
Conc: 46.50 ng/ml



#3 AR-1016-1


R.T.: 4.775 min
Delta R.T.: -0.003 min
Response: 181934708
Conc: 554.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.762 min
Delta R.T.: -0.004 min
Response: 91310307
Conc: 520.26 ng/ml




#4 AR-1016-2

R.T.: 4.795 min
Delta R.T.: -0.003 min
Response: 253880733
Conc: 558.04 ng/ml



#4 AR-1016-2


R.T.: 4.781 min
Delta R.T.: -0.004 min
Response: 130643641
Conc: 519.53 ng/ml



#5 AR-1016-3


R.T.: 4.851 min
Delta R.T.: -0.003 min
Response: 174957072
Conc: 542.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.956 min
Delta R.T.: -0.004 min
Response: 68193575
Conc: 502.49 ng/ml




#6 AR-1016-4

R.T.: 4.971 min
Delta R.T.: -0.004 min
Response: 137944453
Conc: 553.63 ng/ml



#6 AR-1016-4


R.T.: 4.998 min
Delta R.T.: -0.004 min
Response: 56404722
Conc: 493.19 ng/ml



#7 AR-1016-5


R.T.: 5.228 min
Delta R.T.: -0.004 min
Response: 146121301
Conc: 543.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.211 min
Delta R.T.: -0.004 min
Response: 73150003
Conc: 490.70 ng/ml




#31 AR-1260-1

R.T.: 6.268 min
Delta R.T.: -0.004 min
Response: 253096552
Conc: 535.72 ng/ml



#31 AR-1260-1


R.T.: 6.242 min
Delta R.T.: -0.004 min
Response: 121178736
Conc: 492.67 ng/ml



#32 AR-1260-2


R.T.: 6.457 min
Delta R.T.: -0.003 min
Response: 304843235
Conc: 519.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.430 min
Delta R.T.: -0.004 min
Response: 141472132
Conc: 484.91 ng/ml




#33 AR-1260-3

R.T.: 6.824 min
Delta R.T.: -0.004 min
Response: 257540996
Conc: 522.53 ng/ml



#33 AR-1260-3


R.T.: 6.582 min
Delta R.T.: -0.005 min
Response: 130169665
Conc: 480.54 ng/ml



#34 AR-1260-4


R.T.: 7.085 min
Delta R.T.: -0.004 min
Response: 217893666
Conc: 513.39 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.053 min
Delta R.T.: -0.004 min
Response: 96484957
Conc: 480.89 ng/ml



#35 AR-1260-5

R.T.: 7.326 min
Delta R.T.: -0.004 min
Response: 537892944
Conc: 515.48 ng/ml



#35 AR-1260-5

R.T.: 7.294 min
Delta R.T.: -0.003 min
Response: 221279829
Conc: 482.24 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 15:58 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-2	(2)	4.80	4.80	4.70	4.90	0.00
Aroclor-1016-3	(3)	4.85	4.86	4.76	4.96	0.01
Aroclor-1016-4	(4)	4.97	4.98	4.88	5.08	0.01
Aroclor-1016-5	(5)	5.23	5.23	5.13	5.33	0.00
Aroclor-1260-1	(1)	6.27	6.27	6.17	6.37	0.00
Aroclor-1260-2	(2)	6.46	6.46	6.36	6.56	0.00
Aroclor-1260-3	(3)	6.83	6.83	6.73	6.93	0.01
Aroclor-1260-4	(4)	7.08	7.09	6.99	7.19	0.01
Aroclor-1260-5	(5)	7.33	7.33	7.23	7.43	0.00
Tetrachloro-m-xylene		3.69	3.69	3.59	3.79	0.00
Decachlorobiphenyl		8.73	8.73	8.63	8.83	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/10/2025 04/10/2025

Continuing Calib Time: 15:58 Initial Calibration Time(s): 09:36 17:52

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.76	4.77	4.67	4.87	0.01
Aroclor-1016-2	(2)	4.78	4.78	4.68	4.88	0.00
Aroclor-1016-3	(3)	4.96	4.96	4.86	5.06	0.00
Aroclor-1016-4	(4)	5.00	5.00	4.90	5.10	0.00
Aroclor-1016-5	(5)	5.21	5.22	5.12	5.32	0.01
Aroclor-1260-1	(1)	6.24	6.25	6.15	6.35	0.01
Aroclor-1260-2	(2)	6.43	6.43	6.33	6.53	0.00
Aroclor-1260-3	(3)	6.58	6.59	6.49	6.69	0.01
Aroclor-1260-4	(4)	7.05	7.06	6.96	7.16	0.01
Aroclor-1260-5	(5)	7.29	7.30	7.20	7.40	0.01
Tetrachloro-m-xylene		3.68	3.69	3.59	3.79	0.01
Decachlorobiphenyl		8.68	8.68	8.58	8.78	0.00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110861.D Time Analyzed: 15:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.776	4.679	4.879	567.510	500.000	13.5
Aroclor-1016-2	4.795	4.698	4.898	567.550	500.000	13.5
Aroclor-1016-3	4.851	4.755	4.955	552.160	500.000	10.4
Aroclor-1016-4	4.972	4.875	5.075	566.160	500.000	13.2
Aroclor-1016-5	5.229	5.132	5.332	568.900	500.000	13.8
Aroclor-1260-1	6.268	6.172	6.372	552.510	500.000	10.5
Aroclor-1260-2	6.457	6.360	6.560	540.900	500.000	8.2
Aroclor-1260-3	6.825	6.729	6.929	541.450	500.000	8.3
Aroclor-1260-4	7.084	6.988	7.188	532.720	500.000	6.5
Aroclor-1260-5	7.327	7.230	7.430	535.700	500.000	7.1
Decachlorobiphenyl	8.728	8.632	8.832	49.210	50.000	-1.6
Tetrachloro-m-xylene	3.686	3.588	3.788	57.430	50.000	14.9



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/10/2025 04/10/2025

Client Sample No.: CCAL02 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PO110861.D Time Analyzed: 15:58

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.763	4.666	4.866	533.100	500.000	6.6
Aroclor-1016-2	4.781	4.684	4.884	531.290	500.000	6.3
Aroclor-1016-3	4.956	4.860	5.060	513.330	500.000	2.7
Aroclor-1016-4	4.998	4.902	5.102	496.070	500.000	-0.8
Aroclor-1016-5	5.211	5.115	5.315	510.550	500.000	2.1
Aroclor-1260-1	6.242	6.146	6.346	502.200	500.000	0.4
Aroclor-1260-2	6.429	6.334	6.534	495.650	500.000	-0.9
Aroclor-1260-3	6.582	6.487	6.687	487.880	500.000	-2.4
Aroclor-1260-4	7.053	6.957	7.157	485.750	500.000	-2.9
Aroclor-1260-5	7.294	7.197	7.397	488.550	500.000	-2.3
Decachlorobiphenyl	8.679	8.584	8.784	46.590	50.000	-6.8
Tetrachloro-m-xylene	3.683	3.586	3.786	54.630	50.000	9.3

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042925\
 Data File : PO110861.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2025 15:58
 Operator : YP/AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 02:26:07 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 11 02:12:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.686	3.683	502.5E6	272.6E6	57.434	54.630
2) SA Decachlor...	8.728	8.679	388.5E6	89691972	49.213	46.593
Target Compounds						
3) L1 AR-1016-1	4.776	4.763	186.2E6	93564868	567.510	533.104
4) L1 AR-1016-2	4.795	4.781	258.2E6	133.6E6	567.553	531.286
5) L1 AR-1016-3	4.851	4.956	178.1E6	69663894	552.164	513.326
6) L1 AR-1016-4	4.972	4.998	141.1E6	56734615	566.157	496.071
7) L1 AR-1016-5	5.229	5.211	153.0E6	76109735	568.905	510.550
31) L7 AR-1260-1	6.268	6.242	261.0E6	123.5E6	552.508	502.198
32) L7 AR-1260-2	6.457	6.429	317.2E6	144.6E6	540.898	495.646
33) L7 AR-1260-3	6.825	6.582	266.9E6	132.2E6	541.450	487.884
34) L7 AR-1260-4	7.084	7.053	226.1E6	97461414	532.723	485.753
35) L7 AR-1260-5	7.327	7.294	559.0E6	224.2E6	535.705	488.553



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042925\
Data File : PO110861.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2025 15:58
Operator : YP/AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 02:26:07 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 02:12:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.686 min
Delta R.T.: -0.002 min
Response: 502496938
Conc: 57.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.683 min
Delta R.T.: -0.003 min
Response: 272589697
Conc: 54.63 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.728 min
Delta R.T.: -0.004 min
Response: 388536408
Conc: 49.21 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.679 min
Delta R.T.: -0.005 min
Response: 89691972
Conc: 46.59 ng/ml



#3 AR-1016-1


R.T.: 4.776 min
Delta R.T.: -0.003 min
Response: 186224740
Conc: 567.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.763 min
Delta R.T.: -0.003 min
Response: 93564868
Conc: 533.10 ng/ml




#4 AR-1016-2

R.T.: 4.795 min
Delta R.T.: -0.003 min
Response: 258207362
Conc: 567.55 ng/ml



#4 AR-1016-2


R.T.: 4.781 min
Delta R.T.: -0.003 min
Response: 133599849
Conc: 531.29 ng/ml



#5 AR-1016-3


R.T.: 4.851 min
Delta R.T.: -0.003 min
Response: 178087994
Conc: 552.16 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 4.956 min
Delta R.T.: -0.004 min
Response: 69663894
Conc: 513.33 ng/ml




#6 AR-1016-4

R.T.: 4.972 min
Delta R.T.: -0.003 min
Response: 141066143
Conc: 566.16 ng/ml



#6 AR-1016-4


R.T.: 4.998 min
Delta R.T.: -0.004 min
Response: 56734615
Conc: 496.07 ng/ml



#7 AR-1016-5


R.T.: 5.229 min
Delta R.T.: -0.003 min
Response: 153041965
Conc: 568.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.211 min
Delta R.T.: -0.004 min
Response: 76109735
Conc: 510.55 ng/ml




#31 AR-1260-1

R.T.: 6.268 min
Delta R.T.: -0.004 min
Response: 261029889
Conc: 552.51 ng/ml



#31 AR-1260-1


R.T.: 6.242 min
Delta R.T.: -0.004 min
Response: 123522395
Conc: 502.20 ng/ml



#32 AR-1260-2


R.T.: 6.457 min
Delta R.T.: -0.003 min
Response: 317195671
Conc: 540.90 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.429 min
Delta R.T.: -0.005 min
Response: 144605151
Conc: 495.65 ng/ml




#33 AR-1260-3

R.T.: 6.825 min
Delta R.T.: -0.004 min
Response: 266865989
Conc: 541.45 ng/ml



#33 AR-1260-3


R.T.: 6.582 min
Delta R.T.: -0.005 min
Response: 132160031
Conc: 487.88 ng/ml



#34 AR-1260-4


R.T.: 7.084 min
Delta R.T.: -0.004 min
Response: 226099797
Conc: 532.72 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.053 min
Delta R.T.: -0.004 min
Response: 97461414
Conc: 485.75 ng/ml



#35 AR-1260-5

R.T.: 7.327 min
Delta R.T.: -0.004 min
Response: 558992942
Conc: 535.70 ng/ml



#35 AR-1260-5

R.T.: 7.294 min
Delta R.T.: -0.003 min
Response: 224176425
Conc: 488.55 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 18:03 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.67	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.00
Aroclor-1016-4	(4)	5.85	5.85	5.75	5.95	0.00
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.26	7.27	7.17	7.37	0.01
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.41	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.23	10.24	10.14	10.34	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 18:03 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.90	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.35	5.35	5.25	5.45	0.00
Aroclor-1260-1	(1)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-2	(2)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP071557.D Time Analyzed: 18:03

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.665	5.571	5.771	479.350	500.000	-4.1
Aroclor-1016-2	5.686	5.593	5.793	482.270	500.000	-3.5
Aroclor-1016-3	5.748	5.654	5.854	479.730	500.000	-4.1
Aroclor-1016-4	5.846	5.752	5.952	487.050	500.000	-2.6
Aroclor-1016-5	6.139	6.045	6.245	473.110	500.000	-5.4
Aroclor-1260-1	7.257	7.165	7.365	468.870	500.000	-6.2
Aroclor-1260-2	7.511	7.418	7.618	472.380	500.000	-5.5
Aroclor-1260-3	7.869	7.777	7.977	486.210	500.000	-2.8
Aroclor-1260-4	8.094	8.002	8.202	464.380	500.000	-7.1
Aroclor-1260-5	8.414	8.322	8.522	483.210	500.000	-3.4
Decachlorobiphenyl	10.226	10.137	10.337	49.680	50.000	-0.6
Tetrachloro-m-xylene	4.512	4.417	4.617	48.490	50.000	-3.0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL03 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP071557.D Time Analyzed: 18:03

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.895	4.798	4.998	508.180	500.000	1.6
Aroclor-1016-2	4.913	4.816	5.016	513.170	500.000	2.6
Aroclor-1016-3	5.090	4.993	5.193	522.490	500.000	4.5
Aroclor-1016-4	5.131	5.036	5.236	523.600	500.000	4.7
Aroclor-1016-5	5.346	5.250	5.450	524.280	500.000	4.9
Aroclor-1260-1	6.380	6.285	6.485	500.820	500.000	0.2
Aroclor-1260-2	6.569	6.473	6.673	499.190	500.000	-0.2
Aroclor-1260-3	6.721	6.627	6.827	511.360	500.000	2.3
Aroclor-1260-4	7.193	7.098	7.298	524.290	500.000	4.9
Aroclor-1260-5	7.434	7.339	7.539	517.900	500.000	3.6
Decachlorobiphenyl	8.842	8.749	8.949	50.210	50.000	0.4
Tetrachloro-m-xylene	3.810	3.712	3.912	47.590	50.000	-4.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
 Data File : PP071557.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Apr 2025 18:03
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 01:10:04 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.512	3.810	96136716	67040243	48.489	47.592
2) SA Decachlor...	10.226	8.842	71932550	44034229	49.677	50.207
Target Compounds						
3) L1 AR-1016-1	5.665	4.895	32140554	27204817	479.353	508.182
4) L1 AR-1016-2	5.686	4.913	49494353	39187780	482.268	513.173
5) L1 AR-1016-3	5.748	5.090	29573698	22054623	479.726	522.492
6) L1 AR-1016-4	5.846	5.131	24984741	18056780	487.050	523.598
7) L1 AR-1016-5	6.139	5.346	22822119	23302699	473.111	524.277
31) L7 AR-1260-1	7.257	6.380	45022935	36209079	468.869	500.822
32) L7 AR-1260-2	7.511	6.569	67587080	43659759	472.381	499.189
33) L7 AR-1260-3	7.869	6.721	54506219	40084119	486.211	511.358
34) L7 AR-1260-4	8.094	7.193	52412659	33087962	464.384	524.285
35) L7 AR-1260-5	8.414	7.434	109.5E6	78087286	483.212	517.898



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
Data File : PP071557.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2025 18:03
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 29 01:10:04 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.512 min
Delta R.T.: -0.005 min
Response: 96136716
Conc: 48.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.810 min
Delta R.T.: -0.002 min
Response: 67040243
Conc: 47.59 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.226 min
Delta R.T.: -0.011 min
Response: 71932550
Conc: 49.68 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.842 min
Delta R.T.: -0.007 min
Response: 44034229
Conc: 50.21 ng/ml



#3 AR-1016-1


R.T.: 5.665 min
Delta R.T.: -0.006 min
Response: 32140554
Conc: 479.35 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.895 min
Delta R.T.: -0.003 min
Response: 27204817
Conc: 508.18 ng/ml




#4 AR-1016-2

R.T.: 5.686 min
Delta R.T.: -0.006 min
Response: 49494353
Conc: 482.27 ng/ml



#4 AR-1016-2


R.T.: 4.913 min
Delta R.T.: -0.003 min
Response: 39187780
Conc: 513.17 ng/ml



#5 AR-1016-3


R.T.: 5.748 min
Delta R.T.: -0.006 min
Response: 29573698
Conc: 479.73 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.090 min
Delta R.T.: -0.003 min
Response: 22054623
Conc: 522.49 ng/ml




#6 AR-1016-4

R.T.: 5.846 min
Delta R.T.: -0.006 min
Response: 24984741
Conc: 487.05 ng/ml



#6 AR-1016-4


R.T.: 5.131 min
Delta R.T.: -0.004 min
Response: 18056780
Conc: 523.60 ng/ml



#7 AR-1016-5


R.T.: 6.139 min
Delta R.T.: -0.006 min
Response: 22822119
Conc: 473.11 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.346 min
Delta R.T.: -0.004 min
Response: 23302699
Conc: 524.28 ng/ml




#31 AR-1260-1

R.T.: 7.257 min
Delta R.T.: -0.007 min
Response: 45022935
Conc: 468.87 ng/ml



#31 AR-1260-1


R.T.: 6.380 min
Delta R.T.: -0.005 min
Response: 36209079
Conc: 500.82 ng/ml



#32 AR-1260-2


R.T.: 7.511 min
Delta R.T.: -0.007 min
Response: 67587080
Conc: 472.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.569 min
Delta R.T.: -0.005 min
Response: 43659759
Conc: 499.19 ng/ml




#33 AR-1260-3

R.T.: 7.869 min
Delta R.T.: -0.008 min
Response: 54506219
Conc: 486.21 ng/ml



#33 AR-1260-3


R.T.: 6.721 min
Delta R.T.: -0.006 min
Response: 40084119
Conc: 511.36 ng/ml



#34 AR-1260-4


R.T.: 8.094 min
Delta R.T.: -0.007 min
Response: 52412659
Conc: 464.38 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.193 min
Delta R.T.: -0.005 min
Response: 33087962
Conc: 524.29 ng/ml



#35 AR-1260-5

R.T.: 8.414 min
Delta R.T.: -0.008 min
Response: 109482924
Conc: 483.21 ng/ml



#35 AR-1260-5

R.T.: 7.434 min
Delta R.T.: -0.004 min
Response: 78087286
Conc: 517.90 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 23:31 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.68	5.69	5.59	5.79	0.01
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.01
Aroclor-1016-4	(4)	5.84	5.85	5.75	5.95	0.01
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.25	7.27	7.17	7.37	0.02
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.41	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.22	10.24	10.14	10.34	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 23:31 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.34	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-2	(2)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP071572.D Time Analyzed: 23:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.661	5.571	5.771	502.500	500.000	0.5
Aroclor-1016-2	5.683	5.593	5.793	488.520	500.000	-2.3
Aroclor-1016-3	5.745	5.654	5.854	492.490	500.000	-1.5
Aroclor-1016-4	5.843	5.752	5.952	494.900	500.000	-1.0
Aroclor-1016-5	6.135	6.045	6.245	486.570	500.000	-2.7
Aroclor-1260-1	7.254	7.165	7.365	498.210	500.000	-0.4
Aroclor-1260-2	7.508	7.418	7.618	488.990	500.000	-2.2
Aroclor-1260-3	7.866	7.777	7.977	505.180	500.000	1.0
Aroclor-1260-4	8.090	8.002	8.202	488.020	500.000	-2.4
Aroclor-1260-5	8.410	8.322	8.522	497.350	500.000	-0.5
Decachlorobiphenyl	10.223	10.137	10.337	50.350	50.000	0.7
Tetrachloro-m-xylene	4.508	4.417	4.617	49.040	50.000	-1.9



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL04 Date Analyzed: 04/28/2025

Lab Sample No.: AR1660CCC500 Data File : PP071572.D Time Analyzed: 23:31

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.893	4.798	4.998	521.320	500.000	4.3
Aroclor-1016-2	4.912	4.816	5.016	520.920	500.000	4.2
Aroclor-1016-3	5.088	4.993	5.193	527.030	500.000	5.4
Aroclor-1016-4	5.130	5.036	5.236	522.470	500.000	4.5
Aroclor-1016-5	5.344	5.250	5.450	552.040	500.000	10.4
Aroclor-1260-1	6.378	6.285	6.485	531.810	500.000	6.4
Aroclor-1260-2	6.567	6.473	6.673	528.120	500.000	5.6
Aroclor-1260-3	6.720	6.627	6.827	531.940	500.000	6.4
Aroclor-1260-4	7.191	7.098	7.298	532.870	500.000	6.6
Aroclor-1260-5	7.432	7.339	7.539	540.330	500.000	8.1
Decachlorobiphenyl	8.840	8.749	8.949	51.350	50.000	2.7
Tetrachloro-m-xylene	3.809	3.712	3.912	51.300	50.000	2.6

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
 Data File : PP071572.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Apr 2025 23:31
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 01:14:20 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.508	3.809	97227226	72264120	49.039	51.301
2) SA Decachlor...	10.223	8.840	72903095	45036088	50.347	51.349
Target Compounds						
3) L1 AR-1016-1	5.661	4.893	33692750	27907959	502.503	521.317
4) L1 AR-1016-2	5.683	4.912	50135562	39779132	488.516	520.917
5) L1 AR-1016-3	5.745	5.088	30360714	22246250	492.492	527.032
6) L1 AR-1016-4	5.843	5.130	25387243	18018006	494.896	522.474
7) L1 AR-1016-5	6.135	5.344	23471285	24536592	486.569	552.038
31) L7 AR-1260-1	7.254	6.378	47840542	38449326	498.211	531.808
32) L7 AR-1260-2	7.508	6.567	69963827	46190035	488.993	528.119
33) L7 AR-1260-3	7.866	6.720	56632991	41697224	505.182	531.937
34) L7 AR-1260-4	8.090	7.191	55079856	33629700	488.016	532.869
35) L7 AR-1260-5	8.410	7.432	112.7E6	81468994	497.352	540.327



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
Data File : PP071572.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2025 23:31
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 29 01:14:20 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.508 min
Delta R.T.: -0.009 min
Response: 97227226
Conc: 49.04 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.809 min
Delta R.T.: -0.003 min
Response: 72264120
Conc: 51.30 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.223 min
Delta R.T.: -0.015 min
Response: 72903095
Conc: 50.35 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.840 min
Delta R.T.: -0.009 min
Response: 45036088
Conc: 51.35 ng/ml



#3 AR-1016-1


R.T.: 5.661 min
Delta R.T.: -0.010 min
Response: 33692750
Conc: 502.50 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.893 min
Delta R.T.: -0.005 min
Response: 27907959
Conc: 521.32 ng/ml




#4 AR-1016-2

R.T.: 5.683 min
Delta R.T.: -0.009 min
Response: 50135562
Conc: 488.52 ng/ml



#4 AR-1016-2


R.T.: 4.912 min
Delta R.T.: -0.005 min
Response: 39779132
Conc: 520.92 ng/ml



#5 AR-1016-3


R.T.: 5.745 min
Delta R.T.: -0.009 min
Response: 30360714
Conc: 492.49 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.088 min
Delta R.T.: -0.005 min
Response: 22246250
Conc: 527.03 ng/ml




#6 AR-1016-4

R.T.: 5.843 min
Delta R.T.: -0.009 min
Response: 25387243
Conc: 494.90 ng/ml



#6 AR-1016-4


R.T.: 5.130 min
Delta R.T.: -0.005 min
Response: 18018006
Conc: 522.47 ng/ml



#7 AR-1016-5


R.T.: 6.135 min
Delta R.T.: -0.010 min
Response: 23471285
Conc: 486.57 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.344 min
Delta R.T.: -0.006 min
Response: 24536592
Conc: 552.04 ng/ml




#31 AR-1260-1

R.T.: 7.254 min
Delta R.T.: -0.011 min
Response: 47840542
Conc: 498.21 ng/ml



#31 AR-1260-1


R.T.: 6.378 min
Delta R.T.: -0.007 min
Response: 38449326
Conc: 531.81 ng/ml



#32 AR-1260-2


R.T.: 7.508 min
Delta R.T.: -0.011 min
Response: 69963827
Conc: 488.99 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.567 min
Delta R.T.: -0.007 min
Response: 46190035
Conc: 528.12 ng/ml




#33 AR-1260-3

R.T.: 7.866 min
Delta R.T.: -0.011 min
Response: 56632991
Conc: 505.18 ng/ml



#33 AR-1260-3


R.T.: 6.720 min
Delta R.T.: -0.007 min
Response: 41697224
Conc: 531.94 ng/ml



#34 AR-1260-4


R.T.: 8.090 min
Delta R.T.: -0.011 min
Response: 55079856
Conc: 488.02 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.191 min
Delta R.T.: -0.007 min
Response: 33629700
Conc: 532.87 ng/ml



#35 AR-1260-5

R.T.: 8.410 min
Delta R.T.: -0.012 min
Response: 112686714
Conc: 497.35 ng/ml



#35 AR-1260-5

R.T.: 7.432 min
Delta R.T.: -0.006 min
Response: 81468994
Conc: 540.33 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 04:58 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.66	5.67	5.57	5.77	0.01
Aroclor-1016-2	(2)	5.68	5.69	5.59	5.79	0.01
Aroclor-1016-3	(3)	5.75	5.75	5.65	5.85	0.01
Aroclor-1016-4	(4)	5.84	5.85	5.75	5.95	0.01
Aroclor-1016-5	(5)	6.14	6.15	6.05	6.25	0.01
Aroclor-1260-1	(1)	7.25	7.27	7.17	7.37	0.02
Aroclor-1260-2	(2)	7.51	7.52	7.42	7.62	0.01
Aroclor-1260-3	(3)	7.87	7.88	7.78	7.98	0.01
Aroclor-1260-4	(4)	8.09	8.10	8.00	8.20	0.01
Aroclor-1260-5	(5)	8.41	8.42	8.32	8.52	0.01
Tetrachloro-m-xylene		4.51	4.52	4.42	4.62	0.01
Decachlorobiphenyl		10.22	10.24	10.14	10.34	0.02



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Continuing Calib Date: 04/29/2025 Initial Calibration Date(s): 04/22/2025 04/22/2025

Continuing Calib Time: 04:58 Initial Calibration Time(s): 10:29 17:49

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.89	4.90	4.80	5.00	0.01
Aroclor-1016-2	(2)	4.91	4.92	4.82	5.02	0.01
Aroclor-1016-3	(3)	5.09	5.09	4.99	5.19	0.00
Aroclor-1016-4	(4)	5.13	5.14	5.04	5.24	0.01
Aroclor-1016-5	(5)	5.34	5.35	5.25	5.45	0.01
Aroclor-1260-1	(1)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-2	(2)	6.57	6.57	6.47	6.67	0.00
Aroclor-1260-3	(3)	6.72	6.73	6.63	6.83	0.01
Aroclor-1260-4	(4)	7.19	7.20	7.10	7.30	0.01
Aroclor-1260-5	(5)	7.43	7.44	7.34	7.54	0.01
Tetrachloro-m-xylene		3.81	3.81	3.71	3.91	0.00
Decachlorobiphenyl		8.84	8.85	8.75	8.95	0.01



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071587.D Time Analyzed: 04:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.661	5.571	5.771	499.460	500.000	-0.1
Aroclor-1016-2	5.683	5.593	5.793	497.330	500.000	-0.5
Aroclor-1016-3	5.745	5.654	5.854	491.000	500.000	-1.8
Aroclor-1016-4	5.842	5.752	5.952	494.630	500.000	-1.1
Aroclor-1016-5	6.135	6.045	6.245	504.640	500.000	0.9
Aroclor-1260-1	7.254	7.165	7.365	524.920	500.000	5.0
Aroclor-1260-2	7.507	7.418	7.618	503.170	500.000	0.6
Aroclor-1260-3	7.866	7.777	7.977	514.880	500.000	3.0
Aroclor-1260-4	8.090	8.002	8.202	502.310	500.000	0.5
Aroclor-1260-5	8.409	8.322	8.522	515.560	500.000	3.1
Decachlorobiphenyl	10.222	10.137	10.337	52.270	50.000	4.5
Tetrachloro-m-xylene	4.508	4.417	4.617	49.750	50.000	-0.5



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/22/2025 04/22/2025

Client Sample No.: CCAL05 Date Analyzed: 04/29/2025

Lab Sample No.: AR1660CCC500 Data File : PP071587.D Time Analyzed: 04:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.892	4.798	4.998	520.170	500.000	4.0
Aroclor-1016-2	4.911	4.816	5.016	523.850	500.000	4.8
Aroclor-1016-3	5.088	4.993	5.193	534.650	500.000	6.9
Aroclor-1016-4	5.130	5.036	5.236	526.290	500.000	5.3
Aroclor-1016-5	5.344	5.250	5.450	567.310	500.000	13.5
Aroclor-1260-1	6.378	6.285	6.485	527.630	500.000	5.5
Aroclor-1260-2	6.567	6.473	6.673	513.280	500.000	2.7
Aroclor-1260-3	6.719	6.627	6.827	518.650	500.000	3.7
Aroclor-1260-4	7.190	7.098	7.298	511.600	500.000	2.3
Aroclor-1260-5	7.432	7.339	7.539	517.450	500.000	3.5
Decachlorobiphenyl	8.840	8.749	8.949	51.260	50.000	2.5
Tetrachloro-m-xylene	3.808	3.712	3.912	50.380	50.000	0.8

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
 Data File : PP071587.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2025 04:58
 Operator : YP\AJ
 Sample : AR1660CCC500
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 AR1660CCC500

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 06:32:09 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.508	3.808	98630046	70966791	49.747	50.380
2) SA Decachlor...	10.222	8.840	75686412	44958662	52.269	51.261
Target Compounds						
3) L1 AR-1016-1	5.661	4.892	33488935	27846711	499.463	520.173
4) L1 AR-1016-2	5.683	4.911	51040394	40002961	497.332	523.848
5) L1 AR-1016-3	5.745	5.088	30269022	22567659	491.005	534.647
6) L1 AR-1016-4	5.842	5.130	25373407	18149541	494.626	526.288
7) L1 AR-1016-5	6.135	5.344	24343128	25215289	504.642	567.308
31) L7 AR-1260-1	7.254	6.378	50405387	38146893	524.922	527.625
32) L7 AR-1260-2	7.507	6.567	71992443	44892452	503.171	513.283
33) L7 AR-1260-3	7.866	6.719	57720285	40655396	514.881	518.646
34) L7 AR-1260-4	8.090	7.190	56693309	32287501	502.312	511.602
35) L7 AR-1260-5	8.409	7.432	116.8E6	78020402	515.560	517.454



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
Data File : PP071587.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2025 04:58
Operator : YP\AJ
Sample : AR1660CCC500
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 29 06:32:09 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.508 min
Delta R.T.: -0.009 min
Response: 98630046
Conc: 49.75 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#1 Tetrachloro-m-xylene

R.T.: 3.808 min
Delta R.T.: -0.004 min
Response: 70966791
Conc: 50.38 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.222 min
Delta R.T.: -0.015 min
Response: 75686412
Conc: 52.27 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.840 min
Delta R.T.: -0.009 min
Response: 44958662
Conc: 51.26 ng/ml



#3 AR-1016-1


R.T.: 5.661 min
Delta R.T.: -0.010 min
Response: 33488935
Conc: 499.46 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#3 AR-1016-1

R.T.: 4.892 min
Delta R.T.: -0.006 min
Response: 27846711
Conc: 520.17 ng/ml




#4 AR-1016-2

R.T.: 5.683 min
Delta R.T.: -0.010 min
Response: 51040394
Conc: 497.33 ng/ml



#4 AR-1016-2


R.T.: 4.911 min
Delta R.T.: -0.006 min
Response: 40002961
Conc: 523.85 ng/ml



#5 AR-1016-3


R.T.: 5.745 min
Delta R.T.: -0.009 min
Response: 30269022
Conc: 491.00 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#5 AR-1016-3

R.T.: 5.088 min
Delta R.T.: -0.005 min
Response: 22567659
Conc: 534.65 ng/ml




#6 AR-1016-4

R.T.: 5.842 min
Delta R.T.: -0.009 min
Response: 25373407
Conc: 494.63 ng/ml



#6 AR-1016-4


R.T.: 5.130 min
Delta R.T.: -0.006 min
Response: 18149541
Conc: 526.29 ng/ml



#7 AR-1016-5


R.T.: 6.135 min
Delta R.T.: -0.010 min
Response: 24343128
Conc: 504.64 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#7 AR-1016-5

R.T.: 5.344 min
Delta R.T.: -0.006 min
Response: 25215289
Conc: 567.31 ng/ml




#31 AR-1260-1

R.T.: 7.254 min
Delta R.T.: -0.010 min
Response: 50405387
Conc: 524.92 ng/ml



#31 AR-1260-1


R.T.: 6.378 min
Delta R.T.: -0.007 min
Response: 38146893
Conc: 527.63 ng/ml



#32 AR-1260-2


R.T.: 7.507 min
Delta R.T.: -0.011 min
Response: 71992443
Conc: 503.17 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#32 AR-1260-2

R.T.: 6.567 min
Delta R.T.: -0.006 min
Response: 44892452
Conc: 513.28 ng/ml




#33 AR-1260-3

R.T.: 7.866 min
Delta R.T.: -0.011 min
Response: 57720285
Conc: 514.88 ng/ml



#33 AR-1260-3


R.T.: 6.719 min
Delta R.T.: -0.007 min
Response: 40655396
Conc: 518.65 ng/ml



#34 AR-1260-4


R.T.: 8.090 min
Delta R.T.: -0.011 min
Response: 56693309
Conc: 502.31 ng/ml

Instrument :
ECD_P
ClientSampleId :
AR1660CCC500




#34 AR-1260-4

R.T.: 7.190 min
Delta R.T.: -0.008 min
Response: 32287501
Conc: 511.60 ng/ml



#35 AR-1260-5

R.T.: 8.409 min
Delta R.T.: -0.012 min
Response: 116811998
Conc: 515.56 ng/ml



#35 AR-1260-5

R.T.: 7.432 min
Delta R.T.: -0.007 min
Response: 78020402
Conc: 517.45 ng/ml

Analytical Sequence

Client: Kleinfelder

SDG No.: Q1889

Project: Mitchell School

Instrument ID: ECD_O

GC Column: ZB-MR1

ID: 0.32 (mm)

Inst. Calib. Date(s): 04/10/2025 04/10/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	04/10/2025	09:17	PO110348.D	8.73	3.69
AR1660ICC1000	AR1660ICC1000	04/10/2025	09:36	PO110349.D	8.73	3.69
AR1660ICC750	AR1660ICC750	04/10/2025	09:54	PO110350.D	8.73	3.69
AR1660ICC500	AR1660ICC500	04/10/2025	10:13	PO110351.D	8.73	3.69
AR1660ICC250	AR1660ICC250	04/10/2025	10:31	PO110352.D	8.73	3.69
AR1660ICC050	AR1660ICC050	04/10/2025	10:49	PO110353.D	8.73	3.69
AR1221ICC500	AR1221ICC500	04/10/2025	11:08	PO110354.D	8.73	3.69
AR1232ICC500	AR1232ICC500	04/10/2025	11:26	PO110355.D	8.73	3.69
AR1242ICC1000	AR1242ICC1000	04/10/2025	11:44	PO110356.D	8.73	3.69
AR1242ICC750	AR1242ICC750	04/10/2025	12:03	PO110357.D	8.73	3.69
AR1242ICC500	AR1242ICC500	04/10/2025	12:21	PO110358.D	8.73	3.69
AR1242ICC250	AR1242ICC250	04/10/2025	12:39	PO110359.D	8.73	3.69
AR1242ICC050	AR1242ICC050	04/10/2025	12:58	PO110360.D	8.73	3.69
AR1248ICC1000	AR1248ICC1000	04/10/2025	13:16	PO110361.D	8.73	3.69
AR1248ICC750	AR1248ICC750	04/10/2025	13:35	PO110362.D	8.73	3.69
AR1248ICC500	AR1248ICC500	04/10/2025	13:53	PO110363.D	8.73	3.69
AR1248ICC250	AR1248ICC250	04/10/2025	14:11	PO110364.D	8.73	3.69
AR1248ICC050	AR1248ICC050	04/10/2025	14:30	PO110365.D	8.73	3.69
AR1254ICC1000	AR1254ICC1000	04/10/2025	14:48	PO110366.D	8.73	3.69
AR1254ICC750	AR1254ICC750	04/10/2025	15:06	PO110367.D	8.73	3.69
AR1254ICC500	AR1254ICC500	04/10/2025	15:25	PO110368.D	8.73	3.69
AR1254ICC250	AR1254ICC250	04/10/2025	15:43	PO110369.D	8.73	3.69
AR1254ICC050	AR1254ICC050	04/10/2025	16:02	PO110370.D	8.73	3.69
AR1262ICC500	AR1262ICC500	04/10/2025	16:20	PO110371.D	8.73	3.69
AR1268ICC1000	AR1268ICC1000	04/10/2025	16:38	PO110372.D	8.73	3.69
AR1268ICC750	AR1268ICC750	04/10/2025	16:57	PO110373.D	8.73	3.69
AR1268ICC500	AR1268ICC500	04/10/2025	17:15	PO110374.D	8.73	3.69
AR1268ICC250	AR1268ICC250	04/10/2025	17:33	PO110375.D	8.73	3.69
AR1268ICC050	AR1268ICC050	04/10/2025	17:52	PO110376.D	8.73	3.69
AR1660CCC500	AR1660CCC500	04/29/2025	09:24	PO110846.D	8.73	3.69
IBLK	IBLK	04/29/2025	10:40	PO110850.D	8.73	3.69
COMP-1	Q1889-01	04/29/2025	11:16	PO110852.D	8.73	3.69
COMP-1MS	Q1889-01MS	04/29/2025	11:35	PO110853.D	8.73	3.69
COMP-1MSD	Q1889-01MSD	04/29/2025	11:53	PO110854.D	8.73	3.69
AR1660CCC500	AR1660CCC500	04/29/2025	15:58	PO110861.D	8.73	3.69
IBLK	IBLK	04/29/2025	17:29	PO110865.D	8.73	3.69
IBLK	IBLK	04/22/2025	10:13	PP071388.D	10.23	4.51
AR1660ICC1000	AR1660ICC1000	04/22/2025	10:29	PP071389.D	10.23	4.51
AR1660ICC750	AR1660ICC750	04/22/2025	10:45	PP071390.D	10.23	4.51
AR1660ICC500	AR1660ICC500	04/22/2025	11:02	PP071391.D	10.24	4.52
AR1660ICC250	AR1660ICC250	04/22/2025	11:18	PP071392.D	10.23	4.51
AR1660ICC050	AR1660ICC050	04/22/2025	11:34	PP071393.D	10.24	4.52

Analytical Sequence

AR1221ICC500	AR1221ICC500	04/22/2025	11:51	PP071394.D	10.23	4.51
AR1232ICC500	AR1232ICC500	04/22/2025	12:07	PP071395.D	10.23	4.52
AR1242ICC1000	AR1242ICC1000	04/22/2025	12:23	PP071396.D	10.23	4.51
AR1242ICC750	AR1242ICC750	04/22/2025	12:39	PP071397.D	10.24	4.52
AR1242ICC500	AR1242ICC500	04/22/2025	12:56	PP071398.D	10.24	4.51
AR1242ICC250	AR1242ICC250	04/22/2025	13:12	PP071399.D	10.23	4.51
AR1242ICC050	AR1242ICC050	04/22/2025	13:28	PP071400.D	10.24	4.52
AR1248ICC1000	AR1248ICC1000	04/22/2025	13:45	PP071401.D	10.24	4.52
AR1248ICC750	AR1248ICC750	04/22/2025	14:01	PP071402.D	10.24	4.52
AR1248ICC500	AR1248ICC500	04/22/2025	14:17	PP071403.D	10.24	4.51
AR1248ICC250	AR1248ICC250	04/22/2025	14:33	PP071404.D	10.24	4.51
AR1248ICC050	AR1248ICC050	04/22/2025	14:50	PP071405.D	10.24	4.51
AR1254ICC1000	AR1254ICC1000	04/22/2025	15:06	PP071406.D	10.24	4.52
AR1254ICC750	AR1254ICC750	04/22/2025	15:22	PP071407.D	10.24	4.52
AR1254ICC500	AR1254ICC500	04/22/2025	15:38	PP071408.D	10.24	4.52
AR1254ICC250	AR1254ICC250	04/22/2025	15:55	PP071409.D	10.23	4.51
AR1254ICC050	AR1254ICC050	04/22/2025	16:11	PP071410.D	10.24	4.51
AR1262ICC500	AR1262ICC500	04/22/2025	16:27	PP071411.D	10.24	4.52
AR1268ICC1000	AR1268ICC1000	04/22/2025	16:44	PP071412.D	10.24	4.51
AR1268ICC750	AR1268ICC750	04/22/2025	17:00	PP071413.D	10.23	4.52
AR1268ICC500	AR1268ICC500	04/22/2025	17:16	PP071414.D	10.24	4.52
AR1268ICC250	AR1268ICC250	04/22/2025	17:33	PP071415.D	10.24	4.52
AR1268ICC050	AR1268ICC050	04/22/2025	17:49	PP071416.D	10.24	4.51
AR1660CCC500	AR1660CCC500	04/28/2025	18:03	PP071557.D	10.23	4.51
IBLK	IBLK	04/28/2025	19:25	PP071561.D	10.23	4.51
PB167765BL	PB167765BL	04/28/2025	20:15	PP071564.D	10.23	4.51
PB167765BS	PB167765BS	04/28/2025	20:31	PP071565.D	10.23	4.51
COMP-2	Q1889-02	04/28/2025	22:09	PP071571.D	10.23	4.51
AR1660CCC500	AR1660CCC500	04/28/2025	23:31	PP071572.D	10.22	4.51
IBLK	IBLK	04/29/2025	00:52	PP071576.D	10.23	4.51
COMP-3	Q1889-03	04/29/2025	01:09	PP071577.D	10.23	4.51
AR1660CCC500	AR1660CCC500	04/29/2025	04:58	PP071587.D	10.22	4.51
IBLK	IBLK	04/29/2025	06:20	PP071591.D	10.22	4.51

Analytical Sequence

Client: Kleinfelder

SDG No.: Q1889

Project: Mitchell School

Instrument ID: ECD_O

GC Column: ZB-MR2

ID: 0.32 (mm)

Inst. Calib. Date(s): 04/10/2025 04/10/2025

THE ANALYTICAL SEQUENCE OF PERFORMANCE EVALUATION MIXTURES, BLANKS, SAMPLES,
AND STANDARDS IS GIVEN BELOW:

EPA SAMPLE NO.	LAB SAMPLE ID	DATE ANALYZED	TIME ANALYZED	DATAFILE	DCB RT #	TCX RT #
IBLK	IBLK	04/10/2025	09:17	PO110348.D	8.68	3.68
AR1660ICC1000	AR1660ICC1000	04/10/2025	09:36	PO110349.D	8.69	3.69
AR1660ICC750	AR1660ICC750	04/10/2025	09:54	PO110350.D	8.69	3.69
AR1660ICC500	AR1660ICC500	04/10/2025	10:13	PO110351.D	8.68	3.69
AR1660ICC250	AR1660ICC250	04/10/2025	10:31	PO110352.D	8.68	3.69
AR1660ICC050	AR1660ICC050	04/10/2025	10:49	PO110353.D	8.68	3.69
AR1221ICC500	AR1221ICC500	04/10/2025	11:08	PO110354.D	8.68	3.69
AR1232ICC500	AR1232ICC500	04/10/2025	11:26	PO110355.D	8.68	3.69
AR1242ICC1000	AR1242ICC1000	04/10/2025	11:44	PO110356.D	8.68	3.69
AR1242ICC750	AR1242ICC750	04/10/2025	12:03	PO110357.D	8.68	3.69
AR1242ICC500	AR1242ICC500	04/10/2025	12:21	PO110358.D	8.68	3.68
AR1242ICC250	AR1242ICC250	04/10/2025	12:39	PO110359.D	8.68	3.69
AR1242ICC050	AR1242ICC050	04/10/2025	12:58	PO110360.D	8.68	3.68
AR1248ICC1000	AR1248ICC1000	04/10/2025	13:16	PO110361.D	8.68	3.69
AR1248ICC750	AR1248ICC750	04/10/2025	13:35	PO110362.D	8.69	3.69
AR1248ICC500	AR1248ICC500	04/10/2025	13:53	PO110363.D	8.69	3.69
AR1248ICC250	AR1248ICC250	04/10/2025	14:11	PO110364.D	8.68	3.69
AR1248ICC050	AR1248ICC050	04/10/2025	14:30	PO110365.D	8.68	3.68
AR1254ICC1000	AR1254ICC1000	04/10/2025	14:48	PO110366.D	8.68	3.69
AR1254ICC750	AR1254ICC750	04/10/2025	15:06	PO110367.D	8.68	3.69
AR1254ICC500	AR1254ICC500	04/10/2025	15:25	PO110368.D	8.68	3.69
AR1254ICC250	AR1254ICC250	04/10/2025	15:43	PO110369.D	8.68	3.68
AR1254ICC050	AR1254ICC050	04/10/2025	16:02	PO110370.D	8.68	3.69
AR1262ICC500	AR1262ICC500	04/10/2025	16:20	PO110371.D	8.68	3.69
AR1268ICC1000	AR1268ICC1000	04/10/2025	16:38	PO110372.D	8.68	3.68
AR1268ICC750	AR1268ICC750	04/10/2025	16:57	PO110373.D	8.68	3.68
AR1268ICC500	AR1268ICC500	04/10/2025	17:15	PO110374.D	8.68	3.69
AR1268ICC250	AR1268ICC250	04/10/2025	17:33	PO110375.D	8.68	3.69
AR1268ICC050	AR1268ICC050	04/10/2025	17:52	PO110376.D	8.68	3.68
AR1660CCC500	AR1660CCC500	04/29/2025	09:24	PO110846.D	8.68	3.68
IBLK	IBLK	04/29/2025	10:40	PO110850.D	8.68	3.68
COMP-1	Q1889-01	04/29/2025	11:16	PO110852.D	8.68	3.68
COMP-1MS	Q1889-01MS	04/29/2025	11:35	PO110853.D	8.68	3.68
COMP-1MSD	Q1889-01MSD	04/29/2025	11:53	PO110854.D	8.68	3.68
AR1660CCC500	AR1660CCC500	04/29/2025	15:58	PO110861.D	8.68	3.68
IBLK	IBLK	04/29/2025	17:29	PO110865.D	8.68	3.68
IBLK	IBLK	04/22/2025	10:13	PP071388.D	8.85	3.81
AR1660ICC1000	AR1660ICC1000	04/22/2025	10:29	PP071389.D	8.85	3.81
AR1660ICC750	AR1660ICC750	04/22/2025	10:45	PP071390.D	8.85	3.81
AR1660ICC500	AR1660ICC500	04/22/2025	11:02	PP071391.D	8.85	3.81
AR1660ICC250	AR1660ICC250	04/22/2025	11:18	PP071392.D	8.85	3.81
AR1660ICC050	AR1660ICC050	04/22/2025	11:34	PP071393.D	8.85	3.81

Analytical Sequence

AR1221ICC500	AR1221ICC500	04/22/2025	11:51	PP071394.D	8.85	3.81
AR1232ICC500	AR1232ICC500	04/22/2025	12:07	PP071395.D	8.85	3.81
AR1242ICC1000	AR1242ICC1000	04/22/2025	12:23	PP071396.D	8.85	3.81
AR1242ICC750	AR1242ICC750	04/22/2025	12:39	PP071397.D	8.85	3.81
AR1242ICC500	AR1242ICC500	04/22/2025	12:56	PP071398.D	8.85	3.81
AR1242ICC250	AR1242ICC250	04/22/2025	13:12	PP071399.D	8.85	3.81
AR1242ICC050	AR1242ICC050	04/22/2025	13:28	PP071400.D	8.85	3.81
AR1248ICC1000	AR1248ICC1000	04/22/2025	13:45	PP071401.D	8.85	3.81
AR1248ICC750	AR1248ICC750	04/22/2025	14:01	PP071402.D	8.85	3.81
AR1248ICC500	AR1248ICC500	04/22/2025	14:17	PP071403.D	8.85	3.81
AR1248ICC250	AR1248ICC250	04/22/2025	14:33	PP071404.D	8.85	3.81
AR1248ICC050	AR1248ICC050	04/22/2025	14:50	PP071405.D	8.85	3.81
AR1254ICC1000	AR1254ICC1000	04/22/2025	15:06	PP071406.D	8.85	3.81
AR1254ICC750	AR1254ICC750	04/22/2025	15:22	PP071407.D	8.85	3.81
AR1254ICC500	AR1254ICC500	04/22/2025	15:38	PP071408.D	8.85	3.81
AR1254ICC250	AR1254ICC250	04/22/2025	15:55	PP071409.D	8.85	3.81
AR1254ICC050	AR1254ICC050	04/22/2025	16:11	PP071410.D	8.85	3.81
AR1262ICC500	AR1262ICC500	04/22/2025	16:27	PP071411.D	8.85	3.81
AR1268ICC1000	AR1268ICC1000	04/22/2025	16:44	PP071412.D	8.85	3.81
AR1268ICC750	AR1268ICC750	04/22/2025	17:00	PP071413.D	8.85	3.81
AR1268ICC500	AR1268ICC500	04/22/2025	17:16	PP071414.D	8.85	3.81
AR1268ICC250	AR1268ICC250	04/22/2025	17:33	PP071415.D	8.85	3.81
AR1268ICC050	AR1268ICC050	04/22/2025	17:49	PP071416.D	8.85	3.81
AR1660CCC500	AR1660CCC500	04/28/2025	18:03	PP071557.D	8.84	3.81
IBLK	IBLK	04/28/2025	19:25	PP071561.D	8.84	3.81
PB167765BL	PB167765BL	04/28/2025	20:15	PP071564.D	8.84	3.81
PB167765BS	PB167765BS	04/28/2025	20:31	PP071565.D	8.84	3.81
COMP-2	Q1889-02	04/28/2025	22:09	PP071571.D	8.84	3.81
AR1660CCC500	AR1660CCC500	04/28/2025	23:31	PP071572.D	8.84	3.81
IBLK	IBLK	04/29/2025	00:52	PP071576.D	8.84	3.81
COMP-3	Q1889-03	04/29/2025	01:09	PP071577.D	8.84	3.81
AR1660CCC500	AR1660CCC500	04/29/2025	04:58	PP071587.D	8.84	3.81
IBLK	IBLK	04/29/2025	06:20	PP071591.D	8.84	3.81



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

COMP-1MS

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Lab Sample ID: Q1889-01MS Date(s) Analyzed: 04/29/2025 04/29/2025

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO110853.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.775	4.725	4.825	191	189	9.42
COLUMN 1	2	4.794	4.744	4.844	191		
	3	4.851	4.801	4.901	188		
	4	4.971	4.921	5.021	190		
	5	5.228	5.178	5.278	183		
	1	4.761	4.711	4.811	178		
COLUMN 2	2	4.78	4.73	4.83	179	172	
	3	4.955	4.905	5.005	172		
	4	4.997	4.947	5.047	166		
	5	5.21	5.16	5.26	164		
Aroclor-1260	1	6.267	6.217	6.317	188	169	5.47
COLUMN 1	2	6.457	6.407	6.507	181		
	3	6.824	6.774	6.874	155		
	4	7.084	7.034	7.134	166		
	5	7.326	7.276	7.376	155		
	1	6.241	6.191	6.291	172	160	
COLUMN 2	2	6.429	6.379	6.479	167		
	3	6.581	6.531	6.631	167		
	4	7.053	7.003	7.103	152		
	5	7.293	7.243	7.343	144		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

COMP-1MSD

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Lab Sample ID: Q1889-01MSD Date(s) Analyzed: 04/29/2025 04/29/2025

Instrument ID (1): ECD_O Instrument ID (2): ECD_O

GC Column: (1): ZB-MR1 ID: 0.32 (mm) GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PO110854.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	4.775	4.725	4.825	183	180	9.3
COLUMN 1	2	4.795	4.745	4.845	181		
	3	4.851	4.801	4.901	180		
	4	4.971	4.921	5.021	180		
	5	5.228	5.178	5.278	174		
	1	4.761	4.711	4.811	170		
COLUMN 2	2	4.78	4.73	4.83	170	164	
	3	4.956	4.906	5.006	163		
	4	4.998	4.948	5.048	157		
	5	5.21	5.16	5.26	157		
Aroclor-1260	1	6.268	6.218	6.318	177	158	4.53
COLUMN 1	2	6.457	6.407	6.507	169		
	3	6.825	6.775	6.875	145		
	4	7.084	7.034	7.134	155		
	5	7.326	7.276	7.376	145		
	1	6.241	6.191	6.291	162		
COLUMN 2	2	6.429	6.379	6.479	157	151	
	3	6.581	6.531	6.631	157		
	4	7.052	7.002	7.102	143		
	5	7.293	7.243	7.343	134		

IDENTIFICATION SUMMARY
FOR MULTICOMPONENT ANALYTES

SAMPLE NO.

PB167765BS

Contract: POWE02

Lab Code: CHEM

Case No.: Q1889

SAS No.: Q1889

SDG NO.: Q1889

Lab Sample ID: PB167765BS

Date(s) Analyzed: 04/28/2025

04/28/2025

Instrument ID (1): ECD_P

Instrument ID (2): ECD_P

GC Column: (1): ZB-MR1 ID: 0.32 (mm)

GC Column: (2): ZB-MR2 ID: 0.32 (mm)

Data file PP071565.D

ANALYTE	COL	RT	RT WINDOW		CONCENTRATION	MEAN CONCENTRATION	%RPD
			FROM	TO			
Aroclor-1016	1	5.663	5.613	5.713	140	139	2.84
COLUMN 1	2	5.685	5.635	5.735	140		
	3	5.747	5.697	5.797	139		
	4	5.844	5.794	5.894	140		
	5	6.137	6.087	6.187	134		
	1	4.894	4.844	4.944	143		
COLUMN 2	2	4.913	4.863	4.963	143	143	
	3	5.09	5.04	5.14	146		
	4	5.131	5.081	5.181	146		
	5	5.345	5.295	5.395	140		
Aroclor-1260	1	7.255	7.205	7.305	142	134	9.25
COLUMN 1	2	7.509	7.459	7.559	144		
	3	7.868	7.818	7.918	126		
	4	8.092	8.042	8.142	132		
	5	8.412	8.362	8.462	129		
	1	6.38	6.33	6.43	153	147	
COLUMN 2	2	6.568	6.518	6.618	153		
	3	6.721	6.671	6.771	157		
	4	7.192	7.142	7.242	138		
	5	7.433	7.383	7.483	136		



QC SAMPLE DATA

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Mitchell School	Date Received:	
Client Sample ID:	PB167765BL	SDG No.:	Q1889
Lab Sample ID:	PB167765BL	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100 Decanted:
Sample Wt/Vol:	30.02 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB Group1
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071564.D	1	04/28/25 09:05	04/28/25 20:15	PB167765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	3.90	U	3.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
11096-82-5	Aroclor-1260	3.20	U	3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	20.5		32 - 144	103%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.9		32 - 175	100%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
 Data File : PP071564.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Apr 2025 20:15
 Operator : YP\AJ
 Sample : PB167765BL
 Misc :
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB167765BL

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 01:12:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.810	36411530	28870233	18.365	20.495
2) SA Decachlor...	10.228	8.841	27428987	17475328	18.943	19.925

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
Data File : PP071564.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2025 20:15
Operator : YP\AJ
Sample : PB167765BL
Misc :
ALS Vial : 28 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB167765BL

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 29 01:12:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: -0.004 min
Response: 36411530
Conc: 18.37 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB167765BL




#1 Tetrachloro-m-xylene

R.T.: 3.810 min
Delta R.T.: -0.003 min
Response: 28870233
Conc: 20.50 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.228 min
Delta R.T.: -0.009 min
Response: 27428987
Conc: 18.94 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.841 min
Delta R.T.: -0.008 min
Response: 17475328
Conc: 19.93 ng/ml

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/10/25
Project:	Mitchell School	Date Received:	04/10/25
Client Sample ID:	PIBLK-PO110348.D	SDG No.:	Q1889
Lab Sample ID:	I.BLK-PO110348.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110348.D	1		04/10/25	PO041025

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.1		60 - 140	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.2		60 - 140	96%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0041025\
 Data File : PO110348.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 10 Apr 2025 09:17
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 10 19:25:01 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Thu Apr 10 18:44:28 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.687	3.684	158.3E6	90605960	18.096	18.158
2) SA Decachlor...	8.732	8.683	151.3E6	38875455	19.165	20.195

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO041025\
Data File : PO110348.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 10 Apr 2025 09:17
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 10 19:25:01 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Thu Apr 10 18:44:28 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.687 min
Delta R.T.: 0.000 min
Response: 158319301
Conc: 18.10 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.684 min
Delta R.T.: -0.001 min
Response: 90605960
Conc: 18.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.732 min
Delta R.T.: 0.000 min
Response: 151304209
Conc: 19.16 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.683 min
Delta R.T.: 0.000 min
Response: 38875455
Conc: 20.20 ng/ml

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/29/25
Project:	Mitchell School	Date Received:	04/29/25
Client Sample ID:	PIBLK-PO110850.D	SDG No.:	Q1889
Lab Sample ID:	I.BLK-PO110850.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110850.D	1		04/29/25	PO042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.6		60 - 140	83%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.7		60 - 140	84%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042925\
 Data File : PO110850.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2025 10:40
 Operator : YP/AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 02:23:06 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 11 02:12:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.685	3.682	157.3E6	82558016	17.974	16.545
2) SA Decachlor...	8.726	8.678	133.5E6	32185923	16.914	16.720

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042925\
Data File : PO110850.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2025 10:40
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 02:23:06 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 02:12:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.685 min
Delta R.T.: -0.003 min
Response: 157253270
Conc: 17.97 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.682 min
Delta R.T.: -0.004 min
Response: 82558016
Conc: 16.55 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.726 min
Delta R.T.: -0.005 min
Response: 133536169
Conc: 16.91 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.678 min
Delta R.T.: -0.007 min
Response: 32185923
Conc: 16.72 ng/ml

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/29/25
Project:	Mitchell School	Date Received:	04/29/25
Client Sample ID:	PIBLK-PO110865.D	SDG No.:	Q1889
Lab Sample ID:	I.BLK-PO110865.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110865.D	1		04/29/25	PO042925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.9		60 - 140	84%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.9		60 - 140	85%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042925\
Data File : PO110865.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2025 17:29
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 02:27:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 02:12:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.686	3.683	161.0E6	84295389	18.397	16.894
2) SA Decachlor...	8.729	8.680	138.3E6	32608782	17.513	16.940

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042925\
Data File : PO110865.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2025 17:29
Operator : YP/AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 02:27:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 02:12:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR2 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.686 min
Delta R.T.: -0.002 min
Response: 160952706
Conc: 18.40 ng/ml

Instrument :
ECD_O
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.683 min
Delta R.T.: -0.003 min
Response: 84295389
Conc: 16.89 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.729 min
Delta R.T.: -0.003 min
Response: 138260960
Conc: 17.51 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.680 min
Delta R.T.: -0.004 min
Response: 32608782
Conc: 16.94 ng/ml

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/22/25
Project:	Mitchell School	Date Received:	04/22/25
Client Sample ID:	PIBLK-PP071388.D	SDG No.:	Q1889
Lab Sample ID:	I.BLK-PP071388.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071388.D	1		04/22/25	PP042225

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	17.1		60 - 140	85%	SPK: 20
2051-24-3	Decachlorobiphenyl	17.6		60 - 140	88%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
 Data File : PP071388.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 22 Apr 2025 10:13
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 22 19:33:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Apr 22 17:08:14 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.512	3.811	33865777	24238749	17.081	17.207
2) SA Decachlor...	10.233	8.847	25533987	15462668	17.634	17.630

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042225\
Data File : PP071388.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 22 Apr 2025 10:13
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 22 19:33:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Apr 22 17:08:14 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.512 min
Delta R.T.: -0.005 min
Response: 33865777
Conc: 17.08 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.811 min
Delta R.T.: -0.001 min
Response: 24238749
Conc: 17.21 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.233 min
Delta R.T.: -0.005 min
Response: 25533987
Conc: 17.63 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.847 min
Delta R.T.: -0.002 min
Response: 15462668
Conc: 17.63 ng/ml

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/28/25
Project:	Mitchell School	Date Received:	04/28/25
Client Sample ID:	PIBLK-PP071561.D	SDG No.:	Q1889
Lab Sample ID:	I.BLK-PP071561.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071561.D	1		04/28/25	PP042825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.5		60 - 140	78%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.3		60 - 140	81%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
 Data File : PP071561.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Apr 2025 19:25
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 01:11:14 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.513	3.810	30746085	23071154	15.508	16.378
2) SA Decachlor...	10.228	8.841	23576553	15262740	16.282	17.402

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
Data File : PP071561.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2025 19:25
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 29 01:11:14 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.513 min
Delta R.T.: -0.004 min
Response: 30746085
Conc: 15.51 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.810 min
Delta R.T.: -0.002 min
Response: 23071154
Conc: 16.38 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.228 min
Delta R.T.: -0.009 min
Response: 23576553
Conc: 16.28 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.841 min
Delta R.T.: -0.008 min
Response: 15262740
Conc: 17.40 ng/ml

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/29/25
Project:	Mitchell School	Date Received:	04/29/25
Client Sample ID:	PIBLK-PP071576.D	SDG No.:	Q1889
Lab Sample ID:	I.BLK-PP071576.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071576.D	1		04/29/25	PP042825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	15.8		60 - 140	79%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.4		60 - 140	82%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
 Data File : PP071576.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2025 00:52
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 06:28:38 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.512	3.809	31259685	23779769	15.767	16.881
2) SA Decachlor...	10.225	8.840	23774920	15323361	16.419	17.471

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
Data File : PP071576.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2025 00:52
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 29 06:28:38 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.512 min
Delta R.T.: -0.005 min
Response: 31259685
Conc: 15.77 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.809 min
Delta R.T.: -0.003 min
Response: 23779769
Conc: 16.88 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.225 min
Delta R.T.: -0.012 min
Response: 23774920
Conc: 16.42 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.840 min
Delta R.T.: -0.009 min
Response: 15323361
Conc: 17.47 ng/ml

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/29/25
Project:	Mitchell School	Date Received:	04/29/25
Client Sample ID:	PIBLK-PP071591.D	SDG No.:	Q1889
Lab Sample ID:	I.BLK-PP071591.D	Matrix:	WATER
Analytical Method:	SW8082A	% Solid:	0
Sample Wt/Vol:	1000	Units:	mL
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	5030	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071591.D	1		04/29/25	PP042825

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
TARGETS						
12674-11-2	Aroclor-1016	0.097	U	0.097	0.50	ug/L
11097-69-1	Aroclor-1254	0.094	U	0.094	0.50	ug/L
11096-82-5	Aroclor-1260	0.081	U	0.081	0.50	ug/L
SURROGATES						
877-09-8	Tetrachloro-m-xylene	16.1		60 - 140	80%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.7		60 - 140	83%	SPK: 20

Comments:

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 P = Indicates >25% difference for detected concentrations between the two GC columns
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.
 () = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
 Data File : PP071591.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2025 06:20
 Operator : YP\AJ
 Sample : I.BLK
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 I.BLK

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 07:13:51 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.512	3.808	31812118	23665261	16.045	16.800
2) SA Decachlor...	10.224	8.839	24172879	15683140	16.694	17.882

Target Compounds



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
Data File : PP071591.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2025 06:20
Operator : YP\AJ
Sample : I.BLK
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
I.BLK

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 29 07:13:51 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.512 min
Delta R.T.: -0.005 min
Response: 31812118
Conc: 16.05 ng/ml

Instrument :
ECD_P
ClientSampleId :
I.BLK




#1 Tetrachloro-m-xylene

R.T.: 3.808 min
Delta R.T.: -0.004 min
Response: 23665261
Conc: 16.80 ng/ml



#2 Decachlorobiphenyl

R.T.: 10.224 min
Delta R.T.: -0.013 min
Response: 24172879
Conc: 16.69 ng/ml



#2 Decachlorobiphenyl

R.T.: 8.839 min
Delta R.T.: -0.010 min
Response: 15683140
Conc: 17.88 ng/ml

Report of Analysis

Client:	Kleinfelder	Date Collected:	
Project:	Mitchell School	Date Received:	
Client Sample ID:	PB167765BS	SDG No.:	Q1889
Lab Sample ID:	PB167765BS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	100 Decanted:
Sample Wt/Vol:	30.03 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB Group1
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PP071565.D	1	04/28/25 09:05	04/28/25 20:31	PB167765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	143		3.90	17.0	ug/kg
11097-69-1	Aroclor-1254	3.20	U	3.20	17.0	ug/kg
11096-82-5	Aroclor-1260	147		3.20	17.0	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.1		32 - 144	91%	SPK: 20
2051-24-3	Decachlorobiphenyl	19.8		32 - 175	99%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
 Data File : PP071565.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 28 Apr 2025 20:31
 Operator : YP\AJ
 Sample : PB167765BS
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_P
 ClientSampleId :
 PB167765BS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 29 01:12:17 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Wed Apr 23 05:02:06 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	4.510	3.810	35953786	24928335	18.134	17.697
2) SA Decachlor...	10.225	8.842	27087543	17337538	18.707	19.768
Target Compounds						
3) L1 AR-1016-1	5.663	4.894	28221218	22925970	420.899	428.254
4) L1 AR-1016-2	5.685	4.913	43100738	32733572	419.969	428.654
5) L1 AR-1016-3	5.747	5.090	25761815	18490650	417.892	438.059
6) L1 AR-1016-4	5.844	5.131	21595564	15092065	420.981	437.630
7) L1 AR-1016-5	6.137	5.345	19480746	18667996	403.843	420.003
31) L7 AR-1260-1	7.255	6.380	41084091	33132360	427.850	458.267
32) L7 AR-1260-2	7.509	6.568	61727007	40153544	431.424	459.100
33) L7 AR-1260-3	7.868	6.721	42264311	36939841	377.010	471.246
34) L7 AR-1260-4	8.092	7.192	44748259	26129582	396.477	414.028
35) L7 AR-1260-5	8.412	7.433	87445836	61724959	385.950	409.378



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_P\Data\PP042825\
Data File : PP071565.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 28 Apr 2025 20:31
Operator : YP\AJ
Sample : PB167765BS
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_P
ClientSampleId :
PB167765BS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 29 01:12:17 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_P\methods\PP042225.M
Quant Title : GC EXTRACTABLES
QLast Update : Wed Apr 23 05:02:06 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.510 min
Delta R.T.: -0.007 min
Response: 35953786
Conc: 18.13 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB167765BS




#1 Tetrachloro-m-xylene

R.T.: 3.810 min
Delta R.T.: -0.003 min
Response: 24928335
Conc: 17.70 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.225 min
Delta R.T.: -0.013 min
Response: 27087543
Conc: 18.71 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.842 min
Delta R.T.: -0.008 min
Response: 17337538
Conc: 19.77 ng/ml



#3 AR-1016-1


R.T.: 5.663 min
Delta R.T.: -0.008 min
Response: 28221218
Conc: 420.90 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB167765BS




#3 AR-1016-1

R.T.: 4.894 min
Delta R.T.: -0.004 min
Response: 22925970
Conc: 428.25 ng/ml




#4 AR-1016-2

R.T.: 5.685 min
Delta R.T.: -0.008 min
Response: 43100738
Conc: 419.97 ng/ml



#4 AR-1016-2


R.T.: 4.913 min
Delta R.T.: -0.003 min
Response: 32733572
Conc: 428.65 ng/ml



#5 AR-1016-3


R.T.: 5.747 min
Delta R.T.: -0.007 min
Response: 25761815
Conc: 417.89 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB167765BS




#5 AR-1016-3

R.T.: 5.090 min
Delta R.T.: -0.003 min
Response: 18490650
Conc: 438.06 ng/ml




#6 AR-1016-4

R.T.: 5.844 min
Delta R.T.: -0.008 min
Response: 21595564
Conc: 420.98 ng/ml



#6 AR-1016-4


R.T.: 5.131 min
Delta R.T.: -0.005 min
Response: 15092065
Conc: 437.63 ng/ml



#7 AR-1016-5


R.T.: 6.137 min
Delta R.T.: -0.009 min
Response: 19480746
Conc: 403.84 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB167765BS




#7 AR-1016-5

R.T.: 5.345 min
Delta R.T.: -0.004 min
Response: 18667996
Conc: 420.00 ng/ml




#31 AR-1260-1

R.T.: 7.255 min
Delta R.T.: -0.009 min
Response: 41084091
Conc: 427.85 ng/ml



#31 AR-1260-1


R.T.: 6.380 min
Delta R.T.: -0.005 min
Response: 33132360
Conc: 458.27 ng/ml



#32 AR-1260-2


R.T.: 7.509 min
Delta R.T.: -0.010 min
Response: 61727007
Conc: 431.42 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB167765BS




#32 AR-1260-2

R.T.: 6.568 min
Delta R.T.: -0.005 min
Response: 40153544
Conc: 459.10 ng/ml




#33 AR-1260-3

R.T.: 7.868 min
Delta R.T.: -0.009 min
Response: 42264311
Conc: 377.01 ng/ml



#33 AR-1260-3


R.T.: 6.721 min
Delta R.T.: -0.006 min
Response: 36939841
Conc: 471.25 ng/ml



#34 AR-1260-4


R.T.: 8.092 min
Delta R.T.: -0.010 min
Response: 44748259
Conc: 396.48 ng/ml

Instrument :
ECD_P
ClientSampleId :
PB167765BS




#34 AR-1260-4

R.T.: 7.192 min
Delta R.T.: -0.006 min
Response: 26129582
Conc: 414.03 ng/ml



#35 AR-1260-5

R.T.: 8.412 min
Delta R.T.: -0.010 min
Response: 87445836
Conc: 385.95 ng/ml



#35 AR-1260-5

R.T.: 7.433 min
Delta R.T.: -0.005 min
Response: 61724959
Conc: 409.38 ng/ml

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/24/25
Project:	Mitchell School	Date Received:	04/25/25
Client Sample ID:	COMP-1MS	SDG No.:	Q1889
Lab Sample ID:	Q1889-01MS	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	82.5 Decanted:
Sample Wt/Vol:	30.06 Units: g	Final Vol:	10000 uL
Soil Aliquot Vol:	uL	Test:	PCB Group1
Extraction Type:		Injection Volume :	
GPC Factor :	1.0 PH :		
Prep Method :	SW3541B		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110853.D	1	04/28/25 09:05	04/29/25 11:35	PB167765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	189		4.80	20.6	ug/kg
11097-69-1	Aroclor-1254	3.90	U	3.90	20.6	ug/kg
11096-82-5	Aroclor-1260	169		3.90	20.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	19.4		32 - 144	97%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.4		32 - 175	82%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042925\
 Data File : PO110853.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2025 11:35
 Operator : YP/AJ
 Sample : Q1889-01MS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 COMP-1MS

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 02:23:55 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 11 02:12:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.685	3.682	170.0E6	88813564	19.429	17.799
2) SA Decachlor...	8.727	8.678	129.5E6	31008232	16.405	16.108
Target Compounds						
3) L1 AR-1016-1	4.775	4.761	155.4E6	77342475	473.501	440.674
4) L1 AR-1016-2	4.794	4.780	215.7E6	111.6E6	474.077	443.609
5) L1 AR-1016-3	4.851	4.955	150.4E6	57820772	466.452	426.058
6) L1 AR-1016-4	4.971	4.997	117.6E6	47128288	472.169	412.076
7) L1 AR-1016-5	5.228	5.210	122.1E6	60566905	453.826	406.288
31) L7 AR-1260-1	6.267	6.241	220.3E6	104.7E6	466.328	425.791
32) L7 AR-1260-2	6.457	6.429	263.1E6	120.9E6	448.599	414.541
33) L7 AR-1260-3	6.824	6.581	189.5E6	112.2E6	384.425	414.359
34) L7 AR-1260-4	7.084	7.053	174.2E6	75618343	410.511	376.886
35) L7 AR-1260-5	7.326	7.293	402.4E6	163.7E6	385.607	356.653



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042925\
Data File : PO110853.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2025 11:35
Operator : YP/AJ
Sample : Q1889-01MS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
COMP-1MS

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 02:23:55 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 02:12:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.685 min
Delta R.T.: -0.003 min
Response: 169983809
Conc: 19.43 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1MS




#1 Tetrachloro-m-xylene

R.T.: 3.682 min
Delta R.T.: -0.004 min
Response: 88813564
Conc: 17.80 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.727 min
Delta R.T.: -0.005 min
Response: 129518080
Conc: 16.41 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.678 min
Delta R.T.: -0.006 min
Response: 31008232
Conc: 16.11 ng/ml



#3 AR-1016-1


R.T.: 4.775 min
Delta R.T.: -0.004 min
Response: 155376436
Conc: 473.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1MS




#3 AR-1016-1

R.T.: 4.761 min
Delta R.T.: -0.005 min
Response: 77342475
Conc: 440.67 ng/ml




#4 AR-1016-2

R.T.: 4.794 min
Delta R.T.: -0.004 min
Response: 215680668
Conc: 474.08 ng/ml



#4 AR-1016-2


R.T.: 4.780 min
Delta R.T.: -0.004 min
Response: 111552073
Conc: 443.61 ng/ml



#5 AR-1016-3


R.T.: 4.851 min
Delta R.T.: -0.004 min
Response: 150443543
Conc: 466.45 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1MS




#5 AR-1016-3

R.T.: 4.955 min
Delta R.T.: -0.005 min
Response: 57820772
Conc: 426.06 ng/ml




#6 AR-1016-4

R.T.: 4.971 min
Delta R.T.: -0.004 min
Response: 117647814
Conc: 472.17 ng/ml



#6 AR-1016-4


R.T.: 4.997 min
Delta R.T.: -0.005 min
Response: 47128288
Conc: 412.08 ng/ml



#7 AR-1016-5


R.T.: 5.228 min
Delta R.T.: -0.004 min
Response: 122084432
Conc: 453.83 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1MS




#7 AR-1016-5

R.T.: 5.210 min
Delta R.T.: -0.005 min
Response: 60566905
Conc: 406.29 ng/ml




#31 AR-1260-1

R.T.: 6.267 min
Delta R.T.: -0.005 min
Response: 220314689
Conc: 466.33 ng/ml



#31 AR-1260-1


R.T.: 6.241 min
Delta R.T.: -0.005 min
Response: 104729021
Conc: 425.79 ng/ml



#32 AR-1260-2


R.T.: 6.457 min
Delta R.T.: -0.003 min
Response: 263069258
Conc: 448.60 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1MS




#32 AR-1260-2

R.T.: 6.429 min
Delta R.T.: -0.005 min
Response: 120942650
Conc: 414.54 ng/ml




#33 AR-1260-3

R.T.: 6.824 min
Delta R.T.: -0.005 min
Response: 189472439
Conc: 384.42 ng/ml



#33 AR-1260-3


R.T.: 6.581 min
Delta R.T.: -0.005 min
Response: 112243297
Conc: 414.36 ng/ml



#34 AR-1260-4


R.T.: 7.084 min
Delta R.T.: -0.004 min
Response: 174230235
Conc: 410.51 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1MS




#34 AR-1260-4

R.T.: 7.053 min
Delta R.T.: -0.005 min
Response: 75618343
Conc: 376.89 ng/ml



#35 AR-1260-5

R.T.: 7.326 min
Delta R.T.: -0.005 min
Response: 402370220
Conc: 385.61 ng/ml



#35 AR-1260-5

R.T.: 7.293 min
Delta R.T.: -0.004 min
Response: 163652945
Conc: 356.65 ng/ml

Report of Analysis

Client:	Kleinfelder	Date Collected:	04/24/25
Project:	Mitchell School	Date Received:	04/25/25
Client Sample ID:	COMP-1MSD	SDG No.:	Q1889
Lab Sample ID:	Q1889-01MSD	Matrix:	SOIL
Analytical Method:	SW8082A	% Solid:	82.5
Sample Wt/Vol:	30.03	Units:	g
Soil Aliquot Vol:			uL
Extraction Type:		Test:	PCB Group1
GPC Factor :	1.0	PH :	
Prep Method :	SW3541B	Injection Volume :	

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
PO110854.D	1	04/28/25 09:05	04/29/25 11:53	PB167765

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
12674-11-2	Aroclor-1016	180		4.80	20.6	ug/kg
11097-69-1	Aroclor-1254	3.90	U	3.90	20.6	ug/kg
11096-82-5	Aroclor-1260	158		3.90	20.6	ug/kg
SURROGATES						
877-09-8	Tetrachloro-m-xylene	18.9		32 - 144	95%	SPK: 20
2051-24-3	Decachlorobiphenyl	16.6		32 - 175	83%	SPK: 20

Comments:

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

P = Indicates >25% difference for detected concentrations between the two GC columns

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

S = Indicates estimated value where valid five-point calibration was not performed prior to analyte detection in sample.

() = Laboratory InHouse Limit

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0042925\
 Data File : PO110854.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Apr 2025 11:53
 Operator : YP/AJ
 Sample : Q1889-01MSD
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 COMP-1MSD

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Apr 30 02:24:11 2025
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0041025.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Apr 11 02:12:41 2025
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.685	3.682	165.6E6	85915208	18.924	17.218
2) SA Decachlor...	8.726	8.679	131.2E6	31403514	16.622	16.314
Target Compounds						
3) L1 AR-1016-1	4.775	4.761	148.5E6	74035742	452.499	421.833
4) L1 AR-1016-2	4.795	4.780	204.5E6	105.9E6	449.455	420.972
5) L1 AR-1016-3	4.851	4.956	143.6E6	54856844	445.318	404.218
6) L1 AR-1016-4	4.971	4.998	111.2E6	44479232	446.403	388.914
7) L1 AR-1016-5	5.228	5.210	116.3E6	58083049	432.248	389.626
31) L7 AR-1260-1	6.268	6.241	207.7E6	98966137	439.539	402.361
32) L7 AR-1260-2	6.457	6.429	246.1E6	113.3E6	419.700	388.399
33) L7 AR-1260-3	6.825	6.581	177.3E6	105.7E6	359.627	390.191
34) L7 AR-1260-4	7.084	7.052	163.1E6	70954325	384.365	353.640
35) L7 AR-1260-5	7.326	7.293	374.7E6	152.3E6	359.102	332.012



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO042925\
Data File : PO110854.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Apr 2025 11:53
Operator : YP/AJ
Sample : Q1889-01MSD
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
COMP-1MSD

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Apr 30 02:24:11 2025
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO041025.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Apr 11 02:12:41 2025
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 3.685 min
Delta R.T.: -0.003 min
Response: 165571147
Conc: 18.92 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1MSD




#1 Tetrachloro-m-xylene

R.T.: 3.682 min
Delta R.T.: -0.004 min
Response: 85915208
Conc: 17.22 ng/ml




#2 Decachlorobiphenyl

R.T.: 8.726 min
Delta R.T.: -0.005 min
Response: 131227115
Conc: 16.62 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.679 min
Delta R.T.: -0.005 min
Response: 31403514
Conc: 16.31 ng/ml



#3 AR-1016-1


R.T.: 4.775 min
Delta R.T.: -0.003 min
Response: 148484570
Conc: 452.50 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1MSD




#3 AR-1016-1

R.T.: 4.761 min
Delta R.T.: -0.005 min
Response: 74035742
Conc: 421.83 ng/ml




#4 AR-1016-2

R.T.: 4.795 min
Delta R.T.: -0.003 min
Response: 204478747
Conc: 449.45 ng/ml



#4 AR-1016-2


R.T.: 4.780 min
Delta R.T.: -0.004 min
Response: 105859697
Conc: 420.97 ng/ml



#5 AR-1016-3


R.T.: 4.851 min
Delta R.T.: -0.004 min
Response: 143627249
Conc: 445.32 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1MSD




#5 AR-1016-3

R.T.: 4.956 min
Delta R.T.: -0.004 min
Response: 54856844
Conc: 404.22 ng/ml




#6 AR-1016-4

R.T.: 4.971 min
Delta R.T.: -0.004 min
Response: 111227776
Conc: 446.40 ng/ml



#6 AR-1016-4


R.T.: 4.998 min
Delta R.T.: -0.004 min
Response: 44479232
Conc: 388.91 ng/ml



#7 AR-1016-5


R.T.: 5.228 min
Delta R.T.: -0.004 min
Response: 116279711
Conc: 432.25 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1MSD




#7 AR-1016-5

R.T.: 5.210 min
Delta R.T.: -0.004 min
Response: 58083049
Conc: 389.63 ng/ml




#31 AR-1260-1

R.T.: 6.268 min
Delta R.T.: -0.004 min
Response: 207658079
Conc: 439.54 ng/ml



#31 AR-1260-1


R.T.: 6.241 min
Delta R.T.: -0.005 min
Response: 98966137
Conc: 402.36 ng/ml



#32 AR-1260-2


R.T.: 6.457 min
Delta R.T.: -0.003 min
Response: 246122396
Conc: 419.70 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1MSD




#32 AR-1260-2

R.T.: 6.429 min
Delta R.T.: -0.005 min
Response: 113315674
Conc: 388.40 ng/ml




#33 AR-1260-3

R.T.: 6.825 min
Delta R.T.: -0.004 min
Response: 177250154
Conc: 359.63 ng/ml



#33 AR-1260-3


R.T.: 6.581 min
Delta R.T.: -0.005 min
Response: 105696664
Conc: 390.19 ng/ml



#34 AR-1260-4


R.T.: 7.084 min
Delta R.T.: -0.004 min
Response: 163133370
Conc: 384.37 ng/ml

Instrument :
ECD_O
ClientSampleId :
COMP-1MSD




#34 AR-1260-4

R.T.: 7.052 min
Delta R.T.: -0.005 min
Response: 70954325
Conc: 353.64 ng/ml



#35 AR-1260-5

R.T.: 7.326 min
Delta R.T.: -0.004 min
Response: 374713254
Conc: 359.10 ng/ml



#35 AR-1260-5

R.T.: 7.293 min
Delta R.T.: -0.004 min
Response: 152346332
Conc: 332.01 ng/ml



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

PO041025

Instrument

ECD_o

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

PO042925

Instrument

ECD_o

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:	PP042225	Instrument	ECD_p
-----------	----------	------------	-------

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1660ICC050	PP071393.D	AR-1260-3	yogesh	4/23/2025 7:32:38 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1660ICC050	PP071393.D	AR-1260-4	yogesh	4/23/2025 7:32:38 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1660ICC050	PP071393.D	AR-1260-5	yogesh	4/23/2025 7:32:38 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1254ICC750	PP071407.D	AR-1254-5	yogesh	4/23/2025 7:32:40 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1254ICC050	PP071410.D	Tetrachloro-m-xylene	yogesh	4/23/2025 7:32:41 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC1000	PP071412.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:43 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC750	PP071413.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:44 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC500	PP071414.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:46 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC250	PP071415.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:48 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC050	PP071416.D	AR-1268-1	yogesh	4/23/2025 7:32:50 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC050	PP071416.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:50 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICC050	PP071416.D	AR-1268-2	yogesh	4/23/2025 7:32:50 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software
AR1268ICV500	PP071421.D	AR-1268-1 #2	yogesh	4/23/2025 7:32:52 AM	mohammad	4/24/2025 6:33:42	Peak Integrated by Software



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

PP042225

Instrument

ECD_p

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:

PP042825

Instrument

ECD_p

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
AR1254CCC500	PP071545.D	AR-1254-5	yogesh	4/29/2025 8:19:06 AM	 	 	Peak Integrated by Software
AR1254CCC500	PP071560.D	AR-1254-5	yogesh	4/29/2025 10:16:08 AM	 	 	Peak Integrated by Software
AR1254CCC500	PP071575.D	AR-1254-5	yogesh	4/29/2025 10:15:52 AM	 	 	Peak Integrated by Software
AR1254CCC500	PP071596.D	AR-1254-5	yogesh	4/29/2025 10:15:57 AM	 	 	Peak Integrated by Software

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO041025

Review By	yogesh	Review On	4/10/2025 11:27:35 AM
Supervise By	mohammad	Supervise On	4/14/2025 12:40:29 AM
SubDirectory	PO041025	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO110347.D	10 Apr 2025 08:59	YP/AJ	Ok
2	I.BLK	PO110348.D	10 Apr 2025 09:17	YP/AJ	Ok
3	AR1660ICC1000	PO110349.D	10 Apr 2025 09:36	YP/AJ	Ok
4	AR1660ICC750	PO110350.D	10 Apr 2025 09:54	YP/AJ	Ok
5	AR1660ICC500	PO110351.D	10 Apr 2025 10:13	YP/AJ	Ok
6	AR1660ICC250	PO110352.D	10 Apr 2025 10:31	YP/AJ	Ok
7	AR1660ICC050	PO110353.D	10 Apr 2025 10:49	YP/AJ	Ok
8	AR1221ICC500	PO110354.D	10 Apr 2025 11:08	YP/AJ	Ok
9	AR1232ICC500	PO110355.D	10 Apr 2025 11:26	YP/AJ	Ok
10	AR1242ICC1000	PO110356.D	10 Apr 2025 11:44	YP/AJ	Ok
11	AR1242ICC750	PO110357.D	10 Apr 2025 12:03	YP/AJ	Ok
12	AR1242ICC500	PO110358.D	10 Apr 2025 12:21	YP/AJ	Ok
13	AR1242ICC250	PO110359.D	10 Apr 2025 12:39	YP/AJ	Ok
14	AR1242ICC050	PO110360.D	10 Apr 2025 12:58	YP/AJ	Ok
15	AR1248ICC1000	PO110361.D	10 Apr 2025 13:16	YP/AJ	Ok
16	AR1248ICC750	PO110362.D	10 Apr 2025 13:35	YP/AJ	Ok
17	AR1248ICC500	PO110363.D	10 Apr 2025 13:53	YP/AJ	Ok
18	AR1248ICC250	PO110364.D	10 Apr 2025 14:11	YP/AJ	Ok
19	AR1248ICC050	PO110365.D	10 Apr 2025 14:30	YP/AJ	Ok
20	AR1254ICC1000	PO110366.D	10 Apr 2025 14:48	YP/AJ	Ok
21	AR1254ICC750	PO110367.D	10 Apr 2025 15:06	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO041025

Review By	yogesh	Review On	4/10/2025 11:27:35 AM
Supervise By	mohammad	Supervise On	4/14/2025 12:40:29 AM
SubDirectory	PO041025	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PO110368.D	10 Apr 2025 15:25	YP/AJ	Ok
23	AR1254ICC250	PO110369.D	10 Apr 2025 15:43	YP/AJ	Ok
24	AR1254ICC050	PO110370.D	10 Apr 2025 16:02	YP/AJ	Ok
25	AR1262ICC500	PO110371.D	10 Apr 2025 16:20	YP/AJ	Ok
26	AR1268ICC1000	PO110372.D	10 Apr 2025 16:38	YP/AJ	Ok
27	AR1268ICC750	PO110373.D	10 Apr 2025 16:57	YP/AJ	Ok
28	AR1268ICC500	PO110374.D	10 Apr 2025 17:15	YP/AJ	Ok
29	AR1268ICC250	PO110375.D	10 Apr 2025 17:33	YP/AJ	Ok
30	AR1268ICC050	PO110376.D	10 Apr 2025 17:52	YP/AJ	Ok
31	PO041025ICV500	PO110377.D	10 Apr 2025 18:09	YP/AJ	Ok
32	AR1242ICV500	PO110378.D	10 Apr 2025 18:46	YP/AJ	Ok
33	AR1248ICV500	PO110379.D	10 Apr 2025 19:22	YP/AJ	Ok
34	AR1254ICV500	PO110380.D	10 Apr 2025 19:58	YP/AJ	Ok
35	AR1268ICV500	PO110381.D	10 Apr 2025 20:35	YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042925

Review By	yogesh	Review On	4/29/2025 11:07:42 AM
Supervise By		Supervise On	
SubDirectory	PO042925	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PO110845.D	29 Apr 2025 09:06	YP/AJ	Ok
2	AR1660CCC500	PO110846.D	29 Apr 2025 09:24	YP/AJ	Ok
3	AR1242CCC500	PO110847.D	29 Apr 2025 09:45	YP/AJ	Ok
4	AR1248CCC500	PO110848.D	29 Apr 2025 10:03	YP/AJ	Ok
5	AR1254CCC500	PO110849.D	29 Apr 2025 10:21	YP/AJ	Ok
6	I.BLK	PO110850.D	29 Apr 2025 10:40	YP/AJ	Ok
7	PB167765BL	PO110851.D	29 Apr 2025 10:58	YP/AJ	Not Ok
8	Q1889-01	PO110852.D	29 Apr 2025 11:16	YP/AJ	Ok
9	Q1889-01MS	PO110853.D	29 Apr 2025 11:35	YP/AJ	Ok
10	Q1889-01MSD	PO110854.D	29 Apr 2025 11:53	YP/AJ	Ok
11	Q1860-34DL2	PO110855.D	29 Apr 2025 12:11	YP/AJ	Ok,NR
12	Q1904-01	PO110856.D	29 Apr 2025 13:36	YP/AJ	Ok
13	Q1902-02	PO110857.D	29 Apr 2025 13:54	YP/AJ	Ok,NS
14	Q1903-01	PO110858.D	29 Apr 2025 14:12	YP/AJ	Ok
15	Q1903-02	PO110859.D	29 Apr 2025 14:31	YP/AJ	Ok
16	Q1903-03	PO110860.D	29 Apr 2025 14:49	YP/AJ	Ok
17	AR1660CCC500	PO110861.D	29 Apr 2025 15:58	YP/AJ	Ok
18	AR1242CCC500	PO110862.D	29 Apr 2025 16:35	YP/AJ	Ok
19	AR1248CCC500	PO110863.D	29 Apr 2025 16:53	YP/AJ	Ok
20	AR1254CCC500	PO110864.D	29 Apr 2025 17:11	YP/AJ	Ok
21	I.BLK	PO110865.D	29 Apr 2025 17:29	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042925

Review By	yogesh	Review On	4/29/2025 11:07:42 AM
Supervise By		Supervise On	
SubDirectory	PO042925	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1904-01	PO110866.D	29 Apr 2025 17:48	YP/AJ	Ok
23	Q1904-01MS	PO110867.D	29 Apr 2025 18:05	YP/AJ	Ok
24	Q1904-01MSD	PO110868.D	29 Apr 2025 18:23	YP/AJ	Ok
25	Q1906-01	PO110869.D	29 Apr 2025 18:42	YP/AJ	Ok
26	Q1906-05	PO110870.D	29 Apr 2025 19:00	YP/AJ	Ok
27	Q1906-09	PO110871.D	29 Apr 2025 19:19	YP/AJ	Ok,NR
28	Q1906-13	PO110872.D	29 Apr 2025 19:37	YP/AJ	Ok
29	Q1897-01	PO110873.D	29 Apr 2025 19:55	YP/AJ	Ok,NS
30	Q1901-06	PO110874.D	29 Apr 2025 20:14	YP/AJ	Ok
31	Q1897-02	PO110875.D	29 Apr 2025 20:32	YP/AJ	Ok,NS
32	AR1660CCC500	PO110876.D	29 Apr 2025 22:00	YP/AJ	Ok
33	AR1242CCC500	PO110877.D	29 Apr 2025 22:36	YP/AJ	Ok
34	AR1248CCC500	PO110878.D	29 Apr 2025 22:55	YP/AJ	Ok
35	AR1254CCC500	PO110879.D	29 Apr 2025 23:13	YP/AJ	Ok
36	I.BLK	PO110880.D	29 Apr 2025 23:31	YP/AJ	Ok
37	PB167790BL	PO110881.D	29 Apr 2025 23:50	YP/AJ	Ok
38	PB167790BS	PO110882.D	30 Apr 2025 00:08	YP/AJ	Ok
39	Q1909-01	PO110883.D	30 Apr 2025 00:26	YP/AJ	Ok,NS
40	Q1909-02	PO110884.D	30 Apr 2025 00:45	YP/AJ	Ok,NS
41	AR1660CCC500	PO110885.D	30 Apr 2025 02:11	YP/AJ	Ok
42	AR1242CCC500	PO110886.D	30 Apr 2025 02:48	YP/AJ	Ok
43	AR1248CCC500	PO110887.D	30 Apr 2025 03:06	YP/AJ	Ok
44	AR1254CCC500	PO110888.D	30 Apr 2025 03:25	YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042925

Review By	yogesh	Review On	4/29/2025 11:07:42 AM
Supervise By		Supervise On	
SubDirectory	PO042925	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	I.BLK	PO110889.D	30 Apr 2025 03:43	YP/AJ	Ok
----	-------	------------	-------------------	-------	----

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	Sampleld	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071387.D	22 Apr 2025 09:56	YP\AJ	Ok
2	I.BLK	PP071388.D	22 Apr 2025 10:13	YP\AJ	Ok
3	AR1660ICC1000	PP071389.D	22 Apr 2025 10:29	YP\AJ	Ok
4	AR1660ICC750	PP071390.D	22 Apr 2025 10:45	YP\AJ	Ok
5	AR1660ICC500	PP071391.D	22 Apr 2025 11:02	YP\AJ	Ok
6	AR1660ICC250	PP071392.D	22 Apr 2025 11:18	YP\AJ	Ok
7	AR1660ICC050	PP071393.D	22 Apr 2025 11:34	YP\AJ	Ok,M
8	AR1221ICC500	PP071394.D	22 Apr 2025 11:51	YP\AJ	Ok
9	AR1232ICC500	PP071395.D	22 Apr 2025 12:07	YP\AJ	Ok
10	AR1242ICC1000	PP071396.D	22 Apr 2025 12:23	YP\AJ	Ok
11	AR1242ICC750	PP071397.D	22 Apr 2025 12:39	YP\AJ	Ok
12	AR1242ICC500	PP071398.D	22 Apr 2025 12:56	YP\AJ	Ok
13	AR1242ICC250	PP071399.D	22 Apr 2025 13:12	YP\AJ	Ok
14	AR1242ICC050	PP071400.D	22 Apr 2025 13:28	YP\AJ	Ok
15	AR1248ICC1000	PP071401.D	22 Apr 2025 13:45	YP\AJ	Ok
16	AR1248ICC750	PP071402.D	22 Apr 2025 14:01	YP\AJ	Ok
17	AR1248ICC500	PP071403.D	22 Apr 2025 14:17	YP\AJ	Ok
18	AR1248ICC250	PP071404.D	22 Apr 2025 14:33	YP\AJ	Ok
19	AR1248ICC050	PP071405.D	22 Apr 2025 14:50	YP\AJ	Ok
20	AR1254ICC1000	PP071406.D	22 Apr 2025 15:06	YP\AJ	Ok
21	AR1254ICC750	PP071407.D	22 Apr 2025 15:22	YP\AJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	AR1254ICC500	PP071408.D	22 Apr 2025 15:38	YP\AJ	Ok
23	AR1254ICC250	PP071409.D	22 Apr 2025 15:55	YP\AJ	Ok
24	AR1254ICC050	PP071410.D	22 Apr 2025 16:11	YP\AJ	Ok,M
25	AR1262ICC500	PP071411.D	22 Apr 2025 16:27	YP\AJ	Ok
26	AR1268ICC1000	PP071412.D	22 Apr 2025 16:44	YP\AJ	Ok,M
27	AR1268ICC750	PP071413.D	22 Apr 2025 17:00	YP\AJ	Ok,M
28	AR1268ICC500	PP071414.D	22 Apr 2025 17:16	YP\AJ	Ok,M
29	AR1268ICC250	PP071415.D	22 Apr 2025 17:33	YP\AJ	Ok,M
30	AR1268ICC050	PP071416.D	22 Apr 2025 17:49	YP\AJ	Ok,M
31	PP032725ICV500	PP071417.D	22 Apr 2025 18:05	YP\AJ	Ok
32	AR1242ICV500	PP071418.D	22 Apr 2025 18:38	YP\AJ	Ok
33	AR1248ICV500	PP071419.D	22 Apr 2025 19:10	YP\AJ	Ok
34	AR1254ICV500	PP071420.D	22 Apr 2025 19:43	YP\AJ	Ok
35	AR1268ICV500	PP071421.D	22 Apr 2025 20:16	YP\AJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042825

Review By	yogesh	Review On	4/28/2025 11:52:03 AM
Supervise By		Supervise On	
SubDirectory	PP042825	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	HEXANE	PP071527.D	28 Apr 2025 08:12	YP\AJ	Ok
2	AR1660CCC500	PP071528.D	28 Apr 2025 08:28	YP\AJ	Ok
3	AR1242CCC500	PP071529.D	28 Apr 2025 08:47	YP\AJ	Ok
4	AR1248CCC500	PP071530.D	28 Apr 2025 09:03	YP\AJ	Ok
5	AR1254CCC500	PP071531.D	28 Apr 2025 09:20	YP\AJ	Ok
6	I.BLK	PP071532.D	28 Apr 2025 09:36	YP\AJ	Ok
7	PB167737BL	PP071533.D	28 Apr 2025 09:53	YP\AJ	Ok
8	PB167750BL	PP071534.D	28 Apr 2025 10:10	YP\AJ	Ok
9	PB167741BS	PP071535.D	28 Apr 2025 10:26	YP\AJ	Ok
10	Q1860-33	PP071536.D	28 Apr 2025 10:43	YP\AJ	Ok,NS
11	Q1860-35	PP071537.D	28 Apr 2025 10:59	YP\AJ	Dilution
12	Q1875-04	PP071538.D	28 Apr 2025 11:16	YP\AJ	Ok
13	Q1860-42	PP071539.D	28 Apr 2025 11:32	YP\AJ	Ok,NS
14	Q1876-04DL	PP071540.D	28 Apr 2025 11:48	YP\AJ	Ok,NS
15	Q1876-09DL	PP071541.D	28 Apr 2025 12:04	YP\AJ	Ok,NS
16	AR1660CCC500	PP071542.D	28 Apr 2025 12:53	YP\AJ	Ok
17	AR1242CCC500	PP071543.D	28 Apr 2025 13:09	YP\AJ	Ok
18	AR1248CCC500	PP071544.D	28 Apr 2025 13:26	YP\AJ	Ok
19	AR1254CCC500	PP071545.D	28 Apr 2025 13:42	YP\AJ	Ok,NS
20	I.BLK	PP071546.D	28 Apr 2025 13:58	YP\AJ	Ok
21	Q1893-03	PP071547.D	28 Apr 2025 14:14	YP\AJ	Dilution

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042825

Review By	yogesh	Review On	4/28/2025 11:52:03 AM
Supervise By		Supervise On	
SubDirectory	PP042825	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1893-04	PP071548.D	28 Apr 2025 14:31	YP\AJ	Ok,NS
23	Q1893-05	PP071549.D	28 Apr 2025 14:47	YP\AJ	Dilution
24	Q1893-06	PP071550.D	28 Apr 2025 15:03	YP\AJ	Ok,NS
25	Q1893-07	PP071551.D	28 Apr 2025 15:19	YP\AJ	Dilution
26	Q1893-03DL	PP071552.D	28 Apr 2025 15:35	YP\AJ	Ok
27	Q1893-05DL	PP071553.D	28 Apr 2025 15:52	YP\AJ	Ok
28	Q1893-07DL	PP071554.D	28 Apr 2025 16:08	YP\AJ	Ok
29	Q1898-01	PP071555.D	28 Apr 2025 16:41	YP\AJ	Ok,NS
30	Q1898-02	PP071556.D	28 Apr 2025 16:58	YP\AJ	Ok
31	AR1660CCC500	PP071557.D	28 Apr 2025 18:03	YP\AJ	Ok
32	AR1242CCC500	PP071558.D	28 Apr 2025 18:36	YP\AJ	Ok
33	AR1248CCC500	PP071559.D	28 Apr 2025 18:53	YP\AJ	Ok
34	AR1254CCC500	PP071560.D	28 Apr 2025 19:09	YP\AJ	Ok,NS
35	I.BLK	PP071561.D	28 Apr 2025 19:25	YP\AJ	Ok
36	PB167754BL	PP071562.D	28 Apr 2025 19:42	YP\AJ	Ok
37	PB167754BS	PP071563.D	28 Apr 2025 19:58	YP\AJ	Ok
38	PB167765BL	PP071564.D	28 Apr 2025 20:15	YP\AJ	Ok
39	PB167765BS	PP071565.D	28 Apr 2025 20:31	YP\AJ	Ok
40	Q1884-01	PP071566.D	28 Apr 2025 20:47	YP\AJ	Dilution
41	Q1884-02	PP071567.D	28 Apr 2025 21:04	YP\AJ	Dilution
42	Q1889-01	PP071568.D	28 Apr 2025 21:20	YP\AJ	Not Ok
43	Q1889-01MS	PP071569.D	28 Apr 2025 21:36	YP\AJ	Not Ok
44	Q1889-01MSD	PP071570.D	28 Apr 2025 21:53	YP\AJ	Not Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042825

Review By	yogesh	Review On	4/28/2025 11:52:03 AM
Supervise By		Supervise On	
SubDirectory	PP042825	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

45	Q1889-02	PP071571.D	28 Apr 2025 22:09	YPIAJ	Ok
46	AR1660CCC500	PP071572.D	28 Apr 2025 23:31	YPIAJ	Ok
47	AR1242CCC500	PP071573.D	29 Apr 2025 00:03	YPIAJ	Ok
48	AR1248CCC500	PP071574.D	29 Apr 2025 00:20	YPIAJ	Ok
49	AR1254CCC500	PP071575.D	29 Apr 2025 00:36	YPIAJ	Ok,NS
50	I.BLK	PP071576.D	29 Apr 2025 00:52	YPIAJ	Ok
51	Q1889-03	PP071577.D	29 Apr 2025 01:09	YPIAJ	Ok
52	Q1891-01	PP071578.D	29 Apr 2025 01:25	YPIAJ	Ok,NS
53	Q1891-05	PP071579.D	29 Apr 2025 01:42	YPIAJ	Ok,NS
54	Q1892-01	PP071580.D	29 Apr 2025 01:58	YPIAJ	Ok
55	Q1892-05	PP071581.D	29 Apr 2025 02:14	YPIAJ	Ok
56	Q1892-09	PP071582.D	29 Apr 2025 02:31	YPIAJ	Ok
57	Q1895-01	PP071583.D	29 Apr 2025 02:47	YPIAJ	Ok
58	Q1895-03	PP071584.D	29 Apr 2025 03:03	YPIAJ	Ok
59	Q1895-05	PP071585.D	29 Apr 2025 03:20	YPIAJ	Ok,NS
60	Q1898-04	PP071586.D	29 Apr 2025 03:36	YPIAJ	Ok
61	AR1660CCC500	PP071587.D	29 Apr 2025 04:58	YPIAJ	Ok
62	AR1242CCC500	PP071588.D	29 Apr 2025 05:31	YPIAJ	Ok
63	AR1248CCC500	PP071589.D	29 Apr 2025 05:47	YPIAJ	Ok
64	AR1254CCC500	PP071590.D	29 Apr 2025 06:03	YPIAJ	Ok
65	I.BLK	PP071591.D	29 Apr 2025 06:20	YPIAJ	Ok
66	Q1898-03	PP071592.D	29 Apr 2025 06:36	YPIAJ	Ok
67	AR1660CCC500	PP071593.D	29 Apr 2025 07:58	YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042825

Review By	yogesh	Review On	4/28/2025 11:52:03 AM
Supervise By		Supervise On	
SubDirectory	PP042825	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP2		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

68	AR1242CCC500	PP071594.D	29 Apr 2025 08:31	YP\AJ	Ok
69	AR1248CCC500	PP071595.D	29 Apr 2025 08:47	YP\AJ	Ok
70	AR1254CCC500	PP071596.D	29 Apr 2025 09:04	YP\AJ	Ok,NS
71	I.BLK	PP071597.D	29 Apr 2025 09:20	YP\AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO041025

Review By	yogesh	Review On	4/10/2025 11:27:35 AM
Supervise By	mohammad	Supervise On	4/14/2025 12:40:29 AM
SubDirectory	PO041025	HP Acquire Method	HP Processing Method PO041025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO110347.D	10 Apr 2025 08:59		YP/AJ	Ok
2	I.BLK	I.BLK	PO110348.D	10 Apr 2025 09:17		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO110349.D	10 Apr 2025 09:36		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO110350.D	10 Apr 2025 09:54		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO110351.D	10 Apr 2025 10:13		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO110352.D	10 Apr 2025 10:31		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO110353.D	10 Apr 2025 10:49		YP/AJ	Ok
8	AR1221ICC500	AR1221ICC500	PO110354.D	10 Apr 2025 11:08		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO110355.D	10 Apr 2025 11:26		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO110356.D	10 Apr 2025 11:44		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO110357.D	10 Apr 2025 12:03		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO110358.D	10 Apr 2025 12:21		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO110359.D	10 Apr 2025 12:39		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO110360.D	10 Apr 2025 12:58		YP/AJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PO110361.D	10 Apr 2025 13:16		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO110362.D	10 Apr 2025 13:35		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO110363.D	10 Apr 2025 13:53		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO110364.D	10 Apr 2025 14:11		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO041025

Review By	yogesh	Review On	4/10/2025 11:27:35 AM
Supervise By	mohammad	Supervise On	4/14/2025 12:40:29 AM
SubDirectory	PO041025	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PO110365.D	10 Apr 2025 14:30		YP/AJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PO110366.D	10 Apr 2025 14:48		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO110367.D	10 Apr 2025 15:06		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO110368.D	10 Apr 2025 15:25		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO110369.D	10 Apr 2025 15:43		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO110370.D	10 Apr 2025 16:02		YP/AJ	Ok
25	AR1262ICC500	AR1262ICC500	PO110371.D	10 Apr 2025 16:20		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO110372.D	10 Apr 2025 16:38		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO110373.D	10 Apr 2025 16:57		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO110374.D	10 Apr 2025 17:15		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO110375.D	10 Apr 2025 17:33		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO110376.D	10 Apr 2025 17:52		YP/AJ	Ok
31	PO041025ICV500	ICVPO041025	PO110377.D	10 Apr 2025 18:09		YP/AJ	Ok
32	AR1242ICV500	ICVPO041025AR1242	PO110378.D	10 Apr 2025 18:46		YP/AJ	Ok
33	AR1248ICV500	ICVPO041025AR1248	PO110379.D	10 Apr 2025 19:22		YP/AJ	Ok
34	AR1254ICV500	ICVPO041025AR1254	PO110380.D	10 Apr 2025 19:58		YP/AJ	Ok
35	AR1268ICV500	ICVPO041025AR1268	PO110381.D	10 Apr 2025 20:35		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042925

Review By	yogesh	Review On	4/29/2025 11:07:42 AM
Supervise By		Supervise On	
SubDirectory	PO042925	HP Acquire Method	HP Processing Method PO041025

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO110845.D	29 Apr 2025 09:06		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO110846.D	29 Apr 2025 09:24		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO110847.D	29 Apr 2025 09:45		YP/AJ	Ok
4	AR1248CCC500	AR1248CCC500	PO110848.D	29 Apr 2025 10:03		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO110849.D	29 Apr 2025 10:21		YP/AJ	Ok
6	I.BLK	I.BLK	PO110850.D	29 Apr 2025 10:40		YP/AJ	Ok
7	PB167765BL	PB167765BL	PO110851.D	29 Apr 2025 10:58		YP/AJ	Not Ok
8	Q1889-01	COMP-1	PO110852.D	29 Apr 2025 11:16		YP/AJ	Ok
9	Q1889-01MS	COMP-1MS	PO110853.D	29 Apr 2025 11:35		YP/AJ	Ok
10	Q1889-01MSD	COMP-1MSD	PO110854.D	29 Apr 2025 11:53		YP/AJ	Ok
11	Q1860-34DL2	P001-SS028-01DL2	PO110855.D	29 Apr 2025 12:11		YP/AJ	Ok,NR
12	Q1904-01	VNJ-210	PO110856.D	29 Apr 2025 13:36		YP/AJ	Ok
13	Q1902-02	343	PO110857.D	29 Apr 2025 13:54		YP/AJ	Ok,NS
14	Q1903-01	COMP-4	PO110858.D	29 Apr 2025 14:12		YP/AJ	Ok
15	Q1903-02	COMP-5	PO110859.D	29 Apr 2025 14:31		YP/AJ	Ok
16	Q1903-03	COMP-6	PO110860.D	29 Apr 2025 14:49		YP/AJ	Ok
17	AR1660CCC500	AR1660CCC500	PO110861.D	29 Apr 2025 15:58		YP/AJ	Ok
18	AR1242CCC500	AR1242CCC500	PO110862.D	29 Apr 2025 16:35		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042925

Review By	yogesh	Review On	4/29/2025 11:07:42 AM
Supervise By		Supervise On	
SubDirectory	PO042925	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248CCC500	AR1248CCC500	PO110863.D	29 Apr 2025 16:53		YP/AJ	Ok
20	AR1254CCC500	AR1254CCC500	PO110864.D	29 Apr 2025 17:11		YP/AJ	Ok
21	I.BLK	I.BLK	PO110865.D	29 Apr 2025 17:29		YP/AJ	Ok
22	Q1904-01	VNJ-210	PO110866.D	29 Apr 2025 17:48		YP/AJ	Ok
23	Q1904-01MS	VNJ-210MS	PO110867.D	29 Apr 2025 18:05		YP/AJ	Ok
24	Q1904-01MSD	VNJ-210MSD	PO110868.D	29 Apr 2025 18:23		YP/AJ	Ok
25	Q1906-01	WC-4	PO110869.D	29 Apr 2025 18:42		YP/AJ	Ok
26	Q1906-05	WC-5	PO110870.D	29 Apr 2025 19:00		YP/AJ	Ok
27	Q1906-09	WC-6	PO110871.D	29 Apr 2025 19:19		YP/AJ	Ok,NR
28	Q1906-13	WC-7	PO110872.D	29 Apr 2025 19:37		YP/AJ	Ok
29	Q1897-01	40125	PO110873.D	29 Apr 2025 19:55	AR1242 Hit	YP/AJ	Ok,NS
30	Q1901-06	FB04262025	PO110874.D	29 Apr 2025 20:14		YP/AJ	Ok
31	Q1897-02	41425	PO110875.D	29 Apr 2025 20:32		YP/AJ	Ok,NS
32	AR1660CCC500	AR1660CCC500	PO110876.D	29 Apr 2025 22:00		YP/AJ	Ok
33	AR1242CCC500	AR1242CCC500	PO110877.D	29 Apr 2025 22:36		YP/AJ	Ok
34	AR1248CCC500	AR1248CCC500	PO110878.D	29 Apr 2025 22:55		YP/AJ	Ok
35	AR1254CCC500	AR1254CCC500	PO110879.D	29 Apr 2025 23:13		YP/AJ	Ok
36	I.BLK	I.BLK	PO110880.D	29 Apr 2025 23:31		YP/AJ	Ok
37	PB167790BL	PB167790BL	PO110881.D	29 Apr 2025 23:50		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO042925

Review By	yogesh	Review On	4/29/2025 11:07:42 AM
Supervise By		Supervise On	
SubDirectory	PO042925	HP Acquire Method	HP Processing Method PO041025
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

38	PB167790BS	PB167790BS	PO110882.D	30 Apr 2025 00:08		YP/AJ	Ok
39	Q1909-01	KMH358F-1-1	PO110883.D	30 Apr 2025 00:26	AR1254 Hit	YP/AJ	Ok,NS
40	Q1909-02	KMH358F-1-2	PO110884.D	30 Apr 2025 00:45	AR1254 Hit	YP/AJ	Ok,NS
41	AR1660CCC500	AR1660CCC500	PO110885.D	30 Apr 2025 02:11		YP/AJ	Ok
42	AR1242CCC500	AR1242CCC500	PO110886.D	30 Apr 2025 02:48		YP/AJ	Ok
43	AR1248CCC500	AR1248CCC500	PO110887.D	30 Apr 2025 03:06		YP/AJ	Ok
44	AR1254CCC500	AR1254CCC500	PO110888.D	30 Apr 2025 03:25		YP/AJ	Ok
45	I.BLK	I.BLK	PO110889.D	30 Apr 2025 03:43		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071387.D	22 Apr 2025 09:56		YPIAJ	Ok
2	I.BLK	I.BLK	PP071388.D	22 Apr 2025 10:13		YPIAJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP071389.D	22 Apr 2025 10:29		YPIAJ	Ok
4	AR1660ICC750	AR1660ICC750	PP071390.D	22 Apr 2025 10:45		YPIAJ	Ok
5	AR1660ICC500	AR1660ICC500	PP071391.D	22 Apr 2025 11:02		YPIAJ	Ok
6	AR1660ICC250	AR1660ICC250	PP071392.D	22 Apr 2025 11:18		YPIAJ	Ok
7	AR1660ICC050	AR1660ICC050	PP071393.D	22 Apr 2025 11:34		YPIAJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PP071394.D	22 Apr 2025 11:51		YPIAJ	Ok
9	AR1232ICC500	AR1232ICC500	PP071395.D	22 Apr 2025 12:07		YPIAJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP071396.D	22 Apr 2025 12:23		YPIAJ	Ok
11	AR1242ICC750	AR1242ICC750	PP071397.D	22 Apr 2025 12:39		YPIAJ	Ok
12	AR1242ICC500	AR1242ICC500	PP071398.D	22 Apr 2025 12:56		YPIAJ	Ok
13	AR1242ICC250	AR1242ICC250	PP071399.D	22 Apr 2025 13:12		YPIAJ	Ok
14	AR1242ICC050	AR1242ICC050	PP071400.D	22 Apr 2025 13:28		YPIAJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PP071401.D	22 Apr 2025 13:45		YPIAJ	Ok
16	AR1248ICC750	AR1248ICC750	PP071402.D	22 Apr 2025 14:01		YPIAJ	Ok
17	AR1248ICC500	AR1248ICC500	PP071403.D	22 Apr 2025 14:17		YPIAJ	Ok
18	AR1248ICC250	AR1248ICC250	PP071404.D	22 Apr 2025 14:33		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042225

Review By	yogesh	Review On	4/22/2025 1:34:56 PM
Supervise By	mohammad	Supervise On	4/24/2025 6:33:42 AM
SubDirectory	PP042225	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP071405.D	22 Apr 2025 14:50		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PP071406.D	22 Apr 2025 15:06		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PP071407.D	22 Apr 2025 15:22		YPIAJ	Ok,M
22	AR1254ICC500	AR1254ICC500	PP071408.D	22 Apr 2025 15:38		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP071409.D	22 Apr 2025 15:55		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP071410.D	22 Apr 2025 16:11		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP071411.D	22 Apr 2025 16:27		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP071412.D	22 Apr 2025 16:44		YPIAJ	Ok,M
27	AR1268ICC750	AR1268ICC750	PP071413.D	22 Apr 2025 17:00		YPIAJ	Ok,M
28	AR1268ICC500	AR1268ICC500	PP071414.D	22 Apr 2025 17:16		YPIAJ	Ok,M
29	AR1268ICC250	AR1268ICC250	PP071415.D	22 Apr 2025 17:33		YPIAJ	Ok,M
30	AR1268ICC050	AR1268ICC050	PP071416.D	22 Apr 2025 17:49		YPIAJ	Ok,M
31	PP032725ICV500	ICVPP042225	PP071417.D	22 Apr 2025 18:05		YPIAJ	Ok
32	AR1242ICV500	ICVPP042225AR1242	PP071418.D	22 Apr 2025 18:38		YPIAJ	Ok
33	AR1248ICV500	ICVPP042225AR1248	PP071419.D	22 Apr 2025 19:10		YPIAJ	Ok
34	AR1254ICV500	ICVPP042225AR1254	PP071420.D	22 Apr 2025 19:43		YPIAJ	Ok
35	AR1268ICV500	ICVPP042225AR1268	PP071421.D	22 Apr 2025 20:16		YPIAJ	Ok,M

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042825

Review By	yogesh	Review On	4/28/2025 11:52:03 AM
Supervise By		Supervise On	
SubDirectory	PP042825	HP Acquire Method	HP Processing Method PP042225

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369
CCC	PP24332,PP24347,PP24352,PP24357
Internal Standard/PEM	
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP071527.D	28 Apr 2025 08:12		YP\AJ	Ok
2	AR1660CCC500	AR1660CCC500	PP071528.D	28 Apr 2025 08:28		YP\AJ	Ok
3	AR1242CCC500	AR1242CCC500	PP071529.D	28 Apr 2025 08:47		YP\AJ	Ok
4	AR1248CCC500	AR1248CCC500	PP071530.D	28 Apr 2025 09:03		YP\AJ	Ok
5	AR1254CCC500	AR1254CCC500	PP071531.D	28 Apr 2025 09:20		YP\AJ	Ok
6	I.BLK	I.BLK	PP071532.D	28 Apr 2025 09:36		YP\AJ	Ok
7	PB167737BL	PB167737BL	PP071533.D	28 Apr 2025 09:53		YP\AJ	Ok
8	PB167750BL	PB167750BL	PP071534.D	28 Apr 2025 10:10		YP\AJ	Ok
9	PB167741BS	PB167741BS	PP071535.D	28 Apr 2025 10:26		YP\AJ	Ok
10	Q1860-33	P001-SS027-01	PP071536.D	28 Apr 2025 10:43	AR1248+AR1254+AR1260 Hit	YP\AJ	Ok,NS
11	Q1860-35	P001-SS029-01	PP071537.D	28 Apr 2025 10:59	AR1248+AR1254+AR1260 Hit, Need dil, TCMX high in both column	YP\AJ	Dilution
12	Q1875-04	AUD-25-0024	PP071538.D	28 Apr 2025 11:16		YP\AJ	Ok
13	Q1860-42	P001-SS036-01	PP071539.D	28 Apr 2025 11:32	AR1254 Hit	YP\AJ	Ok,NS
14	Q1876-04DL	AUD-25-0061DL	PP071540.D	28 Apr 2025 11:48	AR1254 Hit	YP\AJ	Ok,NS
15	Q1876-09DL	AUD-25-0066DL	PP071541.D	28 Apr 2025 12:04	AR1254 Hit	YP\AJ	Ok,NS
16	AR1660CCC500	AR1660CCC500	PP071542.D	28 Apr 2025 12:53		YP\AJ	Ok
17	AR1242CCC500	AR1242CCC500	PP071543.D	28 Apr 2025 13:09		YP\AJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042825

Review By	yogesh	Review On	4/28/2025 11:52:03 AM
Supervise By		Supervise On	
SubDirectory	PP042825	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1248CCC500	AR1248CCC500	PP071544.D	28 Apr 2025 13:26		YPIAJ	Ok
19	AR1254CCC500	AR1254CCC500	PP071545.D	28 Apr 2025 13:42		YPIAJ	Ok,NS
20	I.BLK	I.BLK	PP071546.D	28 Apr 2025 13:58		YPIAJ	Ok
21	Q1893-03	INTERIOR-1	PP071547.D	28 Apr 2025 14:14	AR1248 Hit, Need 20X	YPIAJ	Dilution
22	Q1893-04	INTERIOR-2	PP071548.D	28 Apr 2025 14:31	AR1248 Hit	YPIAJ	Ok,NS
23	Q1893-05	INTERIOR-3	PP071549.D	28 Apr 2025 14:47	AR1248 Hit, Need 50X, TCMX high 2nd column	YPIAJ	Dilution
24	Q1893-06	INTERIOR-4	PP071550.D	28 Apr 2025 15:03	AR1248 Hit	YPIAJ	Ok,NS
25	Q1893-07	EXTERIOR-1	PP071551.D	28 Apr 2025 15:19	AR1248 Hit, Need 10X	YPIAJ	Dilution
26	Q1893-03DL	INTERIOR-1DL	PP071552.D	28 Apr 2025 15:35	AR1248 Hit	YPIAJ	Ok
27	Q1893-05DL	INTERIOR-3DL	PP071553.D	28 Apr 2025 15:52	AR1248 Hit	YPIAJ	Ok
28	Q1893-07DL	EXTERIOR-1DL	PP071554.D	28 Apr 2025 16:08	AR1248 Hit	YPIAJ	Ok
29	Q1898-01	41525A	PP071555.D	28 Apr 2025 16:41		YPIAJ	Ok,NS
30	Q1898-02	41525B	PP071556.D	28 Apr 2025 16:58		YPIAJ	Ok
31	AR1660CCC500	AR1660CCC500	PP071557.D	28 Apr 2025 18:03		YPIAJ	Ok
32	AR1242CCC500	AR1242CCC500	PP071558.D	28 Apr 2025 18:36		YPIAJ	Ok
33	AR1248CCC500	AR1248CCC500	PP071559.D	28 Apr 2025 18:53		YPIAJ	Ok
34	AR1254CCC500	AR1254CCC500	PP071560.D	28 Apr 2025 19:09		YPIAJ	Ok,NS
35	I.BLK	I.BLK	PP071561.D	28 Apr 2025 19:25		YPIAJ	Ok
36	PB167754BL	PB167754BL	PP071562.D	28 Apr 2025 19:42		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042825

Review By	yogesh	Review On	4/28/2025 11:52:03 AM
Supervise By		Supervise On	
SubDirectory	PP042825	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

37	PB167754BS	PB167754BS	PP071563.D	28 Apr 2025 19:58		YPIAJ	Ok
38	PB167765BL	PB167765BL	PP071564.D	28 Apr 2025 20:15		YPIAJ	Ok
39	PB167765BS	PB167765BS	PP071565.D	28 Apr 2025 20:31		YPIAJ	Ok
40	Q1884-01	P001-SS037-01	PP071566.D	28 Apr 2025 20:47	AR1248 + AR1254 + AR1260 Hit, Need Dilution, DCB High in both col	YPIAJ	Dilution
41	Q1884-02	P001-SS038-01	PP071567.D	28 Apr 2025 21:04	AR1254 + AR1260 Hit, Need dilution, DCB High in both column	YPIAJ	Dilution
42	Q1889-01	COMP-1	PP071568.D	28 Apr 2025 21:20	Check Chromatogram	YPIAJ	Not Ok
43	Q1889-01MS	COMP-1MS	PP071569.D	28 Apr 2025 21:36		YPIAJ	Not Ok
44	Q1889-01MSD	COMP-1MSD	PP071570.D	28 Apr 2025 21:53		YPIAJ	Not Ok
45	Q1889-02	COMP-2	PP071571.D	28 Apr 2025 22:09		YPIAJ	Ok
46	AR1660CCC500	AR1660CCC500	PP071572.D	28 Apr 2025 23:31		YPIAJ	Ok
47	AR1242CCC500	AR1242CCC500	PP071573.D	29 Apr 2025 00:03		YPIAJ	Ok
48	AR1248CCC500	AR1248CCC500	PP071574.D	29 Apr 2025 00:20		YPIAJ	Ok
49	AR1254CCC500	AR1254CCC500	PP071575.D	29 Apr 2025 00:36		YPIAJ	Ok,NS
50	I.BLK	I.BLK	PP071576.D	29 Apr 2025 00:52		YPIAJ	Ok
51	Q1889-03	COMP-3	PP071577.D	29 Apr 2025 01:09		YPIAJ	Ok
52	Q1891-01	MH-C	PP071578.D	29 Apr 2025 01:25		YPIAJ	Ok,NS
53	Q1891-05	MH-D	PP071579.D	29 Apr 2025 01:42		YPIAJ	Ok,NS
54	Q1892-01	MH-G	PP071580.D	29 Apr 2025 01:58		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP042825

Review By	yogesh	Review On	4/28/2025 11:52:03 AM
Supervise By		Supervise On	
SubDirectory	PP042825	HP Acquire Method	HP Processing Method PP042225
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,P P24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP 24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369		
CCC	PP24332,PP24347,PP24352,PP24357		
Internal Standard/PEM			
ICV/I.BLK	PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

55	Q1892-05	MH-H	PP071581.D	29 Apr 2025 02:14		YPIAJ	Ok
56	Q1892-09	MH-U2	PP071582.D	29 Apr 2025 02:31		YPIAJ	Ok
57	Q1895-01	COMP-1	PP071583.D	29 Apr 2025 02:47		YPIAJ	Ok
58	Q1895-03	COMP-2	PP071584.D	29 Apr 2025 03:03		YPIAJ	Ok
59	Q1895-05	COMP-3	PP071585.D	29 Apr 2025 03:20	AR1254 Hit	YPIAJ	Ok,NS
60	Q1898-04	42525B	PP071586.D	29 Apr 2025 03:36		YPIAJ	Ok
61	AR1660CCC500	AR1660CCC500	PP071587.D	29 Apr 2025 04:58		YPIAJ	Ok
62	AR1242CCC500	AR1242CCC500	PP071588.D	29 Apr 2025 05:31		YPIAJ	Ok
63	AR1248CCC500	AR1248CCC500	PP071589.D	29 Apr 2025 05:47		YPIAJ	Ok
64	AR1254CCC500	AR1254CCC500	PP071590.D	29 Apr 2025 06:03		YPIAJ	Ok
65	I.BLK	I.BLK	PP071591.D	29 Apr 2025 06:20		YPIAJ	Ok
66	Q1898-03	42525A	PP071592.D	29 Apr 2025 06:36		YPIAJ	Ok
67	AR1660CCC500	AR1660CCC500	PP071593.D	29 Apr 2025 07:58		YPIAJ	Ok
68	AR1242CCC500	AR1242CCC500	PP071594.D	29 Apr 2025 08:31		YPIAJ	Ok
69	AR1248CCC500	AR1248CCC500	PP071595.D	29 Apr 2025 08:47		YPIAJ	Ok
70	AR1254CCC500	AR1254CCC500	PP071596.D	29 Apr 2025 09:04		YPIAJ	Ok,NS
71	I.BLK	I.BLK	PP071597.D	29 Apr 2025 09:20		YPIAJ	Ok

M : Manual Integration



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/28/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 04/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:37
Out Date: 04/26/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135558

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q1883-01	OU4-PCS-TC-27-042325	1	1.18	10.04	11.22	10.67	94.5	
Q1883-03	OU4-PCS-TC-28-042325	2	1.14	9.97	11.11	10.56	94.5	
Q1883-05	OU4-PCS-TC-29-042325	3	1.19	10.54	11.73	11.28	95.7	
Q1883-07	OU4-PCS-TC-30-042325	4	1.17	10.15	11.32	10.96	96.5	
Q1883-09	OU4-PCS-TC-31-042325	5	1.18	10.48	11.66	11.34	96.9	
Q1883-11	OU4-PCS-TC-32-042325	6	1.18	9.96	11.14	10.87	97.3	
Q1883-13	OU4-PCS-18-042325	7	1.15	10.25	11.4	10.86	94.7	
Q1883-15	OU4-PCS-19-042325	8	1.19	10.54	11.73	11.29	95.8	
Q1884-01	P001-SS037-01	9	1.17	10.30	11.47	10.87	94.2	
Q1884-02	P001-SS038-01	10	1.17	10.32	11.49	11.22	97.4	
Q1888-05	SVOC-GPC-BLANK	11	1.00	1.00	2.00	2.00	100.0	
Q1888-06	PEST-GPC-BLANK	12	1.00	1.00	2.00	2.00	100.0	
Q1888-07	PEST-GPC-BLANK-SPIKE	13	1.00	1.00	2.00	2.00	100.0	
Q1888-08	PCB-GPC-BLANK	14	1.00	1.00	2.00	2.00	100.0	
Q1888-09	PCB-GPC-BLANK-SPIKE	15	1.00	1.00	2.00	2.00	100.0	
Q1888-10	SVOC-GPC2-BLANK	16	1.00	1.00	2.00	2.00	100.0	
Q1888-11	PEST-GPC2-BLANK	17	1.00	1.00	2.00	2.00	100.0	
Q1888-12	PEST-GPC2-BLANK-SPIKE	18	1.00	1.00	2.00	2.00	100.0	
Q1888-13	PCB-GPC2-BLANK	19	1.00	1.00	2.00	2.00	100.0	
Q1888-14	PCB-GCP2-BLANK-SPIKE	20	1.00	1.00	2.00	2.00	100.0	
Q1889-01	COMP-1	21	1.19	10.07	11.26	9.5	82.5	
Q1889-02	COMP-2	22	1.16	10.50	11.66	9.61	80.5	
Q1889-03	COMP-3	23	1.11	10.73	11.84	9.77	80.7	
Q1891-01	MH-C	24	1.15	9.48	10.63	9.8	91.2	
Q1891-02	MH-C-EPH	25	1.16	9.82	10.98	10.15	91.5	
Q1891-03	MH-C-VOC	26	1.19	9.94	11.13	10.2	90.6	
Q1891-05	MH-D	27	1.15	10.17	11.32	9.64	83.5	
Q1891-06	MH-D-EPH	28	1.15	9.97	11.12	8.74	76.1	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/28/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 04/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:37
Out Date: 04/26/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135558

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
Q1891-07	MH-D-VOC	29	1.18	9.46	10.64	8.83	80.9	
Q1892-01	MH-G	30	1.13	9.61	10.74	9.35	85.5	
Q1892-02	MH-G-EPH	31	1.15	10.29	11.44	9.98	85.8	
Q1892-03	MH-G-VOC	32	1.18	10.45	11.63	9.82	82.7	
Q1892-05	MH-H	33	1.14	9.66	10.8	9.92	90.9	
Q1892-06	MH-H-EPH	34	1.12	10.66	11.78	10.68	89.7	
Q1892-07	MH-H-VOC	35	1.18	10.19	11.37	10.28	89.3	
Q1892-09	MH-U2	36	1.18	10.04	11.22	9.96	87.5	
Q1892-10	MH-U2-EPH	37	1.19	10.34	11.53	10.04	85.6	
Q1892-11	MH-U2-VOC	38	1.18	10.22	11.4	10.14	87.7	
Q1893-01	UGGP-1	39	1.00	1.00	2.00	2.00	100.0	TAR SAMPLE
Q1893-02	UGGP-2	40	1.00	1.00	2.00	2.00	100.0	TAR SAMPLE
Q1893-03	INTERIOR-1	41	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1893-04	INTERIOR-2	42	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1893-05	INTERIOR-3	43	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1893-06	INTERIOR-4	44	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1893-07	EXTERIOR-1	45	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1895-01	COMP-1	55	1.14	10.58	11.72	10.97	92.9	
Q1895-03	COMP-2	56	1.14	10.77	11.91	10.96	91.2	
Q1895-05	COMP-3	57	1.18	10.06	11.24	10.56	93.2	
Q1896-01	295-BERGEN-RO	58	1.14	11.31	12.45	10.91	86.4	
Q1896-02	295-BERGEN-RO	59	1.18	10.01	11.19	9.9	87.1	
Q1898-01	41525A	60	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1898-02	41525B	61	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1898-03	42525A	62	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1898-04	42525B	63	1.00	1.00	2.00	2.00	100.0	wipe sample
Q1900-01	WC-1	46	1.14	9.98	11.12	9.56	84.4	
Q1900-02	WC-1-EPH	47	1.14	10.35	11.49	10.19	87.4	



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 4/28/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 17:25
In Date: 04/25/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:37
Out Date: 04/26/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB135558

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q1900-03	WC-1-VOC	48	1.14	10.01	11.15	9.73	85.8	
Q1900-05	WC-2	49	1.19	8.82	10.01	8.76	85.8	
Q1900-06	WC-2-EPH	50	1.19	10.08	11.27	9.76	85.0	
Q1900-07	WC-2-VOC	51	1.17	10.79	11.96	10.33	84.9	
Q1900-09	WC-3	52	1.14	9.61	10.75	9.32	85.1	
Q1900-10	WC-3-EPH	53	1.12	10.38	11.5	9.11	77.0	
Q1900-11	WC-3-VOC	54	1.19	10.71	11.9	10.39	85.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

WORKLIST(Hardcopy Internal Chain)

13558

WorkList Name : %1-042525

WorkList ID : 189135

Department : Wet-Chemistry

Date : 04-25-2025 08:09:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1883-01	OU4-PCS-TC-27-042325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L41	04/23/2025	Chemtech -SO
Q1883-03	OU4-PCS-TC-28-042325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L41	04/23/2025	Chemtech -SO
Q1883-05	OU4-PCS-TC-29-042325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L41	04/23/2025	Chemtech -SO
Q1883-07	OU4-PCS-TC-30-042325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L41	04/23/2025	Chemtech -SO
Q1883-09	OU4-PCS-TC-31-042325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L41	04/23/2025	Chemtech -SO
Q1883-11	OU4-PCS-TC-32-042325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L41	04/23/2025	Chemtech -SO
Q1883-13	OU4-PCS-18-042325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L41	04/23/2025	Chemtech -SO
Q1883-15	OU4-PCS-19-042325	Solid	Percent Solids	Cool 4 deg C	NOBI03	L41	04/23/2025	Chemtech -SO
Q1884-01	P001-SS037-01	Solid	Percent Solids	Cool 4 deg C	ROYF02	L51	04/24/2025	Chemtech -SO
Q1884-02	P001-SS038-01	Solid	Percent Solids	Cool 4 deg C	ROYF02	L51	04/24/2025	Chemtech -SO
Q1888-05	SVOC-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	04/18/2025	Chemtech -SO
Q1888-06	PEST-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	04/18/2025	Chemtech -SO
Q1888-07	PEST-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	04/18/2025	Chemtech -SO
Q1888-08	PCB-GPC-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	04/18/2025	Chemtech -SO
Q1888-09	PCB-GPC-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	04/18/2025	Chemtech -SO
Q1888-10	SVOC-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	04/18/2025	Chemtech -SO
Q1888-11	PEST-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	04/18/2025	Chemtech -SO
Q1888-12	PEST-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	04/18/2025	Chemtech -SO
Q1888-13	PCB-GPC2-BLANK	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	04/18/2025	Chemtech -SO
Q1888-14	PCB-GPC2-BLANK-SPIKE	Solid	Percent Solids	Cool 4 deg C	CHEM02	L31	04/18/2025	Chemtech -SO
Q1889-01	COMP-1	Solid	Percent Solids	Cool 4 deg C	POWE02	L51	04/24/2025	Chemtech -SO

Date/Time 04/25/25 13:20

Raw Sample Received by: 50 (wec)

Raw Sample Relinquished by: 50 (wec)

Date/Time 04/25/25

Raw Sample Received by: 50 (wec)

Raw Sample Relinquished by: 50 (wec)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-042525

WorkList ID : 189135

Department : Wet-Chemistry

Date : 04-25-2025 08:09:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1889-02	COMP-2	Solid	Percent Solids	Cool 4 deg C	POWE02	L51	04/24/2025	Chemtech -SO
Q1889-03	COMP-3	Solid	Percent Solids	Cool 4 deg C	POWE02	L51	04/24/2025	Chemtech -SO
Q1891-01	MH-C	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1891-02	MH-C-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1891-03	MH-C-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1891-05	MH-D	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1891-06	MH-D-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1891-07	MH-D-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1892-01	MH-G	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/24/2025	Chemtech -SO
Q1892-02	MH-G-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/24/2025	Chemtech -SO
Q1892-03	MH-G-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/24/2025	Chemtech -SO
Q1892-05	MH-H	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/24/2025	Chemtech -SO
Q1892-06	MH-H-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/24/2025	Chemtech -SO
Q1892-07	MH-H-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/24/2025	Chemtech -SO
Q1892-09	MH-U2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/25/2025	Chemtech -SO
Q1892-10	MH-U2-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/25/2025	Chemtech -SO
Q1892-11	MH-U2-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L51	04/25/2025	Chemtech -SO
Q1893-01	UGGP-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1893-02	UGGP-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1893-03	INTERIOR-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1893-04	INTERIOR-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO

Date/Time

04/25/25

15420

Raw Sample Received by:

SQ WOLU

Raw Sample Relinquished by:

JD CSM

Date/Time 04/25/25

17:30

Raw Sample Received by:

JD CSM

Raw Sample Relinquished by:

JD CSM

WORKLIST(Hardcopy Internal Chain)

135558

WorkList Name : %1-042525

WorkList ID : 189135

Department : Wet-Chemistry

Date : 04-25-2025 08:09:39

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1893-05	INTERIOR-3	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1893-06	INTERIOR-4	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1893-07	EXTERIOR-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1895-01	COMP-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1895-03	COMP-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1895-05	COMP-3	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1896-01	295-BERGEN-RO	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/25/2025	Chemtech -SO
Q1896-02	295-BERGEN-RO	Solid	Percent Solids	Cool 4 deg C	PSEG03	L31	04/25/2025	Chemtech -SO
Q1898-01	41525A	Solid	Percent Solids	Cool 4 deg C	PSEG03	L12	04/25/2025	Chemtech -SO
Q1898-02	41525B	Solid	Percent Solids	Cool 4 deg C	PSEG03	L12	04/25/2025	Chemtech -SO
Q1898-03	42525A	Solid	Percent Solids	Cool 4 deg C	PSEG03	L12	04/25/2025	Chemtech -SO
Q1898-04	42525B	Solid	Percent Solids	Cool 4 deg C	PSEG03	L12	04/25/2025	Chemtech -SO
Q1900-01	WC-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1900-02	WC-1-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1900-03	WC-1-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1900-05	WC-2	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1900-06	WC-2-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1900-07	WC-2-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1900-09	WC-3	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1900-10	WC-3-EPH	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO
Q1900-11	WC-3-VOC	Solid	Percent Solids	Cool 4 deg C	PSEG03	L41	04/25/2025	Chemtech -SO

04/25/25 13:20

17:30

Raw Sample Received by: 30 wcc

Raw Sample Received by: JDCSM

Raw Sample Relinquished by: JDCSM

Raw Sample Relinquished by: 30 wcc

SOP ID: M3541-ASE Extraction-14

Clean Up SOP #: Acid Cleanup **Extraction Start Date :** 04/28/2025

Matrix : Solid **Extraction Start Time :** 09:05

Weigh By: EH **Extraction By:** RJ **Extraction End Date :** 04/28/2025

Balance check: RJ **Filter By:** RJ **Extraction End Time :** 12:10

Balance ID: EX-SC-2 **pH Meter ID:** N/A **Concentration By:** EH

pH Strip Lot#: N/A **Hood ID:** 3,7 **Supervisor By :** RUPESH

Extraction Method: ☐ Separatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☒ Soxhlet

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	5000 PPB	PP24461
Surrogate	1.0ML	200 PPB	PP24460
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Hexane/Acetone/1:1	N/A	EP2601
Baked Na2SO4	N/A	EP2607
Sand	N/A	E2865
Hexane	N/A	E3928
H2SO4 1:1	N/A	EP2565
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

40 ML Vial lot# 03-40 BTS723, Q1893-01,02 Used Limited volume as samples are Tar matrix.

KD Bath ID: N/A **Envap ID:** NEVAP-02

KD Bath Temperature: N/A **Envap Temperature:** 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
4/28/25	RS (Est Lab)	Y-P-pestiP43
12:15	Preparation Group	Analysis Group

Analytical Method: M3541-ASE Extraction-14

Concentration Date: 04/28/2025

Sample ID	Client Sample ID	Test	g/mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB167765BL	ABLK765	PCB	30.02	N/A	ritesh	Evelyn	10			U2-1
PB167765BS	ALCS765	PCB	30.03	N/A	ritesh	Evelyn	10			2
Q1884-01	P001-SS037-01	PCB	30.07	N/A	ritesh	Evelyn	10			3
Q1884-02	P001-SS038-01	PCB	30.06	N/A	ritesh	Evelyn	10			4
Q1889-01	COMP-1	PCB Group1	30.02	N/A	ritesh	Evelyn	10	E		5
Q1889-01MS	COMP-1MS	PCB Group1	30.06	N/A	ritesh	Evelyn	10	E		6
Q1889-01MS D	COMP-1MSD	PCB Group1	30.03	N/A	ritesh	Evelyn	10	E		U3-1
Q1889-02	COMP-2	PCB Group1	30.05	N/A	ritesh	Evelyn	10	E		2
Q1889-03	COMP-3	PCB Group1	30.02	N/A	ritesh	Evelyn	10	E		3
Q1891-01	MH-C	PCB	30.07	N/A	ritesh	Evelyn	10	E		4
Q1891-05	MH-D	PCB	30.08	N/A	ritesh	Evelyn	10	E		5
Q1892-01	MH-G	PCB	30.04	N/A	ritesh	Evelyn	10	E		6
Q1892-05	MH-H	PCB	30.09	N/A	ritesh	Evelyn	10	E		U6-1
Q1892-09	MH-U2	PCB	30.02	N/A	ritesh	Evelyn	10	E		2
Q1893-01	UGGP-1	PCB	10.03	N/A	ritesh	Evelyn	10	B	Tar Mat	3
Q1893-02	UGGP-2	PCB	10.06	N/A	ritesh	Evelyn	10	B	Tar Mat	4
Q1895-01	COMP-1	PCB	30.05	N/A	ritesh	Evelyn	10	E		5
Q1895-03	COMP-2	PCB	30.08	N/A	ritesh	Evelyn	10	E		6
Q1895-05	COMP-3	PCB	30.01	N/A	ritesh	Evelyn	10	E		U1-1
Q1896-01	295-BERGEN-RO	PCB	30.05	N/A	ritesh	Evelyn	10	E		2
Q1900-01	WC-1	PCB	30.04	N/A	ritesh	Evelyn	10	E		3
Q1900-05	WC-2	PCB	30.07	N/A	ritesh	Evelyn	10	E		4
Q1900-09	WC-3	PCB	30.02	N/A	ritesh	Evelyn	10	E		5

* Extracts relinquished on the same date as received.

RS
4/28

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q1884 WorkList ID : 189180 Department : Extraction Date : 04-28-2025 08:46:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1884-01	P001-SS037-01	Solid	PCB	Cool 4 deg C	ROYF02	L51	04/24/2025	8082A
Q1884-02	P001-SS038-01	Solid	PCB	Cool 4 deg C	ROYF02	L51	04/24/2025	8082A
Q1889-01	COMP-1	Solid	PCB Group1	Cool 4 deg C	POWE02	L51	04/24/2025	8082A
Q1889-02	COMP-2	Solid	PCB Group1	Cool 4 deg C	POWE02	L51	04/24/2025	8082A
Q1889-03	COMP-3	Solid	PCB Group1	Cool 4 deg C	POWE02	L51	04/24/2025	8082A
Q1891-01	MH-C	Solid	PCB	Cool 4 deg C	PSEG03	L41	04/25/2025	8082A
Q1891-05	MH-D	Solid	PCB	Cool 4 deg C	PSEG03	L41	04/25/2025	8082A
Q1892-01	MH-G	Solid	PCB	Cool 4 deg C	PSEG03	L51	04/24/2025	8082A
Q1892-05	MH-H	Solid	PCB	Cool 4 deg C	PSEG03	L51	04/24/2025	8082A
Q1892-09	MH-U2	Solid	PCB	Cool 4 deg C	PSEG03	L51	04/25/2025	8082A
Q1893-01	UGGP-1	Solid	PCB	Cool 4 deg C	PSEG03	L41	04/25/2025	8082A
Q1893-02	UGGP-2	Solid	PCB	Cool 4 deg C	PSEG03	L41	04/25/2025	8082A
Q1895-01	COMP-1	Solid	PCB	Cool 4 deg C	PSEG03	L41	04/25/2025	8082A
Q1895-03	COMP-2	Solid	PCB	Cool 4 deg C	PSEG03	L41	04/25/2025	8082A
Q1895-05	COMP-3	Solid	PCB	Cool 4 deg C	PSEG03	L41	04/25/2025	8082A
Q1896-01	295-BERGEN-RO	Solid	PCB	Cool 4 deg C	PSEG03	L31	04/25/2025	8082A
Q1900-01	WC-1	Solid	PCB	Cool 4 deg C	PSEG03	L41	04/25/2025	8082A
Q1900-05	WC-2	Solid	PCB	Cool 4 deg C	PSEG03	L41	04/25/2025	8082A
Q1900-09	WC-3	Solid	PCB	Cool 4 deg C	PSEG03	L41	04/25/2025	8082A

Date/Time
Raw Sample Received by:
Raw Sample Relinquished by:

4/28/25 9:00
RJ(Ext-194)
RJ(Ext-194)

Date/Time
Raw Sample Received by:
Raw Sample Relinquished by:

4/28/25 9:40
RJ(Ext-194)
RJ(Ext-194)

Prep Standard - Chemical Standard Summary

Order ID : Q1889

Test : PCB Group1

Prepbatch ID : PB167765,

Sequence ID/Qc Batch ID: PO042925,PP042825,

Standard ID :

EP2565,EP2601,EP2607,PP24329,PP24330,PP24331,PP24332,PP24333,PP24334,PP24335,PP24336,PP24337,PP24338,PP24339,PP24340,PP24341,PP24342,PP24343,PP24344,PP24345,PP24346,PP24347,PP24348,PP24349,PP24350,PP24351,PP24352,PP24353,PP24354,PP24355,PP24356,PP24357,PP24358,PP24359,PP24360,PP24361,PP24362,PP24363,PP24364,PP24365,PP24366,PP24367,PP24368,PP24369,PP24370,PP24371,PP24372,PP24373,PP24374,PP24375,PP24376,PP24377,PP24378,PP24379,PP24380,PP24381,PP24382,PP24384,PP24385,PP24386,PP24387,PP24460,PP24461,

Chemical ID :

E2865,E3551,E3804,E3877,E3916,E3917,E3928,M5173,P11522,P12699,P12702,P12931,P12936,P12949,P12955,P12957,P13355,P13356,P13373,P13381,P13589,P13591,P13697,P13702,P13830,P13878,P13883,W3112,W3177,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2565	11/20/2024	05/20/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 11/20/2024

FROM 1000.00000ml of M5173 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
230	1:1ACETONE/HEXANE	EP2601	04/07/2025	10/03/2025	Rajesh Parikh	None	None	Riteshkumar Patel 04/07/2025

FROM 8000.00000ml of E3916 + 8000.00000ml of E3917 = Final Quantity: 8000.000 ml

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2607	04/25/2025	07/01/2025	RUPESHKUMAR SHAH	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 04/25/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
84	Pest/PCB Surrogate Stock 20 PPM	PP24329	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P13356 + 9.00000ml of W3177 = Final Quantity: 10.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
202	AR1660 1000/100 ppb working solution 1st source	PP24330	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13697 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
203	AR1660 750 PPB STD	PP24331	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
204	AR1660 500 PPB STD	PP24332	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24330 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
205	AR1660 250 PPB STD	PP24333	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24330 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
206	AR1660 50 PPB STD	PP24334	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24332 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
213	AR1221 1000 PPB WORKING SOLUTION	PP24335	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13702 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1079	AR1221 750 PPB STD	PP24336	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24335 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
222	AR1221 500 PPB STD	PP24337	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24335 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1080	AR1221 250 PPB STD	PP24338	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24335 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1081	AR1221 50 PPB STD	PP24339	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24337 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
214	AR1232 1000 PPB WORKING SOLUTION	PP24340	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13878 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1063	AR1232 750 PPB STD	PP24341	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24340 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
223	AR1232 500 PPB STD	PP24342	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24340 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1064	AR1232 250 PPB STD	PP24343	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24340 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1065	AR1232 50 PPB STD	PP24344	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24342 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
215	AR1242 1000 PPB WORKING STD	PP24345	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12931 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1067	AR1242 750 PPB STD	PP24346	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.75000ml of PP24345 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
224	AR1242 500 PPB STD	PP24347	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24345 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1068	AR1242 250 PPB STD	PP24348	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24345 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1069	AR1242 50 PPB STD	PP24349	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24347 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
216	AR1248 1000 PPB WORKING STD	PP24350	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P12936 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1075	AR1248 750 PPB STD	PP24351	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24350 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
225	AR1248 500 PPB STD	PP24352	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24350 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1076	AR1248 250 PPB STD	PP24353	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24350 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1077	AR1248 50 PPB STD	PP24354	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24352 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
217	AR1254 1000 PPB WORKING STD	PP24355	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13830 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1071	AR1254 750 PPB STD	PP24356	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24355 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
226	AR1254 500 PPB STD	PP24357	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24355 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1072	AR1254 250 PPB STD	PP24358	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24355 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1073	AR1254 50 PPB STD	PP24359	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24357 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1529	AR1262 1000 PPB Working Solution	PP24360	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13883 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3753	AR1262 750 PPB STD	PP24361	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24360 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1530	AR1262 500 PPB STD	PP24362	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24360 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3754	AR1262 250 PPB STD	PP24363	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24360 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3755	AR1262 50 PPB STD	PP24364	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24362 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1532	AR1268 1000 PPB Working Solution	PP24365	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.10000ml of P13381 + 99.40000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3820	AR1268 750 PPB STD	PP24366	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.25000ml of W3177 + 0.75000ml of PP24365 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1533	AR1268 500 PPB STD	PP24367	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24365 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3821	AR1268 250 PPB STD	PP24368	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.75000ml of W3177 + 0.25000ml of PP24365 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3822	AR1268 50 PPB STD	PP24369	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.90000ml of W3177 + 0.10000ml of PP24367 = Final Quantity: 1.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
404	AR1660 100 PPM Stock Solution 2nd Source	PP24370	03/18/2025	09/18/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 1.00000ml of P12949 + 9.00000ml of E3804 = Final Quantity: 10.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
405	AR1660 1000/100 PPB ICV STD	PP24371	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza 04/03/2025

FROM 98.50000ml of W3177 + 0.50000ml of PP24329 + 1.00000ml of PP24370 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
406	AR1660 500 PPB ICV	PP24372	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24371 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3789	AR1221 1000 PPB WORKING SOL.2ND SOURCE(AGILENT)	PP24373	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13373 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1886	AR1221 500 PPB ICV	PP24374	03/18/2025	08/12/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of E3877 + 0.50000ml of W3177 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1887	AR1232 1000 PPB Working Sol. 2nd Source	PP24375	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12699 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1888	AR1232 500 PPB ICV	PP24376	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24375 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1889	AR1242 1000 PPB Working Sol. 2nd Source	PP24377	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13589 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1891	AR1242 500 PPB ICV	PP24378	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24377 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1890	AR1248 1000 PPB Working Sol. 2nd Source	PP24379	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P13591 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1892	AR1248 500 PPB ICV	PP24380	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24379 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1893	AR1254 1000 PPB Working Sol. 2nd Source	PP24381	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12957 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1894	AR1254 500 PPB ICV	PP24382	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24381 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3757	AR1262 1000 PPB Working Solution second source	PP24384	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P12702 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3758	AR1262 500 PPB STD ICV	PP24385	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24384 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3817	AR1268 1000 ppb Working Soln. 2nd source	PP24386	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 1.00000ml of P11522 + 98.50000ml of W3177 + 0.50000ml of PP24329 = Final Quantity: 100.000 ml

Pest/Pcb STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3823	AR1268 500 PPB STD ICV	PP24387	03/18/2025	08/22/2025	Yogesh Patel	None	None	Abdul Mirza
								04/03/2025

FROM 0.50000ml of W3177 + 0.50000ml of PP24386 = Final Quantity: 1.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
465	200 PPB Pest/PCB Surrogate Spike	PP24460	04/11/2025	10/03/2025	Abdul Mirza	None	None	Yogesh Patel
								04/16/2025

FROM 1.00000ml of P13355 + 999.00000ml of E3917 = Final Quantity: 1000.000 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3857	5000 PPB PCB SPIKE SOLUTION 2ND SOURCE	PP24461	04/11/2025	10/03/2025	Abdul Mirza	None	None	Yogesh Patel 04/16/2025
<u>FROM</u> 0.50000ml of P12955 + 99.50000ml of E3917 = Final Quantity: 100.000 ml								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-3382-05 / Sand, Purified (cs/4x2.5kg)	0000243821	06/30/2025	04/30/2020 / RAJESH	04/28/2020 / RAJESH	E2865

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	9005-05 / Acetone Ultra (cs/4x4L)	24E0761004	11/05/2025	10/01/2024 / Rajesh	09/25/2024 / Rajesh	E3804

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	08/12/2025	02/12/2025 / Rajesh	02/12/2025 / Rajesh	E3877

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	243570	10/03/2025	04/03/2025 / Rajesh	03/31/2025 / Rajesh	E3916

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	10/03/2025	04/03/2025 / Rajesh	03/31/2025 / Rajesh	E3917

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	25C0362005	10/22/2025	04/18/2025 / RUPESH	04/16/2025 / RUPESH	E3928

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	0000281827	06/02/2025	06/01/2022 /	04/05/2022 / william	M5173

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-382-1 / Aroclor 1268	0006587800	09/18/2025	03/18/2025 / yogesh	02/21/2022 / Ankita	P11522

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	91867 / Aroclor 1232 100 ug/mL	020823	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12699

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc	x9166 / Aroclor 1262 100 ug/mL	060523	09/18/2025	03/18/2025 / yogesh	08/07/2023 / Ankita	P12702

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32009 / PCB Mix, Aroclor 1242, 1000ug/mL, Hexane, 1mL/ampul	a0203672	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12931

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32010 / PCB Mix, Aroclor 1248, 1000ug/mL, Hexane, 1mL/ampul	a0202803	09/18/2025	03/18/2025 / yogesh	12/07/2023 / Ankita	P12936

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	09/18/2025	03/18/2025 / yogesh	12/20/2023 / Yogesh	P12949

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	20064 / Aroclor 1016/1260	022023	10/11/2025	04/11/2025 / Abdul	12/20/2023 / Yogesh	P12955

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Absolute Standards, Inc.	/ Arochlor 1254	121823	04/03/2025	10/03/2024 / Ankita	12/20/2023 / Yogesh	P12957

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	10/11/2025	04/11/2025 / Abdul	04/22/2024 / Abdul	P13355

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32000 / Pesticide Mix, CLP method, Pesticide Surrogate Mix, 200ug/mL, Acetone, 1mL	A0206810	09/18/2025	03/18/2025 / yogesh	04/22/2024 / Abdul	P13356

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-292-1 / Aroclor 1221	0006783205	09/18/2025	03/18/2025 / yogesh	05/02/2024 / Ankita	P13373

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32410 / PCB Stock Solution, Aroclor 1268 Std, 1mL, Hexane	A0207475	09/18/2025	03/18/2025 / yogesh	05/03/2024 / Abdul	P13381

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-312-1 / Aroclor 1242	0006665550	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13589

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Agilent Technologies	PP-342-1 / Aroclor 1248	0006726317	09/18/2025	03/18/2025 / yogesh	10/14/2024 / Ankita	P13591

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32039 / PCB Mix, Aroclor 1016/1260, 1000ug/mL, hexane, 1mL/ampul	A0210629	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13697

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32007 / PCB Mix, Aroclor 1221, 1000ug/mL, Hexane, 1mL/ampul	A0215270	09/18/2025	03/18/2025 / yogesh	10/17/2024 / yogesh	P13702

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32011 / PCB Mix, Aroclor 1254, 1000ug/mL, Hexane, 1mL/ampul	A0217391	09/18/2025	03/18/2025 / yogesh	12/09/2024 / Ankita	P13830

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32008 / PCB Mix, Aroclor 1232, 1000ug/mL, Hexane, 1mL/ampul	A0219655	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13878

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	32409 / PCB Stock Solution, Aroclor 1262 Std, 1mL, Hexane	A0220950	09/18/2025	03/18/2025 / yogesh	01/23/2025 / Ankita	P13883

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9262-03 / Hexane, Ultra-Resi (cs/4x4L)	24G1962003	08/22/2025	02/03/2025 / jignesh	01/31/2025 / jignesh	W3177

Sand
Purified
Washed and Ignited



Material No.: 3382-05
Batch No.: 0000243821
Manufactured Date: 2018/04/09
Retest Date: 2025/04/07
Revision No: 1


Certificate of Analysis

Test	Specification	Result
Substances Soluble in HCl	$\leq 0.16\%$	0.01

For Laboratory, Research or Manufacturing Use
Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US
Packaging Site: Paris Mfg Ctr & DC

E 2865


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3

Acetone
CMOS

avantor™



Material No.: 9005-05
Batch No.: 24E0761004
Manufactured Date: 2024-05-02
Retest Date: 2029-05-01
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	≥ 99.5 %	99.8 %
Color (APHA)	≤ 10	< 5
Residue after Evaporation	≤ 5 ppm	< 1 ppm
Titration Acid (μeq/g)	≤ 0.3	0.1
Titration Base (μeq/g)	≤ 0.5	0.1
Water (H ₂ O)	≤ 0.5 %	0.1 %
Solubility in H ₂ O	Passes Test	Passes Test
Chloride (Cl)	≤ 0.2 ppm	< 0.2 ppm
Phosphate (PO ₄)	≤ 0.05 ppm	< 0.05 ppm
Trace Impurities – Aluminum (Al)	≤ 50.0 ppb	< 5.0 ppb
Arsenic and Antimony (as As)	≤ 5.0 ppb	< 5.0 ppb
Trace Impurities – Barium (Ba)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Beryllium (Be)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Bismuth (Bi)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Boron (B)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Cadmium (Cd)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Calcium (Ca)	≤ 25.0 ppb	3.6 ppb
Trace Impurities – Chromium (Cr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Cobalt (Co)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Copper (Cu)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Gallium (Ga)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Germanium (Ge)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Gold (Au)	≤ 20 ppb	< 5 ppb
Trace Impurities – Iron (Fe)	≤ 20.0 ppb	< 1.0 ppb
Trace Impurities – Lead (Pb)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Lithium (Li)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Magnesium (Mg)	≤ 20 ppb	< 1 ppb
Trace Impurities – Manganese (Mn)	≤ 10.0 ppb	< 1.0 ppb

Recd by RP on 9/25/24

E 3804

>>> Continued on page 2 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
Trace Impurities – Molybdenum (Mo)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Nickel (Ni)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Niobium (Nb)	≤ 50.0 ppb	< 1.0 ppb
Trace Impurities – Potassium (K)	≤ 10.0 ppb	< 10.0 ppb
Trace Impurities – Silicon (Si)	≤ 50 ppb	< 10 ppb
Trace Impurities – Silver (Ag)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Sodium (Na)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Strontium (Sr)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Tantalum (Ta)	≤ 50.0 ppb	< 5.0 ppb
Trace Impurities – Thallium (Tl)	≤ 10.0 ppb	< 5.0 ppb
Trace Impurities – Tin (Sn)	≤ 20.0 ppb	< 10.0 ppb
Trace Impurities – Titanium (Ti)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Vanadium (V)	≤ 10.0 ppb	< 1.0 ppb
Trace Impurities – Zinc (Zn)	≤ 20.0 ppb	7.9 ppb
Trace Impurities – Zirconium (Zr)	≤ 10.0 ppb	< 1.0 ppb
Particle Count – 0.5 µm and greater (Rion KS42AF)	≤ 100 par/ml	8 par/ml
Particle Count – 1.0 µm and greater (Rion KS42AF)	≤ 8 par/ml	2 par/ml

>>> Continued on page 3 >>>

Acetone
CMOS

 **avantor™**



Material No.: 9005-05
Batch No.: 24E0761004

Test	Specification	Result
------	---------------	--------

For Microelectronic Use

Country of Origin: USA
Packaging Site: Paris Mfg Ctr & DC



Michelle Bales
Sr. Manager, Quality Assurance

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Recd. by RP on 2/12/25

 E3877

Harout Sahagian - Quality Control Manager - Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.

If there are any questions with this certificate, please call at (800) 227-6701.

*Based on suggested storage condition.

Certificate of Analysis

1 Reagent Lane
 Fair Lawn, NJ 07410
 201.796.7100 tel
 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System
 Standard ISO9001:2015 by SAI Global Certificate Number CERT - 0120633

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	H303	Quality Test / Release Date	11/07/2024
Lot Number	243570		
Description	HEXANES - OPTIMA		
Country of Origin	United States	Suggested Retest Date	Nov/2029
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

N/A

Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	Clear, colorless liquid
ASSAY (N-HEXANE)	%	>= 60	69
ASSAY (SUM C6 HYDROCARBONS)	%	>= 99.9	>99.9
COLOR	APHA	<= 5	<5
DENSITY AT 25 DEGREES C	GM/ML	Inclusive Between 0.653 - 0.673	0.669
EVAPORATION RESIDUE	ppm	<= 1	<1
FLUORESCENCE BACKGROUND	ppb	<= 1	<1
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
OPTICAL ABS AT 195 NM	ABS. UNITS	<= 1	0.74
OPTICAL ABS AT 210 NM	ABS. UNITS	<= 0.25	0.17
OPTICAL ABS AT 220 NM	ABS. UNITS	<= 0.07	0.05
OPTICAL ABS AT 254 NM	ABS. UNITS	<= 0.005	0.001
PESTICIDE RESIDUE ANALYSIS	NG/L	<= 10	<10
REFRACTIVE INDEX @ 25 DEG C		Inclusive Between 1.375 - 1.385	1.379
SUITABILITY FOR GC/MS		= PASS TEST	PASS TEST
SULFUR COMPOUNDS	%	<= 0.005	<0.005
THIOPHENE	PASS/FAIL	= PASS TEST	PASS TEST
WATER (H2O)	%	<= 0.01	<0.01
WATER-SOLUBLE TITRABLE ACID	MEQ/G	<= 0.0003	0.0001

Harout Sahagian

Harout Sahagian - Quality Control Manager - Fair Lawn

Recd by RP on 3/31/25

E 3946

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of this catalog number listed above.
 If there are any questions with this certificate, please call at (800) 227-6701.
 *Based on suggested storage condition.

Acetone

BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H2762008

Manufactured Date: 2024-04-18

Expiration Date: 2027-04-18

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

Recd by RP on 03/31/25

E3917

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis

 **avantors[™]**



Material No.: 9262-03

Batch No.: 25C0362005

Manufactured Date: 2025-01-29

Expiration Date: 2026-04-30

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	6
ECD-Sensitive Impurities (as EthyleneDibromide) - Single Impurity Peak (ng/mL)	≤ 5	5
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	100.0 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	100 %
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	$< 0.01 \%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3928



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials U.S.

Hydrochloric Acid, 36.5–38.0%
 BAKER INSTRA-ANALYZED® Reagent
 For Trace Metal Analysis



Material No.: 9530-33
 Batch No.: 0000281827
 Manufactured Date: 2021/03/30
 Retest Date: 2026/03/29
 Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid–base titrn)	36.5 – 38.0 %	37.6
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	1
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.189
ACS – Bromide (Br)	<= 0.005 %	< 0.005
ACS – Extractable Organic Substances	<= 5 ppm	< 1
ACS – Free Chlorine (as Cl ₂)	<= 0.5 ppm	< 0.5
Phosphate (PO ₄)	<= 0.05 ppm	< 0.03
Sulfate (SO ₄)	<= 0.5 ppm	< 0.3
Sulfite (SO ₃)	<= 0.8 ppm	0.3
Ammonium (NH ₄)	<= 3 ppm	< 1
Trace Impurities – Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities – Aluminum (Al)	<= 10.0 ppb	0.5
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities – Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities – Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities – Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities – Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities – Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities – Calcium (Ca)	<= 50.0 ppb	15.0
Trace Impurities – Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities – Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities – Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities – Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
 Avantor Performance Materials, LLC
 100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Test	Specification	Result
Trace Impurities – Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities – Gold (Au)	<= 4.0 ppb	3.0
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities – Iron (Fe)	<= 15.0 ppb	1.0
Trace Impurities – Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities – Lithium (Li)	<= 1.0 ppb	< 0.2
Trace Impurities – Magnesium (Mg)	<= 10.0 ppb	< 0.4
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	0.2
Trace Impurities – Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities – Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities – Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities – Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities – Selenium (Se), For Information Only	ppb	1.0
Trace Impurities – Silicon (Si)	<= 100.0 ppb	18.0
Trace Impurities – Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities – Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities – Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities – Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities – Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities – Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities – Titanium (Ti)	<= 1.0 ppb	< 0.2
Trace Impurities – Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities – Zinc (Zn)	<= 5.0 ppb	0.4
Trace Impurities – Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use

Product Information (not specifications):

Appearance (clear, fuming liquid)

Meets ACS Specifications

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality



Certificate of Analysis

P11518
↓
P11522
AJ
02/21/22

Product Name: Aroclor 1268 Standard

Product Number: PP-382-1

Lot Issue Date: 09-Feb-2021

Lot Number: 0006587800

Expiration Date: 31-Mar-2029

Description:

This analytical reference material (RM) was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	Concentration ± Uncertainty
Aroclor 1268	011100-14-4	RM00937	100.0 ± 0.5 µg/mL

Matrix: isooctane (2,2,4-trimethylpentane)

Storage Conditions: Store at Room Temperature (15° to 30°C).

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This RM was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Intended Use:

This RM is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Hazards:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this RM.


Expiration of Certification:

The certification of this RM is valid until the expiration date specified above, provided the RM is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the RM is damaged, contaminated, or otherwise modified.

Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with TUV USA Inc registered ISO 9001 Quality Management System. Cert # 56 100 18560026
Page: 1 of 1

www.agilent.com/quality/
CSD-QA-015.1



ISO 17025 Cert
No. AT-1937



CERTIFIED WEIGHT REPORT

Part Number: **91867**
 Lot Number: **020823**
 Description: **WP 037 - Aroclor 1232**
PCB Technical Mixture
 Expiration Date: **020833**
 Recommended Storage: **Ambient (20 °C)**
 Nominal Concentration ($\mu\text{g/mL}$): **100**
 NIST Test ID#: **6UTB**

Solvent(
Aceton


Weight(s) shown below were combined and diluted to (mL): **100.0**
 5E-05 Balance Uncertainty
 0.057 Flask Uncertainty

Compound	RM#	Lot Number	Nominal Conc ($\mu\text{g/mL}$)	Purity (%)	Uncertainty Purity	Target Weight (g)
1. Aroclor 1232	17	45-6A	100	100	0.5	0.01000

- The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
- Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
- Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
- All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
- Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measure Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments

GC3-M1 Analysis by Melissa Stonier
 Column ID SPB-608 30 meter X 0.53mm X 5 μm film thickness
 Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
 Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
 Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 290°C (Time 2 = 13.5 min)
 Rate = 8°C/min, Total run time = 35 min
 Injector temp. = 200°C, FID Temp. = 300°C. FID Signal = Edaq Channel 1
 Standard injection = 1.5 μL , Range=3






110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32009 **Lot No.:** A0203672

Description: Aroclor® 1242 Standard

Aroclor® 1242 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

p12928

→

p12932

AJ

12/27/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty* (95% C.L.; K=2)
1	Aroclor 1242	53469-21-9	01141	—%	1,004.7 µg/mL	+/- 55.7515

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Russ Boothamer

Russ Boothamer - Operations Technician I

Date Mixed: 26-Oct-2023

Balance Serial #

B442140311

Jennifer Polino

Jennifer Polino - Operations Tech III - ARM QC

Date Passed: 06-Nov-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32010 **Lot No.:** A0202803

Description: Aroclor® 1248 Standard

Aroclor® 1248 Standard 1,000µg/mL, Hexane, 1mL/ampul

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: January 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P1219697
P1219697
AF
12/10/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1248	12672-29-6	13897600	---%	1,001.7 µg/mL	+/- 55.5850

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

0.2ul



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Laith Clemente
Laith Clemente - Operations Technician I

Date Mixed: 03-Oct-2023

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 09-Oct-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FW 80397



Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty

Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

Solvent(s):
Hexane

Lot#
273615

Formulated By:
Benson Chan

Reviewed By:
Pedro L. Rentas

DATE
022023

DATE
022023

5E-05 Balance Uncertainty
0.010 Flask Uncertainty


Nominal Conc (µg/mL):
200.0

Weight(s) shown below were combined and diluted to (mL):

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11086-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





Certified Reference Material CRM

CERTIFIED WEIGHT REPORT

Part Number:
20064

Lot Number:
022023

Description:
CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date:
022033

Recommended Storage:
Ambient (20 °C)

Nominal Concentration (µg/mL):
1000

NIST Test ID#:
6UTB

20064

022023

CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

022033

Ambient (20 °C)

1000

6UTB

Formulated By: Benson Chan

Reviewed By: Pedro L. Rentas

DATE: 022023

DATE: 022023

5E-05 Balance Uncertainty

0.010 Flask Uncertainty

200.0

200.0

Weight(s) shown below were combined and diluted to (mL):

200.0

Part Number: 20064

Lot Number: 022023

Description: CLP PCB'S - Aroclor Mix
Aroclors 1016 & 1260

Expiration Date: 022033

Recommended Storage: Ambient (20 °C)

Nominal Concentration (µg/mL): 1000

NIST Test ID#: 6UTB

Formulated By: Benson Chan

Reviewed By: Pedro L. Rentas

DATE: 022023

DATE: 022023

5E-05 Balance Uncertainty

0.010 Flask Uncertainty

200.0

200.0


Weight(s) shown below were combined and diluted to (mL):

200.0

Compound	RM#	Lot Number	Nominal Conc (µg/mL)	Purity (%)	Purity Uncertainty	Target Weight(g)	Actual Weight(g)	Actual Conc (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	CAS#	OSHA PEL (TWA)	LD50
1. Aroclor 1016	15	020491JC	1000	100	0.2	0.20004	0.20060	1002.8	4.0	12674-11-2	N/A	N/A
2. Aroclor 1260	21	020491JC	1000	100	0.2	0.20004	0.20081	1003.9	4.0	11086-82-5	0.5mg/m3	orl-rat 1315mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stortier
Column ID SPB-608 30 meter X 0.53mm X5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150°C (Time 1 = 4 min), Temp 2 = 280°C (Time 2 = 13.5 min)
Rate = 8°C/min, Total run time = 35 min
Injector temp. = 200°C, FID Temp. = 300°C, FID Signal = Etdaq Channel 1
Standard Injection = 1.5µL, Range=3





CERTIFIED WEIGHT REPORT

Part Number: 99139
Lot Number: 121823
Description: Aroclor 1254

Expiration Date: 121833
Recommended Storage: Ambient (20 °C)
Nominal Concentration (µg/mL): 100
NIST Test ID#: 6UTB

Volume(s) shown below were combined and diluted to (mL): 20.0
Note: Aroclor 1254 is a mix of isomers.

Solvent(s): Iso-octane
Lot#: 82227

Formulated By: Anthony Mahoney	121823	DATE
Reviewed By: Pedro L. Rentas	121823	DATE


P12956 / Y.P.
12/18/23
P12957

Compound	Part Number	Lot Number	Dilution Factor	Initial Vol. (mL)	Initial Uncertainty	Pipette (mL)	Conc. (µg/mL)	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	SDS Information		
											(Solvent Safety Info. On Attached pg.)	CAS#	LD50
1. Aroclor 1254	79100	121823	0.10	2.00	0.017	0.017	1003.3	100.1	1.8	11097-69-1	0.5mg/m3 (skin)		or-rat 1295mg/kg

* The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise stated.
* Standards are prepared gravimetrically using balances that are calibrated with weights traceable to NIST (see above).
* Standards are certified (+/-) 0.5% of the stated value, unless otherwise stated.
* All Standards, after opening ampule, should be stored with caps tight and under appropriate laboratory conditions.
* Uncertainty Reference: Taylor, B.N. and Kuyat, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. Government Printing Office, Washington, DC, (1994).

Comments
GC3-M1 Analysis by Melissa Stonier
Column ID SPB-608 30 meter X 0.53mm X 5µm film thickness
Flow rates: Helium (carrier) = 5mL/min, Helium (make-up) = 25mL/min
Hydrogen (make-up) = 30mL/min, Air (make-up) = 350mL/min
Oven Profile: Temp 1 = 150 °C (Time 1 = 4 min), Temp 2 = 280 °C (Time 2 = 13.5 min)
Rate = 8 °C/min, Total run time = 35 min
Injector temp. = 200 °C, FID Temp. = 300 °C, FID Signal = EDAQ Channel 1
Standard injection = 1.5µL, Range=3

Peak No.	Name	FID RT (min.)
1	Total Aroclor 1254	18.12





110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32000 **Lot No.:** A0206810

Description : Pesticide Surrogate Mix
Pesticide Surrogate Mix 200 µg/mL, Acetone, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Contains PCBs - sonicate prior to use. **Ship:** Ambient

P13348
↓
P13357
10
DAUF
04/25/2024

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2,4,5,6-Tetrachloro-m-xylene	877-09-8	RP220407	99%	200.3 µg/mL	+/- 11.1143
2	Decachlorobiphenyl (BZ# 209)	2051-24-3	30638	99%	200.6 µg/mL	+/- 11.1298

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Acetone
CAS # 67-64-1
Purity 99%

Tech Tips:

Decachlorobiphenyl has poor solubility in most organic solvents. The maximum concentration that can be prepared in acetone, hexane, or isooctane is 200µg/mL. Temperature will affect the solubility as well. Storing solutions at reduced temperatures will cause decachlorobiphenyl to precipitate.

Products containing decachlorobiphenyl must be sonicated for a minimum of 10 minutes prior to opening the ampul. Because each ultrasonic bath operates at a different energy level, 10 minutes is a guideline only. Longer sonication time will not affect product quality.

These precautions apply to working solutions prepared in your laboratory as well. The amount of compound that precipitates depends on concentration AND temperature. If you store your standards at a temperature lower than 4°C (even dilute solutions), allow extra sonication time.

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Jan-2024


Balance Serial # 1128360905


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 24-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P 13348
↓
P 13357
10


SAUF
04/25/2025



ISO 17034

Reference Material Certificate

Product Information Sheet

Product Name: Aroclor 1221 Standard
Product Number: PP-292-1
Storage Conditions: Store at Room Temperature (15° to 30°C).

Lot Number: 0006783205
Lot Issue Date: 20-Feb-2024
Expiration Date: 31-Mar-2032

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1221	100.3 ±	0.5 µg/mL	011104-28-2	NT01017

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13342
AJ
05106124
P13343

Page: 1 of 2

CSD-QA-015.2

ISO 17025
Cert No. AT-1937


250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:


Monica Bourgeois
QMS Representative



ISO 17034
Cert No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/
CSD-QA-015.2

ISO 17025
Cert No. AT-1937

250 Smith Street North Kingstown, Rhode Island 02852 www.agilent.com/quality



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32410 **Lot No.:** A0207475

Description : Aroclor® 1268 Standard

Aroclor® 1268 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : May 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1268	11100-14-4	10947000	---%	1,000.0 µg/mL	+/- 55.4925

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P 13380
↓
P 13381
↓
SAZU
05/6/2024

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

Split ratio 500:1

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 06-Feb-2024

Balance Serial # B442140311

Dillon Murphy
Dillon Murphy - Operations Technician I


Date Passed: 09-Feb-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

P13380 } ②
↓
P13381 }
1

[Signature]
05/6/2024

ISO 17034



Agilent

Trusted Answers

Reference Material Certificate

Product Name: Aroclor 1242 Standard

Lot Number: 0006665550

Product Number: PP-312-1

Lot Issue Date: 08-Feb-2022

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 31-Jan-2027

Component Name	CERTIFIED VALUES		CAS#	Analyte Lot
	Concentration	Expanded Uncertainty		
Aroclor 1242	100.4	± 0.5 µg/mL	053469-21-9	NT01020

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13589
↓
P13590AJ
10/14/24

ISO 17034



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO 9001:2015
Quality Management System. Cert# 95T215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1



ISO 17025
Cert No. AT-

Reference Material Certificate
Product Information Sheet

Product Name: Aroclor 1248 Standard

Lot Number: 0006726317

Product Number: PP-342-1

Lot Issue Date: 27-Jan-2023

Storage Conditions: Store at Room Temperature (15° to 30°C).

Expiration Date: 28-Feb-2031

Component Name	Concentration	Uncertainty	CAS#	Analyte Lot
Aroclor 1248	100.3 ±	0.5 µg/mL	012672-29-6	NT01582

Matrix: isooctane (2,2,4-trimethylpentane)

Description:

This document is prepared in accordance with ISO 17034 and Guide 31. This analytical reference material (RM) standard was manufactured and verified in accordance with an ISO 9001 registered quality system and analyte concentrations were verified by an ISO 17025 accredited laboratory. The concentration and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed above. Purity values are taken from approved vendor raw material certificates.

Traceability:

The balances used for these measurements are calibrated with weights traceable to NIST in compliance with ANSI/NCSL Z540.3, ISO 9001, ISO 17025, and ISO 17034. Calibrated Class A glassware is used for volumetric measurements. Thermometers are calibrated against a NIST traceable thermometer in accordance with NIST Special Publication 1088.

Homogeneity:

This analytical reference (RM) standard was unitized according to an in-house procedure and is guaranteed to be homogeneous. There is no minimum sub-sample size required.

Instructions for Use:

Sample aliquots for analysis should be withdrawn at 20°C to 25°C immediately after opening the container and should be processed without delay for the certified values to be valid within the stated uncertainties.

Safety:

Refer to the Safety Data Sheet on www.agilent.com for information regarding this analytical reference material.

Intended Use:

This analytical reference (RM) standard is intended for the preparation of working reference samples for use in routine laboratory analyses, calibration of instruments, validation of analytical methods, assessments of measurement methods, and continuing calibration verification.

Expiration of Certification:

The certification of this analytical reference standard (RM) is valid until the expiration date specified above, provided the material is handled and stored in accordance with the instructions given in this certificate. This certification is nullified if the material is damaged, contaminated, or otherwise modified.

P13591
↓
P13592

AS
10/14/2024



Maintenance of Certification:

If substantive changes are noted that affect the certification before the expiration of this certificate, Agilent will notify the purchaser.

Sample lot approver:

Monica Bourgeois
QMS Representative



ISO 17034 Cert
No. AR-1936

RM was produced in accordance with the TUV/SUD registered ISO
9001:2015 Quality Management System. Cert# 951215321

Page: 2 of 2

www.agilent.com/quality/

CSD-QA-015.1

ISO 17025



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32039 **Lot No.:** A0210629

Description : Aroclor® 1016/1260 Mix
Aroclor® 1016/1260 Mix 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13697
↓
P13701 } Y.P.
10/19/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1016	12674-11-2	07	----%	1,005.3 µg/mL	+/- 55.7809
2	Aroclor 1260	11096-82-5	1320657	----%	1,000.0 µg/mL	+/- 55.4850

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:


10 ml/min.

Inj. Vol

0.2µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Laith Clemente - Operations Technician I

Date Mixed: 22-Apr-2024

Balance Serial # B442140311


Dillan Murphy - Operations Technician I

Date Passed: 24-Apr-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32007 **Lot No.:** A0215270

Description : Aroclor® 1221 Standard

Aroclor® 1221 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2030 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

P13902
P13703 } Y.P.
10/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1221	11104-28-2	14969200	----%	1,005.0 µg/mL	+/- 55.7700

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 16-Aug-2024

Balance Serial # 1128360905

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 20-Aug-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32011 **Lot No.:** A0217391

Description : Aroclor® 1254 Standard

Aroclor® 1254 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : January 31, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1254	11097-69-1	124-191-B	----%	1,004.7 µg/mL	+/- 55.7515

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Hexane
CAS # 110-54-3
Purity 99%

P13830
↓
P13832
AJ
12/09/24

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye
Michael Maye - Operations Tech I

Date Mixed: 02-Oct-2024

Balance Serial # C322230531

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 07-Oct-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 32008

Lot No.: A0219655

Description : Aroclor® 1232 Standard

Aroclor® 1232 Standard 1,000 µg/mL, Hexane, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : March 31, 2031

Storage: 25°C nominal

Handling:

This product contains PCBs.

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1232	11141-16-5	15665-01	---%	1,007.0 µg/mL	+/- 55.8810

Solvent: Hexane

CAS # 110-54-3

Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13878



AJ
01/28/25

P13880

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Michael Maye

Michael Maye - Operations Tech I

Date Mixed: 02-Dec-2024

Balance Serial #

C322230531

Brittany Federhik

Brittany Federhiko - Operations Tech I

Date Passed: 05-Dec-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397




110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No.: 32409 **Lot No.:** A0220950

Description: Aroclor® 1262 Standard

Aroclor® 1262 Standard 1,000 µg/mL, 1mL/ampul, Hexane

Container Size: 2 mL **Pkg Amt:** > 1 mL

Expiration Date: April 30, 2031 **Storage:** 25°C nominal

Handling: This product contains PCBs. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Aroclor 1262	37324-23-5	10849100	----%	1,002.0 µg/mL	+/- 55.6035

Solvent: Hexane

CAS # 110-54-3
Purity 99%

* Expanded Uncertainty displayed in same units as Grav. Conc.

P13882



AJ
01/28/25

P13883

Quality Confirmation Test

Column:

30m x .25mm x .2um
Rtx-CLP II (cat.# 11323)

Carrier Gas:

helium-constant pressure 20 psi.

Temp. Program:

200°C to 300°C
@ 25°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

300°C

Det. Type:


ECD

Split Vent:

300 ml/min.

Inj. Vol

0.2µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Tom Suckal - Mix Technician

Date Mixed: 09-Jan-2025 **Balance Serial #** C322230531


Britiany Federenko - Operations Tech I

Date Passed: 14-Jan-2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

n-Hexane 95%
ULTRA RESI-ANALYZED
For Organic Residue Analysis



Material No.: 9262-03
Batch No.: 24G1962003
Manufactured Date: 2024-05-23
Expiration Date: 2025-08-22
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	3
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	≤ 10	1
ECD-Sensitive Impurities (as Ethylene Dibromide) - Single Impurity Peak (ng/mL)	≤ 5	1
Assay (Total Saturated C ₆ Isomers) (by GC, corrected for water)	$\geq 99.5 \%$	99.7 %
Assay (as n-Hexane) (by GC, corrected for water)	$\geq 95 \%$	98 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.1 ppm
Substances Darkened by H ₂ SO ₄	Passes Test	Passes Test
Water (by KF, coulometric)	$\leq 0.05 \%$	< 0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: USA
Packaging Site: Phillipsburg Mfg Ctr & DC

Jamie Croak
Director Quality Operations, Bioscience Production



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Kleinfelder
ADDRESS: 180 Sheree Blvd Suite 3800
CITY: Exton STATE: PA ZIP: 19341
ATTENTION: Mark Warchol
PHONE: 484-883-3892 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Mitchell School
PROJECT NO.: 4005164.001A LOCATION: Philadelphia, PA
PROJECT MANAGER: Mark Warchol
e-mail: mwarchol@kleinfelder.com
PHONE: 484-883-3892 FAX:

CLIENT BILLING INFORMATION

BILL TO: PO#: Same
ADDRESS: Same
CITY: STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) 5 DAYS*
HARDCOPY (DATA PACKAGE): 5 DAYS*
EDD: 5 DAYS*

*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☒ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

1	2	3	4	5	6	7	8	9
PADEP Historic Parameters								

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		E	E								← Specify Preservatives	
1.	COMP-1	Soil	✓		4/24/25	10:15	4	✓									A-HCl	D-NaOH
2.	COMP-2		↓			10:45	↓	↓									B-HNO3	E-ICE
3.	COMP-3		↓			11:25	↓	↓									C-H2SO4	F-OTHER
4.	SB-1			✓		9:45	1		✓									
5.	SB-2					9:55	↓		↓									
6.	SB-3					10:00	↓		↓									
7.	SB-4					10:10	↓		↓									
8.	SB-5					10:18	↓		↓									
9.	SB-6					10:30	↓		↓									
10.	SB-7					10:35	↓		↓									

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP
1. <u>[Signature]</u>	<u>4/24/25 13:00</u>	1. <u>[Signature]</u>	<u>4.7°C</u>
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	Comments: <u>Hold grab samples SB-1 through SB-12</u>
2. <u>FedEx</u>	<u>4-25-25 1045</u>	2. <u>[Signature]</u>	<u>Adjust Factor +1</u>
RELINQUISHED BY SAMPLER:	DATE/TIME:	RECEIVED BY:	
3. <u>[Signature]</u>		3. <u>[Signature]</u>	<u>ILQW #1</u>
Page <u>1</u> of <u>2</u>			CLIENT: ☐ Hand Delivered ☒ Other
			Shipment Complete ☐ YES ☐ NO

CLIENT INFORMATION

CLIENT PROJECT INFORMATION

CLIENT BILLING INFORMATION

REPORT TO BE SENT TO:

COMPANY: Kleinfelder
ADDRESS: 180 Sheree Blvd Suite 3800
CITY: Exton STATE: PA ZIP: 19341
ATTENTION: Mark Warchol
PHONE: 484-883-3892 FAX:

PROJECT NAME: Mitchell School
PROJECT NO.: 4005164.001A LOCATION: Philadelphia, PA
PROJECT MANAGER: Mark Warchol
e-mail: m.warchol@kleinfelder.com
PHONE: 484-883-3892 FAX:

BILL TO: PO#: Same
ADDRESS: Same
CITY: STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) 5 DAYS*
HARDCOPY (DATA PACKAGE): 5 DAYS*
EOD: 5 DAYS*

*TO BE APPROVED BY CHEMTECH

STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☐ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☒ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other
☐ EDD FORMAT

1	2	3	4	5	6	7	8	9

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER	
1.	SB-8	Soil		✓	4/24/25	10:40	1	✓										
2.	SB-9	↓		↓	↓	11:00	↓	↓										
3.	SB-10	↓		↓	↓	11:05	↓	↓										
4.	SB-11	↓		↓	↓	11:15	↓	↓										
5.	SB-12	↓		↓	↓	11:20	↓	↓										
6.																		
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>[Signature]</u>	DATE/TIME: <u>4/24/25 13:00</u>	RECEIVED BY: 1. <u>[Signature]</u>	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>47°C</u>
RELINQUISHED BY SAMPLER: 2. <u>FedEx</u>	DATE/TIME: <u>4-25-25 1045</u>	RECEIVED BY: 2. <u>[Signature]</u>	Comments: <u>Adjust Factor +1</u> <u>Julian #1</u>
RELINQUISHED BY SAMPLER: 3.	DATE/TIME:	RECEIVED BY: 3.	Page <u>2</u> of <u>2</u> CLIENT: ☐ Hand Delivered ☒ Other <u>FedEx</u> Shipment Complete ☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1889	POWE02	Order Date : 4/25/2025 11:06:00 AM	Project Mgr :
Client Name : Kleinfelder		Project Name : Mitchell School Lincoln High School	Report Type : Results+QC
Client Contact : Mark Warchol		Receive DateTime : 4/25/2025 10:45:00 AM	EDD Type : EXCEL NOCLEANUP
Invoice Name : Kleinfelder		Purchase Order :	Hard Copy Date :
Invoice Contact : Mark Warchol			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1889-01	COMP-1	Solid	04/24/2025	10:15					
					VOCMS Group1		8260D	5 Bus. Days	
Q1889-02	COMP-2	Solid	04/24/2025	10:45					
					VOCMS Group1		8260D	5 Bus. Days	
Q1889-03	COMP-3	Solid	04/24/2025	11:25					
					VOCMS Group1		8260D	5 Bus. Days	

Relinquished By :

Date / Time : 4/25/25 11:34

Received By :

Date / Time : 04/25/25 11:34

Storage Area : VOA Refridgerator Room