

Surrogate Summary

SDG No.: **Q1889**

Client: **Kleinfelder**

Analytical Method: **8082A**

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PO110348.D	PIBLK-PO110348.D	Tetrachloro-m-xylene	1	20	18.1	91		60	140
		Decachlorobiphenyl	1	20	19.2	96		60	140
		Tetrachloro-m-xylene	2	20	18.2	91		60	140
		Decachlorobiphenyl	2	20	20.2	101		60	140
I.BLK-PO110850.D	PIBLK-PO110850.D	Tetrachloro-m-xylene	1	20	18.0	90		60	140
		Decachlorobiphenyl	1	20	16.9	85		60	140
		Tetrachloro-m-xylene	2	20	16.6	83		60	140
		Decachlorobiphenyl	2	20	16.7	84		60	140
Q1889-01	COMP-1	Tetrachloro-m-xylene	1	20	21.5	108		32	144
		Decachlorobiphenyl	1	20	16.8	84		32	175
		Tetrachloro-m-xylene	2	20	19.7	99		32	144
		Decachlorobiphenyl	2	20	16.5	83		32	175
Q1889-01MS	COMP-1MS	Tetrachloro-m-xylene	1	20	19.4	97		32	144
		Decachlorobiphenyl	1	20	16.4	82		32	175
		Tetrachloro-m-xylene	2	20	17.8	89		32	144
		Decachlorobiphenyl	2	20	16.1	81		32	175
Q1889-01MSD	COMP-1MSD	Tetrachloro-m-xylene	1	20	18.9	95		32	144
		Decachlorobiphenyl	1	20	16.6	83		32	175
		Tetrachloro-m-xylene	2	20	17.2	86		32	144
		Decachlorobiphenyl	2	20	16.3	82		32	175
I.BLK-PO110865.D	PIBLK-PO110865.D	Tetrachloro-m-xylene	1	20	18.4	92		60	140
		Decachlorobiphenyl	1	20	17.5	88		60	140
		Tetrachloro-m-xylene	2	20	16.9	84		60	140
		Decachlorobiphenyl	2	20	16.9	85		60	140
I.BLK-PP071388.D	PIBLK-PP071388.D	Tetrachloro-m-xylene	1	20	17.1	85		60	140
		Decachlorobiphenyl	1	20	17.6	88		60	140
		Tetrachloro-m-xylene	2	20	17.2	86		60	140
		Decachlorobiphenyl	2	20	17.6	88		60	140
I.BLK-PP071561.D	PIBLK-PP071561.D	Tetrachloro-m-xylene	1	20	15.5	78		60	140
		Decachlorobiphenyl	1	20	16.3	81		60	140
		Tetrachloro-m-xylene	2	20	16.4	82		60	140
		Decachlorobiphenyl	2	20	17.4	87		60	140
PB167765BL	PB167765BL	Tetrachloro-m-xylene	1	20	18.4	92		32	144
		Decachlorobiphenyl	1	20	18.9	95		32	175
		Tetrachloro-m-xylene	2	20	20.5	103		32	144
		Decachlorobiphenyl	2	20	19.9	100		32	175
PB167765BS	PB167765BS	Tetrachloro-m-xylene	1	20	18.1	91		32	144
		Decachlorobiphenyl	1	20	18.7	94		32	175
		Tetrachloro-m-xylene	2	20	17.7	89		32	144
		Decachlorobiphenyl	2	20	19.8	99		32	175
Q1889-02	COMP-2	Tetrachloro-m-xylene	1	20	19.8	99		32	144

Surrogate Summary

SDG No.: Q1889

Client: Kleinfelder

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
Q1889-02	COMP-2	Decachlorobiphenyl	1	20	17.9	90		32	175
		Tetrachloro-m-xylene	2	20	22.5	113		32	144
I.BLK-PP071576.D	PIBLK-PP071576.D	Decachlorobiphenyl	2	20	18.8	94		32	175
		Tetrachloro-m-xylene	1	20	15.8	79		60	140
		Decachlorobiphenyl	1	20	16.4	82		60	140
		Tetrachloro-m-xylene	2	20	16.9	84		60	140
		Decachlorobiphenyl	2	20	17.5	87		60	140
Q1889-03	COMP-3	Tetrachloro-m-xylene	1	20	19.6	98		32	144
		Decachlorobiphenyl	1	20	19.0	95		32	175
		Tetrachloro-m-xylene	2	20	21.8	109		32	144
		Decachlorobiphenyl	2	20	19.8	99		32	175
I.BLK-PP071591.D	PIBLK-PP071591.D	Tetrachloro-m-xylene	1	20	16.1	80		60	140
		Decachlorobiphenyl	1	20	16.7	83		60	140
		Tetrachloro-m-xylene	2	20	16.8	84		60	140
		Decachlorobiphenyl	2	20	17.9	89		60	140