

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/23/2025 04/23/2025

Continuing Calib Time: 22:11 Initial Calibration Time(s): 11:31 12:26

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	9.11	9.12	9.02	9.22	0.01
Tetrachloro-m-xylene	3.59	3.60	3.50	3.70	0.01
Aldrin	5.31	5.32	5.22	5.42	0.01
Dieldrin	6.39	6.40	6.30	6.50	0.01
4,4'-DDE	6.24	6.24	6.14	6.34	0.00
4,4'-DDD	6.75	6.75	6.65	6.85	0.00
4,4'-DDT	7.06	7.07	6.97	7.17	0.01

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

Continuing Calib Date: 04/28/2025 Initial Calibration Date(s): 04/23/2025 04/23/2025

Continuing Calib Time: 22:11 Initial Calibration Time(s): 11:31 12:26

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.08	8.09	7.99	8.19	0.01
Tetrachloro-m-xylene	2.90	2.91	2.81	3.01	0.01
Aldrin	4.38	4.39	4.29	4.49	0.01
Dieldrin	5.53	5.53	5.43	5.63	0.00
4,4'-DDE	5.39	5.40	5.30	5.50	0.01
4,4'-DDD	5.94	5.95	5.85	6.05	0.01
4,4'-DDT	6.20	6.20	6.10	6.30	0.00