

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL01 Date Analyzed: 04/28/2025

Lab Sample No.: PSTDCCC050 Data File : PL095440.D Time Analyzed: 16:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.748	6.653	6.853	58.710	50.000	17.4
4,4'-DDE	6.239	6.143	6.343	56.090	50.000	12.2
4,4'-DDT	7.064	6.969	7.169	55.970	50.000	11.9
Aldrin	5.314	5.218	5.418	52.060	50.000	4.1
Decachlorobiphenyl	9.116	9.022	9.222	55.670	50.000	11.3
Dieldrin	6.391	6.296	6.496	54.260	50.000	8.5
Tetrachloro-m-xylene	3.593	3.495	3.695	51.230	50.000	2.5

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM **Case No.:** Q1889 **SAS No.:** Q1889 **SDG NO.:** Q1889

GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 04/23/2025 04/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 04/28/2025

Lab Sample No.: PSTDCCC050 **Data File :** PL095440.D **Time Analyzed:** 16:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.944	5.849	6.049	50.710	50.000	1.4
4,4'-DDE	5.392	5.297	5.497	47.450	50.000	-5.1
4,4'-DDT	6.198	6.104	6.304	42.540	50.000	-14.9
Aldrin	4.386	4.290	4.490	49.990	50.000	0.0
Decachlorobiphenyl	8.083	7.989	8.189	54.680	50.000	9.4
Dieldrin	5.527	5.432	5.632	48.720	50.000	-2.6
Tetrachloro-m-xylene	2.904	2.806	3.006	50.270	50.000	0.5