

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL02 Date Analyzed: 04/28/2025

Lab Sample No.: PSTDCCC050 Data File : PL095451.D Time Analyzed: 18:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.748	6.653	6.853	58.810	50.000	17.6
4,4'-DDE	6.238	6.143	6.343	54.780	50.000	9.6
4,4'-DDT	7.063	6.969	7.169	54.390	50.000	8.8
Aldrin	5.314	5.218	5.418	53.180	50.000	6.4
Decachlorobiphenyl	9.114	9.022	9.222	54.790	50.000	9.6
Dieldrin	6.391	6.296	6.496	55.290	50.000	10.6
Tetrachloro-m-xylene	3.592	3.495	3.695	52.860	50.000	5.7

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL02 Date Analyzed: 04/28/2025

Lab Sample No.: PSTDCCC050 Data File : PL095451.D Time Analyzed: 18:58

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.943	5.849	6.049	55.510	50.000	11.0
4,4'-DDE	5.391	5.297	5.497	53.150	50.000	6.3
4,4'-DDT	6.197	6.104	6.304	51.870	50.000	3.7
Aldrin	4.386	4.290	4.490	54.570	50.000	9.1
Decachlorobiphenyl	8.082	7.989	8.189	56.620	50.000	13.2
Dieldrin	5.526	5.432	5.632	53.280	50.000	6.6
Tetrachloro-m-xylene	2.904	2.806	3.006	54.290	50.000	8.6