

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL03 Date Analyzed: 04/29/2025

Lab Sample No.: PSTDCCC050 Data File : PL095475.D Time Analyzed: 13:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.750	6.653	6.853	59.620	50.000	19.2
4,4'-DDE	6.240	6.143	6.343	54.050	50.000	8.1
4,4'-DDT	7.065	6.969	7.169	52.560	50.000	5.1
Aldrin	5.316	5.218	5.418	52.020	50.000	4.0
Decachlorobiphenyl	9.118	9.022	9.222	53.870	50.000	7.7
Dieldrin	6.393	6.296	6.496	55.030	50.000	10.1
Tetrachloro-m-xylene	3.594	3.495	3.695	51.720	50.000	3.4

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL03 Date Analyzed: 04/29/2025

Lab Sample No.: PSTDCCC050 Data File : PL095475.D Time Analyzed: 13:07

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.942	5.849	6.049	51.510	50.000	3.0
4,4'-DDE	5.390	5.297	5.497	48.410	50.000	-3.2
4,4'-DDT	6.196	6.104	6.304	45.710	50.000	-8.6
Aldrin	4.385	4.290	4.490	49.140	50.000	-1.7
Decachlorobiphenyl	8.081	7.989	8.189	53.560	50.000	7.1
Dieldrin	5.525	5.432	5.632	48.570	50.000	-2.9
Tetrachloro-m-xylene	2.903	2.806	3.006	49.110	50.000	-1.8