

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL04 Date Analyzed: 04/29/2025

Lab Sample No.: PSTDCCC050 Data File : PL095484.D Time Analyzed: 17:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.753	6.653	6.853	53.600	50.000	7.2
4,4'-DDE	6.243	6.143	6.343	46.370	50.000	-7.3
4,4'-DDT	7.068	6.969	7.169	43.530	50.000	-12.9
Aldrin	5.318	5.218	5.418	44.960	50.000	-10.1
Decachlorobiphenyl	9.122	9.022	9.222	46.790	50.000	-6.4
Dieldrin	6.396	6.296	6.496	47.690	50.000	-4.6
Tetrachloro-m-xylene	3.596	3.495	3.695	44.610	50.000	-10.8

CALIBRATION VERIFICATION SUMMARY

Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG NO.: Q1889

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 04/23/2025 04/23/2025

Client Sample No.: CCAL04 Date Analyzed: 04/29/2025

Lab Sample No.: PSTDCCC050 Data File : PL095484.D Time Analyzed: 17:42

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.946	5.849	6.049	52.250	50.000	4.5
4,4'-DDE	5.392	5.297	5.497	47.760	50.000	-4.5
4,4'-DDT	6.200	6.104	6.304	42.790	50.000	-14.4
Aldrin	4.386	4.290	4.490	51.290	50.000	2.6
Decachlorobiphenyl	8.084	7.989	8.189	45.080	50.000	-9.8
Dieldrin	5.529	5.432	5.632	48.760	50.000	-2.5
Tetrachloro-m-xylene	2.906	2.806	3.006	50.780	50.000	1.6