

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: POWE02

Lab Code: CHEM

SAS No.: Q1889

SDG NO.: Q1889

Lab File ID: BP024434.D

DFTPP Injection Date: 04/28/2025

Instrument ID: BNA_P

DFTPP Injection Time: 12:53

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	17.1
68	Less than 2.0% of mass 69	0.4 (1.9) 1
69	Mass 69 relative abundance	19.4
70	Less than 2.0% of mass 69	0.1 (0.5) 1
127	10.0 - 80.0% of mass 198	29.6
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	4.5
275	10.0 - 60.0% of mass 198	23
365	Greater than 1% of mass 198	3.7
441	Present, but less than mass 443	15.5
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.1 (19.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BP024435.D	04/28/2025	13:34
COMP-1	Q1889-01	BP024441.D	04/28/2025	17:43
COMP-3	Q1889-03	BP024442.D	04/28/2025	18:24