

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: POWE02

Lab Code: CHEM Case No.: Q1889 SAS No.: Q1889 SDG No.: Q1889

Instrument ID: MSVOA_Y Calibration Date/Time: 04/28/2025 09:59

Lab File ID: VY022021.D Init. Calib. Date(s): 04/22/2025 04/22/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 13:39 16:15

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
cis-1,2-Dichloroethene	0.622	0.571		-8.2	20
1,1,1-Trichloroethane	0.896	0.830		-7.37	20
Benzene	1.387	1.363		-1.73	20
Trichloroethene	0.384	0.372		-3.13	20
Toluene	0.898	0.912		1.56	20
Ethyl Benzene	1.829	1.803		-1.42	20
m/p-Xylenes	0.728	0.739		1.51	20
o-Xylene	0.671	0.674		0.45	20
Isopropylbenzene	3.341	3.228		-3.38	20
1,2-Dichloroethane-d4	0.462	0.422		-8.66	20
Dibromofluoromethane	0.330	0.326		-1.21	20
Toluene-d8	1.245	1.244		-0.08	20
4-Bromofluorobenzene	0.419	0.422		0.72	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.