

Surrogate Summary

SDG No.: **Q1889**

Client: **Kleinfelder**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1889-01	COMP-1	1,2-Dichloroethane-d4	50	54.1	108	63	155
		Dibromofluoromethane	50	52.0	104	70	134
		Toluene-d8	50	48.5	97	74	123
		4-Bromofluorobenzene	50	39.6	79	38	136
Q1889-02	COMP-2	1,2-Dichloroethane-d4	50	56.6	113	63	155
		Dibromofluoromethane	50	51.9	104	70	134
		Toluene-d8	50	48.2	96	74	123
		4-Bromofluorobenzene	50	39.8	80	38	136
Q1889-03	COMP-3	1,2-Dichloroethane-d4	50	54.0	108	63	155
		Dibromofluoromethane	50	50.7	101	70	134
		Toluene-d8	50	48.6	97	74	123
		4-Bromofluorobenzene	50	38.2	76	38	136
VY0428SBL01	VY0428SBL01	1,2-Dichloroethane-d4	50	51.3	103	63	155
		Dibromofluoromethane	50	50.8	102	70	134
		Toluene-d8	50	47.1	94	74	123
		4-Bromofluorobenzene	50	57.4	115	38	136
VY0428SBS01	VY0428SBS01	1,2-Dichloroethane-d4	50	49.5	99	63	155
		Dibromofluoromethane	50	50.8	101	70	134
		Toluene-d8	50	51.9	104	74	123
		4-Bromofluorobenzene	50	50.6	101	38	136
VY0428SBSD01	VY0428SBSD01	1,2-Dichloroethane-d4	50	48.0	96	63	155
		Dibromofluoromethane	50	50.0	100	70	134
		Toluene-d8	50	50.3	101	74	123
		4-Bromofluorobenzene	50	48.8	98	38	136