

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1889

Client: Kleinfelder

Analytical Method: SW8260D

Datafile : VY022023.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0428SBS01	cis-1,2-Dichloroethene	20	19.9	ug/Kg	100			82	123	
	1,1,1-Trichloroethane	20	20.6	ug/Kg	103			80	126	
	Benzene	20	21.0	ug/Kg	105			84	121	
	Trichloroethene	20	20.8	ug/Kg	104			83	122	
	Toluene	20	21.1	ug/Kg	106			83	122	
	Ethyl Benzene	20	20.0	ug/Kg	100			82	124	
	m/p-Xylenes	40	41.2	ug/Kg	103			83	124	
	o-Xylene	20	20.2	ug/Kg	101			83	123	
	Isopropylbenzene	20	19.5	ug/Kg	98			82	124	