

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1889

Client: Kleinfelder

Analytical Method: SW8260D

Datafile : VY022024.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0428SBSD01	cis-1,2-Dichloroethene	20	19.6	ug/Kg	98	2		82	123	20
	1,1,1-Trichloroethane	20	20.2	ug/Kg	101	2		80	126	20
	Benzene	20	20.4	ug/Kg	102	3		84	121	20
	Trichloroethene	20	20.5	ug/Kg	103	1		83	122	20
	Toluene	20	20.7	ug/Kg	104	2		83	122	20
	Ethyl Benzene	20	19.7	ug/Kg	99	1		82	124	20
	m/p-Xylenes	40	40.4	ug/Kg	101	2		83	124	20
	o-Xylene	20	19.9	ug/Kg	100	1		83	123	20
	Isopropylbenzene	20	19.5	ug/Kg	98	0		82	124	20