

Report of Analysis

Client:	ENTACT	Date Collected:	04/24/25 15:00
Project:	540 Degraw St, Brooklyn, NY - E9309	Date Received:	04/25/25
Client Sample ID:	TW-WTS-07	SDG No.:	Q1890
Lab Sample ID:	Q1890-01	Matrix:	WATER
		% Solid:	0

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
BOD5	115		1	0.20	2.00	mg/L		04/25/25 17:40	SM 5210 B-16
Flash Point	>212		1	0	0	o F		04/30/25 10:00	1010B
TSS	4.70		1	1.00	4.00	mg/L		04/30/25 15:45	SM 2540 D-15

Comments: Other method reference for flash point : Pensky-Martens Closed Cup Flash Point ASTM D 93 - IP 34

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits