



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: ENTACT **SDG No.:** Q1890
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1890 **SAS No.:** Q1890
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.17	0.2	86	40 - 160	CV	05/01/2025	12:05	LB135623
CRI01	Cadmium	6.36	6.0	106	40 - 160	P	05/01/2025	14:16	LB135636
	Copper	23.1	20.0	116	40 - 160	P	05/01/2025	14:16	LB135636
	Lead	13.3	12.0	111	40 - 160	P	05/01/2025	14:16	LB135636
	Nickel	41.9	40.0	105	40 - 160	P	05/01/2025	14:16	LB135636
	Zinc	42.5	40.0	106	40 - 160	P	05/01/2025	14:16	LB135636