



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>ENTACT</u>				SDG No.:	<u>Q1890</u>				
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1890</u>		SAS No.:	<u>Q1890</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB112	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	11:49	LB135623	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>ENTACT</u>		SDG No.: <u>Q1890</u>							
Contract: <u>ENTA05</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1890</u>	SAS No.: <u>Q1890</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB62	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	12:01	LB135623
CCB63	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	12:51	LB135623
CCB64	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	13:23	LB135623
CCB65	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	14:02	LB135623
CCB66	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	14:40	LB135623
CCB67	Mercury	0.20	+/-0.20	U	0.20	CV	05/01/2025	15:25	LB135623

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT

SDG No.: Q1890

Contract: ENTA05

Lab Code: CHEM

Case No.: Q1890

SAS No.: Q1890

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	14:11	LB135636
	Copper	20.0	+/-20.0	U	20.0	P	05/01/2025	14:11	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	14:11	LB135636
	Nickel	40.0	+/-40.0	U	40.0	P	05/01/2025	14:11	LB135636
	Zinc	40.0	+/-40.0	U	40.0	P	05/01/2025	14:11	LB135636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: ENTACT		SDG No.: Q1890							
Contract: ENTA05	Lab Code: CHEM	Case No.: Q1890	SAS No.: Q1890						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	15:12	LB135636
	Copper	20.0	+/-20.0	U	20.0	P	05/01/2025	15:12	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	15:12	LB135636
	Nickel	40.0	+/-40.0	U	40.0	P	05/01/2025	15:12	LB135636
	Zinc	40.0	+/-40.0	U	40.0	P	05/01/2025	15:12	LB135636
CCB02	Cadmium	0.60	+/-6.00	J	6.00	P	05/01/2025	16:04	LB135636
	Copper	20.0	+/-20.0	U	20.0	P	05/01/2025	16:04	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	16:04	LB135636
	Nickel	40.0	+/-40.0	U	40.0	P	05/01/2025	16:04	LB135636
	Zinc	40.0	+/-40.0	U	40.0	P	05/01/2025	16:04	LB135636
CCB03	Cadmium	0.88	+/-6.00	J	6.00	P	05/01/2025	17:07	LB135636
	Copper	20.0	+/-20.0	U	20.0	P	05/01/2025	17:07	LB135636
	Lead	3.14	+/-12.0	J	12.0	P	05/01/2025	17:07	LB135636
	Nickel	40.0	+/-40.0	U	40.0	P	05/01/2025	17:07	LB135636
	Zinc	40.0	+/-40.0	U	40.0	P	05/01/2025	17:07	LB135636
CCB04	Cadmium	0.92	+/-6.00	J	6.00	P	05/01/2025	17:53	LB135636
	Copper	20.0	+/-20.0	U	20.0	P	05/01/2025	17:53	LB135636
	Lead	2.63	+/-12.0	J	12.0	P	05/01/2025	17:53	LB135636
	Nickel	40.0	+/-40.0	U	40.0	P	05/01/2025	17:53	LB135636
	Zinc	40.0	+/-40.0	U	40.0	P	05/01/2025	17:53	LB135636
CCB05	Cadmium	0.89	+/-6.00	J	6.00	P	05/01/2025	18:40	LB135636
	Copper	20.0	+/-20.0	U	20.0	P	05/01/2025	18:40	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	18:40	LB135636
	Nickel	40.0	+/-40.0	U	40.0	P	05/01/2025	18:40	LB135636
	Zinc	40.0	+/-40.0	U	40.0	P	05/01/2025	18:40	LB135636
CCB06	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	19:27	LB135636
	Copper	20.0	+/-20.0	U	20.0	P	05/01/2025	19:27	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	19:27	LB135636
	Nickel	40.0	+/-40.0	U	40.0	P	05/01/2025	19:27	LB135636
	Zinc	40.0	+/-40.0	U	40.0	P	05/01/2025	19:27	LB135636
CCB07	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	20:15	LB135636
	Copper	20.0	+/-20.0	U	20.0	P	05/01/2025	20:15	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	20:15	LB135636
	Nickel	40.0	+/-40.0	U	40.0	P	05/01/2025	20:15	LB135636
	Zinc	40.0	+/-40.0	U	40.0	P	05/01/2025	20:15	LB135636
CCB08	Cadmium	6.00	+/-6.00	U	6.00	P	05/01/2025	20:28	LB135636
	Copper	20.0	+/-20.0	U	20.0	P	05/01/2025	20:28	LB135636
	Lead	12.0	+/-12.0	U	12.0	P	05/01/2025	20:28	LB135636
	Nickel	40.0	+/-40.0	U	40.0	P	05/01/2025	20:28	LB135636

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>ENTACT</u>				SDG No.:	<u>Q1890</u>			
Contract:	<u>ENTA05</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1890</u>	SAS No.:	<u>Q1890</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Zinc	40.0	+/-40.0	U	40.0	P	05/01/2025	20:28	LB135636