

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: ENTACT **SDG No.:** Q1890
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1890 **SAS No.:** Q1890
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-4.57	1.0	457	-5	7	05/01/2025	14:35	LB135636
	Copper	5.64	2.0	282	-18	22	05/01/2025	14:35	LB135636
	Lead	-4.15			-12	12	05/01/2025	14:35	LB135636
	Nickel	2.25	2.0	112	-38	42	05/01/2025	14:35	LB135636
	Zinc	1.75			-40	40	05/01/2025	14:35	LB135636
ICSAB01	Cadmium	1080	972	111	826	1120	05/01/2025	14:39	LB135636
	Copper	540	511	106	434	588	05/01/2025	14:39	LB135636
	Lead	52.5	49.0	107	37	61	05/01/2025	14:39	LB135636
	Nickel	1060	954	111	810	1100	05/01/2025	14:39	LB135636
	Zinc	1090	952	114	809	1095	05/01/2025	14:39	LB135636