

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT **SDG No.:** Q1890
Contract: ENTA05 **Lab Code:** CHEM **Case No.:** Q1890 **SAS No.:** Q1890

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167799BS							
Cadmium	ug/L	100	93.2		93	80 - 120	P
Copper	ug/L	150	142		95	80 - 120	P
Lead	ug/L	500	457		91	80 - 120	P
Nickel	ug/L	250	229		92	80 - 120	P
Zinc	ug/L	100	94.3		94	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: ENTACT

SDG No.: Q1890

Contract: ENTA05

Lab Code: CHEM

Case No.: Q1890

SAS No.: Q1890

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB167812BS Mercury	ug/L	4.0	4.11		103	80 - 120	CV