

Surrogate Summary

SW-846

SDG No.: Q1890

Client: ENTACT

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB167798BL	PB167798BL	2-Fluorophenol	150	113	76		15 (10)	110 (139)
		Phenol-d6	150	113	76		15 (10)	110 (134)
		Nitrobenzene-d5	100	80.3	80		30 (49)	130 (133)
		2-Fluorobiphenyl	100	70.4	70		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	118	79		15 (44)	110 (137)
		Terphenyl-d14	100	64.3	64		30 (48)	130 (125)
PB167798BS	PB167798BS	2-Fluorophenol	150	111	74		15 (10)	110 (139)
		Phenol-d6	150	113	75		15 (10)	110 (134)
		Nitrobenzene-d5	100	78.7	79		30 (49)	130 (133)
		2-Fluorobiphenyl	100	69.2	69		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	131	87		15 (44)	110 (137)
		Terphenyl-d14	100	73.5	74		30 (48)	130 (125)
PB167798BSD	PB167798BSD	2-Fluorophenol	150	108	72		15 (10)	110 (139)
		Phenol-d6	150	110	74		15 (10)	110 (134)
		Nitrobenzene-d5	100	80.3	80		30 (49)	130 (133)
		2-Fluorobiphenyl	100	68.2	68		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	130	87		15 (44)	110 (137)
		Terphenyl-d14	100	73.8	74		30 (48)	130 (125)
Q1890-01	TW-WTS-07	2-Fluorophenol	150	56.7	38		15 (10)	110 (139)
		Phenol-d6	150	34.8	23		15 (10)	110 (134)
		Nitrobenzene-d5	100	81.8	82		30 (49)	130 (133)
		2-Fluorobiphenyl	100	73.5	74		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	116	77		15 (44)	110 (137)
		Terphenyl-d14	100	53.6	54		30 (48)	130 (125)