

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1890

Client: ENTACT

Analytical Method: 8270E DataFile: BF142252.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB167798BS	Phenol	50	41.2	ug/L	82				20 (66)	160 (118)	
	1,4-Dichlorobenzene	50	40.4	ug/L	81				70 (76)	130 (103)	
	1,2,4-Trichlorobenzene	50	41.0	ug/L	82				70 (41)	130 (115)	
	Naphthalene	50	40.4	ug/L	81				70 (64)	130 (107)	