

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: ENTA05

Lab Code: CHEM Case No.: Q1890 SAS No.: Q1890 SDG No.: Q1890

Instrument ID: MSVOA_X Calibration Date/Time: 04/28/2025 10:16

Lab File ID: VX045944.D Init. Calib. Date(s): 04/01/2025 04/01/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 17:06 19:02

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Methyl tert-butyl Ether	2.100	2.358		12.29	20
Carbon Tetrachloride	0.518	0.629		21.43	20
Chloroform	1.306	1.408		7.81	20
1,1,1-Trichloroethane	1.105	1.245		12.67	20
Benzene	1.463	1.594		8.95	20
Toluene	0.885	0.977		10.4	20
Tetrachloroethene	0.353	0.392		11.05	20
Ethyl Benzene	1.914	2.229		16.46	20
m/p-Xylenes	0.696	0.812		16.67	20
o-Xylene	0.686	0.800		16.62	20
1,2-Dichloroethane-d4	0.914	0.849		-7.11	20
Dibromofluoromethane	0.355	0.352		-0.85	20
Toluene-d8	1.238	1.156		-6.62	20
4-Bromofluorobenzene	0.451	0.457		1.33	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.