

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: ENTA05
 Lab Code: CHEM Case No.: Q1890 SAS No.: Q1890 SDG NO.: Q1890
 Lab File ID: VX045944.D Date Analyzed: 04/28/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:16
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	95099	5.54	163172	6.75	141994	10.05
UPPER LIMIT	190198	6.044	326344	7.251	283988	10.549
LOWER LIMIT	47549.5	5.044	81586	6.251	70997	9.549
EPA SAMPLE NO.						
TW-WTS-07	72284	5.55	142090	6.76	135017	10.05
VX0428WBL01	63502	5.54	124129	6.76	116546	10.05
VX0428WBS01	81017	5.55	142403	6.76	124900	10.05
VX0428WBSD01	82134	5.54	144508	6.76	128045	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.