

## **CASE NARRATIVE**

### **Portal Partners Tri-Venture**

**Project Name: Amtrak Sawtooth Bridges 2025**

**Project # N/A**

**Order ID # Q1901**

**Test Name: TCLP BNA**

### **A. Number of Samples and Date of Receipt:**

9 Solid samples were received on 04/28/2025.

2 Water samples were received on 04/28/2025.

### **B. Parameters**

According to the Chain of Custody document, the following analyses were requested: Corrosivity, EPH, EPH, Flash Point, Hexavalent Chromium, Ignitability, Mercury, Metals ICP-TAL, METALS-TAL, PCB, pH, RCRA CHARACTERISTICS, Reactive Cyanide, Reactive Sulfide, SVOC-TCL BNA -20, TCLP BNA, TCLP Extraction, TCLP FULL, TCLP Herbicide, TCLP ICP Metals, TCLP Mercury, TCLP Pesticide, TCLP VOA, TCLP ZHE Extraction, TCLP-FULL, TPH GC, Trivalent Chromium, VOC-TCLVOA-10 and VOC-TCLVOA-10. This data package contains results for TCLP BNA.

### **C. Analytical Techniques:**

The samples were analyzed on instrument BNA\_F using GC Column DB-UI 8270D which is 20 meters, 0.18 mm ID, 0.36 um df The samples were analyzed on instrument BNA\_P using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of TCLP BNA was based on method 8270E and extraction was done based on method 3510 and TCLP extraction method was 1311.

### **D. QA/ QC Samples:**

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for B-167-SB02 [2,4,6-Tribromophenol - 141%, Terphenyl-d14 - 181%], B-167-SB02RE [2,4,6-Tribromophenol - 369%, 2-Fluorobiphenyl - 313% and Terphenyl-d14 - 902%], these compounds did not meet the NJDKQP criteria and in-house criteria but Failure sample for surrogate was reanalyzed to confirm the failure and both run were reported in Hard Copy.

The Internal Standards Areas met the acceptable requirements except for B-167-SB02, B-167-SB02RE, Failure sample for Internal Standard was reanalyzed to confirm the failure and both run were reported in Hard Copy.

The Retention Times were acceptable for all samples.



284 Sheffield Street, Mountainside, NJ 07092  
Phone: 908 789 8900 Fax: 908 789 8922

The MS {Q1901-08MS} with File ID: BF142277.D recoveries met the requirements for all compounds except for 1,4-Dichlorobenzene[68%], this compound did not meet the NJDKQP criteria but met the in-house criteria.

The MSD {Q1901-08MSD} with File ID: BF142278.D recoveries met the acceptable requirements except for 1,4-Dichlorobenzene[68%], this compound did not meet the NJDKQP criteria but met the in-house criteria.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The % RSD is greater than 20% in the Initial Calibration (8270-BF043025.M) for 2,4-Dinitrotoluene, this compound is passing on Linear Regression.

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

**E. Additional Comments:**

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

**F. Manual Integration Comments:**

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

---

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature\_\_\_\_\_