



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture
Project: Amtrak Sawtooth Bridges 2025

SDG No.: Q1901
RunNo.: LB135579

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.500	0.5	100	90-110	04/28/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.504	0.5	101	90-110	04/28/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.504	0.5	101	90-110	04/28/2025

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q1901

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB135581

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Reactive Cyanide	mg/L	0.096	0.099	97	85-115	04/28/2025
Sample ID: CCV1 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	04/28/2025
Sample ID: CCV2 Reactive Cyanide	mg/L	0.24	0.25	96	90-110	04/28/2025
Sample ID: CCV3 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	04/28/2025
Sample ID: CCV4 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	04/28/2025
Sample ID: CCV5 Reactive Cyanide	mg/L	0.25	0.25	100	90-110	04/28/2025

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q1901

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB135585

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Corrosivity	ICV	pH	7.01	7	100	90-110	04/29/2025
Sample ID: Corrosivity	CCV1	pH	2.01	2.00	101	90-110	04/29/2025
Sample ID: Corrosivity	CCV2	pH	12.02	12.00	100	90-110	04/29/2025
Sample ID: Corrosivity	CCV3	pH	2.01	2.00	101	90-110	04/29/2025

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q1901

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB135586

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	7.01	7	100	90-110	04/29/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	04/29/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	04/29/2025

Initial and Continuing Calibration Verification

Client: Portal Partners Tri-Venture

SDG No.: Q1901

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB135596

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Hexavalent Chromium	mg/L	0.491	0.5	98	90-110	04/29/2025
Sample ID: CCV1 Hexavalent Chromium	mg/L	0.499	0.5	100	90-110	04/29/2025
Sample ID: CCV2 Hexavalent Chromium	mg/L	0.491	0.5	98	90-110	04/29/2025
Sample ID: CCV3 Hexavalent Chromium	mg/L	0.494	0.5	99	90-110	04/29/2025
Sample ID: CCV4 Hexavalent Chromium	mg/L	0.494	0.5	99	90-110	04/29/2025