

Initial and Continuing Calibration Blank Summary

Client: Portal Partners Tri-Venture

SDG No.: Q1901

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB135579

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	04/28/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	04/28/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	04/28/2025

Initial and Continuing Calibration Blank Summary

Client: Portal Partners Tri-Venture

SDG No.: Q1901

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB135581

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/28/2025
Sample ID: CCB1 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/28/2025
Sample ID: CCB2 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/28/2025
Sample ID: CCB3 Reactive Cyanide	mg/L	0.0011	0.0025	J	0.00096	0.005	04/28/2025
Sample ID: CCB4 Reactive Cyanide	mg/L	< 0.0025	0.0025	U	0.00096	0.005	04/28/2025
Sample ID: CCB5 Reactive Cyanide	mg/L	0.0011	0.0025	J	0.00096	0.005	04/28/2025

Initial and Continuing Calibration Blank Summary

Client: Portal Partners Tri-Venture

SDG No.: Q1901

Project: Amtrak Sawtooth Bridges 2025

RunNo.: LB135596

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	04/29/2025
Sample ID: CCB1 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	04/29/2025
Sample ID: CCB2 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	04/29/2025
Sample ID: CCB3 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	04/29/2025
Sample ID: CCB4 Hexavalent Chromium	mg/L	< 0.0050	0.0050	U	0.0029	0.01	04/29/2025