


LB135623

7470A

INSTRUMENT 1A: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.014	0.014	-7	0.000	-7	0			
0.2	0.200	0.198	-0.002	966	0.0 %	966				
2.5	2.500	2.415	-0.085	12681	0.0 %	12681				
5.0	5.000	5.160	0.160	27185	0.0 %	27185				
7.5	7.500	7.402	-0.098	39029	0.0 %	39029				
10.0	10.000	10.011	0.011	52816	0.0 %	52816				

6/10
 11
 13
 33
 11
 0

LB135555623 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 May 2025 11:10:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-7	-	0.0000	7470A	PPB	01 May 2025 11:17:27	S	Std
0.2 - 1	50.2	966	-	0.2000	7470A	PPB	01 May 2025 11:19:44	S	Std
2.5 - 1	52.5	12681	-	2.5000	7470A	PPB	01 May 2025 11:22:01	S	Std
5.0 - 1	55	27185	-	5.0000	7470A	PPB	01 May 2025 11:26:58	S	Std
7.5 - 1	57.5	39029	-	7.5000	7470A	PPB	01 May 2025 11:31:57	S	Std
10.0 - 1	59	52816	-	10.0000	7470A	PPB	01 May 2025 11:36:52	S	Std
ICV112 - 1	ICV112	20054	3.8106	-7470A	PPB	01 May 2025 11:47:40	U	SMPL	
ICB112 - 1	ICB112	-356	-0.0519	-7470A	PPB	01 May 2025 11:49:58	U	SMPL	
CCV62 - 1	CCV62	25268	4.7973	-7470A	PPB	01 May 2025 12:01:04	U	SMPL	
CCB62 - 1	CCB62	-363	-0.0533	-7470A	PPB	01 May 2025 12:05:54	U	SMPL	
CRA - 1	CRA	824	0.1714	-7470A	PPB	01 May 2025 12:08:09	U	SMPL	
HighStd - 1	HighStd	49868	9.4528	-7470A	PPB	01 May 2025 12:18:24	U	SMPL	
ChkStd - 1	ChkStd	39767	7.5412	-7470A	PPB	01 May 2025 12:28:15	U	SMPL	
PB167806BL - 1	PBW	306	0.0733	-7470A	PPB	01 May 2025 12:30:30	U	SMPL	
PB167806BS - 1	LCSW	17569	3.3403	-7470A	PPB	01 May 2025 12:32:47	U	SMPL	
Q1901-08 - 1	B-167-SB01	-265	-0.0347	-7470A	PPB	01 May 2025 12:35:05	U	SMPL	
Q1901-08DUP - 1	B-167-SB01DUP	-24	0.0109	-7470A	PPB	01 May 2025 12:39:26	U	SMPL	
Q1901-08MS - 1	B-167-SB01MS	20586	3.9113	-7470A	PPB	01 May 2025 12:41:43	U	SMPL	
Q1901-08MSD - 1	B-167-SB01MSD	21549	4.0935	-7470A	PPB	01 May 2025 12:44:01	U	SMPL	
Q1901-09 - 1	B-170-SB01	-223	-0.0268	-7470A	PPB	01 May 2025 12:49:05	U	SMPL	
CCV63 - 1	CCV63	26403	5.0121	-7470A	PPB	01 May 2025 12:51:22	U	SMPL	
CCB63 - 1	CCB63	-486	-0.0765	-7470A	PPB	01 May 2025 12:53:39	U	SMPL	
Q1901-10 - 1	B-167-SB02	-25	0.0107	-7470A	PPB	01 May 2025 12:55:54	U	SMPL	
Q1901-11 - 1	B-170-SB02	-209	-0.0241	-7470A	PPB	01 May 2025 12:58:10	U	SMPL	
Q1904-02 - 1	VNJ-210	-109	-0.0052	-7470A	PPB	01 May 2025 13:00:26	U	SMPL	
Q1905-04 - 1	MH-G	-28	0.0101	-7470A	PPB	01 May 2025 13:02:42	U	SMPL	
Q1905-08 - 1	MH-H	-17	0.0122	-7470A	PPB	01 May 2025 13:04:58	U	SMPL	
Q1906-04 - 1	WC-4	17	0.0186	-7470A	PPB	01 May 2025 13:07:14	U	SMPL	
Q1906-08 - 1	WC-5	11	0.0175	-7470A	PPB	01 May 2025 13:09:31	U	SMPL	
Q1906-12 - 1	WC-6	19	0.0190	-7470A	PPB	01 May 2025 13:11:48	U	SMPL	
Q1906-16 - 1	WC-7	46	0.0241	-7470A	PPB	01 May 2025 13:14:04	U	SMPL	
Q1907-02 - 1	CO-8R-WC	60	0.0268	-7470A	PPB	01 May 2025 13:19:10	U	SMPL	
CCV64 - 1	CCV64	26832	5.0933	-7470A	PPB	01 May 2025 13:23:57	U	SMPL	
CCB64 - 1	CCB64	-101	-0.0037	-7470A	PPB	01 May 2025 13:26:13	U	SMPL	
Q1912-04 - 1	MH-E	-32	0.0094	-7470A	PPB	01 May 2025 13:28:29	U	SMPL	
Q1912-08 - 1	MH-F	-7	0.0141	-7470A	PPB	01 May 2025 13:30:44	U	SMPL	
PB167807BL - 1	PBW	-14	0.0128	-7470A	PPB	01 May 2025 13:42:10	U	SMPL	
PB167807BS - 1	LCSW	21802	4.1414	-7470A	PPB	01 May 2025 13:44:25	U	SMPL	
Q1913-02 - 1	WC-12-A-202504	-316	-0.0444	-7470A	PPB	01 May 2025 13:46:47	U	SMPL	
Q1913-02DUP - 1	WC-12-A-202504DUP	-57	0.0046	-7470A	PPB	01 May 2025 13:49:03	U	SMPL	
Q1913-02MS - 1	WC-12-A-202504MS	16431	3.1249	-7470A	PPB	01 May 2025 13:51:20	U	SMPL	
Q1913-02MSD - 1	WC-12-A-202504MSD	16445	3.1276	-7470A	PPB	01 May 2025 13:53:38	U	SMPL	
Q1913-04 - 1	WC-13-A-202504	-268	-0.0353	-7470A	PPB	01 May 2025 13:55:57	U	SMPL	
PB167812BL - 1	PBW	-25	0.0107	-7470A	PPB	01 May 2025 13:58:13	U	SMPL	
CCV65 - 1	CCV65	25106	4.7667	-7470A	PPB	01 May 2025 14:02:49	U	SMPL	
CCB65 - 1	CCB65	-150	-0.0130	-7470A	PPB	01 May 2025 14:05:08	U	SMPL	
PB167812BS - 1	LCSW	21628	4.1085	-7470A	PPB	01 May 2025 14:07:24	U	SMPL	
Q1890-01 - 1	TW-WTS-07	-103	-0.0041	-7470A	PPB	01 May 2025 14:15:04	U	SMPL	
Q1896-03 - 1	295-BERGEN-FRAC	5385	1.0345	-7470A	PPB	01 May 2025 14:17:26	U	SMPL	
Q1896-03DUP - 1	295-BERGEN-FRACDUP	4924	0.9473	-7470A	PPB	01 May 2025 14:22:23	U	SMPL	
Q1896-03MS - 1	295-BERGEN-FRACMS	20396	3.8753	-7470A	PPB	01 May 2025 14:24:39	U	SMPL	
Q1896-03MSD - 1	295-BERGEN-FRACMSD	20002	3.8008	-7470A	PPB	01 May 2025 14:26:57	U	SMPL	
Q1901-06 - 1	FB04262025	-263	-0.0343	-7470A	PPB	01 May 2025 14:29:16	U	SMPL	
CCV66 - 1	CCV66	24238	4.6024	-7470A	PPB	01 May 2025 14:40:49	U	SMPL	
CCB66 - 1	CCB66	114	0.0370	-7470A	PPB	01 May 2025 14:43:05	U	SMPL	
Q1915-01 - 1		11	0.0175	-7470A	PPB	01 May 2025 14:45:21	U	SMPL	
PB167774TB - 1	PB167774TB	5	0.0164	-7470A	PPB	01 May 2025 14:47:36	U	SMPL	
PB167805TB - 1	PB167805TB	25	0.0202	-7470A	PPB	01 May 2025 14:49:52	U	SMPL	
Q1901-08LX5 - 1		51	0.0251	-7470A	PPB	01 May 2025 15:00:14	U	SMPL	
Q1901-08A - 1		24645	4.6794	-7470A	PPB	01 May 2025 15:02:29	U	SMPL	
Q1913-02LX5 - 1		-368	-0.0542	-7470A	PPB	01 May 2025 15:04:47	U	SMPL	
Q1913-02A - 1		21948	4.1690	-7470A	PPB	01 May 2025 15:07:02	U	SMPL	
Q1896-03LX5 - 1		602	0.1294	-7470A	PPB	01 May 2025 15:12:39	U	SMPL	
Q1896-03A - 1		25697	4.8785	-7470A	PPB	01 May 2025 15:23:09	U	SMPL	
CCV67 - 1	CCV67	24276	4.6096	-7470A	PPB	01 May 2025 15:25:28	U	SMPL	
CCB67 - 1	CCB67	-188	-0.0201	-7470A	PPB	01 May 2025 15:54:28	U	SMPL	
GEB62 - 1	CCB62	0	0.0154	-7470A	PPB				