



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** Q1901
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1901 **SAS No.:** Q1901
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	106	40 - 160	CV	04/30/2025	13:45	LB135611
CRI01	Aluminum	97.5	100	98	40 - 160	P	04/30/2025	16:41	LB135614
	Antimony	54.6	50.0	109	40 - 160	P	04/30/2025	16:41	LB135614
	Arsenic	21.9	20.0	109	40 - 160	P	04/30/2025	16:41	LB135614
	Barium	90.6	100	91	40 - 160	P	04/30/2025	16:41	LB135614
	Beryllium	6.30	6.0	105	40 - 160	P	04/30/2025	16:41	LB135614
	Cadmium	6.34	6.0	106	40 - 160	P	04/30/2025	16:41	LB135614
	Calcium	2030	2000	101	40 - 160	P	04/30/2025	16:41	LB135614
	Chromium	9.88	10.0	99	40 - 160	P	04/30/2025	16:41	LB135614
	Cobalt	30.3	30.0	101	40 - 160	P	04/30/2025	16:41	LB135614
	Copper	23.2	20.0	116	40 - 160	P	04/30/2025	16:41	LB135614
	Iron	90.9	100	91	40 - 160	P	04/30/2025	16:41	LB135614
	Lead	13.5	12.0	112	40 - 160	P	04/30/2025	16:41	LB135614
	Magnesium	2140	2000	107	40 - 160	P	04/30/2025	16:41	LB135614
	Manganese	21.0	20.0	105	40 - 160	P	04/30/2025	16:41	LB135614
	Nickel	41.3	40.0	103	40 - 160	P	04/30/2025	16:41	LB135614
	Potassium	2030	2000	101	40 - 160	P	04/30/2025	16:41	LB135614
	Selenium	16.5	20.0	82	40 - 160	P	04/30/2025	16:41	LB135614
	Silver	10.8	10.0	108	40 - 160	P	04/30/2025	16:41	LB135614
	Sodium	1740	2000	87	40 - 160	P	04/30/2025	16:41	LB135614
	Thallium	44.9	40.0	112	40 - 160	P	04/30/2025	16:41	LB135614
	Vanadium	38.9	40.0	97	40 - 160	P	04/30/2025	16:41	LB135614
	Zinc	42.8	40.0	107	40 - 160	P	04/30/2025	16:41	LB135614
CRA	Mercury	0.17	0.2	86	40 - 160	CV	05/01/2025	12:05	LB135623
CRI01	Aluminum	101	100	101	40 - 160	P	05/01/2025	14:16	LB135636
	Antimony	53.6	50.0	107	40 - 160	P	05/01/2025	14:16	LB135636
	Arsenic	23.3	20.0	116	40 - 160	P	05/01/2025	14:16	LB135636
	Barium	93.6	100	94	40 - 160	P	05/01/2025	14:16	LB135636
	Beryllium	6.02	6.0	100	40 - 160	P	05/01/2025	14:16	LB135636
	Cadmium	6.36	6.0	106	40 - 160	P	05/01/2025	14:16	LB135636
	Calcium	2030	2000	102	40 - 160	P	05/01/2025	14:16	LB135636
	Chromium	10.6	10.0	106	40 - 160	P	05/01/2025	14:16	LB135636
	Cobalt	30.7	30.0	102	40 - 160	P	05/01/2025	14:16	LB135636
	Copper	23.1	20.0	116	40 - 160	P	05/01/2025	14:16	LB135636
	Iron	104	100	104	40 - 160	P	05/01/2025	14:16	LB135636

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Portal Partners Tri-Venture **SDG No.:** Q1901

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1901 **SAS No.:** Q1901

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Lead	13.3	12.0	111	40 - 160	P	05/01/2025	14:16	LB135636
	Magnesium	2150	2000	108	40 - 160	P	05/01/2025	14:16	LB135636
	Manganese	20.7	20.0	104	40 - 160	P	05/01/2025	14:16	LB135636
	Nickel	41.9	40.0	105	40 - 160	P	05/01/2025	14:16	LB135636
	Potassium	2180	2000	109	40 - 160	P	05/01/2025	14:16	LB135636
	Selenium	23.1	20.0	115	40 - 160	P	05/01/2025	14:16	LB135636
	Silver	10.6	10.0	106	40 - 160	P	05/01/2025	14:16	LB135636
	Sodium	2040	2000	102	40 - 160	P	05/01/2025	14:16	LB135636
	Thallium	44.6	40.0	112	40 - 160	P	05/01/2025	14:16	LB135636
	Vanadium	41.6	40.0	104	40 - 160	P	05/01/2025	14:16	LB135636
	Zinc	42.5	40.0	106	40 - 160	P	05/01/2025	14:16	LB135636
CRI01	Aluminum	102	100	102	40 - 160	P	05/06/2025	15:31	LB135689
	Antimony	50.4	50.0	101	40 - 160	P	05/06/2025	15:31	LB135689
	Arsenic	20.3	20.0	101	40 - 160	P	05/06/2025	15:31	LB135689
	Barium	89.4	100	89	40 - 160	P	05/06/2025	15:31	LB135689
	Beryllium	5.91	6.0	98	40 - 160	P	05/06/2025	15:31	LB135689
	Cadmium	5.91	6.0	98	40 - 160	P	05/06/2025	15:31	LB135689
	Calcium	1950	2000	97	40 - 160	P	05/06/2025	15:31	LB135689
	Chromium	9.54	10.0	95	40 - 160	P	05/06/2025	15:31	LB135689
	Cobalt	28.3	30.0	94	40 - 160	P	05/06/2025	15:31	LB135689
	Copper	21.6	20.0	108	40 - 160	P	05/06/2025	15:31	LB135689
	Iron	95.6	100	96	40 - 160	P	05/06/2025	15:31	LB135689
	Lead	10.0	12.0	83	40 - 160	P	05/06/2025	15:31	LB135689
	Magnesium	2050	2000	102	40 - 160	P	05/06/2025	15:31	LB135689
	Manganese	19.8	20.0	99	40 - 160	P	05/06/2025	15:31	LB135689
	Nickel	38.6	40.0	97	40 - 160	P	05/06/2025	15:31	LB135689
	Potassium	1730	2000	87	40 - 160	P	05/06/2025	15:31	LB135689
	Selenium	22.5	20.0	112	40 - 160	P	05/06/2025	15:31	LB135689
	Silver	9.66	10.0	97	40 - 160	P	05/06/2025	15:31	LB135689
	Sodium	1760	2000	88	40 - 160	P	05/06/2025	15:31	LB135689
	Thallium	37.5	40.0	94	40 - 160	P	05/06/2025	15:31	LB135689
	Vanadium	39.4	40.0	98	40 - 160	P	05/06/2025	15:31	LB135689
	Zinc	39.7	40.0	99	40 - 160	P	05/06/2025	15:31	LB135689