

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1901
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1901 **SAS No.:** Q1901
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV112	Mercury	3.81	4.0	95	90 - 110	CV	05/01/2025	11:47	LB135623

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1901
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1901 **SAS No.:** Q1901
Initial Calibration Source: EPA
Continuing Calibration Source: PLASMA-PURE

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV62	Mercury	4.80	5.0	96	90 - 110	CV	05/01/2025	11:58	LB135623
CCV63	Mercury	5.01	5.0	100	90 - 110	CV	05/01/2025	12:49	LB135623
CCV64	Mercury	5.09	5.0	102	90 - 110	CV	05/01/2025	13:19	LB135623
CCV65	Mercury	4.77	5.0	95	90 - 110	CV	05/01/2025	13:58	LB135623
CCV66	Mercury	4.60	5.0	92	90 - 110	CV	05/01/2025	14:29	LB135623
CCV67	Mercury	4.61	5.0	92	90 - 110	CV	05/01/2025	15:23	LB135623

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture SDG No.: Q1901

Contract: PORT06 Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1010	1000	100	90 - 110	P	05/01/2025	13:44	LB135636
	Barium	495	520	95	90 - 110	P	05/01/2025	13:44	LB135636
	Cadmium	508	510	100	90 - 110	P	05/01/2025	13:44	LB135636
	Chromium	506	520	97	90 - 110	P	05/01/2025	13:44	LB135636
	Lead	989	1000	99	90 - 110	P	05/01/2025	13:44	LB135636
	Selenium	1040	1000	104	90 - 110	P	05/01/2025	13:44	LB135636
	Silver	240	250	96	90 - 110	P	05/01/2025	13:44	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1901
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1901 **SAS No.:** Q1901
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	23.7	20.0	119	80 - 120	P	05/01/2025	14:01	LB135636
	Barium	95.8	100	96	80 - 120	P	05/01/2025	14:01	LB135636
	Cadmium	6.39	6.0	107	80 - 120	P	05/01/2025	14:01	LB135636
	Chromium	10.5	10.0	105	80 - 120	P	05/01/2025	14:01	LB135636
	Lead	13.1	12.0	109	80 - 120	P	05/01/2025	14:01	LB135636
	Selenium	22.9	20.0	114	80 - 120	P	05/01/2025	14:01	LB135636
	Silver	10.2	10.0	102	80 - 120	P	05/01/2025	14:01	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1901

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1901 **SAS No.:** Q1901

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5090	5000	102	90 - 110	P	05/01/2025	15:05	LB135636
	Barium	9650	10000	96	90 - 110	P	05/01/2025	15:05	LB135636
	Cadmium	2500	2500	100	90 - 110	P	05/01/2025	15:05	LB135636
	Chromium	1000	1000	100	90 - 110	P	05/01/2025	15:05	LB135636
	Lead	5010	5000	100	90 - 110	P	05/01/2025	15:05	LB135636
	Selenium	5140	5000	103	90 - 110	P	05/01/2025	15:05	LB135636
	Silver	1250	1250	100	90 - 110	P	05/01/2025	15:05	LB135636
CCV02	Arsenic	4930	5000	99	90 - 110	P	05/01/2025	16:00	LB135636
	Barium	10200	10000	102	90 - 110	P	05/01/2025	16:00	LB135636
	Cadmium	2470	2500	99	90 - 110	P	05/01/2025	16:00	LB135636
	Chromium	995	1000	100	90 - 110	P	05/01/2025	16:00	LB135636
	Lead	4980	5000	100	90 - 110	P	05/01/2025	16:00	LB135636
	Selenium	4960	5000	99	90 - 110	P	05/01/2025	16:00	LB135636
	Silver	1220	1250	98	90 - 110	P	05/01/2025	16:00	LB135636
CCV03	Arsenic	5040	5000	101	90 - 110	P	05/01/2025	17:03	LB135636
	Barium	9950	10000	100	90 - 110	P	05/01/2025	17:03	LB135636
	Cadmium	2510	2500	100	90 - 110	P	05/01/2025	17:03	LB135636
	Chromium	988	1000	99	90 - 110	P	05/01/2025	17:03	LB135636
	Lead	5010	5000	100	90 - 110	P	05/01/2025	17:03	LB135636
	Selenium	5060	5000	101	90 - 110	P	05/01/2025	17:03	LB135636
	Silver	1220	1250	97	90 - 110	P	05/01/2025	17:03	LB135636
CCV04	Arsenic	5160	5000	103	90 - 110	P	05/01/2025	17:49	LB135636
	Barium	10000	10000	100	90 - 110	P	05/01/2025	17:49	LB135636
	Cadmium	2470	2500	99	90 - 110	P	05/01/2025	17:49	LB135636
	Chromium	987	1000	99	90 - 110	P	05/01/2025	17:49	LB135636
	Lead	4950	5000	99	90 - 110	P	05/01/2025	17:49	LB135636
	Selenium	5240	5000	105	90 - 110	P	05/01/2025	17:49	LB135636
	Silver	1250	1250	100	90 - 110	P	05/01/2025	17:49	LB135636
CCV05	Arsenic	5090	5000	102	90 - 110	P	05/01/2025	18:36	LB135636
	Barium	9950	10000	100	90 - 110	P	05/01/2025	18:36	LB135636
	Cadmium	2500	2500	100	90 - 110	P	05/01/2025	18:36	LB135636
	Chromium	984	1000	98	90 - 110	P	05/01/2025	18:36	LB135636
	Lead	4980	5000	100	90 - 110	P	05/01/2025	18:36	LB135636
	Selenium	5150	5000	103	90 - 110	P	05/01/2025	18:36	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1901
Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1901 **SAS No.:** Q1901
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV05	Silver	1230	1250	98	90 - 110	P	05/01/2025	18:36	LB135636
CCV06	Arsenic	4990	5000	100	90 - 110	P	05/01/2025	19:23	LB135636
	Barium	9830	10000	98	90 - 110	P	05/01/2025	19:23	LB135636
	Cadmium	2460	2500	98	90 - 110	P	05/01/2025	19:23	LB135636
	Chromium	980	1000	98	90 - 110	P	05/01/2025	19:23	LB135636
	Lead	4920	5000	98	90 - 110	P	05/01/2025	19:23	LB135636
	Selenium	5030	5000	101	90 - 110	P	05/01/2025	19:23	LB135636
	Silver	1240	1250	99	90 - 110	P	05/01/2025	19:23	LB135636
CCV07	Arsenic	5090	5000	102	90 - 110	P	05/01/2025	20:11	LB135636
	Barium	9790	10000	98	90 - 110	P	05/01/2025	20:11	LB135636
	Cadmium	2500	2500	100	90 - 110	P	05/01/2025	20:11	LB135636
	Chromium	995	1000	100	90 - 110	P	05/01/2025	20:11	LB135636
	Lead	4980	5000	100	90 - 110	P	05/01/2025	20:11	LB135636
	Selenium	5120	5000	102	90 - 110	P	05/01/2025	20:11	LB135636
	Silver	1250	1250	100	90 - 110	P	05/01/2025	20:11	LB135636
CCV08	Arsenic	5180	5000	104	90 - 110	P	05/01/2025	20:24	LB135636
	Barium	9880	10000	99	90 - 110	P	05/01/2025	20:24	LB135636
	Cadmium	2490	2500	100	90 - 110	P	05/01/2025	20:24	LB135636
	Chromium	980	1000	98	90 - 110	P	05/01/2025	20:24	LB135636
	Lead	4990	5000	100	90 - 110	P	05/01/2025	20:24	LB135636
	Selenium	5270	5000	105	90 - 110	P	05/01/2025	20:24	LB135636
	Silver	1240	1250	99	90 - 110	P	05/01/2025	20:24	LB135636

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture SDG No.: Q1901

Contract: PORT06 Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Arsenic	1060	1000	106	90 - 110	P	05/02/2025	13:47	LB135654
	Barium	518	520	100	90 - 110	P	05/02/2025	13:47	LB135654
	Cadmium	493	510	97	90 - 110	P	05/02/2025	13:47	LB135654
	Chromium	503	520	97	90 - 110	P	05/02/2025	13:47	LB135654
	Lead	966	1000	97	90 - 110	P	05/02/2025	13:47	LB135654
	Selenium	1040	1000	104	90 - 110	P	05/02/2025	13:47	LB135654
	Silver	230	250	92	90 - 110	P	05/02/2025	13:47	LB135654

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture SDG No.: Q1901

Contract: PORT06 Lab Code: CHEM Case No.: Q1901 SAS No.: Q1901

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Arsenic	19.0	20.0	95	80 - 120	P	05/02/2025	15:03	LB135654
	Barium	81.5	100	82	80 - 120	P	05/02/2025	15:03	LB135654
	Cadmium	5.51	6.0	92	80 - 120	P	05/02/2025	15:03	LB135654
	Chromium	10.1	10.0	101	80 - 120	P	05/02/2025	15:03	LB135654
	Lead	11.1	12.0	93	80 - 120	P	05/02/2025	15:03	LB135654
	Selenium	16.8	20.0	84	80 - 120	P	05/02/2025	15:03	LB135654
	Silver	10.8	10.0	108	80 - 120	P	05/02/2025	15:03	LB135654

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Portal Partners Tri-Venture **SDG No.:** Q1901

Contract: PORT06 **Lab Code:** CHEM **Case No.:** Q1901 **SAS No.:** Q1901

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Arsenic	5450	5000	109	90 - 110	P	05/02/2025	16:01	LB135654
	Barium	9540	10000	95	90 - 110	P	05/02/2025	16:01	LB135654
	Cadmium	2440	2500	98	90 - 110	P	05/02/2025	16:01	LB135654
	Chromium	1010	1000	101	90 - 110	P	05/02/2025	16:01	LB135654
	Lead	4820	5000	96	90 - 110	P	05/02/2025	16:01	LB135654
	Selenium	5460	5000	109	90 - 110	P	05/02/2025	16:01	LB135654
	Silver	1270	1250	102	90 - 110	P	05/02/2025	16:01	LB135654
CCV02	Arsenic	5240	5000	105	90 - 110	P	05/02/2025	16:34	LB135654
	Barium	9720	10000	97	90 - 110	P	05/02/2025	16:34	LB135654
	Cadmium	2390	2500	96	90 - 110	P	05/02/2025	16:34	LB135654
	Chromium	1000	1000	100	90 - 110	P	05/02/2025	16:34	LB135654
	Lead	4740	5000	95	90 - 110	P	05/02/2025	16:34	LB135654
	Selenium	5240	5000	105	90 - 110	P	05/02/2025	16:34	LB135654
	Silver	1260	1250	101	90 - 110	P	05/02/2025	16:34	LB135654