

Continuing Calibration Report for SequenceID : FC050225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068742.D

Aliphatic C9-C12	6488349.000	60.000	3.343	6.645	108139.150	105282.003	-2.714
Aliphatic C12-C16	4261833.000	40.000	6.646	10.046	106545.825	102322.575	-4.127
Aliphatic C16-C21	6185679.000	60.000	10.047	13.415	103094.650	98816.539	-4.329
Aliphatic C21-C28	7845338.000	80.000	13.416	17.079	98066.725	94185.337	-4.121
Aliphatic C28-C40	12251595.000	120.000	17.080	22.085	102096.625	89882.835	-13.589
Aliphatic EPH	37032794.000	360.000	3.343	22.085	102868.872	96276.618	-6.847