

Continuing Calibration Report for SequenceID : FC050225AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC068748.D

Aliphatic C9-C12	5997337.000	60.000	3.343	6.645	99955.617	105282.003	5.059
Aliphatic C12-C16	3867344.000	40.000	6.646	10.046	96683.600	102322.575	5.511
Aliphatic C16-C21	5592959.000	60.000	10.047	13.415	93215.983	98816.539	5.668
Aliphatic C21-C28	7186492.000	80.000	13.416	17.079	89831.150	94185.337	4.623
Aliphatic C28-C40	11519078.000	120.000	17.080	22.085	95992.317	89882.835	-6.797
Aliphatic EPH	34163210.000	360.000	3.343	22.085	94897.806	96276.618	1.432